# **Rocky Flats Independent Verification**

# Sampling and Survey Report

Evaluation and Interpretation of the Residual Radiological Surface Contamination Sampling Results

Buildings 727, 782, and 783 Survey Units 727–01, 727–02, 782–01, and 782–02

Volume I

March 2000

Work Performed Under DOE Contract No. DE-AC13-96GJ87335



**ADMIN RECCRD** 

B779-A-000158

1/250

# **Rocky Flats Independent Verification**

Sampling and Survey Report Buildings 727, 782, and 783 Survey Units 727-01, 727-02, 782-01, and 782-02

March 2000

Prepared by
U.S. Department of Energy
Grand Junction Office
Grand Junction, Colorado

Project Number RFL-131-0001-00-000 Document Number Z00007AA

Work Performed under DOE Contract No. DE-AC13-96GJ87335

# Signature Page

Prepared by			
•	Art Samiljan	Dat	te
	Staff Health Physicist		
Approved by			
	Steve Rima	Dat	te
	Project, Technical Manager		
Approved for			
Release by	Michael Butherus	Dat	_ te
	Manager, Major Projects		

# Contents

1.0	Intro	duction	1-1
	1.1	Background	1-1
	1.2	History	1-2
		1.2.1 Building 727	1–2
		1.2.2 Building 782	1–2
		1.2.3 Building 783	1–2
	1.3	Current Condition of Buildings 727, 782, and 783	
	1.4	Overview of the Assessment and Independent Verification Process	
2.0	Field	I Investigation	
	2.1	Mobilization	2-1
	2.2	Field Selection of Survey Units for Independent Verification	
	2.3	Field Identification of Sample Locations	
	2.4	Sampling Equipment and Procedures	
		2.4.1 Radiological Instrumentation	
		2.4.2 Sampling	2-17
		2.4.2.1 Smear Sampling	
•		2.4.2.2 Surface Media Sampling	
		2.4.3 Laboratory Measurements	
		2.4.4 Field Measurements	
		2.4.4.1 Background Determination	
		2.4.4.2 Direct Static Surface Radioactivity Measurements	2-20
3.0	Sam	pling and Survey Results	
	3.1	Direct Static Field Measurements	3–1
		3.1.1 Background Measurements	3–1
		3.1.1.1 Background Adjustment	
		3.1.2 Field Measurement Data	
		3.1.2.1 Post Surface Media Sampling Measurements	
	3.2	Laboratory Measurements	
		3.2.1 Smear Samples	
		3.2.2 Surface Media Samples	
4.0	Ana	lysis of Sample Plan Results for Compliance	
	4.1	·	
		4.1.1 Direct Static Surface Measurements	4–2
		4.1.2 Smear Samples for Removable Surface Contamination	
		4.1.3 Surface Media Samples	
	4.2	Summary of Field Sampling Data	
5.0	Gran	phic Presentations of the Survey and Sampling Results	
	5.1	Posting Plots—Spatial Contamination Distribution Graphics	5–1
	5.2	Histograms—Concentration Distribution Graphics	
	5.3	High-Low Graphs—Data Variability Graphics	
6.0		lity Control Sampling Results and Analysis	
	6.1	Stage I—Independent Quality Control of the Contractor's Sampling	
		6.1.1 Smear Samples	
		6.1.2 Surface Media Samples	
		6.1.3 Direct Surface Emission Measurements	
	6.2	Stage II—Quality Control of the Independent Verification Sampling	
	-	6.2.1 Smear Samples	
		•	

6.2.2 Surface Media Samples	6–9
6.2.3 Direct Static Measurements	6–13
6.2.3.1 Replicate Field Measurements	
6.2.3.2 Instrument Response Check Data	
7.0 Data Quality Analysis	
7.1 Detection Limit Adequacy	7–1
7.2 Sample Size and Statistical Power	7–2
7.3 Measurement Uncertainty and Data Quality Indicators	7–10
7.4 Overall Quality Assurance and Quality Control	
8.0 Summary and Conclusions	
8.1 Independent Verification Sampling and Survey	8–1
8.2 Independent Review of the Contractor's Final Status Survey Report and	0.0
Conclusions	
9.0 References	9-1
<b>*</b> ***********************************	
Figures	
E' 2 1 G 1: - C .: 1 G II .: 222 01	2.4
Figure 2–1. Sampling Grid—Survey Unit 727–01	
Figure 2–2. Sampling Grid—Survey Unit 727–01	
Figure 2–3. Sample Location Identifier	
Figure 2-4. Direct Static Surface Contamination Measurement Configuration	
Figure 2–5. Photo of Rotary Impact Drill and Bit	
Figure 2–6. Surface Media Sample Collection	
Figure 3-1. Data Evaluation Statistics—"Free Air" Instrument Background, 727-01	
Figure 3–2. Data Evaluation Statistics—Direct Static Surface Measurements, 727–01 Figure 3–3. Data Evaluation Statistics—Direct Static Surface Measurements, 727–01	
Figure 3–4. Data Evaluation Statistics—Uranium Series Activity, Surface Media Samples	
Figure 3–4. Data Evaluation Statistics—Cramum Series Activity, Surface Media Samples Figure 3–5. Data Evaluation Statistics—Transuranic Series Activity,	, <i>3</i> –33
Surface Media Samples	3_37
Figure 5-1. Posting Plot—Direct Static Surface Contamination Measurements	
Figure 5–2. Posting Plot—Surface Media Samples, Transuranic Activity	
Figure 5–2. Posting Plot—Surface Media Samples, Transulante Activity	
Figure 5–4. Histogram—Instrument Background Measurements, 727–01	
Figure 5-5. Histogram—Direct Static Surface Contamination Measurements, 727-01	3–22
Figure 5–6. Histogram—Post Surface Media Sampling Direct Static	5 24
Surface Measurements, 727–01	
Figure 5–7. Histogram—Surface Media Samples, Transuranic Activity, 727–01	
Figure 5–8. Histogram—Surface Media Samples, Uranium Series Activity, 727–01	
Figure 5–9. High-Low Graphs—Direct Static Surface Measurements, 727–01	
Figure 5-10. High-Low Graphs—Surface Media Samples, Transuranic Activity, 727-01	5–33
Figure 5-11. High-Low Graphs—Surface Media Samples, Uranium Series	
Activity, 727–01	
Figure 6-1. Comparison of Duplicate Alpha Isotopic Sample Analysis Results	
Figure 6–2. Comparison Between Replicate Direct Static Measurements	
Figure 6–3. Instrument Response Check Control Chart, Survey Unit 727–01	
Figure 6-4. Instrument Response Check Control Chart, Survey Unit 727-02	
Figure 6-5. Instrument Response Check Control Chart, Survey Unit 782-01	
Figure 6-6. Instrument Response Check Control Chart, Survey Unit 782-02	6–18

Figure 7-1. Retrospective Power of the Sign Test—Direct Static Surface Measurements	7–3
Figure 7–2. Retrospective Power of the Sign Test—Smear Sample Measurements	
Figure 7-3. Retrospective Power of the Sign Test—Smear Media Samples/	
Transuranic Activity	76
Figure 7-4. Retrospective Power of the Sign Test—Surface Media Samples/	
Uranium Series Activity	78
Tables	
Table 2-1. Survey Units Identified for Building 779 Cluster	2–2
Table 2–2. Smear Sample Analytical Method	2–18
Table 2-3. Surface Media Sample Analytical Method	2–19
Table 3-1. Direct Static Measurement Background Data, Buildings 727, 782, and 783	3–2
Table 3–2. Background Data Summary Statistics	
Table 3-3. Static Surface Contamination Measurement MDA Parameters	3–8
Table 3-4. Direct Static Surface Contamination Measurements, Survey Units 727-01,	
727-02, 782-01, and 782-02	3–11
Table 3-5. Summary Statistics, Direct Static Measurements, Buildings 727, 782, and 783	3–16
Table 3-6. Post Surface Media Sample Direct Static Surface Measurements,	
Survey Units 727–01, 727–02, 782–01, and 782–02	3–17
Table 3-7. Summary Statistics, Post Media Sampling Direct Static Measurements	
Table 3-8. Comparison of Direct Static Measurement Data Sets Summary Statistics,	
Survey Unit 727–01	3–18
Table 3-9. Comparison of Direct Static Measurement Data Sets Summary Statistics,	2 22
Survey Unit 727–02	
Table 3–10. Comparison of Direct Static Measurement Data Sets Summary Statistics,	
Table 3-11. Comparison of Direct Static Measurement Data Sets Summary Statistics,	
Table 3–12. Direct Static Surface Measurements Data Sets, Survey Units 727–01, 727–02,	3–25
782–01, and 782–02	5–25
782–01, and 782–02	3_27
Table 3-14. Summary of Analytical Report Data for Smears	3-28
Table 3–15. Surface Media Sample Data, Buildings 727 and 783, Survey Unit 727–01—	
Table 3–16. Summary Statistics, Surface Media Samples	
Table 4–1. Comparison of Direct Static Survey Measurements to Applicable DCGLs	4–3
Table 4–2. Comparison of Background Adjusted Direct Static Survey Results to	
Applicable DCGLs	4–3
Table 4-3. Comparison of Smear Sample Results to Applicable DCGLw	4-4
Table 4-4. Comparison of Surface Media Sample Results to Applicable DCGLs	
Table 6-1. Results of the Contractor's Assay of QC Smear Samples Provided by the IVC	
Table 6-2. Results of the Contractor's Assay of QC Surface Media Samples Provided	
by the IVC	6-4
Table 6-3. Comparison of Response of Instruments Used to Make Direct Surface	
Measurements	66
Table 6-4. Results of the GJO Analytical Laboratory Assay of QC Smear Samples	_
Provided by the IVC	
Table 6-5. Results of the IV Laboratory Internal QC Measurements for Smear Samples	6–10

	esults of the IV Laboratory Internal Blank QC Measurements for	
St	ırface Media Samples	6–11
Table 6-7. R	esults of the IV Laboratory Internal LCS QC Measurements for	
Sı	ırface Media Samples	6-12
Table 6-8. R	esults of the IV Laboratory Internal Duplicate QC Measurements for	
	ırface Media Samples	6–13
Table 6-9. R	esults of Replicate Direct Static Surface QC Measurements	6–15
Table 7-1. A	dequacy of Independent Verification Measurement Detection Limits	7–1
	arget Data Quality Indicators and Findings	
	omparison of Buildings 727, 782, and 783 DCGLs to Highest Observed	
	ompliance Parameters	8-1
	Appendices	
Appendix A	Random Selection Data	A-1
	Manufacturer's Certification	
* *	Analytical Laboratory Reports-Requisition Numbers 16900, 16901, 169	
	16905, 16906, and 16908	•
Appendix D	Background Sample Data	
	Background Data Set	
	Duplicate Sample Data	
	Instrument Calibration Data	
	Raw Data	

# **Acronyms and Abbreviations**

Am-241 americium-241

CCV continuing calibration verification

CDPHE Colorado Department of Public Health and Environment

CLP Contract Laboratory Program.

cpm counts per minute cm<sup>2</sup> square centimeter(s)

D&D decontamination and decomissioning DCGL derived concentration guideline level

DCGL<sub>EMC</sub> derived concentration guideline level-elevated measurement comparison

DCGL<sub>W</sub> derived concentration guideline level-average concentration

DOE U.S. Department of Energy dpm disintegration(s) per minute

DQA data quality analysis
DQI data quality indicator
DQO data quality objective

EPA U.S. Environmental Protection Agency

GJO Grand Junction Office IV independent verification

IVC independent verification contractor IVP independent verification program

LCS laboratory control sample

MARSSIM Multi-Agency Radiation Survey and Site Investigation Manual

MDA minimum detectable activity

mm millimeter(s) m<sup>2</sup> square meter(s)

NIST National Institute of Standards and Technology

PB preparation blank pCi picoCurie(s) Pu-238 plutonium-238 Pu-239 plutonium-239 Pu-240 plutonium-240 QA quality assurance QC quality control

RFETS Rocky Flats Environmental Technology Site

RFFO Rocky Flats Field Office

RMRS Rocky Mountain Remediation Services

SAP Sampling and Analysis Plan

U uranium

UCL<sub>95</sub> 95 percent upper confidence limit

## 1.0 Introduction

## 1.1 Background

This sampling and survey report evaluates the final status survey data collected in Buildings 727, 782, and 783, both by the Rocky Flats Environmental Technology Site's (RFETS) Contractors (Kaiser-Hill, Rocky Mountain Remediation Services, L.L.C., and their subcontractors, hereafter referred to as the Contractor) and by MACTEC-ERS, the independent verification contractor (IVC). Data collected by the IVC is designed to independently assess and verify the RFETS' compliance with the approved derived concentration guideline levels (DCGL) established for the buildings in the 779 Cluster. Data collected by the Contractor has been reviewed by the IVC and compared with the independent verification data collected by the IVC.

The sampling and survey data collected has been compared with the approved surface contamination concentration benchmark values known as DCGLs. The RFETS DCGLs for surface contamination concentration are specified in the Contractor's *Closeout Radiological Survey Plan for the 779 Cluster* (RMRS 1999a). The independent verification DCGLs are specified in the IVC's *Independent Verification Sampling and Analysis Plan for Building 779 Cluster* (DOE 1999a).

Samples collected and surveys performed to obtain independent verification and corroboration of the RFETS sampling and survey results were collected in accordance with the U.S. Environmental Protection Agency (EPA), Colorado Department of Public Health and Environment (CDPHE), and U.S. Department of Energy (DOE) approved *Independent Verification Sampling and Analysis Plan for Building 779 Cluster* (IV SAP) (DOE 1999a). The data is evaluated herein principally on the Multi-Agency Radiation Survey and Site Investigation Manual (MARSSIM) data quality assessment methods, conventional guidance from EPA, and accepted practice and methods used in radiological site assessment and characterization. Principal guidance documents include:

- Multi-Agency Radiation Survey and Site Investigation Manual (EPA 1997)
- Data Quality Objectives Process for Superfund (EPA 1993)
- Guidance for Data Quality Assessment–Practical Methods for Data Analysis (EPA 1998)
- Manual for Conducting Radiological Surveys in Support of License Termination (NRC 1992)
- A Nonparametric Statistical Methodology for the Design and Analysis of Final Status Decommissioning Surveys (NRC 1995)

A common theme in these guidance sources is the use of a seven-step data quality objective (DQO) activity as the foundation for the SAP development and subsequent data evaluation.

Following this introductory background is a discussion of Buildings 727, 782, and 783 histories and an overview of the assessment and independent verification process used. Section 2 describes the field methods and procedures used to collect data. Section 3 presents the sampling results and summary statistics for each subset of data. It also describes the data reduction process

used and statistical comparisons of the data subsets and their significance. Section 4 presents the sampling results in the context of compliance with the benchmark values while Section 5 presents the survey and sampling results in a graphic format. Evaluation of the Contractor's Final Status Radiological Survey, computations, and conclusions are presented in Section 6. The IVC collected data are compared to their respective DQOs in what is termed the Data Quality Analysis in Section 7. Section 8 summarizes the data quality analysis, provides objective assessment of the concentrations of residual contamination found in the buildings, and offers conclusions and recommendations for disposition of the buildings. Appendices are included to provide additional detail where appropriate.

The risk manager and decision maker for this project is DOE-Rocky Flats Field Office (DOE-RFFO).

## 1.2 History

The Building 779 Cluster is located on DOE's Rocky Flats site near Golden, Colorado. The site is a former nuclear weapons production facility. The various process facilities and laboratories were grouped together with their various support buildings and structures and identified as "clusters," with the building number of the principal building as the cluster name (e.g., the Building 779 Cluster). The 779 Cluster was primarily used for research and development activities and supported a number of various operations as part of the research and development mission including: 1) process chemistry technology, 2) physical metallurgy, 3) machining and gauging, 4) joining technology, and 5) hydriding operations. No processes or operations are now active.

# 1.2.1 Building 727

Building 727 is a single story structure constructed in 1973 that housed the emergency diesel generator in support of Building 782. The walls are cinderblock and the roof is fiberboard and tar with ballast material supported by beams.

### 1.2.2 Building 782

Building 782 is a single story structure constructed in 1973 that served as the second plenum building for Building 779. The walls are prefabricated concrete panels and concrete support columns. The roof is fiberboard and tar with ballast material supported by Tee beams.

#### 1.2.3 **Building 783**

Building 783 is a single story structure constructed in 1973 that housed the motor control centers for the Building 779 cooling towers and circulating pumps as well as the circulating pumps themselves. The walls and roof are galvanized steel.

# 1.3 Current Condition of Buildings 727, 782, and 783

Buildings 727, 782, and 783 underwent a decontamination and decommissioning (D&D) process to ready it for final status radiological survey. In the D&D process, the buildings were stripped of utility services, and equipment and penetrations were removed or cut flush with the walls. All penetrations in the slab were grouted and will remain until environmental restoration is

accomplished. All areas where contamination was detected were decontaminated by the contractor prior to conducting the final survey. Buildings 727, 782, and 783 were initially subdivided into five survey units. The IVC randomly selected four units as part of the overall Building 779 independent verification. At the request of the Contractors, the IVC has prepared this stand-alone report for Buildings 727, 782, and 783.

## 1.4 Overview of the Assessment and Independent Verification Process

The approach used to independently determine whether Buildings 727, 782, and 783 met the mean, or average, benchmark release criteria (derived concentration guideline level-average concentration [DCGL<sub>W</sub>]) followed the MARSSIM method. Ten of 49 survey units identified in the Building 779 Cluster were selected for actual measurement by the IVC. In this case, survey units 727–01, 727–02, 782–01 and 782–02 (Buildings 727, 782, and 783) were four of the ten selected for independent verification, thus meeting the contractual requirement to assess 5 to 10 percent of the Contractor's results. The IVC used oversight of the Contractor's scanning surveys and a critical review of the data collected by the Contractor to independently determine compliance with the maximum concentration benchmark release criteria (derived concentration guideline level-elevated measurement comparison [DCGL<sub>EMC</sub>]).

The first step in the process to independently assess the Contractor's basis for decision on the disposition of Buildings 727, 782, and 783 was to review the Contractor's SAP (RMRS 1999a) and associated D&D planning documents. All comments and issues raised by the IVC were reported to DOE–RFFO and were addressed by the RFETS Contractor and implemented in the final status survey plan, as necessary.

The Contractor's SAP establishes the criteria which, when met, represent acceptable levels of risk from exposure to residual contamination which might be present in the building. DOE–RFFO, EPA, and CDPHE agreed upon surface contamination concentration criteria below which further remedial action would not be warranted. These criteria, or DCGLs, serve as the benchmarks against which the building surfaces were to be measured. The Contractor's DCGLs are:

- The mean removable alpha surface contamination concentration in the selected survey unit(s) is below 20 disintegrations per minute (dpm)/100 square centimeters (cm<sup>2</sup>).
- The mean total alpha surface contamination concentration attributable to transuranic radioactivity as measured by direct surface emission in the selected survey unit(s) is below 100 dpm/100 cm<sup>2</sup> (averaged over 1 square meter [m<sup>2</sup>]).
- The mean total alpha surface contamination concentration attributable to uranium series radioactivity<sup>1</sup> as measured by direct surface emission in the selected survey unit(s) is below 1,000 dpm/100 cm<sup>2</sup> (averaged over 1 m<sup>2</sup>).
- The maximum total alpha surface contamination concentration attributable to transuranic radioactivity as measured by direct surface emission in the selected survey unit(s) is below 300 dpm/100 cm<sup>2</sup>.

<sup>&</sup>lt;sup>1</sup>In cases where isotopic composition is not determined, the SAP requires the application of the more restrictive limits associated with the transuranic series radionuclides.

- The maximum total alpha surface contamination concentration attributable to uranium series radioactivity as measured by direct surface emission in the selected survey unit(s) is below 5,000 dpm/100 cm<sup>2</sup>.
- The mean total alpha contamination concentration attributable to transuranic radioactivity on and beneath a surface with a surface coating as measured by collection and analysis of a surface media sample in the selected survey unit(s) is below 100 dpm/100 cm<sup>2</sup>.
- The mean total alpha contamination concentration attributable to uranium series radioactivity on and beneath a surface with a surface coating as measured by collection and analysis of a surface media sample in the selected survey unit(s) is below 1,000 dpm/100 cm<sup>2</sup>.

After reviewing the Contractor's SAP, an IV SAP was constructed. The IV SAP was developed in consultation with DOE-RFFO, EPA, and CDPHE. It was designed to detect and measure the concentration of the radioactive contamination remaining in survey units selected for independent verification such that statistically appropriate analyses could be used to determine whether the results obtained by the Contractor in the same survey unit could be verified or corroborated by the IVC. The IV SAP alone does not collect enough data to make the required decision for the entire building but provides sufficient data for critical comparison with the Contractor's conclusion in a single survey unit. In the case of Buildings 727, 782, and 783, the IVC performed surveys and sampled four of five possible survey units (727–01, 727–02, 782–01, and 782–02) identified by the Contractor.

The next step was to observe and evaluate the Contractor's implementation of the final status survey against the criteria established in the SAP. The IVC Health Physicist is permanently assigned to Rocky Flats and works on site to observe the Contractor's sampling and survey methods and review analytical processes.

The fourth element of the independent verification process was to provide blind matrix samples to the Contractor for inclusion in their sample batches from Building 779 Cluster. The blind samples included both blanks and spikes of smear filter paper matrices and surface media matrices. Blind matrix samples were included in the Contractor's sample batches from Building 779 as the Contractor's manpower and schedule permitted. It is important to note that it is not critical to the sampling objective to introduce Stage-I quality control samples to a particular batch of the Contractor's samples or even while they are sampling a particular building or survey unit being considered for independent verification.

Finally, and with the approved IV SAP, the sampling plan was executed. The IVC collected samples and performed measurements in the selected survey units in order to corroborate the results obtained by the Contractor. The measurements and samples were obtained in accordance with the *Independent Verification Sampling and Analysis Plan for Building 779 Cluster* (DOE 1999a).

The field data was reviewed in the field with representatives from DOE and the Contractor. The EPA and CDPHE have been apprized of the results of independent verification field data collected. Field data was recorded both on paper (Appendix D) and electronically (Appendix H). Following data collection, the data was verified and reduced so that the appropriate comparisons and analyses could be conducted. The presentation of the results of the field sampling are

# 2.0 Field Investigation

#### 2.1 Mobilization

Prior to mobilizing the independent verification sampling team at the site, each member of the team was provided with a copy of the IV SAP and was trained on the field sampling equipment and procedures to be used. The Contractor made detailed measurements of the buildings and supplied simple architectural drawings of the survey units in Buildings 727, 782, and 783 (727–01, 727–02, 782–01, and 782–02) to be used in laying out the sampling grids and sample locations.

Mobilization to the field began the week of January 24, 2000. The sampling team personnel had completed all required RFETS training necessary to support the sampling and survey work during a previous visit to RFETS. Equipment was staged and final details were arranged.

The radiation survey detectors selected for this application were gas-filled, proportional counters made by Eberline; model HP-100. Gas-filled detectors are subject to response and calibration variation with changes in altitude. Consequently, the HP-100 probes were field calibrated on site at the RFETS altitude and using RFETS supplied counting gas. The instrument and probe package was response checked and verified to be in working order and within the parameters established in the SAP. The surface media sampling tool was tested, and test measurements were made confirming the suitability of the measurement protocol. A walk-through of each building was made to assess the condition of the building, to identify any intrinsic safety issues, and to compare the building structure and features with the assumptions made and procedures outlined in the SAP. It was concluded that the assumptions used to develop the SAP and its associated procedures were consistent with the conditions existing in the buildings and that the procedures developed for characterizing the contaminant concentrations in the buildings accounted for these conditions.

# 2.2 Field Selection of Survey Units for Independent Verification

The first step in the sample allocation strategy was to select from among the 49 survey units available in the Building 779 Cluster those survey units to be sampled and surveyed by the IVC. The random selection process assigned greater weighting or priority to survey units with a classification indicating greater potential to exceed the allowable radiological concentration. By assigning weighting factors to the survey units based on radiological classification, the independent verification survey will preferably select survey units which have a higher probability of exceeding the applicable DCGLs. Class 1 survey units (the most likely to be contaminated) are three times more likely to be selected than Class 2 units and six times more likely than Class 3 units. Table 2–1 lists the 49 survey units identified by the Contractor for Building 779 (RMRS 1999b). A simple, commercially available, spreadsheet program with a random number generation feature was used to randomly select the survey units to be independently verified. Appendix A contains a printout of the survey units selected by the computer generated random number method. Survey units 727–01, 727–02, 782–01, and 782–02, (interior and exterior of Buildings 727, 782, and 783) were four of 49 units selected from the Building 779 Cluster.

Table 2-1. Survey Units Identified for Building 779 Cluster

Survey Unit Description	Survey Unit #	Survey Classification	Weight Factor
Rooms 170/172 Floor and Lower Walls plus Mezzanine	779–01	· Class 1	6
Room 171 Vaults	779–02	Class 1	6
Rooms 171/172 Upper Walls and Ceilings	779–03	Class 2	. 3
Stairs, Rooms 270/272, and the Room 170 Dumbwaiter	779–04	Class 2	3
Exterior South/West Walls and Roof	779–05	Class 2	3
Exterior West Wall outside of Room 142, and Roof	779–06	Class 2	3
Exterior Annex A West/North Wall and Roof	779–07	Class 2	3
Dock Walls and Roof	779–08	Class 2	3
Exterior Walls and Roof of Admin Building, Duct Tower	779–09	Class 2	3
2 <sup>nd</sup> Floor Admin Building – Rooms 201 through 214	779–15	Class 3	11
Floor in Room 208	779–16	Class 1	6
1 <sup>st</sup> Floor Admin Building – Rooms 105 through 113	779–17	Class 3	1
Rooms 100, 101, 101A, 104, 116, 116A, 116B, 117	779–18	Class 3	1
Dock and Ramp	779–19	Class 3	1
Rooms 114, 115, 115A, Exhaust Duct Tower	779–20	Class 2	3
Rooms 143, 144, 145, 146, 147, 148, 151	779–21	Class 1	6
Room 150	779–22	Class 1	6
Room 152	779–23	Class 1	6
Rooms 154, 156	779–24	Class 1	6
Rooms 160, 160A	779–25	Class 1	6
Rooms 153, 153A, 153B, 155	779–26	Class 1	,6
Rooms 157, 159	779–27	Class 1	6
Rooms 161, 163, 163A, 167, 167A	779–28	Class 2	3
Rooms 162, 164, 165, 166	779–29	Class 2	3
Hallway to Annex A , Bridge to B777	. 779–30	Class 2	3
Room 217	779–32	Class 1	6
Rooms 219, 221, 221A, 221B, 221C, 223, 225, 229, 230, 231, 232, 233, 235, 271, 273, 274, 275, 277	779–33	Class 2	3
Rooms 215, 218, 220, 224	779–34	Class 1	6
Rooms 222, 222A	779–35	Class 1	6
Rooms 216, 226	779–36	Class 1	6
Room 228	779–37	Class 1	6
Rooms 234, 234A, 234B	779–38	Class 1	6
Rooms 103, 103A, 103B, 118, 120, 121, 121A, 121B, 173	779–39	Class 2	3
Rooms 122, 123, 126, 127	779 <del>-4</del> 0	Class 2	3
Rooms 142,142 Mezzanine	779 <del>-4</del> 1	Class 2	3
Rooms 119, 124, 125, 128, 129, 132, 134, 135, 136, 138	779-42	Class 2	3
Room 001 and Pits	779-43	Class 1	6
Rooms 130, 131, 133	779-44	Class 1	6
Room 137	779-45	Class 1	6
Rooms 139, 140, 140A, 140B	779-46	Class 1	6
Rooms 141, 141A, 141B, 141C	779-47	Class 1	6
B782 Exterior Wall and Roof	782-01	Class 2	3
B782 Plenum Area	782-02	Class 2	3
B782 Basement and Tunnel	782–03	Class 2	3
B727 and B783 Interior	727-01	Class 3	1
B727 and B783 Exterior	727–02	Class 2	3

# 2.3 Field Identification of Sample Locations

Once the survey units to be verified had been selected, the proposed location of each measurement and sample was laid out using the sample allocation protocol specified in the SAP. Drawings of the survey units were created with the walls and ceiling "unfolded" and set flat to assist the process of spatial distribution and sample location recording. The survey units were then divided by a 2-meter sampling grid superimposed over the surface of the survey units. Figure 2–1 shows the sample grid layout for survey units 727–01, 727–02, 782–01, and 782–02.

The selection and distribution of sampling locations within each survey unit was made using the protocol approved in the SAP including:

- Random selection of the sampling starting point within the selected survey unit(s) (Appendix A),
- Systematic distribution of sample locations within the selected survey unit(s) to ensure representative spatial coverage of the survey unit, and
- Personnel safety during the execution of the sampling plan

Drawings of each surface within the survey unit and actual sample locations as determined in the field are shown on Figure 2–2. After the sample locations were allocated, an inspection of each survey unit was conducted to ensure that each sample location selected could be accessed and sampled safely. Selected sample locations that were inaccessible or presented safety hazards were relocated in accordance with the sample relocation protocol outlined in the SAP due to safety issues. The samples relocated are annotated on the drawings in Figure 2–2.

Sample locations were next laid out on the building surfaces within the survey units. Each sample location was measured out and marked on the surface with a permanent marker. Unique alpha-numeric bar codes were affixed to the surface adjacent to the selected sample location. A duplicate bar code was affixed to the field data sheet and the bar code number was recorded on a copy of the survey unit drawings (Figure 2–3). It should be noted that all sample locations were selected without prior knowledge of contaminant concentrations in the area and before radiological survey instruments were employed. In this way, sample locations were not biased.




Figure 2-1. Sampling Grid—Survey Unit 727-01




Figure 2-1 (continued). Sampling Grid—Survey Unit 727-02




Figure 2-1 (continued). Sampling Grid—Survey Unit 782-01




Figure 2–1 (continued). Sampling Grid—Survey Unit 782–01




Figure 2-1 (continued). Sampling Grid—Survey Unit 782-02




Figure 2-1 (continued). Sampling Grid—Survey Unit 782-02

# INDEPENDENT VERIFICATION FOR THE BUILDING 779 CLUSTER SURVEY SURVEY UNIT SAMPLE PLAN

Building: 727 & 783

Survey Unit: 727-01

Number of Sample Locations: 29

Classification: 3

Survey Unit Description: Interior

Grid Size: 2m x 2m

#### SURVEY UNIT 727-01 MAP 1 OF 1




Figure 2-2. Selected Sample Locations—Survey Unit 727-01

# INDEPENDENT VERIFICATION FOR THE BUILDING 779 CLUSTER SURVEY SURVEY UNIT SAMPLE PLAN

Building: 727 & 783

Survey Unit: 727-02


Number of Sample Locations: 29

Classification: 3

Survey Unit Description: Exterior

Grid Size: 2m x 2m

#### SURVEY UNIT 727-02 MAP 1 OF 1








Figure 2–2 (Continued). Selected Sample Locations—Survey Unit 727–02

# INDEPENDENT VERIFICATION FOR THE BUILDING 779 CLUSTER SURVEY SURVEY UNIT SAMPLE PLAN

Building: 782

Survey Unit: 782-01

Number of Sample Locations: 29

Grid Size: 2m x 2m

Classification: 2

Survey Unit Description: Exterior Walls and Roof

## SURVEY UNIT 782-01 MAP 1 OF 2






Figure 2–2 (Continued). Selected Sample Locations—Survey Unit 782–01

# INDEPENDENT VERIFICATION FOR THE BUILDING 779 CLUSTER

SURVEY SURVEY UNIT SAMPLE PLAN

Building: 782

Survey Unit: 782-01

Number of Sample Locations: 29

Grid Size: 2m x 2m

Classification: 1 & 2 Survey Unit Description: Exterior Walls and Roof

SURVEY UNIT 782-01 MAP 2 OF 2






Figure 2-2 (Continued). Selected Sample Locations—Survey Unit 782-01

# INDEPENDENT VERIFICATION FOR THE BUILDING 779 CLUSTER SURVEY SURVEY UNIT SAMPLE PLAN

Building: 782

Survey Unit: 782-02

Number of Sample Locations: 29

Grid Size: 2m x 2m

Classification: 2

Survey Unit Description: Plenum Area




Figure 2–2 (Continued). Selected Sample Locations—Survey Unit 782–02

# INDEPENDENT VERIFICATION FOR THE BUILDING 779 CLUSTER SURVEY SURVEY UNIT SAMPLE PLAN

Building: 782

Survey Unit: 782-02

Number of Sample Locations: 29

Grid Size: 2m x 2m

Classification: 2

Survey Unit Description: Plenum Area

SURVEY UNIT 782-02 MAP 2 OF 2






Figure 2–2 (Continued). Selected Sample Locations—Survey Unit 782–02




Figure 2-3. Sample Location Identifier

## 2.4 Sampling Equipment and Procedures

#### 2.4.1 Radiological Instrumentation

The field measurement instrument used for measuring surface deposited radiological contamination was the Eberline, E-600 Smart Portable Multi-purpose Radiation Survey Instrument with a modified Eberline HP-100 gas proportional detector probe. The detector was fitted with an Eberline "Smart Pack" to convert the conventional detector to be compatible with the microprocessor based E-600 and to electronically store the probe's calibration data. The probe's alpha channel was calibrated to a plutonium-239 (Pu-239) National Institute of Standards and Technology (NIST) traceable calibration source. The calibration certificate for the source is provided in Appendix B and the calibration data sheets for the instruments are provided in Appendix G. Figure 2–4 shows the configuration used to measure the alpha surface emission activity on the surfaces in the survey unit. The direct measurement data was collected in accordance with the procedure outlined in the SAP.




Figure 2-4. Direct Static Surface Contamination Measurement Configuration

### 2.4.2 Sampling

#### 2.4.2.1 Smear Sampling

Smears were used to wipe the surfaces in order to measure the potential for removable radioactivity on the survey unit surfaces. The IVC chose to use 47 millimeter (mm) disc smears made of a duck cloth material rather than the typical paper or cellulose fiber filters commonly used since many of the surfaces requiring measurement are very rough. The duck cloth smears are very durable and will pick-up loose contaminants from even very rough or abrasive surfaces without disintegrating. The smear samples were collected after the direct static surface measurements were obtained. The technician wiped the surface within the 100 cm<sup>2</sup> sample area applying moderate pressure. Each smear was placed individually into a glassine envelope to prevent cross contamination and static charge induced migration of contaminants. Each glassine envelope containing a smear sample was then over-packed in a small sealable plastic bag and then in a manila sample envelope. The envelope was then marked with a bar code label linking it to the sample location from which it was obtained, and entered into a sample custody system to preserve sample integrity for subsequent analysis at the Grand Junction Office (GJO) Analytical Laboratory. The smear samples were secured in a sample box sealed with tamper-evident custody seals at the sample site until the field sampling was complete and then transported to the GJO Analytical Laboratory.

In all, 29 smear samples were collected from each survey unit—one at each of the 29 survey locations. These were submitted to the GJO Analytical Laboratory for radiological analysis. The results of these measurements are summarized in Section 3 and the laboratory analytical reports are contained in Appendix C.

### 2.4.2.2 Surface Media Sampling

Paint and other surface coatings or residues present on the surfaces of the buildings may present an obstruction to detection and measurement of the radioactive surface contamination that might be present. To assess the potential for, and measure the concentration of contaminants which might be present in and/or beneath painted or coated surfaces, a "veneer" of the surface (including any surface coating or residue) is removed from those sample locations that are painted or otherwise coated. When there is no surface coating or residue present, but the radiological measurement of the surface exceeds the *a priori* estimate of the critical detection level of 22 dpm/100 cm², a veneer of the substrate is collected to assess the potential for a near-surface contamination layer embedded in a porous surface (DOE 1999a). A heavy duty, rotary impact drill fitted with a special bit designed to pulverize the surface without drilling into it was used to obtain the media samples (Figure 2–5). The bit was inserted through a port in the containment. Only the bit penetrated the containment. The impact tool was moved over the surface removing the thinnest possible layer until all surface coating within the 100 cm² sample area was removed.




Figure 2-5. Photo of Rotary Impact Drill and Bit

The surface material removed (the sample) was collected in the bottom of the plastic containment. The technician collected the removed media as a sample. The sample was then transferred to a small sealable plastic vial. A bar code label linking the sample location from which it was obtained was affixed to the vial and entered into a sample custody system to preserve sample integrity for subsequent analysis at the GJO Analytical Laboratory (see Figure 2–6). The vials were placed in a sealable plastic bag and secured in a sample box sealed with tamper-evident custody seals at the sample site until the field sampling was complete and then transported to the analytical laboratory.

#### 2.4.3 Laboratory Measurements

Smears and surface media samples were processed and analyzed at the GJO Analytical Laboratory using the methods and procedures identified in Tables 2–2 and 2–3 and prescribed in the IV SAP.

Table 2-2. Smear Sample Analytical Method

Laboratory Method—Gross Alpha Radioactivity					
Counting method	Counting method Gas Proportional Low-Background Alpha/Beta Counting System				
Instrumentation	Instrumentation Canberra Model 2404				
Procedure(s)	Procedure RC-8, "Gross Alpha/Beta Analysis" (WASTREN-GJ)				
Laboratory procedures are governed by QA/QC procedures specified in Handbook of Analytical and Sample-Preparation Procedures (WASTREN-GJ).					




Figure 2-6. Surface Media Sample Collection

Table 2-3. Surface Media Sample Analytical Method

Laboratory Method—Alpha Radioactivity by Isotope Specific Species				
Counting method				
Instrumentation	Instrumentation 1" PIPS with Canberra Alpha Management Software (AMS) Model 48-0721, Ver. 1.0			
Procedure(s)  Solids Digestion, Chemical Extraction, Sample Precipitation, and Sample Counting Procedure RC–19, "Alpha Spectrometry" (WASTREN–GJ).				
Laboratory procedures are governed by QA/QC procedures specified in Handbook of Analytical and Sample-Preparation Procedures (WASTREN-GJ).				

Based on the EPA's terminology, the methods described in Tables 2–2 and 2–3 are categorized as Analytical Level V because they are non-conventional in the EPA's Contract Laboratory Program (CLP) (EPA 1988). However, comparing the level of quality assurance and quality control (QA/QC) embodied in these procedures, they are comparable to EPA's CLP Analytical Level IV.

#### 2.4.4 Field Measurements

### 2.4.4.1 Background Determination

Background was determined in the survey unit being surveyed before, and at least every 2 hours during, each sampling shift. Additionally, background measurements were collected immediately prior to, and immediately after, changing out a detector probe. In the center of the survey unit, a masonite hardboard surface (the back of a clipboard) that has essentially no inherent alpha radioactivity and which was clearly "unaffected" (i.e., clearly not part of the potentially contaminated buildings within the 779 Cluster) was used for establishing background according

to the procedure detailed in the SAP. This method establishes the background associated with the instrument only. Background measurements were recorded both electronically and by hand.

### 2.4.4.2 Direct Static Surface Radioactivity Measurements

All 29 sample locations identified within each of the survey units were directly measured to assess the alpha radioactivity deposited on the surface. Direct static field measurements were made using the approved procedure in the IV SAP. Each measurement was collected for 90 seconds, in the instrument's "SCALER" operating mode, and at a fixed distance of approximately 0.125 inch (1/8th inch) from the surface. When the acquisition count time was complete, the result was read, manually recorded, and electronically logged into the instrument's memory. In cases where surface media samples were taken, a second direct static measurement was made at the same location following removal of the surface veneer. Often, the direct measurement readings obtained subsequent to the veneer removal was unchanged or greater indicating the likelihood that the substrate material (typically concrete) contained an appreciable and measurable alpha background radioactivity that was attenuated by the veneer. In these circumstances, a field decision was made as to the need for the collection of additional surface media to determine compliance with the DCGLs. Pertinent observations regarding the nature of the surface, substrate material, or instrument response were recorded. No anomalies were noted during the direct static measurement process.

<sup>&</sup>lt;sup>1</sup>To avoid the need for making reference survey unit measurements to characterize and quantify natural radioactivity, background has been narrowly defined in the Contractor's Closeout Radiological Survey Plan to include only radiation measured by the instrument system operating in "free air". This definition excludes radioactivity which might be present in the building materials but which has not been contributed or added by DOE. All naturally occurring radioactivity measured during Final Status Survey is to be considered "contributed" or attributable to DOE activities and compared to the applicable DCGLs.

# 3.0 Sampling and Survey Results

Sampling and survey results are divided into four basic categories for discussion, analysis, and comparison with the applicable DCGLs. The categories correspond to the three fundamental samples or measurements employed in the independent verification: Smear sampling, Direct Static Measurements, and Surface Media Samples. The fourth category is for QC data. Quality Control data is presented in Section 6.0 of this report.

#### 3.1 Direct Static Field Measurements

Direct measurements of the radioactivity emission from surfaces were made using static, 90-second counting intervals, over which the total counts were integrated. The measurements recorded were gross values normalized to dpm/100 cm<sup>2</sup>. In the context of this sampling evolution, a "gross measurement" means a measurement made with a radiation detection instrument to which no background correction has been applied. Raw or gross data is important when measurements will be used to make statistical inferences, since not all data will necessarily have the same correction factors applied to properly reduce them to meaningful numbers. Reporting gross or raw data also permits one to analyze the functionality of the instrument with which the measurement was made, and to verify the appropriateness of the data reduction process. The data reduction process for the field measurement data collected in this sampling evolution involves corrections for the efficiency of the radiation detector to the subject radiation and the instrument response to background sources of radiation (excluding surface media contribution to background).

The use of the Eberline E-600 Smart Portable Multi-purpose Radiation Survey Instrument in this application provides a platform for accommodating the probe specific factors including efficiency, high voltage, discrimination thresholds, crossover correction factors, and calibration set up parameters within the detector's associated "smart pack" microchip. These correction factors are common to all of the direct field measurements made with the E-600 and HP-100 detector for this SAP. As a result of incorporating these factors, the instrument reads out and electronically logs data points directly in units of dpm/100 cm<sup>2</sup>. These readings were not, however, field corrected for background radiation.

#### 3.1.1 Background Measurements

The assessment of an instrument's response to background radiation is important from two perspectives. First, it permits the assessment of the minimum sensitivity (detection limit) for the instrument and measurement process in the presence of background radiation. The *a posteriori* minimum detectable activity (MDA) is calculated from this actual background data. Second, by assessing the instrument's response to background radiation in terms of the units that field data will be collected, a correction can be applied to the field measurement data to permit determination of radioactivity present in excess of background. Because the naturally occurring concentrations of background radioactivity in building materials used in the construction of the buildings in the 779 Cluster were expected to be below and well within the DCGL benchmarks for radioactive contamination on building surfaces, the Contractor chose to assign all building material background radioactivity as part of the DOE contributed activity for comparison against the DCGL. As a result, no attempt was made to measure the concentrations of naturally occurring radioactivity measurable on surfaces in a "reference survey unit" or unaffected area. Still, there was the need to measure and account for the instrument's response to other sources of

background radiation (e.g., cosmic radiation) which could otherwise not be distinguished from the contaminant of concern.

To accommodate the need for correcting the instrument data for sensitivity to background radiation, excluding that present in the substrate of the surfaces being measured, instrument background measurements were collected periodically over the sampling period. In all, 48 measurements of the alpha background radiation level were recorded over the sampling period in accordance with the procedure for determining background (DOE 1999a). Each background measurement made during the sampling period is presented in Table 3–1.

Table 3-1. Direct Static Measurement Background Data,	Buildings 727.	782, and 783
, ablo o 1. biloot otatio moadalomont baongroana bata,		, o., a., a , o o

Samula	Survey	Unit 727-01	Survey	Unit 727-02	Survey	Unit 782-01	Survey	Unit 782–02
Sample Location	Date	Value (dpm/100 cm²)	Date	Value (dpm/100 cm²)	Date	Value (dpm/100 cm²)	Date	Value (dpm/100 cm²)
BACKGROUND	2/2/00	7.93	1/25/00	15,10	2/3/00	1.30	2/2/00	4.62
BACKGROUND	2/2/00	11.30	1/25/00	8.45	2/3/00	1.13	2/2/00	14.70
BACKGROUND	2/2/00	7.75	1/25/00	18.70	2/3/00	4.80	2/2/00	7.95
BACKGROUND	2/2/00	8.24	1/25/00	14.90	2/3/00	4.84	2/2/00	8.42
BACKGROUND	2/2/00	8.32	1/25/00	15.60	2/3/00	4.70	2/2/00	22.20
BACKGROUND	2/2/00	4.76	1/25/00	22.50	2/3/00	4.61	2/2/00	22.10
BACKGROUND	2/3/00	4.76	1/25/00	12.30	2/3/00	1.19	2/2/00	14.30
BACKGROUND	2/3/00	4.46	1/25/00	10.30	2/3/00	14.70	2/2/00	7.67
BACKGROUND	2/3/00	4.42	1/25/00	15.10	2/3/00	7.93	2/2/00	28.10
BACKGROUND	2/3/00	4.73			2/3/00	15.80	2/2/00	7.87
BACKGROUND	2/3/00	4.99			2/3/00	12.30	2/2/00	8.14
BACKGROUND	2/3/00	7.56			2/3/00	15.70	2/2/00	4.61
BACKGROUND							2/2/00	11.50
BACKGROUND							2/2/00	8.32
BACKGROUND						2	2/2/00	4.94


From the measurements presented in Table 3–1, it was determined that background did not change appreciably over the duration of each sampling period. When the direct static measurement background data is analyzed both graphically and with goodness-of-fit tests (Figure 3–1), it shows that the measurements are better represented by, or fit to, a log-normal distribution. This is the expected condition for instrument response to alpha background radiation due to the naturally low alpha background count rate and the Poisson distribution associated with low-level radiation counting. The direct static field measurements collected in the survey units also fit a log-normal distribution (see Section 3.1.2). The variance in the recorded background data was small and within the range expected for a gas proportional counter measuring alpha background radiation (see Appendix E for complete background data set).


#### **DATA EVALUATION STATISTICS**

#### **Data Description**

"Free Air" Instrument Background Building 779 Cluster Independent Verification Project Buildings 727 & 783, Survey Unit 727-01

Critical Level	22	
UNITS - dpm/100 cm <sup>2</sup>		
Sample Data		
4.42	Descriptive Statistics	
4.46	Number of Samples	12.000
4.73	Mean	6.602
4.76	Median	6.275
4.76	Standard Deviation	2.214
4.99	cv	0.335443
7.56	Range	6.880
7.75	Minimum	4.420
7.93	Maximum	11.300
8.24	GM	6.284
8.32	GSD	1.383
11.30	Mean of LN(Data)	1.838
	SD of LN(Data)	0.325
	Percent > Critical Level	0.000
	Normal Statistics	
	Mean	6.602
	UCL(Mean) - Z	7.855
	LCL(Mean) - Z	5.349
	95%ile - Z	10.244
	Percent > Critical Level	0.000
	W Test (Data)	0.83973
	Normal (a=0.05)?	No
	Lognormal Statistics GM	6.284
	GSD	1,383
	AM of data	6.602
	AM - MVUE	6,594
	AM - MLE	6,624
	UCL - Norm t stats	8.009
	LCL - Norm t stats	5.195
	UCL LogNorm t	8.141
	LCL LogNorm t	5,390
	UCL - Modified Cox	8,151
	LCL - Modified Cox UCL - "Exact"	5.334
	LCL - "Exact"	
	95%ile	10.717
	UTL 95%, 95%	15.270
	Percent > Critical Level	0.006
1	PEP (Upper)	0.353
	PEP (Lower)	1.23E-07
`	W Test (In Data)	0.84612
	Lognorm (a=0.05)?	No









Figure 3-1. Data Evaluation Statistics-"Free Air" Instrument Background, 727-01


#### DATA EVALUATION STATISTICS

#### **Data Description**

"Free Air" Instrument Background Building 779 Cluster Independent Verification Project Buildings 727 & 783, Survey Unit 727-02

Critical Level	22	
UNITS - dpm/100 cm <sup>2</sup>		
Sample Data		
8.45	Descriptive Statistics	
10.30	Number of Samples	9.000
12.30	Mean	14.772
14.90	Median	15.100
15.10	Standard Deviation	4.215
15.10	cv	0.285349
15.60	Range	14.050
18.70	Minimum	8.450
22.50	Maximum	22.500
	GM	14.227
	GSD	1.344
,	Mean of LN(Data)	2.655
	SD of LN(Data)	0.296
	Percent > Critical Level	11.111
	Normal Statistics	
	Mean	14.772
	UCL(Mean) - Z	17.526
	LCL(Mean) - Z	12.018
	95%ile - Z	21.706
	Percent > Critical Level	4.320
	W Test (Data)	0.957868
	Normal (a=0.05)?	Yes
	Lognormal Statistics	
	GM	14.227
	GSD	1.344
	AM of data	14.772
	AM - MVUE	14.788
	AM - MLE	14.862
	UCL - Norm t stats	18.012
	LCL - Norm t stats	11.532
	UCL LogNorm t	18.655
	LCL LogNorm t	11.841
	UCL - Modified Cox	18.665
	LCL - Modified Cox	11.717
	UCL - "Exact"	
	LCL - "Exact"	90 400
	95%ile	23.139
	UTL 95%, 95%	34.861
	Percent > Critical Level	7.020 21.604
	PEP (Upper) PEP (Lower)	0.349277
	W Test (In Data)	0.957372
	vv rest (in Data) Lognorm (a≈0.05)?	0.95/3/2 Yes
	Eognotti (a-0.00)1	140










Figure 3–1 (continued) Data Evaluation Statistics—"Free Air" Instrument Background, 727–02

# **Data Description**

"Free Air" Instrument Background Building 779 Cluster Independent Verification Project Building 782, Survey Unit 782-01

Critical Level	22
	22
UNITS - dpm/100 cm <sup>2</sup>	
Sample Data	<u> </u>
1.13	Descriptive Statistics
1.19	Number of Samples 12.000
1.30	Mean 7.417
4.61	Median 4.820
4.70	Standard Deviation 5.725
4.80	CV 0.771878
4.84	Range 14.670
7.93	Minimum 1.130
12.30	Maximum 15.800
14.70	GM 5.105
15.70	GSD 2.708
15.80	Mean of LN(Data) 1.630
	SD of LN(Data) 0.996
	Percent > Critical Level 0.000
	Normal Statistics
	Mean 7.417
'	UCL(Mean) - Z 10.656
	LCL(Mean) - Z 4.178
	• •
	95%ile - Z 16.834 Percent > Critical Level 0.543
	Normal (a=0.05)? No
	Lognormal Statistics
	GM 5.105
	GSD 2.708
	AM of data 7.417
	AM - MVUE 7.926
	AM - MLE 8.383
	UCL - Norm t stats 11.054
	LCL - Norm t stats 3.779
	UCL LogNorm t 15.786
	LCL LogNorm t 4.452
	UCL - Modified Cox 17.389
	LCL - Modified Cox 3.613
	UCL - "Exact"
	LCL - "Exact"
	95%ile 26,277
	UTL 95%, 95% 77.900
	Percent > Critical Level 7.124
	PEP (Upper) 19.338
	PEP (Lower) 0.767154
	W Test (In Data) 0.866555
	Lognorm (a=0.05)? Yes











Figure 3–1 (continued) Data Evaluation Statistics—"Free Air" Instrument Background, 782–01

#### **Data Description**

"Free Air" Instrument Background Building 779 Cluster Independent Verification Project Building 782, Survey Unit 782-02










Figure 3-1 (continued) Data Evaluation Statistics--"Free Air" Instrument Background, 782-02

Table 3–2 shows the background data summary statistics.

Table 3-2. Background Data Summary Statistics

Statistic	Survey Unit 727–01	Survey Unit 727–02	Survey Unit 782–01	Survey Unit 782–02
Number of Measurements	12	9	12	15
Arithmetic Mean	6.6	14.8	7.4	11.7
Standard Deviation (sample)	2.2	4.2	5.7	7.2
Coefficient of Variation	0.34	0.29	0.77	0.62
Max	11.3	22.5	15.8	28.1
Median	6.3	15.1	4.8	8.3
Minimum	4.4	8.5	1.1	4.6
Range	6.9	14.1	14.7	23.5
Geometric Mean	6.3	14.2	5.1	10.0
UCL (log-normal "t", a=0.05)	8.1	18.7	15.8	16.2
LCL (log-normal "t", b=0.05)	5.4	11.8	4.5	8.6

### 3.1.1.1 Background Adjustment

Because the background and survey unit sample sets were each log-normally distributed, it was decided that the geometric (or log-normal) mean background value recorded over the sampling period (6.3, 14.2, 5.1, and 10.0 dpm/100 cm², respectively) would be used to correct the gross direct static surface contamination measurements for subsequent comparison to the applicable DCGL. In some situations, a graphical or visual distinction can be made between measurements containing background only and those with added radioactivity. For example, when elevated or contributed activity is present, a graphical distinction can be clearly seen between the population of measurements containing only background response and those containing elevated or contributed activity. This is not typically the case with alpha radioactivity. No graphic distinction between measurement results attributable to background and those with activity in excess of background is clearly discernable with this data set.

Calculational methods are needed to assess the surface activity above background that could be distinguished with statistical significance from background. As discussed earlier, the geometric mean instrument background measurements over the field sampling period were 6.3, 14.2, 5.1, and 10.0 dpm/100 cm², respectively. With the E-600 instruments used, the background radiation influence on the instrument's readings was processed with efficiency corrections and probe size corrections such that background measurements and sample measurements alike read out in units of dpm/100 cm². In order to calculate the statistically significant surface activity, which could be distinguished from background (a posteriori MDA), it was necessary to convert the background measurement units from dpm/100 cm² to units of counts per minute (cpm). In this case, the more conservative metric, the arithmetic mean, was chosen to calculate the detection sensitivity achieved to prevent overstating the actual sensitivity achieved. The converted mean background count rates for the sampling periods are 1.29, 2.88, 1.44, and 2.28 cpm, respectively (Table 3–3). Using the actual instrument field measurement parameters, a calculation of the actual field measurement MDA can be determined by solving Equation 3–1.

Table 3-3. Static Surface Contamination Measurement MDA Parameters

	Parameter		Value	Used		Remarks
	raiailielei	779–21	779–21	779–23	779–21	Kemarks
Сь	Background Counts	1.93	4.33	2.16	3.42	Values used are 6.6, 14.8, 7.4, and 11.7 dpm/100 cm <sup>2</sup> converted to units of counts (cpm × T <sub>s</sub> )
Ts	Sample count time in minutes	1.5	1.5	1.5	1.5	Count time programmed into the calibrated instrument specifically for this sampling event
Ap	Probe size	100	100	100	100	cm <sup>2</sup>
€⊤	Instrument system efficiency in counts per disintegration	0.1949	0.1949	0.1949	0.1949	Actual efficiency for the individual probe is programmed into the memory chip of the probes' smart pack and for the probe used was 19.49%.

The following calculations define the a posteriori MDA.

$$MDA = \frac{3 + 4.65\sqrt{C_b}}{T_s \times \frac{A_p}{100 \text{cm}^2} \times \epsilon_T}$$

$$(3-1)$$

Where: MDA = the minimum surface activity concentration above background radioactivity (in dpm/100 cm<sup>2</sup>) that can be detected with 95 percent confidence.

 $C_b$  = the total number of background counts over the sample count period  $(T_s)$ .

 $T_s$  = sample count time (in minutes).

 $A_P$  = probe size (in cm<sup>2</sup>).

 $\in_T$  = counting system efficiency in count/disintegration.

# Survey Unit 727-01

# Survey Unit 727-02

$$MDA = \frac{3 + 4.65\sqrt{1.93}}{1.5 \times 1 \times 0.1949}$$

$$MDA = \frac{3 + 4.65\sqrt{4.33}}{1.5 \times 1 \times 0.1949}$$
 (3-2)

# Survey Unit 782-01

### Survey Unit 782-02

$$MDA = \frac{3 + 4.65\sqrt{2.16}}{1.5 \times 1 \times 0.1949}$$

$$MDA = \frac{3 + 4.65\sqrt{3.42}}{1.5 \times 1 \times 0.1949}$$

# Survey Unit 727-01

# Survey Unit 727-02

$$MDA = \frac{9.46}{0.2924} = 32 \text{ dpm/100 cm}^2$$

$$MDA = \frac{12.68}{0.2924} = 43 \text{ dpm/}100 \text{ cm}^2$$
 (3-3)

## Survey Unit 782-01

## Survey Unit 782-02

MDA = 
$$\frac{9.83}{0.2924}$$
 = 34 dpm/100 cm<sup>2</sup> MDA =  $\frac{11.60}{0.2924}$  = 40 dpm/100 cm<sup>2</sup>

$$MDA = \frac{11.60}{0.2924} = 40 \text{ dpm/}100 \text{ cm}^2$$

Therefore the "gross" field instrument readings, using the procedures identified in the Building 779 Cluster IV SAP, which can be distinguished as different from background (the adjusted gross MDA) are:

Survey Unit 727-01

Survey Unit 727-02

$$7 + 32 = 39 \text{ dpm}/100 \text{ cm}^2$$

$$15 + 43 = 58 \text{ dpm}/100 \text{ cm}^2$$

(3-4)

Survey Unit 782-01

Survey Unit 782-02

 $7 + 33 = 40 \text{ dpm}/100 \text{ cm}^2$ 

 $12 + 40 = 52 \text{ dpm}/100 \text{ cm}^2$ 

Having identified the a posteriori MDA for the field sampling measurements and the adjusted gross MDAs, a simple sort of the gross field measurement data points was performed to identify those measurements from survey units 727-01, 727-02, 782-01, and 782-02 which were greater than 39 dpm/100 cm<sup>2</sup>, 58 dpm/100 cm<sup>2</sup>, 33 dpm/100 cm<sup>2</sup>, and 52 dpm/100 cm<sup>2</sup>, respectively. Those locations with gross surface activity greater than the adjusted gross MDA are credited as positive indicators of added radioactivity, while those less than the adjusted gross MDA are statistically indistinguishable from background values.

Rather than correct each individual measurement for background, the gross measurement data set was statistically analyzed. The data set was treated as log-normally distributed, the best fit for the data set collected. This treatment conforms to standard EPA methodology for data evaluation statistics, and generally yields conservative estimates of the upper confidence intervals and percentiles values. To correct for the instrument's response to background, the geometric mean background, 6.3, 14.2, 5.1, and 10.0 dpm/100 cm<sup>2</sup>, respectively, was subtracted from the geometric mean of the total surface activity measured by surface emission data set of interest. When comparisons of other metrics (e.g., the median) are provided for information, the comparable background metric is also used to correct the reading for background radiation influence. For example, when the net (background corrected) median direct static surface contamination metric is reported, the median value of the background data set has been subtracted from the median value of the gross direct static surface contamination measurement data set.

### 3.1.2 Field Measurement Data

Direct static measurements were made at the 29 selected sample locations in each of the four survey units (727-01, 727-02, 782-01, and 782-02). Figure 2-2 shows the layout of both survey units and the sample locations selected in accordance with the sample allocation protocol identified in the IV SAP (DOE 1999a). These measurements were made prior to the collection of a smear sample, but subsequent to the collection of media samples. In this way, the "total" surface deposited activity emission rate, whether from fixed or removable radioactivity, was accounted for. Direct static measurements were taken adjacent to the sample locations where media samples had been taken. In all, 33 (34 in 727-02) direct static surface measurements were

made in each survey unit. Four of these, in each survey unit (five for 727–02), were replicate measurements collected as part of the overall QA/QC as described in the SAP. For data reduction purposes, the arithmetic mean of a replicate measurement and the corresponding initial measurement was used as the reported value for a specific sample location at which a replicate measurement was made. Consequently, there are a total of 29 data points (Table 3–4) for each survey unit included in the overall characterization of the building's mean residual surface contamination level as measured by direct surface emission. Further information about the duplicate samples and the assurance of precision and variability is presented in Sections 6.0 and 7.0.

A number of statistical tests of the data were performed to assess the data sets. A key test of the data set is for goodness-of-fit. It is important because it identifies the underlying distribution of the data set and permits the analyst as well as the decision makers and risk managers to compare appropriate metrics calculated from the data. The W-test was used to measure the relative goodness of the fit of the observed data distribution to the normal and log-normal standard distributions. Other distributions were not entertained for this data set since the data were expected to be either normally or log-normally distributed (based on knowledge of radioactivity distribution in the environment and in background) and because the probability plots and histograms generated gave no evidence that other than normal or log-normal distributions might be present. For the direct static measurement data set, the W-test identified the log-normal distribution as the best fit. The data evaluation statistics are provided in Figure 3–2. Table 3–5 summarizes the direct surface measurement data, uncorrected for background, collected in survey units 727–01, 727–02, 782–01, and 782–02.

From Table 3–5 and the data evaluation and summary statistics, it is evident that for survey unit 727–01, approximately 45 percent of all the measurements are below the geometric mean background value of 6.3 dpm/100 cm<sup>2</sup> and approximately 93 percent are below the critical detection level of 22 dpm/100 cm<sup>2</sup>. All measurements are below the adjusted gross minimum detectable activity (MDA<sub>Gross</sub>) of 32 dpm/100 cm<sup>2</sup> for the field measurement process.

For survey unit 727–02, 17 percent of the measurements taken are below the geometric mean background value of 14.2 dpm/100 cm<sup>2</sup> and 76 percent are at or below the critical detection level of 22 dpm/100 cm<sup>2</sup>. All measurements are below the MDA<sub>Gross</sub> of 43 dpm/100 cm<sup>2</sup> for the field measurement process.

For survey unit 782–01, 21 percent of the measurements taken are below the geometric mean background value of 5.1 dpm/100 cm<sup>2</sup> and 83 percent are at or below the critical detection level of 22 dpm/100 cm<sup>2</sup>. Only one measurement (34.5 dpm/100 cm<sup>2</sup> at survey location IVP0000301) exceeded the MDA<sub>Gross</sub> of 34 dpm/100 cm<sup>2</sup> for the field measurement process.

For survey unit 782–02, 38 percent of the measurements taken are below the geometric mean background value of 10.0 dpm/100 cm<sup>2</sup> and all but one are below the critical detection level of 22 dpm/100 cm<sup>2</sup>. All measurements are below the MDA<sub>Gross</sub> of 40 dpm/100 cm<sup>2</sup> for the field measurement process.


Table 3-4. Direct Static Surface Contamination Measurements, Survey Units 727-01, 727-02, 782-01, and 782-02


Survey L	Unit 727-01	Survey Unit 727-02	nit 727-02	Survey U	Survey Unit 782-01	Survey U	Survey Unit 782-02	Instrument		Background
Sample Location	Recorded dpm/100 cm <sup>2</sup>	Sample Location	Recorded dpm/100 cm <sup>2</sup>	Sample Location	Recorded dpm/100 cm <sup>2</sup>	Sample Location	Recorded dpm/100 cm <sup>2</sup>	Operating Mode	Channel Selected	Compensation
IVP0000381	8.94	IVP0000351	12.30	IVP0000301	34.50	IVP0000421	10.40	Scaler	Alpha	Gross
IVP0000382	2.04	IVP0000352	8.82	IVP0000302	31.10	IVP0000422	24.10	Scaler	Alpha	Gross
IVP0000383	5.46	IVP0000353	21.50	IVP0000303	13.80	IVP0000423	20.60	Scaler	Alpha	Gross
IVP0000384	5.05	IVP0000354	20.10	IVP0000304	24.10	IVP0000424	6.94	Scaler	Alpha	Gross
IVP0000385	1.85	IVP0000355	15.70	IVP0000305	17.50	IVP0000425	17.50	Scaler	Alpha	Gross
IVP0000386	9.40	IVP0000356	17.30	IVP0000306	12.25	IVP0000426	14.00	Scaler	Alpha	Gross
IVP0000387	20.90	IVP0000357	21.90	IVP0000307	24.20	IVP0000427	3.96	Scaler	Alpha	Gross
IVP0000388	2.03	IVP0000358	20.50	IVP0000308	24.10	IVP0000428	7.26	Scaler	Alpha	Gross
IVP0000389	1.84	IVP0000359	21.10	IVP0000309	14.10	IVP0000429	7.15	Scaler	Alpha	Gross
IVP0000390	17.70	IVP0000360	20.60	IVP0000310	7.36	IVP0000430	7.23	Scaler	Alpha	Gross
IVP0000391	1.90	IVP0000361	12.40	IVP0000311	10.90	IVP0000431	11.10	Scaler	Alpha	Gross
IVP0000392	8.95	IVP0000362	22.40	IVP0000312	10.95	IVP0000432	9.45	Scaler	Alpha	Gross
IVP0000393	8.87	IVP0000363	19.40	IVP0000313	7.08	IVP0000433	1.10	Scaler	Alpha	Gross
IVP0000394	2.02	IVP0000364	13.40	IVP0000314	10.80	IVP0000434	7.71	Scaler	Alpha	Gross
IVP0000395	1.91	IVP0000365	29.90	IVP0000315	17.50	IVP0000435	21.50	Scaler	Alpha	Gross
IVP0000396	14.40	IVP0000366	19.10	IVP0000316	10.90	IVP0000436	18.30	Scaler	Alpha	Gross
IVP0000397	11.10	IVP0000367	22.00	IVP0000317	3.73	IVP0000437	15.00	Scaler	Alpha	Gross
IVP0000398	4.49	IVP0000368	14.70	IVP0000318	1.20	IVP0000438	18.35	Scaler	Alpha	Gross
IVP0000399	24.70	IVP0000369	24.90	IVP0000319	1.34	IVP0000439	12.00	Scaler	Alpha	Gross
IVP0000340	3.69	IVP0000370	18.90	IVP0000320	8.01	IVP0000440	10.50	Scaler	Alpha	Gross
IVP0000341	7.25	IVP0000371	36.30	IVP0000321	4.66	IVP0000441	3.55	Scaler	Alpha	Gross
IVP0000342	7.50	IVP0000372	29.40	IVP0000322	11.00	IVP0000442	3.44	Scaler	Alpha	Gross
IVP0000343	31.40	IVP0000273	24.70	IVP0000323	11.50	IVP0000443	17.40	Scaler	Alpha	Gross
IVP0000344	10.70	IVP0000374	15.80	IVP0000324	7.79	IVP0000444	19.05	Scaler	Alpha	Gross
IVP0000345	14.60	IVP0000375	15.80	IVP0000325	4.68	IVP0000445	21.00	Scaler	Alpha	Gross
IVP0000346	14.60	IVP0000376	24.80	IVP0000326	18.30	IVP0000446	10.60	Scaler	Alpha	Gross
IVP0000347	4.36	IVP0000377	8.82	IVP0000327	14.90	IVP0000447	17.30	Scaler	Alpha	Gross
IVP0000348	17.60	IVP0000378	19.10	IVP0000328	8.20	IVP0000448	4.16	Scaler	Alpha	Gross
IVP0000349	4.08	IVP0000379	19.40	IVP0000329	1.34	IVP0000449	11.20	Scaler	Alpha	Gross

### **Data Description**

Direct Static Surface Measurements Building 779 Cluster Independent Verification Project Building 727 & 783, Survey Unit 727-01

building 121 & 105,	, Survey Offic 121-01
DCGL	100
UNITS - dpm/100 cm <sup>2</sup>	
Sample Data	•
1.84	Descriptive Statistics
1.85	Number of Samples 29.000
1.90	Mean 9.287
1.91	Median 7.500
2.02	Standard Deviation 7.590
2.03	CV 0.817245
2.04	Range 29.560
3.69	Minimum 1.840
4.08	Maximum 31.400
4.36	GM 6.591
4.49	GSD 2.414
5.05	Mean of LN(Data) 1.886 SD of LN(Data) 0.881
5.46 7.25	SD of LN(Data) 0.881 Percent > DCGL 0.000
7.25 7.50	Percent > DCGL 0.000
7.50 8.87	
8.94	Normal Statistics
8.95	Mean 9.287
9.40	UCL(Mean) - Z 12.050
10.70	LCL(Mean) - Z 6.525
11.10	95%ile - Z 21.773
14.40	Percent > DCGL 0.000
14.60	W Test (Data) 0.869142
14.60	Normal (a=0.05)? No
17.60	
17.70	Lognormal Statistics
20.90	GM 6.591
24.70 31.40	GSD 2.414
31.40	AM of data 9,287
	AM - MVUE 9.544
	AM - MLE 9.717
j	UCL - Norm t stats 12.174
	LCL - Norm t stats 6.400
ļ	UCL LogNorm t 13.585
	LCL LogNorm t 6.950
j ·	UCL - Modified Cox 14,193
	LCL - Modified Cox 6.418
	UCL - "Exact"
	LCL - "Exact"
]	
	95%ile 28.081
	UTL 95%, 95% 47.202
	Percent > DCGL 0.101
	PEP (Upper) 0.739
	PEP (Lower) 0.003219
	W Test (In Data) 0.928377
	Lognorm (a=0.05)? Yes










Figure 3–2. Data Evaluation Statistics—Direct Static Surface Measurements, 727–01

#### **Data Description**

Direct Static Surface Measurements
Building 779 Cluster Independent Verification Project
Building 727 & 783, Survey Unit 727-02

DCGL	100
	100
UNITS - dpm/100 cm <sup>2</sup>	
Sample Data 8.82	December Statistics
8.82	Descriptive Statistics Number of Samples 29.000
12.30	Mean 19.691
12.40	Median 19.400
13.40	Standard Deviation 6.092
14.70	CV 0.309383
15.70	Range 27.480
15.80	Minimum 8.820
15.80	Maximum 36.300
17.30	GM 18.754
18.90	GSD 1.386
19.10	Mean of LN(Data) 2.931
19.10 19.40	SD of LN(Data) 0.326  Percent > DCGL 0.000
19.40	Percent > DCGL 0.000
20.10	
20.50	Normal Statistics
20.60	Mean 19.691
21.10	UCL(Mean) - Z 21.908
21.50	LCL(Mean) - Z 17.474
21.90	95%ile - Z 29.713
22.00	Percent > DCGL 0.000
22.40	W Test (Data) 0.964049
24.70	Normal (a=0.05)? Yes
24.80	
24.90	Lognormal Statistics
29.40 29.90	GM 18.754 GSD 1.386
36.30	AM of data 19.691
00.00	AM - MVUE 19.743
	UCL - Norm t stats 22.008
	LCL - Norm t stats 17.374
	UCL LogNorm t 22.397
	LCL LogNorm t 17.471
	UCL - Modified Cox 22.430
	LCL - Modified Cox 17,378
	UCL - "Exact"
	LCL - "Exact"
	95%ile 32.089
	UTL 95%, 95% 38.898
	•
	Percent > DCGL 0.000
	PEP (Upper) 0.000
	PEP (Lower) 0
	W Test (In Data) 0.95569
	Lognorm (a=0.05)? Yes










Figure 3-2 (continued). Data Evaluation Statistics—Direct Static Surface Measurements, 727-02

### **Data Description**

Direct Static Surface Measurements Building 779 Cluster Independent Verification Project Building 782, Survey Unit 782-01

Building 782, Surve	y Unit 782-01
DCGL	100
	100
UNITS - dpm/100 cm <sup>2</sup>	· ·
Sample Data	D
1,20	Descriptive Statistics
1.34 1.34	Number of Samples 29.000 Mean 12.682
3.73	Median 10.950
4.66	Standard Deviation 8.526
4.68	CV 0.672258
7.08	Range 33.300
7.36	Minimum 1.200
7.79	Maximum 34.500
8,01	GM 9.532
8.20	GSD 2.408
10.80	Mean of LN(Data) 2.255
10.90	SD of LN(Data) 0.879
10.90	Percent > DCGL 0.000
10.95	
11.00	Name of Otalian
11.50	Normal Statistics
12.25	Mean 12.682
13.80 14.10	UCL(Mean) - Z 15.786 LCL(Mean) - Z 9.579
14.10 14.90	LCL(Mean) - Z 9.579 95%∦e - Z 26.707
17.50	Percent > DCGL 0.000
17.50	W Test (Data) 0.924105
18,30	Normal (a=0.05)? No
24.10	
24.10	Lognormal Statistics
24.20	GM 9.532
31.10	GSD 2.408
34.50	AM of data 12.682
	AM - MVUE 13.778
	AM - MLE 14.025
	UCL - Norm t stats 15.925
	LCL - Norm t stats 9.439
	UCL LogNorm t 19.593
	-
	LCL LogNorm t 10.040
	UCL - Modified Cox 20.462
	LCL - Modified Cox 9.277
,	UCL - "Exact"
	LCL - "Exact"
	95%ile 40.462
	UTL 95%, 95% 67.921
,	Percent > DCGL 0.374
	PEP (Upper) 1.810
	PEP (Lower) 0.024998
	W Test (In Data) 0.898559
	, ,
	Lognorm (a=0.05)? No
	·
,	










Figure 3–2 (continued). Data Evaluation Statistics—Direct Static Surface Measurements, 782–01

#### **Data Description**

Direct Static Surface Measurements Building 779 Cluster Independent Verification Project Building 782, Survey Unit 782-02

DCGL	100
UNITS - dpm/100 cm <sup>2</sup>	100
Sample Data	·
1.10	Descriptive Statistics
3.44	Number of Samples 29,000
3.55	Mean 12.133
3.96	Median 11.100
4.16	Standard Deviation 6.376
6.94	CV 0.525542
7.15	Range 23,000
7.23	Minimum 1.100
7.26	Maximum 24.100
7.71	GM 10.048
9.45	GSD 2.032
10.40	Mean of LN(Data) 2.307
10.50	SD of LN(Data) 0.709
10.60	Percent > DCGL 0.000
11.10	
11.20	
12.00	Normal Statistics
14.00	Mean 12.133
15.00 17.30	UCL(Mean) - Z 14.453
17.40	LCL(Mean) - Z 9.812 95%ile - Z 22.622
17.50	95%ile - Z 22.622 Percent > DCGL 0.000
18.30	W Test (Data) 0.954488
18.35	Normal (a=0.05)? Yes
19.05	103 (u=0.05)?
20.60	Lognormal Statistics
21.00	GM 10.048
21.50	GSD 2.032
24.10	AM of data 12.133
	AM - MVUE 12,784
	AM - MLE 12.920
	UCL - Norm t stats 14.558
	LCL - Norm t stats 9.707
	UCL LogNorm t 16.921
	LCL LogNorm t 9.866
	UCL - Modified Cox 17.306
	LCL - Modified Cox 9.444
	UCL - "Exact"
	LCL - "Exact"
	95%ile 32.261
	UTL 95%, 95% 48,999
	•
,	Percent > DCGL 0.060
	PEP (Upper) 0.515
	PEP (Lower) 0.001399
	W Test (In Data) 0.894174
	Lognorm (a=0.05)? No








Figure 3-2 (continued). Data Evaluation Statistics-Direct Static Surface Measurements, 782-02

Statistic	Survey Unit 727–01 Value	Survey Unit 727–02 Value	Survey Unit 782–01 Value	Survey Unit 782–02 Value
Number of Measurements	29	29	29	29
Arithmetic Mean	9.29	19.69	12.68	12.13
Standard Deviation (sample)	7.59	6.09	8.53	6.38
Coefficient of Variation	0.82	0.31	0.67	0.53
Max	31.4	36.3	34.5	24.1
Median	7.5	19.4	10.95	11.1
Minimum	1.84	8.82	1.2	1.1
Range	29.56	27.48	33.3	23.0
Geometric Mean	6.59	18.75	9.53	10.05
UCL (log-normal "t", a=0.05)	13.59	22.40	19.59	16.92
LCL (log-normal "t", b=0.05)	6.95	17.47	10.04	9.87

Table 3-5. Summary Statistics, Direct Static Measurements, Buildings 727, 782, and 783

Those measurements that exceeded the critical level detection level of 22 dpm/100 cm<sup>2</sup> give some evidence of activity above background but not enough to be quantified with certainty.

The test of DCGL<sub>w</sub> for total surface contamination concentration as measured by direct surface emission is the mean (the geometric mean) since all data sets are determined to be log-normally distributed. Section 4.0 provides detailed analysis of the data sets in comparison to the applicable DCGL values.

# 3.1.2.1 Post Surface Media Sampling Measurements

Follow-up direct static measurements were made after each surface media sample was collected in an effort to assure that all of the contaminant, which might have been present beneath the immediate surface, was removed by the physical sampling process. This measurement was necessary to validate the assumption that any contaminant that may have been deposited beneath a paint layer or embedded in the porous substrate was limited to only shallow deposition and would be collected and measured by the surface media sampling. Evidence of elevated radioactivity by direct measurement after a thin surface veneer had been removed might call into question the validity of that assumption, requiring further investigation. The results, however, are not considered in the data set used to evaluate compliance with the DCGL<sub>W</sub> for total surface contamination as measured by direct surface emission. Nonetheless, the post surface media sampling measurements are considered important since they might detect radioactivity that is potentially "hidden" from detection by direct surface emission measurements made before removal of the surface coating or veneer.

A total of 69 surface media samples (Table 3–6) were collected from all accessible sample locations in Buildings 727, 782, and 783. A total of 71 direct static surface measurements were made at the 69 surface media sample locations subsequent to collecting surface samples. Two of these were replicate measurements collected as part of the overall QA/QC as described in the SAP. Where the replicate measurements were made, the arithmetic mean of the replicate measurement and the corresponding initial measurement was used as the reported value for the specific sample location.

Table 3-6. Post Surface Media Sample Direct Static Surface Measurements, Survey Units 727-01, 727-02, 782-01, and 782-02

	Survey L	Survey Unit 727-02	Survey U	Survey Unit 782-01	Survey U	Survey Unit 782-02	Instrument	Channel	Background
Sample Location		Recorded dpm/100 cm <sup>2</sup>	Sample Location	Recorded dpm/100 cm <sup>2</sup>	Sample Location	Recorded dpm/100 cm²	Operating Mode	Selected	Compensation Mode
IVP0000367		20.00	IVP0000301	27.60	IVP0000421	17.30	Scaler	Alpha	Gross
IVP0000368		15.10	IVP0000302	24.00	IVP0000422	3.50	Scaler	Alpha	Gross
IVP0000369		25.50	IVP0000303	20.70	IVP0000423	27.30	Scaler	Alpha	Gross
IVP0000370		16.30	IVP0000304	41.30	IVP0000424	31.00	Scaler	Alpha	Gross
IVP0000371		29.50	IVP0000305	13.90	IVP0000426	17.30	Scaler	Alpha	Gross
IVP0000372		15.70	IVP0000306	20.90	IVP0000427	5.53	Scaler	Alpha	Gross
IVP0000373		14.70	IVP0000307	24.30	IVP0000430	17.20	Scaler	Alpha	Gross
IVP0000374		22.60	IVP0000308	41.90	IVP0000431	17.80	Scaler	Alpha	Gross
IVP0000375		9.03	IVP0000309	20.90	IVP0000432	2.63	Scaler	Alpha	Gross
IVP0000376		22.00	IVP0000310	17.50	IVP0000433	14.70	Scaler	Alpha	Gross
IVP0000377		8.80	IVP0000311	21.10	IVP0000434	17.90	Scaler	Alpha	Gross
IVP0000378		15.80	IVP0000312	21.30	IVP0000435	18.15	Scaler	Alpha	Gross
IVP0000379		12.30	IVP0000313	27.90	IVP0000436	18.20	Scaler	Alpha	Gross
	**.5		IVP0000314	14.20	IVP0000437	8.09	Scaler	Alpha	Gross
	*		IVP0000315	14.10	IVP0000438	4.79	Scaler	Alpha	Gross
			IVP0000316	7.49	IVP0000440	3.55	Scaler	Alpha	Gross
	33				IVP0000441	20.80	Scaler	Ałpha	Gross
					IVP0000442	10.60	Scaler	Alpha	Gross
					IVP0000443	13.40	Scaler	Alpha	Gross
	100	ACRES TO THE STATE OF THE			IVP0000444	4.67	Scaler	Alpha	Gross
	1 1			1945 Jane	IVP0000445	7.13	Scaler	Alpha	Gross
of Marchine forth					IVP0000446	7.23	Scaler	Alpha	Gross
	42.5				IVP0000447	3.83	Scaler	Alpha	Gross

Again, a number of statistical tests of the data were performed to assess the data set. The W-test was used to measure the relative goodness of the fit of the observed data distribution. The W-test and histogram showed the survey units 727–01, 727–02, 782–01, and 782–02 data sets to be lognormally distributed. The data evaluation statistics are provided in Figure 3–3. Table 3–7 summarizes the post surface media sampling direct surface measurement data, uncorrected for background, collected in these survey units.

Statistic	Survey Unit 727–01	Survey Unit 727–02	Survey Unit 782–01	Survey Unit 782–02
Number of Measurements	17	13	16	23
Arithmetic Mean	15.0	17.5	22.4	12.9
Standard Deviation (sample)	7.8	6.2	9.2	7.8
Coefficient of Variation	0.52	0.35	0.41	0.60
Max	31.5	29.5	41.9	31.0
Median	15.8	15.8	21.0	13.4
Minimum	4.2	8.8	7.5	3.5
Range	27.3	20.7	34.4	27.5
Geometric Mean	12.9	16.5	20.7	10.6
UCL (normal "t", a=0.05)	21.1	22.0	28.4	17.9
LCL (normal "t", b=0.05)	11.4	14.1	18.1	9.9

The most telling presentation of the post surface media sampling surface measurements is a side by side comparison of the data set summary statistics with the summary statistics from the direct surface measurements made prior to sampling and the instrument background data collected during the sampling process. Tables 3–8, 3–9, 3–10, and 3–11 separately compare units 727–01, 727–02, 782–01, and 782–02 summary statistics from each of these three data sets.

Table 3-8. Comparison of Direct Static Measurement Data Sets Summary Statistics, Survey Unit 727-01

Statistic	Pre-Surface Media Sampling Measurements Value	Post-Surface Media Sampling Measurements Value	Background Measurements Value
Number of Measurements	29	17	12
Arithmetic Mean	9.29	15.02	6.6
Standard Deviation (sample)	7.59	7.79	2.2
Coefficient of Variation	0.82	0.52	0.34
Max	31.4	31.5	11.3
Median	7.5	15.8	6.3
Minimum	1.84	4.16	4.4
Range	29.56	27.34	6.9
Geometric Mean	6.59	12.91	6.3
UCL (log-normal "t", a=0.05)	13.59	21.07	8.1
LCL (log-normal "t", b=0.05)	6.95	11.35	5.4

The data evaluation and summary statistics for survey unit 727–01 indicate that the post-surface media sampling direct static measurements are slightly higher than those collected prior to sampling and for background. However, the maximum activity measured during pre- and post-


#### **Data Description**


Direct Static Surface Measurements (Post Surface Media Sampling)

**Building 779 Cluster Independent Verification Project** 

Building 727 & 783, Survey Unit 727-01

DCGL	100	
UNITS - dpm/100 cr	n²	
Sample Data		
4.16	Descriptive Statistics	
4.32	Number of Samples	17.000
7.56	Mean	15.015
7.66	Median	15.800
7.71	Standard Deviation	7.799
8.80	CV	0.519431
12.30	Range	27.340
14.15	Minimum	4.160
15.80	Maximum	31.500
17.40	GM	12.909
17.57	GSD	1.825
17.60	Mean of LN(Data)	2.558
18.02	SD of LN(Data)	0.601
20.80	Percent > DCGL	0.000
24.80		
25.10		
31.50	Normal Statistics	
	Mean	15.015
	UCL(Mean) - Z	18.722
	LCL(Mean) - Z	11.307
	95%ile - Z	27.844
	Percent > DCGL	0.000
	W Test (Data)	0.950901
	Normal (a=0.05)?	Yes
	Lognormal Statistics	
	GM	12.909
	GSD	1.825
	AM of data	15.015
	AM - MVUE	15.280
	AM - MLE	15,468
	UCL - Norm t stats	19.025
	LCL - Norm t stats	11.005
	UCL LogNorm t	21,073
	LCL LogNorm t	11.354
	UCL - Modified Cox	21.417
	LCL - Modified Cox	10.902
	UCL - "Exact"	
	LCL - "Exact"	
	95%ile	34.717
	UTL 95%, 95%	57.569
	Percent > DCGL	0.033
l	PEP (Upper)	0.606
]	PEP (Lower)	6.74E-05
	W Test (In Data)	0.931372
· •	Lognorm (a=0.05)?	Yes
	• • • • • • • • • • • • • • • • • • • •	








Figure 3–3. Data Evaluation Statistics—Direct Static Surface Measurements, 727–01 (Post Surface Media Sampling)


# **Data Description**


Direct Static Surface Measurements (Post Surface Media Sampling)

Building 779 Cluster Independent Verification Project

Building 727 & 783, Survey Unit 727-02

DCGL	100	
NITS - dpm/100 c	cm²	
ample Data		
8.80	Descriptive Statistics	
9.03	Number of Samples	13.000
12.30	Mean	17.487
14.70	Median	15.800
15.10	Standard Deviation	6,165
15.70	CV	0.352573
15.80	Range	20,700
16.30	Minimum	8.800
20.00	Maximum	29.500
22.00	GM	16,470
22.60	GSD	1,443
25.50	Mean of LN(Data)	2,802
29.50	SD of LN(Data)	0.367
	Percent > DCGL	0.000
1		
	Normal Statistics	
	Mean	17.487
1	UCL(Mean) - Z	20.838
1	LCL(Mean) - Z	14.135
1	95%ile - Z	27.629
1	Percent > DCGL	0.000
1	W Test (Data)	0.954924
1	Normal (a=0.05)?	Yes
j	Lognormal Statistics	
	GM	16.470
l	GSD	1.443
I	AM of data	17.487
	AM - MVUE	17.522
- 1	AM - MLE	17.617
1	UCL - Norm t stats	21.213
[	LCL - Norm t stats	13.761
1	UCL LogNorm t	21.992
	LCL LogNorm t	14.113
ľ	UCL - Modified Cox	22.047
	LCL - Modified Cox UCL - "Exact"	13.925
	LCL - "Exact"	
ļ	95%ile	30.122
1	UTL 95%, 95%	43.880
ľ	Percent > DCGL	0.000
j	PEP (Upper)	0.014
1	PEP (Lower)	1.39E-12
ł	W Test (In Data)	0.954201
Ì	Lognorm (a=0.05)?	Yes
l		
]		








Figure 3–3 (continued). Data Evaluation Statistics—Direct Static Surface Measurements, 727–02 (Post Surface Media Sampling)


#### **Data Description**


Direct Static Surface Measurements (Post Surface Media Sampling)

**Building 779 Cluster Independent Verification Project** 

Building 782, Survey Unit 782-01

building 762, Sur	vey 01111 762-01	
DCGL	100	
UNITS - dpm/100 cm	1-	
Sample Data		
7.49	Descriptive Statistics	
13.90	Number of Samples	16.000
14.10	Mean	22.443
14.20	Median	21.000
17.50	Standard Deviation	9.187
20.70	CV	0.409324
20.90	Range	34,410
20.90	Minimum	7.490
21.10	Maximum	41.900
21.30	GM	20.718
24.00	GSD	1,529
24.30	Mean of LN(Data)	3.031
27.60	SD of LN(Data)	0.425
27.90	Percent > DCGL	0.000
41.30		
41.90		
	Normal Statistics	00.140
	Mean	22.443
	UCL(Mean) - Z	26.945
	LCL(Mean) - Z	17.942
	95%ile - Z	37.555
	Percent > DCGL	0.000
	W Test (Data)	0.90424
	Normal (a=0.05)?	Yes
	Lognormal Statistics	
	GM	20.718
	GSD	1.529
	AM of data	22.443
	AM - MVUE	22.536
	AM - MLE	22.672
	UCL - Norm t stats	27.338
	LCL - Norm t stats	17.548
	UCL LogNorm t	28.429
	LCL LogNorm t	18.081
	UCL - Modified Cox	28.560
	LCL - Modified Cox	17.782
	UCL - "Exact"	
	LCL - "Exact"	
	95%ile	41.659
	UTL 95%, 95%	60.483
	Percent > DCGL	0.010
	PEP (Upper)	0.328
	PEP (Lower)	5.18E-06
	W Test (In Data)	0.938614
	Lognorm (a=0.05)?	Yes
I I		








Figure 3–3 (continued). Data Evaluation Statistics—Direct Static Surface Measurements, 782–01 (Post Surface Media Sampling)


#### **Data Description**


Direct Static Surface Measurements (Post Surface Media Sampling)

Building 779 Cluster Independent Verification Project

Building 782, Survey Unit 782-02

DCGL	100	
NITS - dpm/100 cr	n <sup>2</sup>	
ample Data		
3.50	Descriptive Statistics	
3.55	Number of Samples	23.000
3.83	Mean	12.939
4.67	Median	13.400
4.79	Standard Deviation	7.776
5.53	CV	0.600959
7.13	Range	27.500
7.23	Minimum	3.500
7.63	Maximum	31.000
8.09	GM	10.589
10.60	GSD	1.974
13.40	Mean of LN(Data)	2.360
14.70	SD of LN(Data)	0.680
17.20	Percent > DCGL	0.000
17.30		
17.30		
17.80	Normal Statistics	
17.90	Mean	12,939
18.15	UCL(Mean) - Z	16,117
18.20	LCL(Mean) - Z	9.761
20.80	95%ile - Z	25.730
27.30	Percent > DCGL	0.000
31.00	W Test (Data)	0.909892
	Normal (a=0.05)?	No
	Lognormal Statistics	
	GM	10.589
	GSD	1.974
	AM of data	12.939
	AM - MVUE	13.184
	AM - MLE	13.344
	UCL - Norm t stats	16.302
	LCL - Norm t stats	9.577
	UCL LogNorm t	17.905
	LCL LogNorm t	9.944
	UCL - Modified Cox	18.296
ŀ	LCL - Modified Cox UCL - "Exact"	9.500
į.	LCL - "Exact"	
	95%ile	32,410
	UTL 95%, 95%	51,604
1	Percent > DCGL	0.048
l	PEP (Upper)	0.554
1	PEP (Lower)	0.00048
	W Test (In Data)	0.00048
	Lognorm (a=0.05)?	Yes
1		








Figure 3–3 (continued). Data Evaluation Statistics—Direct Static Surface Measurements, 782–02 (Post Surface Media Sampling)

media sampling is the same. The arithmetic mean, geometric mean and median for the background and pre-media sampling measurements are virtually the same.

Table 3-9. Comparison of Direct Static Measurement Data Sets Summary Statistics, Survey Unit 727-02

Statistic	Pre-Surface Media Sampling Measurements Value	Post-Surface Media Sampling Measurements Value	Background Measurements Value
Number of Measurements	29	13	9
Arithmetic Mean	19.69	17.5	14.8
Standard Deviation (sample)	6.09	6.2	4.2
Coefficient of Variation	0.31	0.35	0.29
Max	36.3	29.5	22.5
Median	19.4	15.8	15.1
Minimum	8.82	8.8	8.5
Range	27.48	20.7	14.1
Geometric Mean	18.75	16.5	14.2
UCL (normal "t", a=0.05)	22.40	22.0	18.7
LCL (normal "t", b=0.05)	17.47	14.1	11.8

The data evaluation and summary statistics for survey unit 727–02 indicate that there is no appreciable difference between the pre- and post-surface media sampling direct static measurements and the background.

Table 3–10. Comparison of Direct Static Measurement Data Sets Summary Statistics, Survey Unit 782–01

Statistic	Pre-Surface Media Sampling Measurements Value	Post-Surface Media Sampling Measurements Value	Background Measurements Value
Number of Measurements	29	16	12
Arithmetic Mean	12.68	22.4	7.4
Standard Deviation (sample)	8.53	9.2	5.7
Coefficient of Variation	0.67	0.41	0.77
Max	34.5	41.9	15.8
Median	10.95	21.0	4.8
Minimum	1.2	7.5	1.1
Range	33.3	34.4	14.7
Geometric Mean	9.53	20.7	5.1
UCL (normal "t", a=0.05)	19.59	28.4	15.8
LCL (normal "t", b=0.05)	10.04	18.1	4.5

The data evaluation and summary statistics for survey unit 782–01 indicate that the post-surface media sampling direct static measurements are slightly higher than those collected prior to sampling and for background. It is also evident that the pre-media sampling measurements are also slightly higher than those of the background.

Table 3–11. Comparison of Direct Static Measurement Data Sets Summary Statistics, Survey Unit 782–02

Statistic	Pre-Surface Media Sampling Measurements Value	Post-Surface Media Sampling Measurements Value	Background Measurements Value
Number of Measurements	29	23	15
Arithmetic Mean	12.13	12.9	11.7
Standard Deviation (sample)	6.38	7.8	7.2
Coefficient of Variation	0.53	0.60	0.62
Max	24.1	31.0	28.1
Median	11.1	13.4	8.3
Minimum	1.1	3.5	4.6
Range	23.0	27.5	23.5
Geometric Mean	10.05	10.6	10.0
UCL (normal "t", a=0.05)	16.92	17.9	16.2
LCL (normal "t", b=0.05)	9.87	9.9	8.6

The data evaluation and summary statistics for survey unit 782\_02 indicate that there is no appreciable difference between the pre- and post-surface media sampling direct static measurements and the background.

The fact that the data from two of the survey units yielded slightly higher direct static measurements of activity once the surface layer had been removed and that the corresponding surface media sample yielded no appreciable measure of radioactivity tends to support the possibility that the building materials have a measurable concentration of naturally occurring radionuclides<sup>1</sup>. For sample locations where media samples were collected, both the pre- and post-surface media sampling direct static measurement data set are presented side-by-side in Table 3–12.

# 3.2 Laboratory Measurements

The GJO Analytical Laboratory was used to assay all smear and surface media samples collected for independent verification from survey units 727–01, 727–02, 782–01, and 782–02. The GJO Analytical Laboratory was selected because of its method capabilities, quality program, autonomy, and ability to meet the detection limits specified in the SAP (DOE 1999a). In each case, the laboratory met or exceeded the contract required detection limit specified in the SAP. Results of samples analyzed are summarized in the following sections below.

<sup>&</sup>lt;sup>1</sup>Although indications point to the possibility of measurable concentrations of naturally occurring radionuclides, particularly in the concrete materials used in the building construction, no credit is taken by attempting to subtract these from the measured vales in the building. Instead, all radioactivity measured (other than the instrument background) is assumed to be DOE contributed values and is compared against the applicable DCGL to determine compliance with the DQOs.

Table 3–12. Direct Static Surface Measurements Data Sets, Survey Units 727–01, 727–02, 782–01, and 782–02

	·			7	<b>,</b>	_	_			_					_									
:-02	Post Media Sampling		3.50	27.30	31.00	17.30	5.53	17.20	17.80	7.63	14.70	17.90	18.15	18.20	8.09	4.79	3.55	20.80	10.60	13.40	4.67	7.13	7.23	3.83
Survey Unit 782-02	Pre Media Sampling dpm/100 cm²	10.40	24.10	20.60	6.94	14.00	3.96	7.23	11.10	9.45	1.10	7.71	21.50	18.30	15.00	18.35	10.50	3.55	3.44	17.40	19.05	21.00	10.60	17.30
nS	Sample Location	IVP0000421	IVP0000422	IVP0000423	IVP0000424	IVP0000426	IVP0000427	IVP0000430	IVP0000431	IVP0000432	IVP0000433	IVP0000434	IVP0000435	IVP0000436	IVP0000437	IVP0000438	IVP0000440	IVP0000441	IVP0000442	IVP0000443	IVP0000444	IVP0000445	IVP0000446	IVP0000447
01	Post Media Sampling dpm/100 cm²	27.60	24.00	20.70	41.30	13.90	20.90	24.30	41.90	20.90	17.50	21.10	21.30	27.90	14.20	14.10	7.49							4.
Survey Unit 782-01	Pre Media Sampling dpm/100 cm²		31.10	13.80	24.10	17.50	12.25	24.20	24.10	14.10	7.36	10.90	10.95	7.08	10.80	17.50	10.90							TO WASHINGTON
nS	Sample Location	IVP0000301	IVP0000302	IVP0000303	IVP0000304	IVP0000305	IVP0000306	IVP0000307	IVP0000308	IVP0000309	IVP0000310	IVP0000311	IVP0000312	IVP0000313	IVP0000314	IVP0000315	IVP0000316		20.00	24.15				100 March 100 100 100 100 100 100 100 100 100 10
-02	Post Media Sampling dpm/100 cm <sup>2</sup>	20.00	15.10	25.50	16.30	29.50	15.70	14.70	22.60	9.03	22.00	8.80	15.80	12.30		1.00					1.00			7.50 (2.47) (3.47)
Survey Unit 727-02	Pre Media Sampling dpm/100 cm²	22.00	14.70	24.90	18.90	36.30	29.40	24.70	15.80	15.80	24.80	8.82	19.10	19.40				10 4 7 10 10 10 4 10 10 10 10 10 10 10 10 10 10 10 10 10	20.22.4 1. 1. 1.					600 1 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5 5
nS	Sample Location	IVP0000367	IVP0000368	IVP0000369	IVP0000370	IVP0000371	IVP0000372	IVP0000373	IVP0000374	IVP0000375	IVP0000376	IVP0000377	IVP0000378	IVP0000379										
-01	Post Media Sampling dpm/100 cm²	17.40	20.80	17.60	7.56	4.32	18.02	4.16	14.15	17.57	24.80	8.80	15.80	12.30	7.66	7.71	31.50	25.10					1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1. 1	
Survey Unit 727-01	Pre Media Sampling dpm/100 cm²	9.40	20.90	17.70	14.40	11.10	4.49	24.70	3.69	7.25	7.50	31.40	10.70	14.60	14.60	4.36	17.60	4.08						
nS		IVP0000386	IVP0000387	IVP0000390	IVP0000396	IVP0000397	IVP0000398	IVP0000399	IVP0000400	IVP0000401	IVP0000402	IVP0000403	IVP0000404	IVP0000405	IVP0000406	IVP0000407	IVP0000408	IVP0000409						<b>*</b>

## 3.2.1 Smear Samples

Smear samples were collected at each of the 29 designated sample locations from each survey unit. Smear samples were collected following the initial direct static surface measurement by wiping the surface with an absorbent smear filter media using moderate pressure. The smears were packaged and delivered to the GJO Analytical Laboratory for counting. The 29 smear samples from each survey unit were submitted to the GJO Analytical Laboratory along with seven blank and five "spiked" QC smears for analysis. Table 3–13 is provided to aid the reader to keep the sample identification numbers straight. Results and conclusions relative to the quality control smear samples are provided in Section 6.0 of this report.

With the exception of the spiked QC samples submitted, the analytical results showed no measurable radioactivity indicating that there is very little likelihood that the DCGL<sub>W</sub> for removable surface contamination might be exceeded in the survey unit. Since every sample result was below the detection limit for the analysis (MDA), no statistical inferences can be made for the data set. However, since the method detection limit was significantly below the DCGL<sub>W</sub> for removable alpha radioactivity, and each smear sample was shown to have activity less than the detection limit, statistical treatment of the data set is not necessary in order to measure compliance. Table 3–14 summarizes the pertinent information gleaned from the complete analytical report (Requisitions 16900, 16901, and 16908). The entire analytical report is provided in Appendix C.

# 3.2.2 Surface Media Samples

At the request of DOE, surface media samples were collected prior to taking direct static measurements. Surface media samples were collected at locations without consideration of the trigger criteria outlined in the SAP (DOE 1999a). In all, 69 (17 from unit 727–01, 13 from unit 727–02, 16 from unit 782–01, and 23 from unit 782–02) surface media samples were collected. Since the media sample inclusion criteria for a painted or coated surface and/or exceeding the *a priori* estimate of the critical detection level of 22 dpm/100 cm<sup>2</sup> were not assessed prior to taking the media samples, samples were collected at all accessible survey locations except for those on the metal walls and roof of Building 783 and the roof of Building 782.

The walls and roof of Building 783 were constructed of galvanized steel with a thin factory coat of paint. No other paint or coating was ever applied to these surfaces. The collection of media samples from these steel surfaces was determined to be impractical since direct static measurements could quantify any significant fixed surface contamination. The roof material of Building 782 consisted of gravel stone ballast, tar, fiberboard, and 3 inches of insulation. This material was removed by the Contractor prior to the final survey. The remaining vapor barrier covering the concrete roof support was considered a bare surface. Subsequent direct static measurements of all the sample locations on the roof of Building 782 were less than 22 dpm/100 cm<sup>2</sup>.

Table 3-13. Smear Sample Identification Crosswalk, Survey Units 727-01, 727-02, 782-01, and 782-02

S	Survey Unit 727-01	01	Sur	rvey Unit 727-02	12	nS Su	Survey Unit 782-01	5	INS	Survey Unit 782-02	02
Field Sample Location ID#		Laboratory Assigned Sample#	Sample Location ID#	Unique Sample ID#	Laboratory Assigned Sample #	Field Sample Location ID#	Unique Sample ID#	Laboratory Assigned Sample #	Sample Location ID#	Unique Sample ID#	Laboratory Assigned Sample #
IVP0000381	-	264806	IVP0000351	SMR0000351	264838	IVP0000301	SMR0000301	264945	IVP0000421	SMR0000421	264913
IVP0000382		264807	IVP0000352	SMR0000352	264839	IVP0000302	SMR0000302	264946	IVP0000422	SMR0000422	264914
IVP0000383		264808	IVP0000353	SMR0000353	264840	IVP0000303	SMR0000303	264947	IVP0000423	SMR0000423	264915
IVP0000384	4 SMR0000384	264809	IVP0000354	SMR0000354	264841	IVP0000304	SMR00000304	264948	IVP0000424	SMR0000424	264916
IVP0000385		264810	IVP0000355	SMR0000355	264842	1VP0000305	SMR0000305	264949	IVP0000425	SMR0000425	264917
IVP0000386	SMR0000386	264811	IVP0000356	SMR0000356	264843	1VP0000306	SMR0000306	264950	IVP0000426	SMR0000426	264918
IVP0000387		264812	IVP0000357	SMR0000357	264844	IVP0000307	SMR0000307	264951	IVP0000427	SMR0000427	264919
IVP0000388		264813	IVP0000358	SMR0000358	264845	IVP0000308	SMR0000308	264952	IVP0000428	SMR0000428	264920
IVP0000389		264814	IVP0000359	SMR00000359	264846	IVP0000309	8MR0000309	264953	IVP0000429	SMR0000429	264921
IVP0000390	-4	264815	IVP0000360	SMR0000360	264847	IVP0000310	SMR0000310	264954	IVP0000430	SMR0000430	264922
IVP0000391		264816	IVP0000361	SMR0000361	264848	IVP0000311	SMR0000311	264955	IVP0000431	SMR0000431	264923
IVP0000392	2 SMR0000392	264817	IVP0000362	SMR0000362	264849	IVP0000312	SMR0000312	264956	IVP0000432	SMR0000432	264924
IVP0000393		264818	IVP0000363	SMR00000363	264850	IVP0000313	SMR0000313	264957	IVP0000433	SMR0000433	264925
IVP0000394	-	264819	IVP0000364	SMR0000364	264851	IVP0000314	SMR0000314	264958	IVP0000434	SMR0000434	264926
IVP0000395		264820	IVP0000365	SMR00000365	264852	IVP0000315	SMR0000315	264959	IVP0000435	SMR0000435	264927
IVP0000396	SMR0000396	264821	IVP0000366	SMR00000366	264853	IVP0000316	SMR0000316	264960	IVP0000436	SMR0000436	264928
IVP0000397	-	264822	IVP0000367	SMR0000367	264854	IVP0000317	SMR0000317	264961	IVP0000437	SMR0000437	264929
IVP0000398		264823	IVP0000368	SMR00000368	264855	IVP0000318	SMR0000318	264962	IVP0000438	SMR0000438	264930
IVP0000399		264824	IVP0000369	SMR00000369	264856	IVP0000319	SMR0000319	264963	IVP0000439	SMR0000439	264931
IVP0000400		264825	IVP0000370	SMR00000370	264857	IVP0000320	SMR0000320	264964	IVP0000440	SMR0000440	264932
IVP0000401		264826	IVP0000371	SMR0000371	264858	IVP0000321	SMR0000321	264965	IVP0000441	SMR0000441	264933
IVP0000402		264827	IVP0000372	SMR0000372	264859	IVP0000322	SMR0000322	264966	IVP0000442	SMR0000442	264934
IVP0000403		264828	IVP0000373	SMR0000373	264860	IVP0000323	SMR0000323	264967	IVP0000443	SMR0000443	264935
IVP0000404	-4	264829	IVP0000374	SMR0000374	264861	IVP0000324	SMR0000324	264968	IVP0000444	SMR0000444	264936
IVP0000405		264830	IVP0000375	SMR0000375	264862	IVP0000325	SMR0000325	264969	IVP0000445	SMR0000445	264937
IVP0000406		264831	IVP0000376	SMR0000376	264863	IVP0000326	SMR0000326	264970	IVP0000446	SMR0000446	264938
IVP0000407		264832	IVP0000377	SMR0000377	264864	IVP0000327	SMR0000327	264971	IVP0000447	SMR0000447	264939
IVP0000408		264833	IVP0000378	SMR0000378	264865	IVP0000328	SMR0000328	264972	IVP0000448	SMR0000448	264940
IVP0000409		264834	IVP0000379	SMR0000379	264866	IVP0000329	SMR0000329	264973	IVP0000449	SMR0000449	264941
QC SPIKE	SMR0000261	264835	OC BLANK	SMR0000264	264867	QC BLANK	SMR0000454	264974	QC BLANK	SMR0000451	264942
QC SPIKE	SMR0000262	364836	QC SPIKE	SMR0000265	264868	QC BLANK	SMR00000455	264975	QC BLANK	SMR0000452	264943
QC SPIKE	SMR0000263	264837	QC SPIKE	SMR0000266	264869	QC BLANK	SMR0000456	264976	QC BLANK	SMR0000453	264945

6 minutes

Sample Count Time

Survey Unit 727-02 Survey Unit 727-01 29 (264806 through 264834) 29 (264838 through 264866) **Number of Samples** 4.72 dpm/100 cm<sup>2</sup> 4.72 dpm/100 cm<sup>2</sup> Method Detection Limit ( $\alpha$ =0.05) All smear samples were less Removable Alpha Surface All smear samples were less Radioactivity (dpm/100 cm<sup>2</sup>) than (<) MDA than (<) MDA 6 minutes 6 minutes Sample Count Time Survey Unit 782-01 Survey Unit 782-02 **Number of Samples** 29 (264945 through 264973) 29 (264913 through 264941) 4.84 dpm/100 cm<sup>2</sup> Method Detection Limit ( $\alpha$ =0.05) 4.84 dpm/100 cm<sup>2</sup> All smear samples were less All smear samples were less Removable Alpha Surface Radioactivity (dpm/100 cm<sup>2</sup>) than (<) MDA than (<) MDA

Table 3-14. Summary of Analytical Report Data for Smears

In order to achieve the required detection sensitivity and to distinguish between transuranic and uranium series radionuclides, alpha spectroscopy analysis was chosen to assay the surface media samples. Distinction between transuranic and uranium series nuclides is important because radionuclide series specific DCGLs were established for the surface media samples. Sample masses ranged from approximately 5 to 20 grams and were collected over a 100 cm<sup>2</sup> surface area. The laboratory analyzed each sample for the following radionuclide sets:

6 minutes

Transuranic Series Isotopes Pu-238, Pu-239/240, and Am-241

• Uranium Series Isotopes U-234, U-235, and U-238

In spectrometric analysis, each isotope has its own counting statistics and detection limit. Many of the sample measurements resulted in one or more of the isotope specific values below the detection limit. When this occurred, a value equal to one half of the sample specific detection limit was used to calculate the <u>total</u> radionuclide series activity. Collated data has been derived from the complete analytical report (Requisitions 16904, 16905, and 16906) and presented in Table 3–15. The entire analytical reports are provided in Appendix C.

From summary data presented in Table 3–15 several features are apparent:

- Isotopic assay of the contaminants found on and beneath surfaces in Buildings 727, 782, and 783 indicate the persistent presence of uranium series radionuclides. All but the three samples collected from the roof of Building 727 (fiberboard and tar) showed detectable concentrations of both of the two uranium isotopes found most abundantly in nature, U-234 and U-238. This is indicative of the presence of background contributions of these nuclides and is consistent with the background contributions expected in concrete and cinder block materials (NRC 1997).
- No clear presence of transuranic radioactivity was indicated in any of the 69 samples collected.

• The detection limit for a given isotope varied for each sample. This is due to the variation in the total sample mass collected. The laboratory was limited by the mass of sample that could efficiently be processed to extract the nuclides of interest. As a result, the laboratory fractioned a relatively consistent aliquot of the total mass submitted to actually perform the analysis. The larger the sample collected and submitted, the smaller the fraction represented by the aliquot. Thus, the detection limit increased (got poorer) as the total mass collected increased. In each case, however, actual field sampling procedure collected a sample from a 100 cm<sup>2</sup> area until the entire surface had been stripped of the paint layer or surface veneer.

Surface media sample data evaluation statistics for each survey unit is presented for the uranium series and the transuranic series in Figures 3–4 and 3–5, respectively. Table 3–16 presents the summary statistics for the surface media sample data set, with the transuranic and uranium series nuclides presented as independent subsets.

Table 3–15. Surface Media Sample Data, Buildings 727 and 783, Survey Unit 727–01—
Alpha Isotopic Analysis

			<del>,</del>										Total Transuranic	Total Uranium
		ns)	An	1-241	Pu-	238	Pu-23	9/240	U-234	U-2	235	U-238	Activity	Activity
Sample Location	Lab Sample ID#	Sample Weight (grams)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	dpm/100 cm²	dpm/100 cm²
MED0000386	264771	11.03	0.88	0.88	1.29	0.65	1.81	0.91	31.11	1,86	0.93	30.98	2.43	63.02
MED0000387	264772	13.70	1.36	1.36	0.51	0.26	1.07	0.54	60.67	2.61	1.31	59.34	2.15	121.32
MED0000390	264773	11.08	1.45	1.45	0.82	0.41	1.11	0.56	29.24	2.26	2.26	30.58	2.42	62.08
MED0000396	264774	14.91	0.57	0.57	1.16	0.58	2.03	1.02	22.16	2.55	1.28	24.68	2.17	48.12
MED0000397	264775	15.84	1.72	1.72	1.07	0.54	1.07	0.54	26.33	1.99	1.00	26.77	2.79	54.10
MED0000398	264776	13.45	1.20	0.60	0,51	0.26	1.98	0.99	19.85	1.17	0.59	19.77	1.85	40.21
MED0000399	264777	13.74	1.44	1.44	1.00	0.50	1.00	0.50	20.26	2.15	1.08	18.08	2.44	39.42
MED0000400	264778	7.52	2.23	2.23	0.57	0.29	0.78	0.39	13.44	1.41	0.71	13.26	2.91	27.41
MED0000401	264779	3.71	0.91	0.91	0.31	0.16	1.34	1.34	3.17	0.46	0.23	3.43	2.41	6.83
MED0000402	264780	15.31	1.55	1.55	0.50	0.25	1.26	0.63	19.05	2.51	1.26	23.17	2.43	43.48
MED0000403	264781	14.57	1.49	0.75	1,15	0.58	1.56	0.78	20.18	3.15	1.58	18.90	2.10	40.66
MED0000404	264782	6.16	2.04	2.04	0.75	0.38	1.39	1.39	6.04	1.08	0.54	6.67	3.81	13.25
MED0000405	264783	11.00	0.85	0.85	0.91	0.46	1.46	0.73	12.13	1.87	0.94	14.08	2.04	27.15
MED0000406	264784	13.10	1.63	1.63	0.44	0.22	0.91	0.46	21.99	1.69	0.85	19.92	2.31	42.76
MED0000407	264785	17.59	2.68	2.68	1.74	0.87	1.97	0.99	31.75	3.11	1.56	32.05	4.54	65.36
MED0000408	264786	10.28	0.90	0.90	1.32	0.66	1.22	0,61	16.23	1.89	0.95	14.27	2.17	31.45
MED0000409	264787	11.29	1.54	1.54	1.07	0.54	1.44	0.72	19.07	1,25	0.63	15.14	2.80	34.84
Shaded cells in	ndicate v	alues b	elow th	e detecti	on limit. T	he report	ed value	is the sa	ample-sp	ecific N	IDA.			

Table 3–15 (continued). Surface Media Sample Data, Buildings 727 and 783, Survey Unit 727–02— Alpha Isotopic Analysis

		ıs)	Am	-241	Pu-	238	Pu-23	9/240	U-234	U-:	235	U-238	Total Transuranic Activity	Total Uranium Activity
Sample Location	Lab Sample ID#	Sample Weight (grams)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	orte Orte	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	dpm/100 cm²	dpm/100 cm²
MED0000367	264788	17.16	2.02	2.02	0.63	0.32	2.43	1.22	42.31	2.47	2.47	37.69	3.55	82.47
MED0000368	264789	19.23	2.20	2.20	1.05	0.53	1.28	0.64	28.33	3.10	1.55	28.55	3.37	58.43
MED0000369	264790	13.04	1.05	0.53	0.46	0.23	1.34	0.67	0.75	1.63	0.82	0.74	1.43	2.31
MED0000370	264791	16.95	2.27	2.27	0.57	0.29	2.76	2.76	35.07	2.41	2.41	37.52	5.32	75.00
MED0000371	264792	11.64	1.51	1.51	0.35	0.18	1.00	0.50	23.21	1.92	0.96	24.34	2.19	48.51
MED0000372	264793	15.31	1.73	1.73	1.20	0.60	1.37	0.69	35.29	3.43	1.72	37.86	3.02	74.87
MED0000373	264794	7.01	0.59	0.30	0.61	0.31	0.54	0.27	0.43	0.84	0.42	0.43	0.87	1.28
MED0000374	264795	15.42	1.62	1.62	1.30	0.65	1.76	0.88	35.06	3.30	3.30	32.89	3.15	71.25
MED0000375	264796	13.25	4.92	4.92	0.43	0.22	3.07	3.07	18.81	1.69	0.85	18.41	8.21	38.07
MED0000376	264797	5.01	0.38	0.19	0.40	0.20	0.45	0.23	0.30	0.76	0.38	1.00	0.62	1.68
MED0000377	264798	15.08	2.53	2.53	1.69	0.85	2.11	1.06	27.94	3.20	1.60	31.02	4.43 .	60.56
MED0000378	264799	18.60	1.25	1.25	1.85	0.93	2.48	1.24	49.91	3.36	3.36	48.71	3.42	101.98
MED0000379	264800	15.20	1.87	1.87	0.61	0.31	1.26	0.63	33.85	2.55	1.28	32.93	2.81	68.06

Table 3–15 (continued). Surface Media Sample Data, Building 782, Survey Unit 782–01— Alpha Isotopic Analysis


		(St	An	n-241	Pu-	238	Pu-23	9/240	U-234	U-:	235	U-238	Total Transuranic Activity	Total Uranium Activity
Sample Location	Lab Sample ID#	Sample Weight (grams)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	. dpm/100 cm²	dpm/100 cm²
MED0000301	264870	8.07	1.20	1.20	0.35	0.18	0.39	0.20	30.61	1.92	1.92	31.96	1.57	64.49
MED0000302	264871	11.30	0.73	0.73	0.10	0.05	0.72	0.36	22.99	1.64	0.82	22.87	1.14	46.68
MED0000303	264872	11.12	0.74	0.37	0.26	0.13	0.55	0.28	25.33	1.43	1.43	27.72	0.78	54.48
MED0000304	264873	12.10	1.19	0.60	0.21	0.11	0.53	0.27	25.81	7.02	7.02	24.13	0.97	56.96
MED0000305	264874	10.47	0.66	0.33	0.16	80.0	0.26	0.13	18.08	0.92	0.46	17.25	0.54	35.79
MED0000306	264875	8.24	0.82	0.82	0.08	0.04	0.54	0.27	28.21	1.36	1.36	27.63	1.13	57.20
MED0000307	264876	13.70	2.38	2.38	80.0	0.04	0.47	0.24	26.33	0.55	0.28	27.30	2.66	53.91
MED0000308	264877	10.62	1.04	0.52	0.29	0.15	0.74	0.37	32.45	1.59	1.59	30.92	1.04	64.96
MED0000309	264878	8.18	0.70	0.70	0.15	0.08	0.26	0.13	17.89	1.18	1.18	19.37	0.91	38.44
MED0000310	264879	15.16	1.81	1.81	0.33	0.17	1.41	0.71	10.69	0.15	0.08	8.14	2.68	18.91
MED0000311	264880	9.24	1.79	1.79	0.10	0.05	0.00	0.00	21.53	2.25	2.25	19.08	1.84	42.86
MED0000312	264881	11.94	0.76	0.38	0.00	0.00	0.32	0.16	24.39	1.62	0.81	27.08	0.54	52.28
MED0000313	264882	11.11	1.20	1.20	0.21	0.11	0.57	0.29	36.84	1.38	0.69	37.13	1.59	74.66
MED0000314	264883	10.10	1.55	1.55	0.10	0.05	0.35	0.18	18.28	1.29	1.29	18.78	1.78	38.35
MED0000315	264884	9.16	0.72	0.72	0.05	0.03	0.84	0.42	32.07	1.36	1.36	32.36	1.17	65.79
MED0000316	264885	9.19	1.89	1.89	0.36	0.18	0.64	0.32	21.61	1.67	1.67	20.27	2.39	43.55
Shaded cells in	ndicate va	Shaded cells indicate values below the detection limit. The reported value is the sample-specific MDA.												


Table 3–15 (continued). Surface Media Sample Data, Building 782, Survey Unit 782–02— Alpha Isotopic Analysis

		(SI	An	n-241	Pu-	238	Pu-23	9/240	U-234	U-:	235	U-238	Total Transuranic Activity	Total Uranium Activity
Sample Location	Lab Sample ID#	Sample Weight (grams)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	Reported Value	Reported Value (w/ samples less than MDA @ 0.5 MDA)	Reported Value	dpm/100 cm²	dpm/100 cm²
MED0000421	264888	10.49	0.93	0.47	0.10	0.05	0.15	80.0	19.75	1.45	1.45	22.24	0.59	43.44
MED0000422	264889	7.87	0.90	0.45	0.08	0.04	0.48	0.24	13.79	1.18	1.18	14.73	0.73	29.70
MED0000423	264890	10.61	1.09	1.09	0.14	0.07	0.92	0.46	15.48	1.37	0.69	15.05	1.62	31.22
MED0000424	264891	4.52	1.33	1.33	0.07	0.04	2.15	2.15	6.74	0.81	0.41	3.41	3.52	10.56
MED0000426	264892	2.82	0.88	0.88	0.00	0.00	0.72	0.36	2.64	0.39	0.20	2.45	1.24	5.29
MED0000427	264893	3.82	1.08	1.08	0.12	0.06	1.30	1.30	2.31	0.12	0.12	2.23	2.44	4.66
MED0000430	264894	4.49	1.88	1.88	0.07	0.04	1.03	0.52	5.24	0.07	0.04	4.15	2.43	9.43
MED0000431	264895	6.62	1.75	1.75	0.08	0.04	3.00	3.00	7.10	0.09	0.05	7.40	4.79	14.55
MED0000432	264896	8.24	0.81	0.41	0.06	0.03	0.50	0.25	15.36	2.06	2.06	14.61	0.69	32.03
MED0000433	264897	9.21	1.23	1.23	0.17	0.09	0.04	0.02	15.59	0.68	0.34	19.37	1.34	35.30
MED0000434	264898	8.15	0.48	0.24	0.03	0.02	0.72	0.36	16.33	0.92	0.46	14.82	0.62	31.61
MED0000435	264899	5.99	0.05	0.03	0.06	0.03	0.17	0.09	15.43	0.88	0.88	15.49	0.14	31.80
MED0000436	264900	11.24	1.36	1.36	0.30	0.15	0.03	0.02	40.22	1.79	0.90	38.10	1.53	79.22
MED0000437	264901	11.01	0.59	0.30	0.19	0.10	0.83	0.83	19.45	1,24	0.62	21.32	1.22	41.39
MED0000438	264902	7.97	0.65	0.33	0.09	0.05	0.03	0.02	25.96	1.89	0.95	29.11	0.39	56.02
MED0000440	264903	8.63	0.69	0.69	0.19	0.10	80.0	0.04	10.17	0.41	0.21	10.25	0.83	20.63
MED0000441	264904	6.75	0.33	0.33	0.02	0.01	0.06	0.03	9.03	1.06	1.06	9.55	0.37	19.64
MED0000442	264905	5.76	0.43	0.43	0.44	0.22	0.37	0.19	6.36	0.30	0.15	6.09	0.84	12.60
MED0000443	264906	8.97	0.79	0.40	0.02	0.01	0.32	0.16	11.75	0.80	0.40	11.98	0.57	24.13
MED0000444	264907	8.70	1.31	1.31	0.11	0.06	0.39	0.20	10.76	0.38	0.19	11.93	1.56	22.88
MED0000445	264908	8.84	1.02	1.02	80.0	0.04	0.08	0.04	10.01	0.65	0.33	8.29	1.10	18.63
MED0000446	264909	7.09	0.57	0.29	0.02	0.01	0.04	0.02	7.61	0.87	0.44	8.64	0.32	16.69
MED0000447	264910	10.25	1.22	1.22	0.21	0.11	0.21	0.11	13.09	0.34	0.17	12.93	1.43	26.19
Shaded cells in	dicate va	lues bel	ow the	detectio	n limit. Th	e reported	value is	the sam	ple-spe	cific MI	DA.			

Data Description
Uranium Series Activity, Surface Media Samples
Building 779 Cluster, Independent Verification Project
Building 727 & 783, Survey Unit 727-01

Dullding 727 & 700,	duvey one 121-01
DCGL	5000
Sample Data	
UNITS - dpm/100 cm2	Descriptive Statistics
6.83	Number of Samples 17.000
13.25	Mean 44.792
27.15	Median 40.660
27.41	Standard Deviation 25.504
31,45	CV 0.569401
34.84	Range 114.490
39.42	Minimum 6.830
40.21	Maximum 121.320
40.66	GM 37.987
42.76	GSD 1.908
43.48	Mean of LN(Data) 3.637
48.12	SD of LN(Data) 0.646
54.10	Percent > DCGL 0.000
62.08	
63.02	
65.36	Normal Statistics
121.32	Mean 44.792
	UCL(Mean) - Z 56.916
	LCL(Mean) - Z 32.668
	95%ile - Z 86.747
	Percent > DCGL 0.000
	W Test (Data) 0.871687
	Normal (a=0.05)? No
	Lognormal Statistics
	GM 37.987
	GSD 1.908
	AM of data 44.792
	AM - MVUE 46.136
	AM - MLE 46.803
	UCL - Norm t stats 57.905
	LCL - Norm t stats 31.679
	UCL LogNorm t 65.245
	LCL LogNorm t 33.574
	UCL - Modified Cox 66.604
·	LCL - Modified Cox 31.958
	UCL - "Exact"
	LCL - "Exact"
	95%ile 109.953
	UTL 95%, 95% 189.316
	Percent > DCGL 0.000
	PEP (Upper) 0.000
	PEP (Lower) 0
	W Test (in Data) 0.903053
,	Lognorm (a=0.05)? Yes
	·











Figure 3-4. Data Evaluation Statistics-Uranium Series Activity, Surface Media Samples

Data Description
Uranium Series Activity, Surface Media Samples
Building 779 Cluster, Independent Verification Project
Building 727 & 783, Survey Unit 727-02












Figure 3-4 (continued). Data Evaluation Statistics-Uranium Series Activity, Surface Media Samples

Data Description Uranium Series Activity, Surface Media Samples Building 779 Cluster, Independent Verification Project Building 782, Survey Unit 782-01

DCGL	5000	
Sample Data		
UNITS - dpm/100 cm <sup>2</sup>	Descriptive Statistics	
18.91	•	16.000
35.79	•	50.582
38.35	Median 5	53.095
38.44	Standard Deviation	14.067
42.86	CV 0.2	78098
43.55	Range	55.750
46.68	Minimum 1	18.910
52.28	Maximum 7	74.660
53.91	GM 4	18.360
54.48	GSD	1.393
56.96	Mean of LN(Data)	3.879
57.20	SD of LN(Data)	0.331
64.49	Percent > DCGL	0.000
64.96		
65.79		
74.66	Normal Statistics	
	Mean 5	50.582
	UCL(Mean) - Z	57.475
•	LCL(Mean) - Z	13.689
	95%ile - Z 7	73.722
	Percent > DCGL	0.000
	W Test (Data) 0.9	72788
	Normal (a≈0.05)?	Yes
	Lognormal Statistics	
	GM 4	8.360
	GSD	1.393
	AM of data 5	50.582
	AM - MVUE	50.904
	AM - MLE 5	51.087
	UCL - Norm t stats 5	58.078
	LCL - Norm t stats 4	13.086
	UCL LogNorm t 6	60. <del>94</del> 9
	LCL LogNorm t . 4	2.821
	UCL - Modified Cox 6	31.041
	LCL - Modified Cox 4	12.451
	UCL - "Exact"	
	LCL - "Exact"	
	95%ile 8	33,394
	UTL 95%, 95% 11	11.544
	Percent > DCGL	0.000
	PEP (Upper)	
	PEP (Lower)	
	, ,	88412
	Lognorm (a=0.05)?	No










Figure 3-4 (continued). Data Evaluation Statistics—Uranium Series Activity, Surface Media Samples

Data Description Uranium Series Activity, Surface Media Samples Building 779 Cluster, Independent Verification Project Building 782, Survey Unit 782-02

DCGL	5000
<u></u>	
Sample Data	
UNITS - dpm/100 cm2	Descriptive Statistics
4.66	Number of Samples 23.000
5.29	Mean 26.853
9.43	Median 24.130
10.56	Standard Deviation 17.080
12.60	CV 0.636077
14.55	Range 74.560
16.69	Minimum 4.660
18.63	Maximum 79.220
19.64	GM 21.918
20.63	GSD 1.999
22.88	Mean of LN(Data) 3.087
24.13	SD of LN(Data) 0.693
26.19	Percent > DCGL 0.000
29.70	•
31.22	
31.61	Normal Statistics
31.80	Mean 26,853
32.03	UCL(Mean) - Z 33.833
35,30 41,39	LCL(Mean) - Z 19.872 95%ile - Z 54.950
41.39	95766 - Z 54,950 Percent > DCGL 0.000
56.02	W Test (Data) 0.900223
79.22	Normal (a=0.05)? No
19,22	Normar (a=0.05)?
	Lognormal Statistics
	GM 21.918
	GSD 1.999
	AM of data 26.853
	AM - MVUE 27.515
	AM - MLE 27.864
	UCL - Norm t stats 34.239
1	LCL - Norm t stats 19.467
	UCL LogNorm t 37.597
Ĭ	LCL LogNorm t 20.650
}	UCL - Modified Cox 38.469
	LCL - Modified Cox 19.681
	UCL - "Exact"
	LCL - "Exact"
	95%ile 68.514
	UTL 95%, 95% 110.051
	Percent > DCGL 0.000
	PEP (Upper)
	PEP (Lower)
j i	W Test (In Data) 0.962573
	Lognorm (a=0.05)? Yes
1	
1	










Figure 3-4 (continued). Data Evaluation Statistics-Uranium Series Activity, Surface Media Samples

Data Description Transuranic Activity, Surface Media Samples Building 779 Cluster, Independent Verification Project Buildings 727 and 783, Survey Unit 727-01

DCGL	100		
Sample Data			
UNITS - dpm/100 cm2		Descriptive Statistics	
1.85		Number of Samples	17.000
2.04		Mean	2.575
2.10		Median	2.420
2.15		Standard Deviation	0.677
2.17		CV	0.262766
2.17		Range	2.690
2.31		Minimum	1.850
2.41 2.42		Maximum GM	4.540
2.42		GSD	2.507 1.254
2.43		Mean of LN(Data)	0.919
2.44		SD of LN(Data)	0.226
2.79		Percent > DCGL	0.000
2.80		i disalik - DOGL	0.000
2.91			
3.81		Normal Statistics	
4.54		Mean	2.575
		UCL(Mean) - Z	2.896
		LCL(Mean) - Z	2.253
	•	95%ile - Z	3.688
		Percent > DCGL	0.000
		W Test (Data)	0.776405
		Normal (a=0.05)?	No
		Lognormal Statistics	
		GM	2.507
		GSD	1.254
		AM of data	2.575
		AM - MVUE	2.568
		AM - MLE	2.572
		UCL - Norm t stats	2.923
•		LCL - Norm t stats	2.227
		UCL LogNorm t	2.889
		LCL LogNorm t	2.290
		UCL - Modified Cox	2.889
		LCL - Modified Cox	2.283
		UCL - "Exact"	
		LCL - "Exact" 95%ile	2.027
		95%1e UTL 95%, 95%	3.637 4.398
		Percent > DCGL	
		PEP (Upper)	0.000
		PEP (Upper) PEP (Lower)	
		W Test (In Data)	0.865499
		Lognorm (a=0.05)?	0.005499 No
•		Lognomi (a-0.03)!	140
1			










Figure 3–5. Data Evaluation Statistics—Transuranic Series Activity, Surface Media Samples

**Data Description** 

Transuranic Activity, Surface Media Samples Building 779 Cluster, Independent Verification Project Buildings 727 and 783, Survey Unit 727-02

Buildings /2/ and /	53, Survey Unit 727-02
DCGL	100
Sample Data	
UNITS - dpm/100 cm2	Descriptive Statistics
0.62	Number of Samples 13.000
0.87	Mean 3.261
1.43	Median 3.150
2.19	Standard Deviation 1.993
2.81	CV 0.611333
3.02	Range 7.590
3.15	Minimum 0.620
3.37	Maximum 8.210
3.42	GM 2.674
3.55	G\$D 2.035
4.43	Mean of LN(Data) 0.984
5.32	SD of LN(Data) 0.710
8.21	Percent > DCGL 0.000
	Normal Statistics
	Mean 3.261
	UCL(Mean) - Z 4.344
	LCL(Mean) - Z 2.177
	95%ile - Z 6.540
	Percent > DCGL 0.000
	W Test (Data) 0.910396
	Normal (a≈0.05)? Yes
	Lognormal Statistics
	GM 2.674
	GSD 2.035
	AM of data 3.261
	AM - MVUE 3.363
	AM - MLE 3.441
	UCL - Norm t stats 4.465
	LCL - Norm t stats 2.056
	UCL LogNorm t 5.286
	LCL LogNorm t 2.240
	UCL - Modified Cox 5.458
	LCL - Modified Cox 2.072
	UCL - "Exact"
	LCL - "Exact"
	95%ile 8.602
	UTL 95%, 95% 17.814
	Percent > DCGL 0.000
	PEP (Upper) 0.000
	PEP (Lower) 0
	W Test (In Data) 0.927231
	Lognorm (a=0.05)? Yes
	<u> </u>








Figure 3-5 (continued). Data Evaluation Statistics—Transuranic Activity, Surface Media Samples


**Data Description** 


Transuranic Activity, Surface Media Samples

Building 779 Cluster, Independent Verification Project

Building 782, Survey Unit 782-01

DCGL	100		
<del> </del>			
Sample Data			
UNITS - dpm/100 cm <sup>2</sup>		Descriptive Statistics	
0.54		Number of Samples	16.000
0.54	-	<i>l</i> ean	1.421
0.78		Median	1.155
0.91		Standard Deviation	0.693
0.97		CV	0.488157
1.04		Range	2.140
1.13		Ainimum	0.540
1.14		/laximum	2.680
1.17		SM	1.267
1.57		SSD	1.655
1.59		flean of LN(Data)	0.236
1.78		D of LN(Data)	0.504
1.84	,	Percent > DCGL	0.000
2.39			
2.66			
2.68		Iormal Statistics	
		Aean	1.421
		JCL(Mean) - Z	1.760
		.CL(Mean) - Z	1.081
	_	5%ie - Z	2,561
		Percent > DCGL	0.000
	•	V Test (Data)	0.911903
	r	lormal (a=0.05)?	Yes
·	L	ognormal Statistics	
	C	SM .	1.267
	0	SSD	1.655
	A	M of data	1.421
	A	M - MVUE	1.425
•		M - MLE	1.438
		JCL - Norm t stats	1.790
	Ł	.CL - Norm t stats	1.051
	į.	JCL LogNorm t	1.881
		CL LogNorm t	1.099
		JCL - Modified Cox	1.897
		CL - Modified Cox	1.071
		ICL - "Exact"	
		.CL - "Exact"	
		5%ile	2.901
		JTL 95%, 95%	4.515
		Percent > DCGL	0.000
		PEP (Upper)	
		EP (Lower)	•
		V Test (In Data)	0.952173
	L	ognorm (a=0.05)?	Yes









Figure 3-5 (continued). Data Evaluation Statistics—Transuranic Activity, Surface Media Samples


**Data Description** 


Transuranic Activity, Surface Media Samples

Building 779 Cluster, Independent Verification Project

Building 782, Survey Unit 782-02










Figure 3-5 (continued). Data Evaluation Statistics—Transuranic Activity, Surface Media Samples

Table 3-16. Summary Statistics, Surface Media Samples

-		nit 727–01	Survey Unit 727-02				
Statistic	Transuranic Surface Activity Value	Uranium Series Surface Activity Value	Transuranic Surface Activity Value	Uranium Series Surface Activity Value			
Number of Measurements	17	17	13	13			
Arithmetic Mean	2.58	44.79	3.26	52.65			
Standard Deviation	0.68	25.50	1.99	32.91			
Coefficient of Variation	0.26	0.57	0.61	0.63			
Max	4.54	121.32	8.21	101.98			
Median	2.42	40.66	3.15	60.56			
Minimum	1.85	6.83	0.62	1.28			
Range	2.69	114.49	7.59	100.70			
Geometric Mean	2.51	37.99	2.67	28.29			
UCL (log-normal "t", a=0.05)	2.89	65.25	5.29	281.85			
	Survey U	nit 782–01	Survey Unit 782-02				
Statistic	Transuranic Surface Activity Value	Uranium Series Surface Activity Value	Transuranic Surface Activity Value	Uranium Series Surface Activity Value			
Number of Measurements	16	16	23	23			
Arithmetic Mean	1.42	50.58	1.32	26.85			
Standard Deviation	0.69	14.07	1.10	17.08			
Coefficient of Variation	0.49	0.28	0.83	0.64			
Max	2.68	74.66	4.79	79.22			
Median	1.16	53.10	1.10	24.13			
Minimum	0.54	18.91	0.14	4.66			
Range	2.14	55.75	4.65	74.56			
Geometric Mean	1.27	48.36	0.98	21.92			
UCL (log-normal "t", a=0.05)	1.88	60.95	1.94	37.60			

### 4.0 Analysis of Sample Plan Results for Compliance

In accordance with MARSSIM and other EPA guidance (EPA 1997) (EPA 1993), the Building 779 Cluster IV SAP identified the decision rules which provide the basis for independently verifying and assessing the RFETS Contractor's conclusions and recommendations for risk management actions in Building 779 (DOE 1999a). To accomplish this objective, the IVC was tasked with performing independent measurements of a representative fraction of the Contractor's survey, such that a statistically valid, yet independent conclusion could be drawn. In order to obtain a data set robust enough to allow statistically valid comparisons with the decision rules, the IVC selected and sampled four of the five designated survey units in Buildings 727, 782, and 783 or four of 49 designated survey units in the Building 779 Cluster. The first decision rule supports this decision objective. The IVC was also tasked with reviewing and verifying the Contractor's Closeout Radiological Survey Report and its conclusions. Since the Contractor's decision basis is applied independently to each survey unit, a sampling and statistical test with power comparable to that used by the Contractor was needed in order to compare with the conclusions reached by the Contractor. The second decision rule supports this objective. The decision rules which define compliance for the independent verification of the Building 779 Cluster surveys are specified in the IV SAP (DOE 1999a) and were reviewed by the EPA and approved by DOE and CDPHE. The IVC's SAP specifies the following two decision rules:

If the independent verification survey concludes that, in the selected survey unit(s), the mean (or median) removable surface contamination concentration is below 20 dpm/100 cm² gross alpha activity, and the mean (or median) total alpha surface contamination concentration as measured by direct surface emission is below 100 dpm/100 cm², and the maximum total alpha surface contamination concentration as measured by direct surface emission is below 300 dpm/100 cm², and the mean (or median) contamination concentration on and beneath a surface with a surface coating as measured by collection and analysis of a surface media sample is below 100 dpm/100 cm² for all transuranic nuclides combined and below 5,000 dpm/100 cm² for all uranium series nuclides combined, then conclude that the survey unit meets the release criterion.

If the IVC survey conclusion disagrees with the Contractor's final status survey conclusion, then refute the Contractor's conclusion for the survey unit and consult with the DOE-RFFO contact for direction on discrepancy resolution.

Demonstrating compliance with the decision rules for independent verification provides DOE with assurance that a substantial and credible case exists for releasing the buildings from further radioactive contamination controls during demolition or disposal.

The first decision rule forms the basis for the five DCGLs, the benchmarks against which measured values are compared to determine compliance. Each component of the decision rule can be reduced to a specific DCGL. The DCGLs for Buildings 727, 782, and 783 Closeout Radiological Survey are:

- 20 dpm/100 cm<sup>2</sup> for removable alpha surface contamination
- 100 dpm/100 cm<sup>2</sup> (mean or median) total alpha surface contamination as measured by direct surface emission

- 300 dpm/100 cm<sup>2</sup> (maximum) total alpha surface contamination as measured by direct surface emission
- 100 dpm/100 cm² (mean or median) total transuranic surface contamination on and beneath a surface with a surface coating as measured by collection and analysis of a surface media sample
- 5,000 dpm/100 cm² (mean or median) total uranium series surface contamination on and beneath a surface with a surface coating as measured by collection and analysis of a surface media sample

#### 4.1 Survey and Sampling Results Compared to the DCGLs

The following sections address each component of the sampling performed and compare the results to the applicable DCGLs. While the data sets collected by the IVC have been shown to best fit both normal and log-normal distributions, the DCGLw values as stated by the RFETS Contractor (RMRS 1999a), do not specify whether the compliance benchmark assumes the arithmetic mean or some other estimate of central tendency appropriate to the distribution. For example, in the case of log-normally distributed data, the log-normal average (i.e., geometric mean) is a more appropriate indicator of the central tendency. When the distribution is not well known or abnormally skewed, the median value generally provides a good estimate of the central tendency for the data set. For comparison purposes in this report, the arithmetic (or normal) mean, the log-normal mean, and the median value are provided for each data set along with the maximum values observed. These provide the risk managers and decision maker with the range of plausible values that might be encountered and considerable evidence, regardless of the underlying distribution, for comparison with the DCGL benchmarks.

#### 4.1.1 Direct Static Surface Measurements

Table 4–1 presents the gross direct static surface measurement results obtained in survey units 727–01, 727–02, 782–01, and 782–02. In this table, no correction for instrument background has been made in order to provide the risk managers and decision makers with the information needed to compare corrected and uncorrected results in the survey unit and correlate the measured residual radioactivity in the survey unit not only with the DCGL but also with the comparable measure of background.

Table 4–2 compares the background adjusted survey unit measurement results to the applicable DCGL. The background adjustment for direct static measurements is made by simply subtracting the central tendency estimate of the background measurements made over the sampling period from the comparable central tendency estimate of the gross, or unadjusted values collected and recorded in the field (see Section 3.1 for detailed discussion of background correction methods employed).

Table 4-1. Comparison of Direct Static Survey Measurements to Applicable DCGLs

Survey Units 727–01 and 727–02 Unadjusted (Gross) Measurements (dpm/100 cm²)												
Contamination		Arithmetic Geometric Mean		Median		Arithmetic UCL <sub>95</sub>		Log-normal UCL <sub>95</sub>		Maximum		
by direct surface emission)	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02
100 dpm/100 cm <sup>2</sup>	9.29	19.69	6.59	18.75	7.50	19.40	12.17	22.01	13.59	22.40		=
300 dpm/100 cm <sup>2</sup>											31.40	36.30
	Survey Units 782–01 and 782–02  Unadjusted (Gross) Measurements  (dpm/100 cm²)											
DCGL Value (Total Surface Contamination		metic an	Geon	Log-normal Geometric Median Mean		Arithmetic UCL <sub>95</sub>		Log-normal UCL <sub>95</sub>		Maximum		
by direct surface emission)	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02
100 dpm/100 cm <sup>2</sup>	12.68	12.13	9.53	10.05	10.95	11.10	15.93	14.56	19.59	16.92		
300 dpm/100 cm <sup>2</sup>											34.50	24.10

Table 4-2. Comparison of Background Adjusted Direct Static Survey Results to Applicable DCGLs

Survey Units 727–01 and 727–02												
Background Adjusted (Net) Measurements <sup>a</sup> (dpm/100 cm²)												
DCGL Value (Total Surface Contamination	Arithmetic Mean		Lognormal Geometric Mean		Median		Arithmetic UCL <sub>95</sub>		Log-normal UCL <sub>95</sub>		Maxi	imum
by direct surface emission)	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02
100 dpm/100 cm <sup>2</sup>	2.69 <sup>b</sup>	4.92 <sup>b</sup>	0.31 <sup>b</sup>	4.52 <sup>b</sup>	1.22 <sup>b</sup>	4.30 <sup>b</sup>	4.16 <sup>b</sup>	4.00 <sup>b</sup>	5.45 <sup>b</sup>	5.74 <sup>b</sup>		
300 dpm/100 cm <sup>2</sup>											20.1 <sup>b</sup>	13.8 <sup>b</sup>
Survey Units 782–01 and 782–02												
Background Adjusted (Net) Measurements •												
				(d	pm/100	cm²)						
DCGL Value (Total Surface Contamination	Arithmetic Mean		Log-normal Geometric Mean		Median		Arithmetic UCL <sub>95</sub>		Log-normal UCL <sub>95</sub>		Maximum	
by direct surface emission)	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02
100 dpm/100 cm <sup>2</sup>	5.25 <sup>b</sup>	0.43 <sup>b</sup>	4.42 <sup>b</sup>	0.53 <sup>b</sup>	6.13 <sup>b</sup>	2.78 <sup>b</sup>	4.88 <sup>b</sup>	-1.13 <sup>b</sup>	3.80 <sup>b</sup>	0.77 <sup>b</sup>		-
300 dpm/100 cm <sup>2</sup>											18.7 <sup>b</sup>	4.00 <sup>b</sup>
<ul> <li>Background corrected values are provided for information. They are well below the DCGL values and just slightly above background.</li> <li>The following background values were used to adjust the direct static measurement values:</li> </ul>												
				727-02	782-0		2-02					
Arithmetic Mean Log-normal Geome	7.4 5.1		.70 .97									
Median = Arithmetic UCL =	eulo iviea		5.28 6.28 8.01	14.23 15.10 18.01	4.8 4.8 11.0	2 8	1.97 3.32 5.69					
Log-normal UCL = Maximum =	<b>=</b>	1	8.14 1.30	16.66 22.50	15.7 15.8	9 16	.15 .10					

From the above data, it is evident that the surface contamination as measured by direct surface emission from the building surfaces in survey units 727–01, 727–02, 782–01, and 782–02 is well below the DCGL<sub>W</sub>. The IVC did not employ a scanning survey method in the independent verification sampling plan. Rather, the IVC has evaluated the scanning data collected by the Contractor to determine if the data supports the conclusions reached by the Contractor with respect to the DCGL<sub>EMC</sub>. However, it is interesting to note the maximum concentrations measured using direct static methods as they compare to the DCGL<sub>EMC</sub>. The maximum value measured in survey units 727–01, 727–02, 782–01, and 782–02 is substantially below the DCGL<sub>EMC</sub> and the background adjusted mean concentrations are significantly less. The independent verification data cannot exclude the possibility that localized concentrations of surface contamination might exist above the DCGL<sub>EMC</sub> value. But, given the number of measurements made, knowledge about the nature of the distribution of the data, and the large differences between the data metrics and the DCGL<sub>EMC</sub>, it can be inferred that the likelihood of encountering even moderately sized areas with concentrations exceeding the DCGL<sub>EMC</sub> is small.

#### 4.1.2 Smear Samples for Removable Surface Contamination

Smear samples are not subject to the influence of background radiation at the site, but the radiation counting instruments used to assay these samples are subject to background radiation levels at the counting laboratory and have inherent instrument backgrounds which are corrected by the laboratory processing the samples. Since the background corrections performed are not relevant to the conditions encountered in the survey units, only the background adjusted values are provided here in Table 4–3. The raw counting data can be referenced in the analytical laboratory report for the smear samples contained in Appendix C.

Table 4-3. Comparison of Smear Sample Results to Applicable DCGL<sub>W</sub>.

				Smear S		1 and 7 Result cm²)					
DCGL Value	Arithmetic Mean		Lognormal Geometric Mean		Median		Arithmetic UCL <sub>95</sub>	Lognormal UCL <sub>95</sub>	Maximum		
	727-01	727-02	727-01	727-02	727-01	727-02			727-01	727-02	
20 dpm/100 cm <sup>2</sup> Removable Surface Contamination	<4.72	<4.72	<4.72	<4.72	<4.72	<4.72	NA	NA	<4.72	<4.72	
				Smear S		1 and 7 Result cm²)					
Arithmetic Geometric Median UCL <sub>95</sub> Arithmetic Lognormal UCL <sub>95</sub> Maximum											
	782-01	782-02	782-01	782-02	782-01	782-02			782-01	782-02	
20 dpm/100 cm <sup>2</sup> Removable Surface Contamination	<4.84	<4.84	<4.84		<4.84	<4.84	NA	NA	<4.84	<4.84	
All 58 smear sample limit is presented for					w the de	tection li	mit for the anal	ysis. The metho	od detect	tion	

From the above data, it is evident that the removable surface contamination, as measured by smear sampling of the surfaces in survey units 727–01, 727–02, 782–01, and 782–02, is well below the DCGL.

#### 4.1.3 Surface Media Samples

As with smear samples, surface media samples are not subject to the influence of background radiation at the site, but have been corrected for the background present at the laboratory by the laboratory processing the samples. Again, since the background corrections performed are not relevant to the conditions encountered in the survey units, only the background adjusted values are provided here in Table 4–4. The raw counting data can be referenced in the analytical laboratory report for the surface media samples contained in Appendix C.

Table 4-4. Comparison of Surface Media Sample Results to Applicable DCGLs

				y Units ace Me (dp		nple R						
DCGL Value	Arithmetic Mean		Log-normal Geometric Mean		Median		Arithmetic UCL <sub>95</sub>		Log-normal UCL <sub>95</sub>		Maximum	
	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02	727-01	727-02
100 dpm/100 cm² Total Transuranic Activity by surface media sample	2.58	3.26	2.51	2.67	2.42	3.15	2.92	4.47	2.89	5.29	4.54	8.21
5,000 dpm/100 cm <sup>2</sup> Total Uranium Series Activity by surface media sample	44.79	52.65	37.99	28.29	40.66	60.56	57.91	72.54	65.25	281.8 5	121.3 2	101.9 8
				y Units ace Me (dp		nple R						
DCGL Value Arithmetic Geometric Median A Mean									Log-normal UCL <sub>95</sub>		Maximum	
	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02	782-01	782-02
100 dpm/100 cm² Total Transuranic Activity by surface media sample	1.42	1.32	1.27	0.98	1.16	1.10	1.79	1.79	1.88	1.94	2.68	4.79
5,000 dpm/100 cm² Total Uranium Series Activity by surface media sample	50.58	26.85	48.36	21.92	53.10	24.13	58.08	34.24	60.95	37.60	74.66	79.22

From the above data, it is notable that the maximum total transuranic activity contained on and in a thin veneer beneath the surface sampled is significantly below the allowable <u>mean</u> value. This is notable in that, as the maximum total transuranic activity collected within the entire survey unit, this measurement represents less than 5 percent of the allowable mean value. This data indicates that it is extremely unlikely that the building contains any added radioactivity and is considered safe for unrestricted release.

In fact, most of the samples measured for transuranic activity resulted in measured concentrations below the method detection limit for the analysis. Most of the transuranic activity

reported is attributed to americium-241 (Am-241) owing to interference in the energy window for Am-241 rather than americium activity. (See Appendix C for the method blank data indicating the activity showing up as Am-241 even when no americium is present.)

The total uranium series activity was consistently measured at concentrations exceeding the method detection limit even though the concentrations measured were consistently significantly below the applicable DCGL. The presence of detectable concentrations of uranium series nuclides does not, however, necessarily indicate that the activity is DOE contributed activity. In fact, isotopic ratios present in the samples support the position that the uranium series activity is naturally occurring radioactivity present in the construction materials from which the building was made. Nonetheless, because a decision was made during sampling plan design to avoid the need to make reference survey unit comparisons in order to statistically verify this assumption, all of this activity is herein assumed to be DOE contributed and is compared directly to the applicable DCGL. Even with this conservative assumption, it is clear that the residual uranium series activity on and in a thin veneer beneath the surface sampled is well below the DCGL.

#### **Summary of Field Sampling Data**

As evidenced above, each metric—the arithmetic average, logarithmic average, their respective UCL<sub>95</sub> estimates and the median value—is well below the applicable DCGL<sub>w</sub> concentration value. Moreover, the maximum value for each data set is well below not only the applicable DCGL<sub>EMC</sub> but also below the DCGL<sub>W</sub>. Based on the direct static measurements, removable smears sample results, and surface media sample results collected in the survey units selected for independent verification (727-01, 727-02, 782-01, and 782-02), there is no evidence of radiological surface contamination levels exceeding the selected DCGLs.

Thus, the first of the tests of the DQO decision rule—the residual radioactivity must not exceed the applicable DCGLs—has been verified.

### 5.0 Graphic Presentations of the Survey and Sampling Results

Graphics are a powerful and valuable tool used in reviewing the data collected. Graphic presentations—Normal Probability, Log Probability, and Probability Density Function Plots—have already been provided in Section 3.0 in support of the determination of the underlying distribution of each data set. In addition to these graphical treatments of each of the data sets collected, additional pictorial presentations are provided in Section 6.0 to assist the risk manager and decision maker in evaluating the data. Each form of graphic presentation provides a unique perspective or advantage in the data evaluation process.

#### 5.1 Posting Plots—Spatial Contamination Distribution Graphics

Posting plots are presented for visualizing the spatial contaminant distribution within the survey unit sampled and surveyed by the IVC. Trends in spatial distribution become evident when data is plotted in this manner. The results of each data set, normalized to units of dpm/100 cm<sup>2</sup>, are superimposed over the building surfaces. The walls and ceilings in the building are "unfolded" to form a contiguous surface segment, as when a cardboard box is unfolded and laid flat. Three posting plots are provided for each survey unit. One plot, (Figure 5–1) displays the 29 direct static surface contamination measurements made in survey units 727–01, 727–02, 782–01, and 782–02. The data used to generate these posting plots are "gross" measurements (not corrected for the mean background of 6.6, 14.8, 7.4, and 11.7 dpm/100 cm<sup>2</sup>, respectively) to avoid negative numbers. Figures 5–2 and 5–3 display the surface media sample results from each survey unit for transuranic and uranium series activities, respectively.

The posting plots confirm that no substantial spatial trends in residual activity are present.

#### 5.2 Histograms—Concentration Distribution Graphics

One of the oldest methods used for analyzing data set distributions is the histogram (or frequency plot). The data are divided into units, or bins, representing increments of activity. The data set is than sorted into these bins and the number of data points occurring in each bin (the frequency) is counted and then plotted using a bar graph. This presentation is designed to provide for visual means of assessing the symmetry and variability of the data set. When constructed correctly, the histogram will indicate if the data are skewed and will show the direction of skewness (EPA 1998). Figures 5–4, 5–5, 5–6, 5–7, and 5–8 display the histograms (technically frequency plots) for the background, direct static surface measurement, post-surface media sampling direct static surface measurement, surface media samples for transuranics, and surface media samples for uranium data sets, respectively from each survey unit.

### **Posting Plot: Direct Static Measurements**

SURVEY UNIT 727-01 MAP 1 OF 1





Figure 5–1. Posting Plot—Direct Static Surface Contamination Measurements

### **Posting Plot: Direct Static Measurements**

#### SURVEY UNIT 727-02 MAP 1 OF 1






Figure 5–1 (continued). Posting Plot—Direct Static Surface Contamination Measurements

## **Posting Plot: Direct Static Measurements**

SURVEY UNIT 782-01 MAP 1 OF 2




Figure 5-1 (continued). Posting Plot—Direct Static Surface Contamination Measurements

### **Posting Plot: Direct Static Measurements**

SURVEY UNIT 782-01 MAP 2 OF 2




Figure 5-1 (continued). Posting Plot—Direct Static Surface Contamination Measurements

### **Posting Plot: Direct Static Measurements**




Figure 5–1 (continued). Posting Plot—Direct Static Surface Contamination Measurements

## **Posting Plot: Direct Static Measurements**

SURVEY UNIT 782-02 MAP 2 OF 2




Figure 5–1 (continued). Posting Plot—Direct Static Surface Contamination Measurements

### Posting Plot: Surface Media Samples, Transuranic

SURVEY UNIT 727-01 MAP 1 OF 1





Figure 5-2. Posting Plot—Surface Media Samples, Transuranic Activity

### Posting Plot: Surface Media Samples, Transuranic

#### SURVEY UNIT 727-02 MAP 1 OF 1






Figure 5-2 (continued). Posting Plot—Surface Media Samples, Transuranic Activity

### Posting Plot: Surface Media Samples, Transuranic

SURVEY UNIT 782-01 MAP 1 OF 2




Figure 5-2 (continued). Posting Plot—Surface Media Samples, Transuranic Activity

### Posting Plot: Surface Media Samples, Transuranic

SURVEY UNIT 782-01 MAP 2 OF 2




Figure 5-2 (continued). Posting Plot—Surface Media Samples, Transuranic Activity

### Posting Plot: Surface Media Samples, Transuranic




Figure 5–2 (continued). Posting Plot—Surface Media Samples, Transuranic Activity

### Posting Plot: Surface Media Samples, Transuranic

SURVEY UNIT 782-02 MAP 2 OF 2




Figure 5-2 (continued). Posting Plot—Surface Media Samples, Transuranic Activity

### Posting Plot: Surface Media Samples, Uranium

SURVEY UNIT 727-01 MAP 1 OF 1





Figure 5-3. Posting Plot-Surface Media Samples, Uranium Series Activity

Posting Plot: Surface Media Samples, Uranium

#### SURVEY UNIT 727-02 MAP 1 OF 1






Figure 5-3 (continued). Posting Plot—Surface Media Samples, Uranium Series Activity

### Posting Plot: Surface Media Samples, Uranium

SURVEY UNIT 782-01 MAP 1 OF 2




Figure 5-3 (continued). Posting Plot—Surface Media Samples, Uranium Series Activity

### Posting Plot: Surface Media Samples, Uranium

SURVEY UNIT 782-01 MAP 2 OF 2




Figure 5-3 (continued). Posting Plot-Surface Media Samples, Uranium Series Activity

### Posting Plot: Surface Media Samples, Uranium




Figure 5–3 (continued). Posting Plot—Surface Media Samples, Uranium Series Activity

### Posting Plot: Surface Media Samples, Uranium

SURVEY UNIT 782-02 MAP 2 OF 2




Figure 5-3 (continued). Posting Plot-Surface Media Samples, Uranium Series Activity




Figure 5-4. Histogram-Instrument Background Measurements, 727-01




Figure 5-4 (continued). Histogram—Instrument Background Measurements, 727-02




Figure 5-4 (continued). Histogram—Instrument Background Measurements, 782-01




Figure 5-4 (continued). Histogram—Instrument Background Measurements, 782-02




Figure 5-5. Histogram—Direct Static Surface Contamination Measurements, 727-01




Figure 5–5 (continued). Histogram—Direct Static Surface Contamination Measurements, 727–02




Figure 5–5 (continued). Histogram—Direct Static Surface Contamination Measurements, 782–01




Figure 5–5 (continued). Histogram—Direct Static Surface Contamination Measurements, 782–02




Figure 5-6. Histogram—Post Surface Media Sampling Direct Static Surface Measurements, 727-01




Figure 5–6 (continued). Histogram—Post Surface Media Sampling Direct Static Surface Measurements, 727–02




Figure 5–6 (continued). Histogram—Post Surface Media Sampling Direct Static Surface Measurements, 782–01




Figure 5–6 (continued). Histogram—Post Surface Media Sampling Direct Static Surface Measurements, 782–02




Figure 5-7. Histogram—Surface Media Samples, Transuranic Activity, 727-01




Figure 5-7 (continued). Histogram—Surface Media Samples, Transuranic Activity, 727-02




Figure 5-7 (continued). Histogram—Surface Media Samples, Transuranic Activity, 782-01




Figure 5–7 (continued). Histogram—Surface Media Samples, Transuranic Activity, 782–02




Figure 5-8. Histogram—Surface Media Samples, Uranium Series Activity, 727-01




Figure 5-8 (continued). Histogram-Surface Media Samples, Uranium Series Activity, 727-02




Figure 5-8 (continued). Histogram—Surface Media Samples, Uranium Series Activity, 782-01




Figure 5-8 (continued). Histogram—Surface Media Samples, Uranium Series Activity, 782-02

Each of the histogram plots provides evidence of left shifted skewness in the data set with most data clustered around a non-discrete central concentration which is substantially below the applicable DCGL. This distribution and skewness is typical of environmental radioactivity data and supports the conclusion that the data distributions are best estimated by the log-normal distribution for the most part.

## 5.3 High-Low Graphs—Data Variability Graphics

A key element in the evaluation of the sampling and survey data is the variation within the data set. As the data variability increases, the ability of the risk manager to confidently make decisions about the true state of radiological contamination in the survey unit or building in relation to the applicable DCGL and null hypothesis decreases. When variability is small (or excessively large) relative to the difference between the mean and the DCGL, the risk managers can be confident in the decisions made using the data set provided. When evaluating data variability, it is important to know, first, that the data set contains a sufficiently large sample population (number of measurements). Retrospective power curves, demonstrating the "power" of the sign test to reject the null hypothesis with the actual sample size collected, are presented in Section 8.0. High-Low graphs are simple presentations showing the range between the upper and lower 95 percent confidence intervals about the geometric mean. Figures 5–9, 5–10, and 5–11 depict the variability observed in each type of data analyzed.

The pattern of a comparatively low central tendency, and small measure of data variability in each of the data sets presented provide substantial evidence that the estimates of the true mean residual radioactive concentrations presented are below the DCGLs. No DCGL is included in the 95 percent confidence intervals about the mean. The lack of significant variability in any of the data sets is also indicative of a lack of discretely distributed activity (supporting the conclusions of the posting plots above) and excellent precision in the analytical methods employed in the sampling and measurements processes. By presenting the three data sets made with the same instruments and procedures (background, direct static measurements, and post-surface media sampling direct static measurements), it is also evident that they report essentially equivalent measures of activity. In other words, the best estimates of surface activity as measured by direct surface emission are statistically indistinguishable from background.

Other visual presentations of the data are possible and may have been indicated if the data sets available were less robust than they actually are. However, the graphic treatment of the data presented here and in other sections is sufficient to enable the risk managers and decision maker to make confident determinations respecting the data.




Figure 5–9. High-Low Graphs—Direct Static Surface Measurements, 727–01




Figure 5–9 (continued). High-Low Graphs—Direct Static Surface Measurements, 727–02




Figure 5-9 (continued). High-Low Graphs-Direct Static Surface Measurements, 782-01




Figure 5-9 (continued). High-Low Graphs—Direct Static Surface Measurements, 782-02




Figure 5-10. High-Low Graphs-Surface Media Samples, Transuranic Activity, 727-01




Figure 5–10 (continued). High-Low Graphs—Surface Media Samples, Transuranic Activity, 727–02




Figure 5-10 (continued). High-Low Graphs-Surface Media Samples, Transuranic Activity, 782-01




Figure 5–10 (continued). High-Low Graphs—Surface Media Samples, Transuranic Activity, 782–02




Figure 5–11. High-Low Graphs—Surface Media Samples, Uranium Series Activity, 727–01




Figure 5–11(continued). High-Low Graphs—Surface Media Samples, Uranium Series Activity, 727–02




Figure 5-11(continued). High-Low Graphs-Surface Media Samples, Uranium Series Activity, 782-01




Figure 5–11(continued). High-Low Graphs—Surface Media Samples, Uranium Series Activity, 782–02

# 6.0 Quality Control Sampling Results and Analysis

An important aspect of any sampling plan is the effort made to assure the quality of data collected. The independent verification process as a whole is a quality assurance method in itself. Thus, it was critical to assure the quality of all of the independent verification data through quality checks and controls, calibrations, training, and qualification of laboratories and services used. The objective of independent verification for the Building 779 Cluster final status radiological surveys, added an element of quality assurance to the design of the sampling evolution. In addition to designing quality checks and controls into the independent verification sampling, the IVC provided for quality control checks to assess the quality of the Contractor's data.

The IV SAP distinguished these two principle quality control objectives in the design of the sampling plan. Stage I quality control sampling was designed to assess the quality of the data collected by the Contractor. Stage II quality control sampling was designed to assess the quality of the data collected by the IVC. In each case where QC samples were used, either for Stage I or Stage II sampling, the samples were maintained under chain-of-custody control from the time they were prepared until they were introduced to either the Contractor's or IVC's sample batch. Tamper seals and locked storage were employed when samples were not in the physical custody of the IVC's Field Team Leader.

The IV SAP specifies quality control sampling to be performed over the duration of the Final Status Survey performance for all survey units in the Building 779 Cluster rather than for each specific building. This report, specific to Buildings 727, 782, and 783, does not contain every element of QC data planned for the cluster of buildings as a whole. The costs associated with implementing separate QC sampling for each building in the 779 Cluster was determined to be prohibitive and unnecessary. Instead, a cumulative assessment and presentation of quality control data is presented with each successive Independent Verification Report of the Contractor's Final Status Survey.

This section of the report presents the quality control data collected and measures employed to assure that quality objectives in the design of the sampling plan were achieved. Section 7.0 assesses the overall data quality against the published or industry accepted data quality indicators.

## 6.1 Stage I—Independent Quality Control of the Contractor's Sampling

#### 6.1.1 Smear Samples

The IVC provided smear samples to the Contractor for measurement and analysis by the Contractor's selected instrumentation and methods. An unopened package of smear sample media was obtained from the Contractor prior to the start of the independent verification of Buildings 727, 782, and 783. The IVC assigned a series of these as "blanks" and spiked a second series of smears with three different concentrations of an alpha emitting transuranic nuclide (one which is part of the nuclide mix identified as a contaminant of concern for the 779 Cluster).

The spikes were not certified as containing traceable concentrations of the nuclide added. Thus, the spikes do not provide a measure of accuracy directly. Accuracy is established for the instrument measuring the activity on the smears by the RFETS (or contract laboratory, if used)

calibration and analysis procedures. Instead, the spikes serve to provide a comparison between the results achieved by the Contractor and those achieved by the IVC.

The blank smears test the ability of the counting instrument used to distinguish between background and added activity as well as the ability of the counting technician and sample handling process to prevent cross-contamination.

The IVC introduced 20 blank smears and 21 spiked smears into the Contractor's smear sampling batch during the final status survey of the Building 779 Cluster. The QC samples were packaged and identified exactly as the Contractor's procedure dictated. Because the Contractor used the same technician to collect and analyze the smear samples they collected, it was not possible to present a double blind set of QC samples. He would know when samples other than those he personally collected were introduced. They were presented as a single blind set of QC samples. The Contractor's counting technician was not aware of the objective of the samples, nor the fact that some were blanks and some were spikes.

The blank smears were prepared by wiping a clean, unaffected, and uncontaminated surface. Each was then packaged individually, assigned a unique QC sample number, and physically controlled to ensure custody and integrity. The spiked smears were prepared by pipetting liquid standard concentrations onto a smear filter disc.

Spiked smears were prepared with three different quantities of radioactivity to provide a range of gross alpha radioactivity concentrations over the range expected to be encountered in the Final Status Survey. This range is necessarily small and near zero for a Final Status Survey.

The QC samples provided to the Contractor were measured and reported to the IVC. Table 6–1 provides a crosswalk between the IVC and Contractor assigned sample numbers and presents a summary of the results obtained by the Contractor and the GJO Analytical Laboratory which establishes a cumulative cross comparison as a measure of the relative accuracy of the instruments and methods employed.

From Table 6–1, it is seen that all but one of the 20 blank smears returned results indicating no detectable radioactivity or activity less than the reported MDA. As reported in the IVC Sampling and Survey Report, Building 729 (DOE 1999b), sample #2C was identified by the Contractor as having 4.5 dpm/100 cm<sup>2</sup>.

Each of the 21 spiked smear samples yielded measurable radioactivity. The 21 spiked activities were differentiated from one another in the Contractor's assay. Yet, samples #6C, #13C, and #16C were reported to have activity below the MDA for the method. As reported in the IVC Sampling and Survey Report, Building 729 (DOE 1999b), this might be caused by too short of a counting time (background, sample, or both) to adequately distinguish between background and low-level activity. Smears #6C and #13C were spiked to the lowest concentration of the three and were expected to return the lowest measurement of activity among the three. Likewise, the highest spiked activity returned the highest measurement of activity among the three. Another possible explanation for the lower than expected result on these smear samples is in the method used to prepare the spiked samples. As mentioned earlier, a liquid standard containing an alpha emitting transuranic nuclide was pipetted onto the smear and then allowed to dry. It is likely that a significant fraction of the total activity deposited migrated to a depth in the smear matrix that the activity was attenuated by the sample media itself.

Table 6-1. Results of the Contractor's Assay of QC Smear Samples Provided by the IVC

IVC Sample ID#	Contractor Assigned Sample ID#	Sample Type	Contractor Reported MDA (dpm)	IVC Reported MDA (dpm)	Contractor Reported Results (dpm)	IVC Reported Results (dpm)
NED428	1C	Blank	4.1	8	0.0	.8
NED429	2C	Blank	4.1	8	4.5	a
NED430	3C	Blank	4.1	a	0.0	á
NDL431	4C	Blank	10.3	ā	0.3	а
NED432	5C	Blank	8.5	a	-0.6	8
NDL289	6C	Spike	4.1	ā	1.5	a
NDL290	7C	Spike	8.5	5.3	9.9	24.1
NDL291	8C	Spike	4.1	а	28.5	a
NDL292	9C	Spike	8.5	5.4	30.9	54.8
NDL293	10C	Spike	4.1	8	30.9	<b>a</b>
NDL294	12C	Spike	10.3	5.5	52.9	96.5
259740	42C	Blank	8.5	4.9	-0.6	1.1
259741	43C	Blank	8.5	4.9	-0.6	0.3
259742	44C	Blank	8.5	4.9	-0.6	0.3
259743	45C	Blank	7.5	4.8	0.9	1.1
259744	46C	Blank	7.5	4.9	0.9	0.3
259745	47C	Blank	7.5	4.9	-0.6	0.3
259746	48C	Blank	7.5	4.9	0.9	0.3
259747	49C	Blank	7.5	4.9	0.9	-0.4
259748	50C	Blank	8.8	4.9	0.0	0.3
259749	66C	Blank	8.8	4.9	0.0	0.3
259750	67C	Blank	8.8	4.9	0.0	0.3
259751	68C	Blank	8.8	4.9	0.0	0.3
259752	69C	Blank	8.8	5.0	1.5	-0.5
259753	70C	Blank	8.8	4.9	1.5	0.3
259754	81C	Blank	8.8	4.9	1.5	1.9
259770	13C	Spike	8.5	4.9	6.9	7.3
259771	14C	Spike	8.5	5.0	12.9	12.0
259772	15C	Spike	7.5	4.9	8.4	12.8
259773	16C	Spike	7.5	5.0	3.9	20.6
259774	17C	Spike	7.5	5.0	12.9	12.8
259775	18C	Spike	8.5	5.0	23.4	31.5
259776	19C	Spike	10.3	5.1	27.3	38.4
259777	20C	Spike	8.8	5.0	19.7	36.9
259778	21C	Spike	8.8	5.0	34.8	32.2
259779	22C	Spike	8.8	5.0	30.3	26.8
259780	28C	Spike	10.3	5.2	73.9	96.8
259781	29C	Spike	10.3	5.2	84.4	96.1
259782	30C	Spike	8.8	5.4	87.9	87.4
259783	26C	Spike	8.8	5.3	75.8	94.5
259784	27C	Spike	8.8	5.3	77.3	76.5

<sup>&</sup>lt;sup>a</sup> Smear samples were disposed of by the Contractor before being counted by the GJO Analytical Laboratory. The Contractor counted the smears with an Eberline model SAC-4 alpha smear counters (serial #1069, 1406, and 814). The measured background was 0.2, 0.4, and 0.3 cpm respectively. The efficiency was established at 33.3 percent.

#### 6.1.2 Surface Media Samples

A total of 23 Stage I and 23 Stage II surface media QC samples have been analyzed for a cumulative comparison. Surface Media QC samples were made in duplicate (one to be introduced in the Contractor's sample batch and the other in the IVC's sample batch). Before being introduced by either the Contractor or IVC, they are analyzed by the GJO Analytical Laboratory. A comparison of results between the initial count by the GJO Analytical Laboratory, the Contractor's Laboratory, and the GJO Analytical Laboratory when the QC sample was recounted when introduced with the IVC sample batch is presented in Table 6–2.

Identification numbers assigned to samples provided to the Contractor on October 9, 1999 (samples 259868 – 259872) were incorrectly transferred as Contractor ID #s 99R0317–021, 99R0317–022, 99R0317–018, 99R0317–019, and 99R0317–020, respectively. Samples 259868 – 259872 should have been transferred as 99R0317–018, 99R0317–019, 99R0317–020, 99R0317-021 and 99R0317–022, respectively. Analytical results are correctly matched to the corresponding sample identification number in Table 6–2 of this report.

Table 6-2. Results of the Contractor's Assay of QC Surface Media Samples Provided by the IVC

Surface Me	dia Sample	Am-241	Pu-238	Pu- 239/240	U-234	U-235	U238
	•			P	Ci/g		
GJO Lab ID#	259851	0.01	0.02	0.03	0.56	0.07	0.53
IVC ID#	MED0000213	0.01	0.01	0.02	0.55	0.04	0.56
Contractor ID#	99A9022-019	0.12	a	0.13	0.68	0.14	0.39
GJO Lab ID#	259852	0.03	0.03	0.02	0.81	0.05	0.77
IVC ID#	MED0000162	0.03	0.02	0.02	0.88	0.06	0.75
Contractor ID#	99A8940-019	0.04	8	0.05	0.59	0.03	0.80
GJO Lab ID#	259853	0.24	0.01	0.51	2.07	0.09	2.04
IVC ID#	MED0000163	0.29	0.02	0.49	2.26	0.13	2.12
Contractor ID#	99A8940-020	0.28	a	0.69	1.78	0.19	2.14
GJO Lab ID#	259854	0.30	0.01	0.46	0.71	0.04	0.76
IVC ID#	MED0000214	0.30	0.01	0.35	0.62	0.04	0.66
Contractor ID#	99A9009-020	0.09	а	0.54	0.41	-0.01	0.38
GJO Lab ID#	259855	0.18	0.04	0.37	0.53	0.04	0.54
IVC ID#	MED0000271	0.20	0.04	0.39	0.36	0.06	0.49
Contractor ID#	99A9009-019	0.28	а	0.35	0.28	-0.01	0.33
GJO Lab ID#	259856	0.21	0.02	0.34	0.99	0.05	0.86
IVC ID#	MED0000243	0.17	0.01	0.35	0.71	0.06	0.87
Contractor ID#	99A9460-021	0.20	а	0.36	0.69	0.02	1.08
GJO Lab ID#	259857	0.31	0.01	0.43	0.99	0.08	0.89
IVC ID#	MED0000200	0.18	0.02	0.33	0.88	0.05	0.90
Contractor ID#	99A9460-022	0.16	a a	0.26	0.80	-0.01	0.76

Table 6–2 (continued). Results of the Contractor's Assay of QC Surface Media Samples Provided by the IVC

Surface Media Sample		Am-241	Pu-238	Pu- 239/240	U-234	U-235	U238		
	· · · · · · · · · · · · · · · · · · ·	PCi/g							
GJO Lab ID#	259858	0.83	0.02	1.40	1.31	0.08	1.30		
IVC ID#	MED0000180	0.78	0.02	1.43	1.56	0.06	1.34		
Contractor ID#	99A9460-023	0.63	8	1.16	1.03	0.00	1.04		
CONTRACTOR ID#	99/19-100-023	0.03	<u> </u>	1.10	1.03	0.09	1.04		
GJO Lab ID#	259859	0.82	0.02	1.66	1.40	0.05	1.33		
IVC ID#	MED0000189	0.91	0.02	1.70	1.34	0.06	1.34		
Contractor ID#	99A9460-024	0.83	a	1.74	1.56	0.06	1.30		
<del></del>	<u> </u>	<del></del>		<del> </del>		<del>! </del>	<u> </u>		
GJO Lab ID#	259860	0.55	0.02	1.23	0.83	0.07	0.68		
IVC ID#	MED0000215	1.37	0.02	1.11	0.71	0.04	0.72		
Contractor ID#	99A9460-025	0.72	a	0.92	0.57	-0.01	0.44		
GJO Lab ID#	259861	1.12	0.01	2.36	1.02	0.06	1.03		
IVC ID#	MED0000164	1.21	0.01	1.97	0.95	0.04	0.95		
Contractor ID#	99A9460-026	0.80	a	2.21	0.67	0.05	0.62		
							·		
GJO Lab ID#	259862	0.49	0.03	0.78	1.42	0.07	1.18		
IVC ID#	MED0000216	0.60	0.03	1.05	1.67	0.11	1.62		
Contractor ID#	99A9460-027	0.29	8	0.53	1.17	0.03	1.05		
0101.1.10#	1050000					r			
GJO Lab ID#	259863	1.00	0.08	1.50	0.91	0.10	1.04		
IVC ID#	MED0000217	1.33	0.02	2.40	0.83	0.08	0.78		
Contractor ID#	99A9460-028	1.08		2.31	0.57	0.02	0.58		
GJO Lab ID#	259864	1.71	0.04	3.08	0.58	0.10	0.47		
IVC ID#	MED0000165	1.65	0.01	2.99	0.58	0.04	0.56		
Contractor ID#	99A9460-029	1.66	a a	3.41	0.52	0.03	0.32		
O THE OCCUPANT OF THE OCCUPANT	1 00/10/00/020	1.00		0.41	0.02	0.00			
GJO Lab ID#	259865	0.20	0.03	0.66	1.40	0.13	1.03		
IVC ID#	MED0000244	0.10	0.01	0.24	1.21	0.01	0.85		
Contractor ID#	99A9460-030	0.76	a	1.20	0.88	-0.00	0.84		
				<u> </u>		·			
GJO Lab ID#	259866	1.27	0.04	2.60	0.59	0.09	0.38		
IVC ID#	MED0000451	1.21	-0.00	2.44	0.51	0.05	0.48		
Contractor ID#	99R0317-016	1.10	a	1.72	0.68	0.06	0.65		
GJO Lab ID#	259867	1.15	0.05	2.46	1.63	0.11	1.64		
IVC ID#	MED0000272	0.66	0.02	2.08	1.22	0.06	1.16		
Contractor ID#	99R0317-017	0.80	a	2.61	1.52	0.24	1.27		
0101-1-10"	Lacasas								
GJO Lab ID#	259868	1.17	0.02	3.28	0.64	0.08	0.64		
IVC ID#	MED0000166	0.69	0.02	1.69	0.53	0.04	0.49		
Contractor ID#	99R0317-021	0.75	a	1.32	0.57	0.16	0.75		

Table 6–2 (continued). Results of the Contractor's Assay of QC Surface Media Samples

Provided by the IVC

Surface Me	dia Sample	Am-241	Pu-238	Pu- 239/240	U-234	U-235	U238
	_			PC	Ci/g		
GJO Lab ID#	259869	3.39	0.04	5.27	0.56	0.07	0.61
IVC ID#	MED0000452	4.70	-0.01	10.94	0.79	0.05	0.57
Contractor ID#	99R0317-022	3.11	a	6.01	0.45	0.03	0.67
GJO Lab ID#	259870	4.02	0.04	8.59	1.63	0.11	1.60
IVC ID#	MED0000273	7.08	0.02	11.93	1.64	0.07	1.49
Contractor ID#	99R0317-018	5.03	a	а	1.41	0.32	1.77
GJO Lab ID#	259871	8.73	0.05	16.32	2.02	0.09	1.84
IVC ID#	MED0000218	7.30	0.01	12.92	1.93	0.11	1.86
Contractor ID#	99R0317-019	6.90	а	14.2	2.41	0.16	2.30
GJO Lab ID#	259872	14.48	0.04	28.93	2.44	0.10	2.28
IVC ID#	MED0000245	13.38	0.04	26.22	2.32	0.13	2.06
Contractor ID#	99R0317-020	14.20	8	22.20	2.50	0.19	2.42
GJO Lab ID#	259873	3.90	0.06	6.62	1.58	0.10	1.65
IVC ID#	MED0000274	3.30	0.01	6.78	1.61	0.45	1.41
Contractor ID#	99A9022-020	5.02	å	8.28	1.14	0.76	1.83
<sup>a</sup> The Contractor	did not analyze fo	or this radion	uclide.	·			· ·

#### 6.1.3 Direct Surface Emission Measurements

The Contractor and the IVC chose to utilize the same response check source to test the response of instruments used to make direct surface emission measurements (Table 6–3). This enabled the comparison of routine instrument response checks using the same isotope, geometry, and activity. Three instrument systems were employed to make direct surface emission measurements during the Final Status Survey of Buildings 727, 782, and 783. The Contractor employed a subcontractor, Millennium Services, who used a proprietary system (SCM/SIMS) developed by Shonka Research Associates to perform the scan surveys. The SCM is fundamentally a gas proportional counter and the SIMS is the survey information management software. The Contractor also used a NE Electra with a DP-6 dual phosphor scintillation probe to make direct static surface measurements for comparison with the DCGL<sub>W</sub>. The IVC used the Eberline model E-600 multi-purpose survey instrument with a HP-100 gas proportional probe.

Table 6-3. Comparison of Response of Instruments Used to Make Direct Surface Measurements

Parameter	Millennium SCM/SIMS	NE Electra w/ DP-6	EberlineE-600 w/HP-100
Number of Measurements	20	20	20
2π Source Surface Emission Rate	1604	1604	1604
Mean Activity Observed	1315	1261	1302
Standard Deviation	280	83	53
Coefficient of Variation	0.22	0.07	0.04

# 6.2 Stage II—Quality Control of the Independent Verification Sampling

Stage II QC sampling is associated specifically with the IVC's field sampling and serve to establish confidence in the independent verification sampling results rather than correlate them with the Contractor's results.

#### 6.2.1 Smear Samples

The IVC provided smear samples to the GJO Analytical Laboratory for measurement and analysis. Smear sample media was reserved by the IVC prior to the start of the independent verification of Buildings 727, 782, and 783. A series of these were assigned as "blanks" and a second series of smears was spiked with three different concentrations of an alpha emitting transuranic isotope of the nuclide mix identified as a contaminant of concern for the 779 Cluster.

As with the spikes prepared for the Contractor, the spikes were not certified as containing traceable concentrations of the nuclide added. Thus, the spikes do not provide a direct measure of accuracy. Instead, the spikes serve to provide a measure of confidence in the laboratory's ability to detect radioactivity and to establish a basis for subsequent comparison between the results achieved by the Contractor and those achieved by the IVC.

The IVC introduced seven blank and five spiked smears into the smear sampling batch during the independent verification survey of survey units 727–01, 727–02, 782–01, and 782–02. The entire batch of smears was then provided to the GJO Analytical Laboratory. The QC samples were packaged and identified exactly as those samples collected in the survey unit and were not distinguishable to the analyst. Because the IVC used an independent laboratory to assay smears, and because the technician collecting the field smears was not involved with preparing, handling, or counting smears, it was possible to present the QC samples along with the field samples as a double blind set.

The blank smears were prepared by wiping a clean, unaffected, and uncontaminated surface. It was then packaged individually, assigned a unique QC sample number, and physically controlled to ensure custody and integrity. The spiked smears were prepared by pipetting liquid standard concentrations onto a smear filter disc.

Spiked smears were prepared with three different quantities of radioactivity to provide a range of gross alpha radioactivity concentrations over the range expected to be encountered in the Independent verification of the Final Status Survey. The range was, again, small and near zero. The QC samples provided to the IVC's laboratory were measured and reported to the IVC (Appendix F). Table 6-4 provides a crosswalk between the IVC and GJO Analytical Laboratory assigned sample numbers and presents a summary of the cumulative results obtained by the laboratory.

From Table 6–4 it is seen that each blank smear returned results indicating no detectable radioactivity in excess of the method detection limit. Each spiked smear sample except sample 259759 yielded measurable radioactivity. They were differentiated from one another and the lowest result corresponded to the smear with the lowest spiked concentration while the highest result corresponded to the smear with the highest spiked concentration. The initial laboratory measurement of sample 259759 was 5.75 dpm with a counting error of 4.48 dpm. It is

Table 6-4. Results of the GJO Analytical Laboratory Assay of QC Smear Samples Provided by the IVC

IVC QC	IVC Transfer	Laboratory Assigned	Sample	Reported Re	sults (Gross α, dpm) <sup>a</sup>
Sample ID#	Sample ID#	Sample ID#	Type	MDA	Measured Activity
NDL 290	SMR0000260	263395	Spike	5.33	24.05
NDL 292	SMR0000261	263396	Spike	5.41	54.84
NDL 294	SMR0000262	263397	Spike	5.49	96.51
NDL 295	SMR0000593	258235	Spike	5.17	5.68
NDL 296	SMR0000168	261284	Spike	4.88	5.00
NDL 297	SMR0000592	258234	Spike	5.19	15.81
NDL 298	SMR0000169	261285	Spike	5.00	18.98
NDL 299	SMR0000591	258233	Spike	5.35	53.92
NDL 300	SMR0000170	261286	Spike	5.11	41.54
SMR0000594	SMR0000594	258236	Blank	5.13	Op
SMR0000595	SMR0000595	258237	Blank	5.17	O <sup>b</sup>
SMR0000596	SMR0000596	258238	Blank	5.11	O <sub>p</sub>
259725	SMR0000167	261283	Blank	4.88	O <sub>p</sub>
259726	SMR0000166	261282	Blank	4.91	O <sup>b</sup>
259727	SMR0000165	261281	Blank	4.91	0,
259728	SMR0000591	263362	Blank	5.30	O <sup>b</sup>
259729	SMR0000592	263363	Blank	5.34	0 <sup>b</sup>
259730	SMR0000264	264867	Blank	4.67	0°
259731	SMR0000451	264942	Blank	4.62	O <sup>b</sup>
259732	SMR0000452	264943	Blank	4.65	O <sup>b</sup>
259733	SMR0000598	263827	Blank	4.73	О <sup>Б</sup>
259734	SMR0000599	263828	Blank	4.66	O°
259735	SMR0000454	264974	Blank	4.66	<b>0</b> b
259736	SMR0000455	264975	Blank	4.65	O <sub>p</sub>
259737	SMR0000456	264976	Blank	4.68	O <sub>p</sub>
259738	SMR0000599	261638	Blank	4.91	O <sub>p</sub>
259739	SMR0000453	264944	Blank	4.66	O <sub>p</sub>
259755	SMR0000164	261280	Spike	4.88	5.77
259756	SMR0000598	261639	Spike	4.97	7.30
259757	SMR0000597	261640	Spike	4.91	6.55
259758	SMR0000596	261641	Spike	4.91	7.33
259759	SMR0000263	264837	Spike	4.62	O <sup>6</sup>
259760	SMR0000262	264836	Spike	4.79	26.62
259761	SMR0000163	261279	Spike	4.86	23.71
259762	SMR0000266	264869	Spike	4.82	14.02
259763	SMR0000593	263364	Spike	5.39	25.85
259764	SMR0000594	263365	Spike	5.46	41.23
259765	SMR0000595	263361	Spike	5.39	35.82
259766	SMR0000265	264868	Spike	4.84	46.17
259767	SMR0000261	264835	Spike	4.79	38.66
259768	SMR0000597	263826	Spike	4.93	52.04
259769	SMR0000162	261278	Spike	5.13	56.34
*The analytical re-	nort presented results	in pCi per sample. The results h	ava been con	verted to dom for n	constation in this table. Since

The analytical report presented results in pCi per sample. The results have been converted to dpm for presentation in this table. Since the smears were collected over a 100 cm<sup>2</sup> area, the results in dpm are equivalent to dpm/100 cm<sup>2</sup>.

The GJO Analytical Laboratory counted the smears with a Canberra low background automated scaler, model 2404 smear counter. The measured background was 0.097 cpm over 60 minutes. Sample count time was 6 minutes. The alpha efficiency was established at 21.39 percent.

<sup>&</sup>lt;sup>b</sup>The laboratory formally reported values as less than the detection limit (<MDA) but provided the raw supporting data in the complete analytical report. Each of the blank QC samples resulted in a negative net count rate.

likely that a significant fraction of the total activity deposited migrated to a depth in the smear matrix that the activity was attenuated by the sample media itself.

Again, no statement can be made about the accuracy of the results reported by the IVC's laboratory subcontractor but results were in line with those expected. The fact that the GJO Analytical Laboratory apparently used longer counting times than did the Contractor when measuring smears adds weight to the possibility that insufficient counting time may be at the root of the disparity in the results on QC smear samples reported by the Contractor.

As reported in the IVC Sampling and Survey Report, Building 729 (DOE 1999b) several smear samples submitted to the GJO Analytical Laboratory and to the Contractor were disposed of by the laboratories before they could be collected and provided to the other laboratory for a cross comparison of laboratory measurements. To remedy this situation, measured sets of QC samples (one for the Contractor and the other for the IVC) were prepared by the GJO Analytical Laboratory and were introduced to the Contractor's and IVC's sampling protocol. This provided a cumulative cross comparison of the relative accuracy of the instruments and methods employed to assay smears.

Aside from the QC measures interjected by the IVC blind to the laboratory, the independent verification sampling benefits from the internal quality control applied to the measurement process within the laboratory. Three measurement quality controls are employed for each batch of smears. The laboratory inserts a preparation blank (PB), a laboratory control sample (LCS) and a continuing calibration verification (CCV) for each batch of 21 smears. In this case, because each batch was larger than 21 smears, eight sets of laboratory initiated QC measurements were made. Table 6–5 summarizes the internal QC measurements made for the smears from survey units 727–01, 727–02, 782–01, and 782–02.

The internal QC data presented in Table 6–5 shows excellent agreement with the results expected.

## 6.2.2 Surface Media Samples

Stage II QC samples for the surface media sample sets are presented in Table 6–2. In addition to the Stage II QC samples, the IVC's laboratory performed internal quality control measurements to assess the quality of the data produced. Three measurement quality controls were employed for each of the three element groups (Am, Pu, U) processed for each survey unit. The laboratory inserted one PB, one LCS, and processed a duplicate of one randomly selected field sample for each sample batch. In all, there were 15 PB, 15 LCS, and five duplicate measurements made. Table 6–6 summarizes the PBs QC measurements made. Table 6–7 summarizes the LCS measurements and Table 6–8 presents the duplicate sample measurements. A regression analysis of the cumulative laboratory measurements was performed to assess the comparability between the first and duplicate measurements and is graphically presented in Figure 6–1.

Table 6-5. Results of the IV Laboratory Internal QC Measurements for Smear Samples

	Expected Results	MDA	Measured Activity
QC Sample Type		(Gross α, dpi	m)
Preparation Blank	<mda< td=""><td>4.64</td><td>0.33</td></mda<>	4.64	0.33
Preparation Blank	<mda< td=""><td>4.64</td><td>-0.42</td></mda<>	4.64	-0.42
Preparation Blank	<mda< td=""><td>4.66</td><td>-0.42</td></mda<>	4.66	-0.42
Preparation Blank	<mda< td=""><td>4.68</td><td>-0.42</td></mda<>	4.68	-0.42
Preparation Blank	<mda< td=""><td>4.84</td><td>7.79<sup>a</sup></td></mda<>	4.84	7.79 <sup>a</sup>
Preparation Blank	<mda< td=""><td>4.71</td><td>-0.44</td></mda<>	4.71	-0.44
Preparation Blank	<mda< td=""><td>4.66</td><td>-0.42</td></mda<>	4.66	-0.42
Preparation Blank	<mda< td=""><td>4.64</td><td>0.33</td></mda<>	4.64	0.33
Laboratory Control Sample	475	8.57	423
Laboratory Control Sample	475	8.46	419
Laboratory Control Sample	475	8.37	440
Laboratory Control Sample	475	8.44	410
Laboratory Control Sample	475	8.19	454
Laboratory Control Sample	475	8.68	448
Laboratory Control Sample	475	8.52	451
Laboratory Control Sample	475	8.57	423
Continuing Calibration Verification	2,220	22.27	2,182
Continuing Calibration Verification	2,220	22.04	2,210
Continuing Calibration Verification	2,220	22.00	2,265
Continuing Calibration Verification	2,220	22.02	2,221
Continuing Calibration Verification	2,220	22.24	2,343
Continuing Calibration Verification	2,220	22.27	2,294
Continuing Calibration Verification	2,220	22.36	2,165
Continuing Calibration Verification	2,220	22.27	2,182

The analytical report presented results in pCi per sample. The results have been converted to dpm for presentation in this table. Since the smears were collected over a 100 cm<sup>2</sup> area, the results in dpm are equivalent to dpm/100 cm<sup>2</sup>.

 $<sup>^{\</sup>rm a}$  The uncertainty of this measurement is  $\pm$  5.04 dpm.

Table 6-6. Results of the IV Laboratory Internal Blank QC Measurements for Surface Media Samples

Preparation Blank QC Sample (PB)	Expected Results	MDA	Measured Activity
Freparation Blank QC Sample (FB)	·	dpm / samp	
Am-241	<mda< td=""><td>0.02</td><td>0.04<sup>a</sup></td></mda<>	0.02	0.04 <sup>a</sup>
Am-241	<mda< td=""><td>0.01</td><td>0.03<sup>a</sup></td></mda<>	0.01	0.03 <sup>a</sup>
Am-241	<mda< td=""><td>0.04</td><td>0.03</td></mda<>	0.04	0.03
Am-241	<mda< td=""><td>0.04</td><td>0.07<sup>a</sup></td></mda<>	0.04	0.07 <sup>a</sup>
Am-241	<mda< td=""><td>0.03</td><td>0.03</td></mda<>	0.03	0.03
Pu-238	<mda< td=""><td>0.03</td><td>0.01</td></mda<>	0.03	0.01
Pu-238	<mda< td=""><td>0.02</td><td>-0.003</td></mda<>	0.02	-0.003
Pu-238	<mda< td=""><td>0.04</td><td>-0.004</td></mda<>	0.04	-0.004
Pu-238	<mda< td=""><td>0.04</td><td>0.006</td></mda<>	0.04	0.006
Pu-238	<mda< td=""><td>0.02</td><td>0.007</td></mda<>	0.02	0.007
Pu-239/240	<mda< td=""><td>0.04</td><td>0.01</td></mda<>	0.04	0.01
Pu-239/240	<mda< td=""><td>0.02</td><td>0.009</td></mda<>	0.02	0.009
Pu-239/240	<mda< td=""><td>0.04</td><td>0.002</td></mda<>	0.04	0.002
Pu-239/240	<mda< td=""><td>0.05</td><td>0.01</td></mda<>	0.05	0.01
Pu-239/240	<mda< td=""><td>0.02</td><td>0.02</td></mda<>	0.02	0.02
U-234	<mda< td=""><td>0.04</td><td>0.007</td></mda<>	0.04	0.007
U-234	<mda< td=""><td>0.04</td><td>0.005</td></mda<>	0.04	0.005
U-234	<mda< td=""><td>0.04</td><td>0.02</td></mda<>	0.04	0.02
U-234	<mda< td=""><td>0.04</td><td>0.04</td></mda<>	0.04	0.04
U-234	<mda< td=""><td>0.04</td><td>0.02</td></mda<>	0.04	0.02
U-235	<mda< td=""><td>0.04</td><td>0.009</td></mda<>	0.04	0.009
U-235	<mda< td=""><td>0.05</td><td>-0.01</td></mda<>	0.05	-0.01
U-235	<mda< td=""><td>0.04</td><td>0.005</td></mda<>	0.04	0.005
U-235	<mda< td=""><td>0.06</td><td>0.004</td></mda<>	0.06	0.004
U-235	<mda< td=""><td>0.05</td><td>0.009</td></mda<>	0.05	0.009
U-238	<mda< td=""><td>0.04</td><td>0.02</td></mda<>	0.04	0.02
U-238	<mda< td=""><td>0.04</td><td>0.02</td></mda<>	0.04	0.02
U-238	<mda< td=""><td>0.04</td><td>0.02</td></mda<>	0.04	0.02
U-238	<mda< td=""><td>0.03</td><td>0.03</td></mda<>	0.03	0.03
U-238	<mda< td=""><td>0.04</td><td>0.02</td></mda<>	0.04	0.02
The error in the reported result includes the M	MDA for the measurement.	<u> </u>	

Table 6-7. Results of the IV Laboratory Internal LCS QC Measurements for Surface Media Samples

Laboratory Control QC Sample (LCS)	Expected Results	MDA	Measured Activity
	pCi/mL (Am	and Pu)	pCi/L (U)
Am-241	4.74	0.02	4.70
Am-241	4.74	0.05	4.82
Am-241	4.74	0.03	4.47
Am-241	4.74	0.08	4.83
Am-241	4.74	0.04	4.80
Pu-238	10.67	0.05	10.78
Pu-238	10.67	0.05	10.50
Pu-238	10.67	0.06	11.40
Pu-238	10.67	0.08	10.66
Pu-238	10.67	0.05	10.40
Pu-239/240	10.5	0.07	10.74
Pu-239/240	10.5	0.06	10.30
Pu-239/240	10.5	0.07	11.35
Pu-239/240	10.5	0.09	10.35
Pu-239/240	10.5	0.07	10.81
U-234	16.6	0.11	16.78
U-234	16.6	0.12	15.04
U-234	16.6	0.08	16.27
U-234	16.6	0.07	16.10
U-234	16.6	0.11	15.58
U-235	0.77	0.11	0.88
U-235	0.77	0.13	1.01
U-235	0.77	0.07	0.97
U-235	0.77	0.08	0.75
U-235	0.77	0.13	1.00
U-238	16.6	0.09	16.32
U-238	16.6	0.08	16.49
U-238	16.6	0.08	16.63
U-238	16.6	0.07	15.39
U-238	16.6	0.08	16.37

The units reported for the LCS measurements are different from those used in the rest of the analytical report. The selection of units of pCi/mL and pCi/L was based on convenience since the known value of the standard used is reported in pCi/mL and pCi/L. The function and utility of the LCS sample (comparing a measurement result with a known standard) are not compromised by using units other than those used to report sample results.

The Internal QC data presented in Tables 6–5, 6–6, and 6–7 provide substantial indication that the data quality achieved in the surface media sample analysis is excellent.

Table 6-8. Results of the IV Laboratory Internal Duplicate QC Measurements for Surface Media Samples

Duplicate QC	Measured Activity (dpm/sample)							
Samples Lab Sample ID#	Am-241	Pu-238	Pu-239/240	U-234	U-235	U-238		
264775	1.72ª	0.22ª	0.52ª	26.33	1.93ª	26.77		
264775D	2.5ª	-0.16ª	1.02 <sup>a</sup>	22.87	2.35ª	22.87		
264789	2.2ª	0.30 <sup>a</sup>	0.60 <sup>a</sup>	28.33	1.73ª	28.55		
264789D	2.9ª	0.25	1.86ª	31.50	1.85ª	27.32		
264876	2.38 <sup>a</sup>	0.08 <sup>a</sup>	0.47ª	26.33	0.55	24.30		
264876D	0.9ª	0.29ª	0.07 <sup>a</sup>	27.77	3.26	29.21		
264896	0.81	-0.06ª	0.50 <sup>a</sup>	15.36	2.06ª	14.61		
264896D	1.05	0.17 <sup>a</sup>	0.90 <sup>a</sup>	16.51	1.13ª	19.99		
264900	1.36ª	0.30°	0.03ª	40.22	1.79	38.10		
264900D	0.7ª	0.03ª	0.51 <sup>a</sup>	42.19	2.12	43.29		

Either the reported value is less than the MDA for the analysis or the error in the reported result includes the MDA. In this case, reproducibility between duplicate samples is not achievable with high confidence since relative error is high at sample concentrations near the MDA. That the duplicate samples yield results, which are consistently at or near the MDA for the analysis provides evidence, in a qualitative sense, that duplicate measurements are comparable.




Figure 6–1. Comparison of Duplicate Alpha Isotopic Sample Analysis Results
Linear Regression Fit Plot

#### 6.2.3 Direct Static Measurements

Two sets of data collected by the IVC are pertinent to the assessment of direct static surface measurement data quality. They are replicate field measurement data and instrument response check data.

#### 6.2.3.1 Replicate Field Measurements

The second of the two data sets contains the replicate measurements periodically made over the duration of the sampling period. In all, 17 replicate measurements were made in survey units 727–01, 727–02, 782–01, and 782–02. Table 6–9 summarizes the cumulative paired replicate measurement results collected from survey units 729–01, 779–04, 779–17, 779–21, 779–23, 779–35, 727–01, 727–02, 782–01, and 782–02. A regression analysis was performed to assess the comparability between the initial and replicate measurements and is graphically presented in Figure 6–2.




Figure 6–2. Comparison Between Replicate Direct Static Measurements
Linear Regression Fit Plot

#### 6.2.3.2 Instrument Response Check Data

The first of the two data sets used to present the quality of direct static surface measurements is the response of the instruments (E600 with a HP-100 probe) to a planar source with a known amount of radioactivity. The source used was the same source used by the Contractor. It is an anodized surface source containing 2,398 dpm of Pu-239 radioactivity. The source was manufactured and certified to be NIST traceable by AEA Technology and assigned a unique ID# GM-785 (see copy of manufacturer's certification in Appendix B).

Prior to initiating a survey each day, periodically (\*every 2 hours), and at the end of a survey each day, the survey instrument in use was used to make a measurement on the known concentration source. The data sheets are provided for the probe used by the IVC during the

Table 6–9. Results of Replicate Direct Static Surface QC Measurements

Sample Location	Measured Act	ivity(dpm/100 cm²)
Cample Location	Initial Measurement	Replicate Measurement
IVP0000104	11.5	10.9
IVP0000111	7.5	8.1
IVP0000118	26.0	19.6
IVP0000119	5.0	5.0
IVP0000122	6.0	6.3
IVP0000125	27.2	29.2
IVP0000128	12.1	14.9
IVP0000129	30.2	26.4
IVP0000135	13.2	13.1
IVP0000136	7.0	13.8
IVP0000149	21.1	14.0
IVP0000152	20.5	17.0
IVP0000153	34.4	20.6
IVP0000155	10.6	10.8
IVP0000159	16.1	16.1
IVP0000176	38.5	32.1
IVP0000182	19.3	3.2
IVP0000187	12.8	22.4
IVP0000194	22.4	22.4
IVP0000206	16.60	16.60
IVP0000212	20.70	29.00
IVP0000218	10.60	13.90
IVP0000226	41.40	16.60
IVP0000236	20.50	27.50
IVP0000242	14.30	11.10
IVP0000248	13.70	20.50
IVP0000254	10.60	10.50
IVP0000259	17.70	7.59
IVP0000276	8.28	8.28
IVP0000282	12.40	12.40
IVP0000288	6.57	10.00
IVP0000294	20.00	16.60
IVP0000299	16.90	10.00
IVP0000386	11.4	7.4
IVP0000392	8.94	8.97
IVP0000401	7.31	7.20
IVP0000407	4.31	4.38
IVP0000356	15.8	18.7
IVP0000362	22.0	22.7
IVP0000368	15.4	14.0
IVP0000374	15.8	15.8
IVP0000379	19.4	19.4
IVP0000306	13.9	10.6
IVP0000312	11.1	10.8
IVP0000318	1.1	1.2
IVP0000324	7.7	7.9
IVP0000426	14.01	14.02
IVP0000432	11.2	7.7
IVP0000438	18.3	18.4
IVP0000444	20.7	17.4

independent verification of survey units 727–01, 727–02, 782–01, and 782–02. A total of 48 response check measurements were made with the probe during the survey period.

A control chart is provided for the probe for each survey unit (Figures 6–3, 6–4, 6–5, and 6–6) to graphically portray the steadfastness of the instrument's responses to the source over the sampling period. Notable is the relatively tight band within which the response checks fall. No degradation of the instrument response was observed over the period it was used indicating that the 2 hour maximum use constraint on a fresh counting gas charge is adequate and might provide justification for a longer allowable period between purge and charge cycles.

## Instrument Response Check Control Chart HP-100 Probe #S15564




Figure 6-3. Instrument Response Check Control Chart, Survey Unit 727-01-HP-100 Probe #S15564

#### Instrument Response Check Control Chart HP-100 Probe #S15564




Figure 6-4. Instrument Response Check Control Chart, Survey Unit 727-02-HP-100 Probe #S15564

#### Instrument Response Check Control Chart HP-100 Probe #\$15564




Figure 6–5. Instrument Response Check Control Chart, Survey Unit 782–01—HP-100 Probe #S15564

# Instrument Response Check Control Chart HP-100 Probe #S15564




Figure 6-6. Instrument Response Check Control Chart, Survey Unit 782-02-HP-100 Probe #\$15564

## 7.0 Data Quality Analysis

The purpose of this data quality analysis (DQA) is to evaluate the data collected from the field in light of its intended use in decision making. Decision makers should obtain an understanding of the verity of the data used in the verification process from reading this section. The DQA uses guidance from *MARSSIM* (EPA 1997), *Guidance for Data Useability in Risk Assessment* (EPA 1992), information from the IV SAP (DOE 1999a), and professional judgement.

## 7.1 Detection Limit Adequacy

Each of the three measurement methods used to assess the residual radioactive contamination in Buildings 727, 782, and 783 have measurement sensitivities which limit the ability of the measurement to detect and quantify radioactivity. A key concern and design element of the SAP was to assure that sufficiently low detection sensitivities were achieved. Assumptions had to be made about the environment and response of the instrumentation and preparation methods in order to estimate the detection sensitivity before the fact. Now that the measurements have actually been made, assessment of the actual detection sensitivity achieved is possible. Section 6.0 presented data which demonstrated that the detection sensitivities achieved were adequate to identify and quantify radioactivity at a fraction of the applicable limit or DCGL. The target detection sensitivity planned for in the SAP was ≈50 percent of the applicable DCGL. Method detection limits obtained in both the field measurements and the laboratory measurements used were adequate to compare to the associated DCGL as indicated in Table 7–1, and met or exceeded the data quality target for measurement sensitivity.

Table 7-1. Adequacy of Independent Verification Measurement Detection Limits

Measurement	Analytical	DCGL Benchmark	Detection Sensitivity Achieved (dpm/100 cm²)			
	Method	(dpm/100 cm <sup>2</sup> )	727–01	727–02	m/100 cm²) 12 78201  ≈5  33°  ≈2  ≈4	782–02
Average removable surface contamination concentration	Smear counting	20	≈5	≈5	≈5	≈5
Average transuranic surface contamination concentration as measured by direct surface emission.	90 Second Direct Static	100	32ª	43ª	33"	40ª
Maximum transuranic surface contamination concentration as measured by direct surface emission.	Surface Emission Count	300	32	43		40
Average surface transuranic contamination concentration in and beneath surface coatings as measured by surface media sampling.	Alpha	100	≈2	≈2	≈2	
Maximum surface transuranic contamination concentration in and beneath surface coatings as measured by surface media sampling.	Spectroscopy	300				≈2
Average surface uranium contamination concentration in and beneath surface coatings as measured by surface media sampling.	Alpha	5,000	≈4	. 4	≈4	
Maximum surface uranium contamination concentration in and beneath surface coatings as measured by surface media sampling.	Spectroscopy	15,000	≈4	≈4		≈4
The detection sensitivity reported is net MDA. To 22 dpm/100 cm <sup>2</sup> , respectively).	he adjusted gross I	MDA is equal to	the MDA +	background	(39, 58, 40	, and

If detection limits had exceeded the DCGL metrics, then declarations based on measurements made using that method could not have been substantiated. As evidenced by comparing the decision limits as represented by the DCGLs with the MDA associated with the measurement method employed in assessing the residual contamination in Buildings 727, 782, and 783, each detection limit obtained was more than adequate to detect, observe, and make risk management decisions with confidence.

## 7.2 Sample Size and Statistical Power

According to the SAP, sample sizes were specified to ensure a false positive error rate (alpha error) and a false negative error rate (beta error) of no greater than 5 percent when measurement data sets were compared to the DCGL. For each sample media set—direct surface emission measurements, smears, and surface media samples—a sample size of 29 (allowing for a 20 percent contingency) was specified in the IV SAP (DOE 1999a). In the field, 116 direct surface emission measurements, 116 smears, and 69 surface media samples were actually collected from designated locations in Buildings 727, 782, and 783.

Based on the results of each of the data sets, retrospective power curves were developed. Figures 7–1 through 7–4 illustrate the power of the sign test to conclude whether the null hypothesis should be rejected by measuring the probability that a survey unit meets the DCGL. Values of both error types (Type-I and Type-II) can be derived from the power curve at any possible concentration of residual contaminant. Type-I errors (falsely concluding that the DCGL is not exceeded when it actually is exceeded) are those that concern the risk manager and decision maker most. The actual and critical sample size (N) are both presented for each of the four data sets evaluated. The retrospective power curve is calculated using the actual sample size obtained. The boundary of the gray region represents the concentrations between which there is insufficient power at the prescribed alpha and beta error rate, given the sample size obtained and the variability observed in the data set.

Inspection of Figure 7–1 for survey unit 727–01 illustrates that the Type-I error rate drops below 5 percent (the error rate is 1-Power) when the true mean surface contamination concentration is at the DCGL of 100 dpm/100 cm<sup>2</sup>, the sample size is 29, and the standard deviation is 7.59 dpm/100 cm<sup>2</sup> (the actual standard deviation). Alternately, the power to reject the null hypothesis when the mean surface contamination concentration is as high as 92 dpm/100 cm<sup>2</sup> is 95 percent. The critical sample size required to provide the power necessary to meet the sampling objectives outlined in the SAP was determined to be 19. The actual sample size (29) was much higher than that required, thus the actual power was much higher than required by the sample design. Note that the estimate of the central tendency, the geometric mean, is plotted against the power curve. This concentration is significantly less than the concentration at which the power begins to wane (the lower boundary of the gray region). The power to reject the null hypothesis at the observed mean concentration in the survey unit is effectively 100 percent.




Figure 7–1. Retrospective Power of the Sign Test
Direct Static Surface Measurements, 727–01




Figure 7–1 (continued). Retrospective Power of the Sign Test Direct Static Surface Measurements, 727–02




Figure 7–1 (continued). Retrospective Power of the Sign Test Direct Static Surface Measurements, 782–01




Figure 7–1 (continued). Retrospective Power of the Sign Test Direct Static Surface Measurements, 782–02




Figure 7–2. Retrospective Power of the Sign Test Smear Sample Measurements, 727–01 and 727–02




Figure 7–2 (continued). Retrospective Power of the Sign Test Smear Sample Measurements, 782–01 and 782–02




Figure 7–3. Retrospective Power of the Sign Test Surface Media Samples/Transuranic Activity, 727–01




Figure 7–3 (continued). Retrospective Power of the Sign Test Surface Media Samples/Transuranic Activity, 727–02




Figure 7–3 (continued). Retrospective Power of the Sign Test Surface Media Samples/Transuranic Activity, 782–01




Figure 7–3 (continued). Retrospective Power of the Sign Test Surface Media Samples/Transuranic Activity, 782–02




Figure 7–4. Retrospective Power of the Sign Test Surface Media Samples/Uranium Series Activity, 727–01




Figure 7–4 (continued). Retrospective Power of the Sign Test Surface Media Samples/Uranium Series Activity, 727–02




Figure 7–4 (continued). Retrospective Power of the Sign Test Surface Media Samples/Uranium Series Activity, 782–01




Figure 7–4 (continued). Retrospective Power of the Sign Test Surface Media Samples/Uranium Series Activity, 782–02

The same results are observed in Figures 7–2 through 7–4. Rigorous statistical tests of the data sets are not justified since it is known that every data point comprising each of the data sets was less than the applicable DCGL. When this occurs, the sign test will always conclude that the null hypothesis should be rejected, provided that a sufficient number of measurements have been included in the data set (i.e., actual sample size is greater than or equal to the critical sample size). Thus, risk managers can be assured that the data collected is sufficiently robust to decide that the residual surface contamination concentration in the survey unit measured is below the DCGL.

### 7.3 Measurement Uncertainty and Data Quality Indicators

As discussed in the IV SAP (DOE 1999a), measurement uncertainty stems from two sources: field sampling variation, and instrument/laboratory measurement variation. Of the two sources, field sampling variation was noted as the greatest contributor to overall uncertainty because of the inherent logistics of sample collection and the one-of-a-kind aspect of sampling the building. The field measurement methods used in the building survey were standard Health Physics instrument techniques and were governed by approved procedures used in the field sampling process. Laboratory procedures were also utilized by the GJO Analytical Laboratory to assess the radioactivity associated with both smear samples and surface media samples. Surface media samples were weighed prior to sample preparation to minimize error due to sample mass loss during sample preparation. An additional control feature utilized to minimize variability and error in the surface media samples was to homogenize the sample by grinding the surface veneer material removed to a fine powder. In this way, any aliquot of the sample selected for analysis could be confidently expected to yield comparable results.

As discussed in the SAP (DOE 1999a), an important activity in determining the usability of the data based on sampling is assessing the effectiveness of the sampling program (EPA 1998, EPA 1992). Data Quality Indicators (DQIs) were identified as guidelines for the DQA process to provide quantitative and qualitative measures of overall data quality and usability. For comparative purposes, Table 7–2 repeats the target DQIs from the IV SAP and summarizes the post-sampling data quality assessment.

It is important to note that the Quality Object for instrument precision ( $r^2$  of  $\approx 0.75$ ) was not achieved due to the absence of residual contamination and high radon levels; however, the instrument precision obtained in the low background environment is determined to be adequate to meet the intent of this Quality Objective. Most of the sampling area did not contain residual contamination greater than background. When developing the SAP it was assumed that a significant amount to residual contamination above background would be present and a value of 0.75 for  $r^2$  would be achievable. As a result of rigorous decontamination most measurement results were at or below background activity levels, thus statistical variability was high. This Quality Object will be changed to reflect a more realistic object based on the low activity levels found in the Building 779 Cluster.

Inspection of Table 7–2 indicates that the DQIs are achieved and the data is regarded as having sufficient quality to be useable for verification of the DCGL and for assessing the results and conclusions obtained by the Contractor.

## 7.4 Overall Quality Assurance and Quality Control

Based on the forgoing analysis and observed practices in the field, it is apparent that overall project QA/QC goals were obtained. The key technical features of the project included:

The DCGL derivation and SAP development processes were performed in accordance with EPA guidance for DQOs (EPA 1997 and EPA 1993).

Field operations were conducted in accordance with the SAP. Modifications to the sample locations which were either inaccessible or involved appreciable personnel safety hazards were made in accordance with the approved sample relocation procedure outlined in the SAP.

Data analysis was conducted as prescribed by the SAP and in general agreement with EPA guidance (EPA 1997 and EPA 1992).

There were no significant problems or incidents that would compromise the findings. The data collected from the building survey is regarded as useable.

Table 7-2. Target Data Quality Indicators and Findings

DQI	Quality Objective (DOE 1999)	Significance	Action/Remark	Finding
Completeness	90 percent completeness	Less than complete data set could decrease confidence	One hundred and sixteen direct surface emission measurements of the 116 scheduled were obtained (100 percent).	DQI accepted.
		in supporting information.	One hundred and sixteen of the 116 scheduled smear samples were collected (100 percent).	DQI accepted.
			Sixty-nine of the potential 116 surface media samples were collected (60 percent). Less than 29 surface media samples were collected from each survey unit because many of the selected locations did not meet the inclusion criteria for sampling. An assessment of the a posteriori power provided by the surface media sample data sets provide evidence that the sample size of each survey unit is sufficient to be considered complete.	DQI accepted.
Comparability	1) Comparability between instrument efficiencies (~±10 percent)  2) Common or equivalent sampling procedure used.  3) Professional judgement and field observations.	Affects ability to combine data sets produced using different sampling and/or analytical methods.	No measurement data sets were combined for the independent verification of the Final Status Survey of Buildings 727, 782, and 783. Consistent methods, both sampling and analytical, were used throughout the sampling and survey process.	DQI accepted.
Representativeness	Sample allocation approach followed to ensure unbiased sample location selection and spatial distribution of the sampling locations.	Non-representativeness increases or decreases Type I error depending on the bias and results in the need to collect additional samples to improve representativeness.	Sample allocation used in the field strictly followed the approach outlined in the SAP. Two of the one hundred and sixteen sample locations selected at random had to be relocated for personnel safety or location accessibility reasons. Both of them were relocated using the relocation protocol outlined and approved in the SAP and maintained the spatial and unbiased objectives of the sample allocation objective. Each of the 116 sample locations was selected without prior knowledge and is unbiased. The sample locations selected meet the intent of the SAP design and are considered representative of conditions in the buildings. There are no analytical or measurement effects (e.g., holding times or compositing effects) affecting representativeness.	DQI accepted.

Table 7–2 (continued). Target Quality Data Indicators and Findings

DQI	Quality Objective (DOE 1999)	Significance	Action/Remark	Finding
Precision	Field and laboratory processes will be governed by procedures.  Replicate and split samples are used to assess variability as an indicator of precision.	Lack of precision affects the accuracy or confidence in the accuracy of the reported results.	All sampling, field measurement, and laboratory analysis processes were controlled by approved written procedures. Replicate direct static surface measurements made in the field showed adequate precision at the low count rates encountered at most sample locations (most were below the detection limit for the method).	DQI accepted.
	< 10% difference between replicate and split samples.		Regression analysis on the paired data yielded a regression coefficient of determination of 0.52 and a standard error estimate of 4.9%.	
	Overall r² of ≈0.75 or better on paired data sets.  Standard error of the regression estimate (SSE ±10%).		Field instrument response checks and laboratory control standards and continuing calibration verification measurements demonstrated the precision of the laboratory analytical methods showing less than 10% error when control samples were measured more than once.	
			Caution must be exercised when attempting to measure precision on replicate measurements with activity near and below the detection limit. Statistical variability at near zero activity limits the likelihood that measurements results will be precise even when sampling and analytical methods are in fact precise and suitable at concentrations approaching the DCGL.	
			Overall sampling variability is another measure of precision. Quantitative metrics describing measurement precision are all acceptable. CV's range from 0.29 to 0.82 and the Cl <sub>95</sub> for each data set was more than 50% below the applicable DCGL.	
Accuracy	Field and laboratory processes will be governed by procedures.  Response to samples containing known amounts of radioactivity should be within ±10%.	Accuracy is affected by bias and precision. A lack of accuracy can affect Type I and Type II errors depending on the bias.	All procedures were implemented. Spikes and Blanks returned expected results. Responses to samples (or sources) containing known amounts of radioactivity were consistently within ±10% for every analytical measurement method used. Field responses to a low-level source containing a known amount of radioactivity were consistently within the acceptable range of ±20%. As shown above, precision was acceptable	DQI accepted.
	QC Blank samples should return results below detection limit. QC spike samples should return results indicating the presence of the radioactivity of interest.			

CV = Coefficient of Variation Confidence Interval

Cl = r = coefficient of determination

## 8.0 Summary and Conclusions

## 8.1 Independent Verification Sampling and Survey

On the basis of the analysis presented in Sections 3.0 through 7.0 of this report, the IVC has demonstrated that survey units 727–01, 727–02, 782–01, and 782–02 have met each of the compliance benchmarks, or DCGLs. These results show that residual surface radiological contamination is well below the agreed upon benchmarks for the Building 779 Cluster Decontamination and Decommissioning project applicable to Buildings 727, 782, and 783. Table 8–1 provides a summary review of the DCGLs compared to the appropriate compliance parameter. The actual reported values in Table 8–1 represent the highest measurement recorded in all four survey units for each of the metrics. The independent verification sampling and survey results are highly reliable and consistent with the field sampling and survey design. No unexpected results or trends are evident in the data. The sampling and survey results determined that residual radiological contamination in Buildings 727, 782, and 783 is very minimal and, for the most part, barely above background levels. Thus, the IVC concludes that the null hypothesis for survey units 727–01, 727–02, 782–01, and 782–02 (that residual radiological surface contamination exists in concentrations above the DCGLs) should be rejected.

Table 8–1. Comparison of Buildings 727, 782, and 783 DCGLs to Highest Observed Compliance
Parameters

	Surface	e Radioactivity	(dpm/100 cm <sup>2</sup> )	
Metric	DCGL	A	ctual	Pass/Fail
	DCGL	(UCL <sub>95</sub> )	Maximum	
Mean surface contamination as measured by direct surface emission	100	22.0		Pass
Maximum surface contamination as measured by direct surface emission	300		36.3	Pass
Mean removable surface contamination	20	<4.84		Pass
Mean total transuranic surface contamination on and beneath a surface with a surface coating as measured by surface media sample	100	3.3		Pass
Maximum total transuranic surface contamination on and beneath a surface with a surface coating as measured by surface media sample	300		8.2	Pass
Mean total uranium series surface contamination on and beneath a surface with a surface coating as measured by surface media sample	5000	53		Pass
Maximum total uranium series surface contamination on and beneath a surface with a surface coating as measured by surface media sample	15,000		121	Pass

## 8.2 Independent Review of the Contractor's Final Status Survey Report and Conclusions

The IVC has completed a comprehensive review of the Contractor's Closeout Radiological Survey Report for Buildings 727, 782, and 783 (RMRS 1999c) and concurs with the conclusion reached by the Contractor—that each survey unit in Buildings 727, 782, and 783 met the applicable DCGLs and that the building should be released from further radiological controls.

## 9.0 References

- Rocky Mountain Remediation Services, L.L.C. 1999a. Closeout Radiological Survey Plan for the 779 Cluster, RF/RMRS-97-123.UN, Revision 2, Golden, Colorado, December.
- —, 1999b. 779 Cluster Final Survey Breakdown Structure, Revision 0, Golden, Colorado, May.
- ——, 1999c. Closeout Radiological Survey Report For Buildings 727, 782, and 783, RF/RMRS-2000-019476.UN, Revision 0, Golden, Colorado, February.
- ——, 1998. Statement of Work for 779 Closure Project Independent Verification Support, prepared by Safe Sites of Colorado, L.L.C., and Rocky Mountain Remediation Services, L.L.C. for the U.S. Department of Energy Rocky Flats Field Office, Golden, Colorado, October.
- ——, 1997. Decommissioning Operations Plan For The 779 Cluster Interim Measure/Interim Remedial Action, RF/RMRS-97-085.UN, Revision 0, Golden, Colorado, October.
- U.S. Department of Energy, 1999a. *Independent Verification Sampling and Analysis Plan for Building 779 Cluster*, prepared by MACTEC-ERS for the U.S. Department of Energy, Grand Junction and Rocky Flats Operations Offices, Colorado.
- ——, 1999b. Sampling and Survey Report, Evaluation and Interpretation of the Residual Radiological Surface Contamination Sampling Results, Building 729, prepared by MACTEC-ERS for the U.S. Department of Energy, Grand Junction and Rocky Flats Operations Offices, Colorado.
- U.S. Environmental Protection Agency, 1998. *Guidance for Data Quality Assessment—Practical Methods for Data Analysis*, EPA QA/G-9, EPA/600/R-96/084, Office of Research and Development, Washington, DC.
- ——, 1997. Multi-Agency Radiation Survey and Site Investigation Manual (MARSSIM), EPA 402-R-97-016, Office of Air and Radiation, Washington, DC.
- ——, 1994. Guidance for the Data Quality Objectives Process, EPA QA/G-4, EPA/600/R-96/055, Office of Research and Development, Washington, DC.
- ——, 1993. Data Quality Objectives Process for Superfund, EPA 540-R-93-071, Office of Solid Waste and Emergency Response, Washington, DC.
- ——, 1992. Guidance for Data Usability in Risk Assessment, PB 9285.7-09A, Office of Emergency and Remedial Response, Washington, DC.
- —, 1988. Guidance for Conducting Remedial Investigations and Feasibility Studies Under CERCLA, EPA/540/G-89/004, Office of Emergency and Remedial Response, Washington, DC.

U.S. Nuclear Regulatory Commission, 1997. Minimum Detectable Concentrations with Typical Radiation Survey Instruments for Various Contaminants and Field Conditions, NUREG-1507, Office of Nuclear Regulatory Research, Washington, DC.

——, 1995. A Nonparametric Statistical Methodology for the Design and Analysis of Final Status Decommissioning Surveys, Draft Report for Comment, NUREG-1505, Office of Nuclear Regulatory Research, Washington, D.C.

———, 1992. Manual for Conducting Radiological Surveys in Support of License Termination, Draft Report for Comment, NUREG/CR-5849, Office of Nuclear Regulatory Research, Washington, D.C.

WASTREN-GJ, (updated continually). *Handbook of Analytical and Sample-Preparation Procedures, Volumes I, II, and III*, U.S. Department of Energy, Grand Junction Office, Grand Junction, Colorado.

## Random Selection Program to Select Survey Units for IV

Two survey units are to be selected for this building: One exterior and one interior.

CTRL - ALT - F9 to recalculate

Random	Number	Selected	=	5
Random	Number	Selected	=	1

Buildin	g	7	82	2
Survey	ı.	ln	iŧ	ID

Survey Unit ID#		Class (1, 2, 3)	weignting		
782-01	(interior)	2	3	1.712.7	782-01
782-02	(exterior walls)	2	3	2	782-01
782-03	(exterior roof)	2	3	·3	782-01
				. 4	782-02
				5	782-02
				6	782-02
•				7	782-03
,				8	782-03
				9	782-03
	•	•	9		

NOTE: Survey Unit 782-01 was changed by the Contractor to include the exterior walls and roof. Survey Unit 782-02 was changed to include the Plenum Area (interior).

## Random Selection Program to Select Survey Units for IV

Two survey units are to be selected for this building: One exterior and one interior. Since only two survey units have been identified for this building group, both will be <u>surveyed</u>.

CTRL - ALT - F9 to recalculate

 Random	Number	Selected	=	1
Random	Number	Selected	=	3

Buildings 783, 727, 780, 780A, 780B

Survey Unit ID#		Class (1, 2, 3)	Weighting	
727-01	(interior)	3	1	1 727-01
727-02	(exterior)	2	3	2 727-02
				3 727-02
				4 727-02



## CERTIFICATE OF CONFORMANCE

AEA Technology QSA Inc.

40 North Avenue Burlington, MA 01803

Telephone (781) 272-2000

Telephone (800) 815-1383

Facsimile (781) 273-2216

TO: MACTEC-ERS, LLC

For the U.S. DoE 2597 B-3/4 Road

Grand Junction, CO 81503

This is to certify that the items listed below, which were ordered against purchase order number 21764, meet AEA Technology QSA Inc's catalogue specifications and that they comply with the requirements specified in the purchase order. AEA Technology QSA Inc certifies that all materials were produced and controlled in accordance with our documented Quality Assurance Program.

Item No.	Quantity	Product Code	Description	Serial No.
1	1	PIR07012	Pu-239 Anodized aluminum source, AD-100x150mm, OD-120x170mm, NIST traceable*, Overall uncertainty +/-6% at 95% confidence level	GM-785

<sup>\*</sup>Calibration test records are on file in our measurement laboratory and are available for contractor's review, if required.

Ross Jones

Technical Sales Manager

13th April 1999

## DEUTSCHER KALIBRIERDIENST DKD

Kalibrierlaboratorium für Meßgrößen der Radioaktivität Calibration laboratory for measurements of radioactivity

AKKREDITIERT DURCH DIE

## PHYSIKALISCH-TECHNISCHE BUNDESANSTALT (PTB)



AEA Technology QSA GmbH

Postfach 58 42

Gieselweg 1

D-38049 Braunschweig

D-38110 Braunschweig

Tel. +49 (0) 5307 932-0 Fax +49 (0) 5307 932-194

Source no. GM 785

Kalibrierzeichen Calibration mark

der

mit

Darstellung

einstimmung

scheine.

Einheitensystem (SI).

Dieser Kalibrierschein dokumentiert die

Rückführung auf nationale Normale zur

Der Deutsche Kalibrierdienst ist Unter-

zeichner des multilateralen Übereinkommens der European cooperation for

Accreditation of Laboratories (EAL) zur

gegenseitigen Anerkennung der Kalibrier-

Für die Einhaltung einer angemessenen

05872 DKD-K-06501 99-03

Einheiten in Überdem Internationalen

Kalibrierschein Calibration Certificate

Gegenstand Object

Alpha Wide Area Reference Source

Hersteller

Manufacturer

AEA Technology QSA GmbH

Тур

Type

PIR07012

Strahler-Nr.

Source number

**GM 785** 

Auftraggeber

Customer

AEA Technology QSA, Inc. **USA-Burlington, MA 01803** 

Auftragsnummer Order No.

110746

Anzahl der Seiten des Kalibrierscheines Number of pages of the certificate

2

Referenzdatum Reference date

23 February 1999

Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich. This calibration certificate documents the traceability to national standards, which

realize the units of measurement according

to the International System of Units (SI).

The Deutscher Kalibrierdienst is signatory to the multilateral agreement of the European cooperation for Accreditation of Laboratories (EAL) for the mutual recognition of calibration certificates.

The user is obliged to have the object recalibrated at appropriate intervals.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Physikalisch-Technischen Bundesanstalt als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift und Stempel haben keine Gültigkeit.

This calibration certificate may not be reproduced other than in full except with the permission of both the Physikalisch-Technische Bundesanstalt and the issuing laboratory, Calibration certificates without signature and seal are not valid.

eutsche Stempel Seal DKD-K-06501

brierdie

Datum

Date

March 1999

Lejtér dés Kalibrierlaboratoriums Head of the calibration laboratory

r. Thieme

Stellvertreter

Schott

Bearbeiter Person in charge

Linke / Schott / Schüler

03-23-00 AN AComments in Comments in Co 02-02-00 54,44 (ance File Index Number START 60 12 X 12 to Calibration Expires: Signature: 7,74 9.32 8.23 11.34 Date: 7.93 12/00 Instrument ID Number: 2 15622 (#321) Operator Name: TAK CHOKEON 815564 10-626 1.5 Minute Survey Unit: Direct Static Background Measurement 1652 1528 1525 1530 Survey Location: RFETS, 779 Cluster / Building Eberline, HP-100 Instrument Model Number: Eberline, E 600 BACKGROUND BACKGROUND BACKGROUND BACKGROUND Supervisory Review: \_\_ Form IVP-1001, July 1999 BACKGROUND Detector Probe Type:\_ BACKGROUND

IN VERIEICATION INSTRUMENT BACKGROUND BATA SUEET

Signature

Start of Servery Comments ... 330 nclude reason forbackground Calibration Expires: 03.23-00 01-25-10 File Index Number GROSS Instrument
FReading
(qpm/100 cm?) 14.9 Signature: Date: Date NOTIFICATION OF THE CAME IN THE RUNGER BLANKER (50/ 100) 5255/S Instrument ID Number: 57 / 56 23 227-02 Signature 1.5 Minute Operator Name: Survey Unit: \_\_\_ Direct Static Background Measurement 12201 1 8060 4060 9201 Print Name Survey Location: RFETS, 779 Cluster / Building pabo paco1 ₹<u>:</u> Eberline, HP-100 Instrument Model Number: Eberline, E 600 BACKGROUND BACKGROUND BACKGROUND Form IVP-1001, July 1999 Supervisory Review: BACKGROUND Detector Probe Type: BACKGROUND BACKGROUND BACKGROUND

E V

S

Survey Location: RFETS, 779 Cluster / Building	er / Building	Survey Unit:	#:	101	Date:	01-62-10
Instrument Model Number: Eberline.	ne, E 600	Instrumen	Instrument ID Number:	123 (#321	Calibration Expires: <u></u>	ires: 03-23-00
	Eberline, HP-100	Operator	Name:	monbon	Signature:	d Smale-
•		-		The second secon		
Semplo IDNOS e a semplo	dmo	To Measurement	Static Count Time	HP 100 Probe ID#	GROSS Instrument F Reading P (4) (dpm/100 cm²)	Comments (include reason forbackground frimeasurement)
BACKGROUND		Dir. Me	1.5 Minute	\$ 15564 (#109) 06/30/00	10.3	End of Survey
BACKGROUND BACKGROUND	8/11	Direct Static Background Measurement	1.5 Minute	$\rightarrow$	15.1	
BACKGROUND		Direct Static Background Measurement	1.5 Minute			
BACKGROUND		Direct Static Background Measurement	1.5 Minute			
BACKGROUND BACKGROUND		Direct Static Background Measurement	1.5 Minute	-		
BACKGROUND		Direct Static Background Measurement	1.5 Minute			
BACKGROUND		Direct Static Background Measurement	1.5 Minute			
Form IVP-1001, July 1999						File Index Number
Supervisory Review:	Print Name		Signature		Date	

Survey Location: _RFETS, 779 Cluster / Building	Building	Survey Unit	it: 782-0	10	Date:	02/03/00
Instrument Model Number: <u>Eberline.</u>	E 600	Instrument	(1584) - 15622 (#321)	622 (432	Calibration Expires:	oires: 03/23/00
Detector Probe Type: Eberline, HP-100	HP-100	Operator Name:	3314	Comeren	Signature:	Jay Conso
Somacollovo (Chircopycolastrandorum) Goto) (	Tilmo	Meegyterment: type:	SEITE COUNTY THE	(ନ୍ତ୍ରା ଅନ୍ୟର୍ଥ)	GROSSIIN ITAMENI Recepto (danikooemi)	Gomments (Include ceron (en endigeound) menonemi)
	1034	Direct Static Background Measurement	1.5 Minute	\$15564 #109 06/30/00	1.3	START OF Shift
BACKGROUND BACKGROUND	1036	Direct Static Background Measurement	1.5 Minute		/ /	
BACKGROUND BACKGROUND	1037	Direct Static Background Measurement	1.5 Minute		8 %	
BACKGROUND BEILD B	1143	Direct Static Background Measurement	1.5 Minute		4.8	END OF Shift
BACKGROUND BEILDING B	1145	Direct Static Background Measurement	1.5 Minute		4.7	
BACKGROUND BACKGROUND	9711	Direct Static Background Measurement	1.5 Minute		9.7	<b>\</b>
BACKGROUND BUILDING (	r240	Direct Static Background Measurement	1.5 Minute		1.2	start of shif

Form IVP-1001, July 1999

Supervisory Review: A. Sample Brull

1 02/03/00

File Index Number\_

Survey Location: RFETS, 779 Cluster / Building	er / Building	Survey Unit:		185-01	Date:	02/03/00
Instrument Model Number: Eberline.	ne, E 600	instrument		10 Number: 5/5/22 (#321)	Calibration Expires:	olres: 03/23/00
	Eberline, HP-100	Operator Name:	Name: JAY	CAMERON	Signature:	Jay Camers
SomboliDNo (Krik ositeroralstemponitish (Goto)	Time (	Me surament	និជ្ជាខេ ៤៦ហូបការ	ැමු පැහැදු(ම) දුල්	(ড়াইউড্.ড)।চ: វិករាកជាវិ :ইয়েবীনে; (টোগেই (৩০ংকু)) ১,	Comminity (Inelutova conforted ground meecutoning)
BACKGROUND	1242	Direct Static Background Measurement	1.5 Minute	515564 #109 06/30/00	14.66	5-0-5-(conf) 2/3
BACKGROUND BACKGROUND	1244	Direct Static Background Measurement	1.5 Minute		7.9	
BACKGROUND BUT THE BUT	8581	Direct Static Background Measurement	1.5 Minute		15.8	END OF 541.67
BACKGROUND	00+1	Direct Static Background Measurement	1.5 Minute		12.3	
BACKGROUND BELLEMENT BELLE	1071	Direct Static Background Measurement	1.5 Minute	<u> </u>	15,7	
BACKGROUND		Direct Static Background Measurement	1.5 Minute			
BACKGROUND		Direct Static Background Measurement	1.5 Minute			
Form IVP-1001, July 1999 Supervisory Review:	SAM/	June 1	Signature	102	63/00 Date	File Index Number

Comments (\*) (Include reason for background 03-23-00 Thr check 00 02-02-00 5020BM ENVO OF 54,47 57.387 Calibration Expires: GROSS Instrument GROSS Instrument Reading \*(dpm/100, cm?) Signature: 14 34 7.95 8,42 Date: 22. F HP-100 Probe ID# Operator Name: TAK CHONEROW 801 F 495518 06/30 Instrument ID Number: 215622 20-786 1.5 Minute Survey Unit: Direct Static Background Measurement 的沙 23, 1137 Survey Location: RFETS, 779 Cluster / Building 6250 273 Eberline, HP-100 0931 cha) Instrument Model Number: Eberline, E 600 BACKGROUND BEING BACKGROUND BACKGROUND BACKGROUND BACKGROUND BACKGROUND BACKGROUND Form IVP-1001, July 1999 Detector Probe Type:\_ BACKGROUND

MATICAL INSTALLING BANKEROUND DATA SHEET

FER

DEFENDER

File Index Number

SAMITA

Supervisory Review: \_\_\_

168

Survey Location: RFETS, 779 Cluster / Building	r / Building	Survey Un	Unit: 737-1	-0.7-	Date:	00-70-70
Instrument Model Number: Eberline.	9. E 600	Instrumen	Instrument ID Number:	15622 (#321)	Calibration Expires:	res: 03-23-00
	400	Onerator Name		MOREN	Signature:	Jay Comes
Detector Probe Type:Eperime	EDerline, nr-100	- Company				
Samptalline. (Atticoning of the State of the State of the Sector)	Times	Messurement Noo	हाहाहिन्छुमारामाण (minutes)	്യായുമ്മുതുംപ്രു	GROSS In trumont Renging (dpm*100cmn):	Gomments: (Include reason (or background) (messurement)
BACKGROUND	1139	Direct Static Background Measurement	1.5 Minute	20/05/NO \$01# \$255/18	7.66	ENO OF Shift
BACKGROUND	114c	Direct Static Background Measurement	1.5 Minute		23,0	<b>→</b>
BACKGROUND BACKGROUND	1317	Direct Static Background Measurement	1.5 Minute		7.87	START OF
BACKGROUND BACKGROUND	1319	Direct Static Background Measurement	1.5 Minute		8,13	
BACKGROUND BACKGROUND	1328	Direct Static Background Measurement	1.5 Minute		4.6	<b>→</b>
BACKGROUND	1453	Direct Static Background Measurement	1.5 Minute		11,54	END OF 5hitt
BACKGROUND BACKGROUND	1500	Direct Static Background Measurement	1.5 Minute	<b></b>	8:31	
Form IVP-1001, July 1999 Supervisory Review:	HM/ATH		Signature	6 1	0.3/6.2/6.C	File Index Number

Survey Location: RFETS, 779 Cluster / Building	er / Building	Survey Unit:	nit: /34-0	7 2	Date:	02-02-00
Instrument Model Number:Eberline.	1e, E 600	Instrumer	Instrument ID Number: 27/	5622 (#321)	Calibration Expires:	bires: 03-23-00
Detector Drobe Tone	Fberline, HP-100	Operator	Name: TAY C	AMERON	Signature:	gay france
Semple libilo. (XVIIIx on Recording advandulen (Codo)	omfr.	Mesucinate Taya	නියාප ලෝගා මතාව (සාබ්ග්පා)	) IP-(100Prate (D1)	GROSSMatrument Republica (dam/100/em));	Comments (Include treson (or beet ground) The surement)
BACKGROUND	1507	Direct Static Background Measurement	1.5 Minute	00/02/10 50/F 1955/5	hb 't	ero ar sr.tr
BACKGROUND		Direct Static Background Measurement	1.5 Minute		·	
BACKGROUND		Direct Static Background Measurement	1.5 Minute		·	
BACKGROUND BACKGROUND		Direct Static Background Measurement	1.5 Minute			
BACKGROUND	•	Direct Static Background Measurement	1.5 Minute			
BACKGROUND		Direct Static Background Measurement	-1.5 Minute			<i>7</i> 3
BACKGROUND BACKGROUND		Direct Static Background Measurement	1.5 Minute			· .
Form IVP-1001, July 1999 Supervisory Review: A. 5	Print Name		( Family Signature	100	102/00 Date	File Index Number

000 Shift 5.1407 J 2 V 7 02-02-00 File Index Number O. ŧ η **Z**0 Signature: ⊱価値 Calibration Expires: 7 1570 1577 6241 1564 Date: THY CAMBRON  $\omega$ 51556 601 A Instrument ID Number: 5156221 727-0, Signature 1.5 Minute Operator Name: Survey Unit: Direct Static Measurement Response Check 1831 1535 Survey Location: RFETS, 779 Cluster / Building Print Name Eberline, HP-100 Eberline, E 600 1534 RESP/CHECK RESP/CHECK RESP/CHECK RESP/CHECK RESPICE THE RESPIC Instrument Model Number:\_\_\_ Form IVP-1002, July 1999 Supervisory Review: Detector Probe Type:

AT

무

SN

RES

MM

SUITE

FR

TOEF TOEN

# INDEPENDENT VERIFICATION INSTRUMENT RESPONSE CHECK DATA SHEET

Survey Location: RFETS, 779 Cluster / Building	ter / Building	Survey Ur	Unit: 727-0	01	Date: 6	22/03	100
Instrument Model Number: Eberline	ne, E 600	Instrumen	Instrument ID Number:	5622 (H3	Calibration Expires:	pires:	3/23/00
Detector Probe Type: Eberli	Eberline, HP-100	Operator Name:	Name: THY	CAMEZEON	Signature:	Stay	ameer
Sample IDNO (Affixor Record Background Bar (Cods)	e camp.	Mersucomonia.	Strite Extra Taro (minus)	්ලායන්නුවූගනමා	්ල්දීමයිනි(ගෙයනාර්යාව (අදියාවල්) (අදියා <sup>ර</sup> ්ල්වලේදී)	Accopt	(inditial comments (inditial cosson for cospoins) වර්ග (indicestrancial)
RESP/CHECK	000	Direct Static Measurement Response Check	1.5 Minute	\$15564 (4109) 06/30/00	6/1/	>	Start of Shift
RESP/CHECK	2580	Direct Static Measurement Response Check	1.5 Minute		(533	>	
RESP/CHECK	2854	Direct Static Measurement Response Check	1.5 Minute		1891	>	
RESP/CHECK	000/	Direct Static Measurement Response Check	1.5 Minute		1495	7	sno or sn.tt
RESP/CHECK	2001	Direct Static Measurement Response Check	1.5 Minute		0/h/	7	
RESP/CHECK	tea	Direct Static Measurement Response Check	1.5 Minute		1472	2	
RESP/CHECK		Direct Static Measurement Response Check	1.5 Minute				
Form IVP-1002, July 1999 Supervisory Review:	S 4771 (	me ' E	Signature	1 00	7/03/00 Date	File Inc	File Index Number

in check J. END OF SUBERY 54,44 アンダント Calibration Expires: 23-23-00 File Index Number 07-5-00 ET TA Signature: 1530 1520 1570 1441 0191 CK Date: 1421 Date TOTAL Instrument ID Number: 515622 #321 06/30/00 815564 401 F 227-02 1.5 Minute Operator Name: Survey Unit: PERNIT Direct Static Measurement Response Check 1280 920 6859 9101 Survey Location: RFETS, 779 Cluster / Building Eberline, HP-100 8101 550 Eberline, E 600 Instrument Model Number:\_\_\_ RESP/CHECK RESP/CHECK RESP/CHECK Form IVP-1002, July 1999 Supervisory Review: Detector Probe Type: RESP/CHECK

173

Signature

Print Name

# INDEPENDENT VERIFICATION INSTRUMENT RESPONSE CHECK DATA SHEET

Date

Signature

Print Name

Supervisory Review: \_\_

7

	ĺ	i	ľ		1		<del></del>		<del></del> T	T		
	00	3/23/00	Canos	ල්කෝආපි (freligo:cosemior rosponso atcakimozogramoni)	START oc Shift			54.77		->	START OF Shit	File Index Number
A SUCET	24/03	0	Aga,	Accopti	7	7	7	2	7	2	7	File
TION INSTRUMENT BESPONSE CHECK PATA SHEFT	Date:	Calibration Expires:	Signature:	ලබලමෙනි (ගනයා ආරක්) , අපත්වලන් (ලෝක 'අම්ම නෝ)	1571	1551	1547	1393	8251	1671	1567	3/03/00 Date
JENT BESPON	10-	5622 (43	CAMERON	රමුයෙන්වාමන්දැහ	57,5564 #109 04/30/03						$\rightarrow$	
	it: 782	Instrument ID Number:	lame: JAY	Strik-Corm Nor (mogas)	1.5 Minute	Manuel Signature						
INDEPENDENT VERIFICA	Survey Unit:	Instrument	Operator Name:	in (Moreumann) Media	Direct Static Measurement Response Check	me I						
INDEPE	r / Building	9, E 600	Eberline, HP-100	e Cinilly	601	1ho1	7701	1147	6711	1150	9571	S.4.7711
	Survey Location: RFETS, 779 Cluster / Building	Instrument Model Number:Eberline.	Detector Probe Type: Eberlin	Sample IDNO (Afrix or Record Background Bar Code)	RESP/CHECK	Supervisory Review:						

# INDEPENDENT VERIFICATION INSTRUMENT RESPONSE CHECK DATA SHEET

Survey Location: RFETS, 779 Cluster / Building	er / Building	Survey Ur	Unit:		Date:	0403/00	
Instrument Model Number <u>Eberline.</u>	ie, E 600	Instrumen	Instrument ID Number:	5622 (H3	(アンノ) Calibration Expires:	pires: 03/23/00	
Delector Probe Type: Eberlin	Eberline, HP-100	Operator Name:	Name: JAY	CAMERON	Signature:	Jed mus	
					<b>)</b>		
Sample IDVO (Affix or Record Background Bar (Cdb)	. Contraction of the contraction	nomonation.	, Seile Count Turr	රමුංගුවූමුම්මන්වූම	ල්බල්ලන්න (msstyrument) (වෙන්ගේ) (ලේකේ (ල්රිනේක්)	Accepts Gommants (Includona resonitor Accepts)    A N	001ES
RESP/CHECK	1248	Direct Static Measurement Response Check	1.5 Minute	515564 4109 01/30/03	1516	574RT 05 5h.it	
RESP/CHECK	0571	Direct Static Measurement Response Check	1.5 Minute		1666		
RESP/CHECK	1321	Direct Static Measurement Response Check	1.5 Minute		8871	Shift.	
RESP/CHECK	1384	Direct Static Measurement Response Check	1.5 Minute		1455		
RESP/CHECK	1356	Direct Static Measurement Response Check	1.5 Minute	->	1445-	>	
RESP/CHECK		Direct Static Measurement Response Check	1.5 Minute			1	
RESP/CHECK		Direct Static Measurement Response Check	1.5 Minute				
Supervisory Review:	S 4771 (S	me le company	Signature		2/03/00 0 Date	File Index Number	

SINK TON SURVEY 03-23-00 Check 12/ 02-07-00 Such MAEPEMBENT WERIFICATION WSTRUMENT BESPONSE GUECK DATA SUFET Calibration Expires: 7 Signature: 6671 1502 1443 Date: 1591 THY CAMBROW Instrument ID Number: 5 15622 (#32) 06/30/00 815518 601A 787-02 1.5 Minute Operator Name: Survey Unit: Direct Static Measurement Response Check Measurement Response Check Measurement Response Check Direct Static Direct Static 1132 0933 6160 Survey Location: RFETS, 779 Cluster / Building 1034 1035 6931 Eberline, HP-100 1637 Instrument Model Number: Eberline, E 600 RESPICE THE RESPIC RESP/CHECK RESP/CHECK RESP/CHECK \*\*\* Sample IDNo. Form IVP-1002, July 1999 Detector Probe Type:

File Index Number

Signature

Supervisory Review: \_

# INDEPENDENT VERIFICATION INSTRUMENT RESPONSE CHECK DATA SHEET

				20-786	Date:	20-20	7-09
Survey Location: RFETS, 779 Cluster / Building	r / Building	Survey Unit:	± 1	1000		0	3-23-00
Instrument Model Number:Eberline	e, E 600	Instrument	Instrument ID Number:	15642("341		16	Succession
Detector Probe Type: Eberlin	Eberline, HP-100	Operator Name:	vame: JAM	(AMBRON	Signature:	Kris	
			Section Control of the Control of th		The second secon	The state of the s	Comments
Samplo IDNos 11. (Artkon Recording Parking Par	-e cuito	ivi -Mersunaments - t. Mer	<u>රක්ල දෙන</u> ාව මාත (ක්ලේලා)	(HPK00/RPS50ID::	GROSSING HAUM (Recalling (Gjon (100 cm))	<b>Z-©</b> ≿-Ш©	((πείμαδι τα εκοπίος το εκροπεύ αλοσιλιπία εκυταποπί)
				フルーナ			Encl of
RESP/CHECK	1135	Direct Static Measurement Response Check	Boundary C. I	4/03	1502	7	512,77
RESP/CHECK	1137	Direct Static Measurement Response Check	1.5 Minute	·	1549	7	->
							30, 10 ATA
RESP/CHECK	1333	Direct Static Measurement Response Check	1.5 Minute		1547	7	5/11/4
		Direct Static	1.5 Minute		191676		
RESP/CHECK	1324	Measurement Response Check			test.	2	
THE CONTRACTOR OF THE CONTRACT	1326	Direct Static Measurement Response Check	1.5 Minute		1506	7	
						1	70 01.0
RESP/CHECK	1500/	Direct Static Measurement Response Check	1.5 Minute		1403	7	54,47
	,		1 & Minute		6.		
RESP/CHECK	1651	Neasurement Response Check			1548	>	$\rightarrow$
Form IVP-1002, July 1999			17 11.		1.0/10		cile Index Number
Supervisory Review:	Samil Vame	Man	Signature	2	Date Date	2	

00' and By たとう 00-File Index Number 3 3 40-40 20 INDEPENDENT WERIGICATION INSTRUMENT RESPONSE CHECK DAIL 4 SHEEL Calibration Expires: \_\_\_ Signature: Date: TAY CAMBROW Instrument ID Number: 5 15622 (1452) 00/08/90 (501 A) 195518 782-02 1.5 Minute Operator Name: Survey Unit: Direct Static Measurement Response Check 1508 Survey Location: RFETS, 779 Cluster / Building Eberline, HP-100 Instrument Model Number: Eberline, E 600 RESP/CHECK RESP/CHECK RESP/CHECK RESP/CHECK Supervisory Review: \_ Form IVP-1002, July 1999 Detector Probe Type:

Survey Location: RFETS, 779.Cluster / Building	ster/Buildir		Survey Unit:	17/-	,		1	. 1
Eberline.	ine. E 600		Instrument ID No	ID Number: 52.55	122 (#3	Callbr	Calibration Expires: 27-33	3-00
Instrument Model Number.	1		; 	0	and free	Sonature	gay (mass	
Detector Probe Type: Eber	Eberline, HP-100		Operator Name:	1970	AMERON			
			•	A WANTED	A TOTAL SANGER CONTRACTOR			
Simply Most Head?	vellero	Striptons Microsonado Invest	ભાગમાં ક	gol(Orthinas)?	ુકાર્યાં (કારણો પ્રાથમ	ઉરેશકોમાં દબભાગ કોપ્રાપ્રાધણ (લુકાવે/ઉશેહભા)	monaretteilisenektorik Ditaliseteilisenektorikorik Malasiaretteilisenektorikorik	लागुट्यस्य प्रतिस्थान
	8560		867	20/06 20/06	Tre Windia	8.93	On media	1 4
	2060	Direct Static Messurement	<b>\</b>		1.5 Minute	2.04		
	9060	Direct Static Measurement			1.5 Minute	5,45		
	1160	Direct Static Measurement		•	1.5 Minute	5.05		
	0913	Direct Static Measurement			1.5 Minute	8 1		
	5160	Direct Static Measurement			1,5 Minute	11.36	ement	
Dugdick 386	8160	Direct Static Measurement			1.5 Minute	7.4	enoute	
com IVP-1000, July 1999 Supervisory Review:	Print Name	2) Sm		Signature		1 02/03.	/oo File Index Number	90

Print Name

Survey Location: RFETS, 779 Cluster / Bullding	uster / Bulld	lna	Survey Unit:	141-0	/	Date:	02/02/00
Instrument Model NumberEbe	Eberline, E 600		Instrument ID Number: 5	KY	5622 (#	32/) Calibrat	Calibration Expires: 23-00
	; ;			4	2 military	J. Shooting	and and
Detector Probe Type:Ebi	Eberline, HP-100	00	Operator Name:	7470	AMERON	30	
(Cichecantellation (Cichecantellation)	A CULINO	Exmisible (Attriction)	Elegyzett	ં લામાં મુખ્યાલ	सम्बद्धाः अम्बद्धाः अम्बद्धाः	(938) (1931) (1931) (1931) (1931) (1931)	Gennichter (Ingleitetzeitengebongenfeitetzeitzeitzierung) viteltzeorgischen, exponetium)
	0,919	Direct Statio Measurement	1 40	00/	S Minis	S. 05	De Mala Merry oftet Concrete
	CEND	Direct Static Measurement			1.6 Minute	2.02	mital
\\ \\ \\ \\ \\ \\ \\ \\ \\ \\ \\ \\ \\	9760	Direct Static Measurement			1.5 Minute	6.1	mtal
96500000VI	0630	Direct Static Measurement			1.5 Minute	17.7	lmente
1 VP0000391	0933	Direct Static Measurement			1.5 Minute	6.9	meteb
I VP8888392	5660	Direct Static Measurement			1.5 Minute	pb'8	metal
Duplish 392	0937	Direct Static Measurement	<b></b>		1.5 Minute	8.97	metal
*om IVP-1000, July 1999 Supervisory Review:	SAM.	) BAN		Atmile Signature	1	1 02/03/0	20 File Index Number

INDEPENDENT VERIFICATION SOURCE  Survey Unit: 2220/	Instrument ID Number: STSL22 (#321) Calibration Expires: 25.22	Eberline, HP-100 Operator Name: H. JAM CAMERON	ECTRING EXTRIBUTION OF THE PROPERTY OF THE PRO	Direct Static Static Measurement	Direct Static	od ul Messurement		Direct Static Measurement	Direct Static Measurement	Direct Static Messurement	1999 1999 1999 1999 19w: A. Samilander   Admily   02/03/00 File Index Number   Signature   Signature
		- 1	2000		1VPddddsss Odd	1 VP8888394					com IVP-1000, July 1999

## INDEPENDENT VERIFICATION SURVEY DATA SHEET -2000 - 1000

Sept. Building	ster / Buildin	-	Survey Unit: 727-01	(Blag .	C Date:	
Survey Location.				\$ 15672 (#	#32 ) Calibratio	Calibration Expires: 3/23/00
Instrument Model Number. Eberline.	line, E 600		Instrument ID Number:		J	77.6
Detector Probe Type: Ebel	Eberline, HP-100		Operator Name:	ameron	Signature:	"
Samplehocentook Samplehocentook	लामञ	Semiploor Mattigonie	) parmanas (Dutumisa)	Struce Security Vilia	GRESSIMMENT RADERS (GDIAMOSEE)	inatecretranplenofronteramonerisis) modeleempenionrooping
	1550	Direct Static Measurement	$\sim M$	1.5 Minute	14.37	pre media sangling
1 VP0000397	155i	Direct Static Measurement	00/30/00	1.5 Minute	11.12	Cinder Rupschaf Lack
. VP8689398	1534	Direct Static Measurement		1.5 Minute	4.48	8 poetst.
1 VP0000399	2091	Direct Static Measurement		1.5 Minute	24.7	Myselad &
	8091	Direct Static Measurement		1.5 Minute	3.69	inne
	2111	Orect Static Measurement		1.5 Minute	7.31	Lonnite
dup you	7171	Direct Static Measurement		1.5 Minute	7,20	Somethe
com IVP-1000, July 1999 Supervisory Review:	2 Samue	n 15m	Signature	19	/ 02/62 / Date	60 File Index Number

Print Name

3/23/00 5AMD/1428 2/2/00 pre media Calibration Expires: Signature: Date: 14.55 4.31 4.38 10,67 7.49 14.0 Instrument ID Number: 5/5622 (#321) Survey Unit: 727-01 (B/dq, 727) J. Campron 1.5 Minute S15564 (4014) Operator Name: 1 b UD | Direct Static Direct Static Measurement Survey Location: RFETS, 779 Cluster / Building Eberline, HP-100 1881 Instrument Model Number: Eberline, E 600 1637 1630 1634 2 2 -Cop of of orm IVP-1000, July 1999 Detector Probe Type:\_ I VP0000404 | VP8686487

08/

SAMI (THI

Supervisory Review:

File Index Number

60/20/26

## 

Survey Unit: 727-01 (6/43,727) Date: 2/82/00	Instrument ID Numb	19-100 Operator Name: S. C. AMCNON Signature: AM LETTLE REST	Semidoor (Induce et camploor)   Induce et camploor (Induce et camploor)   (Induce et camplo	$\begin{array}{c ccccccccccccccccccccccccccccccccccc$	Direct Static Measurement	Direct Static Measurement	Direct Static Measurement	Direct Static Measurement	Direct Static Measurement	Direct Static Measurement	Company O2/02/00 File Index Number
	Eberline E 600	Eberline, HP-100	Tilino.	#3	. ~		Direct Static Measurement	Direct Static Measurement	Direct Static Measurement	Direct Static Measurement	) - C - C - C
10 044 94999	Survey Location: KrE12, 113 Cluster, Building Instrument Model Number: Eberline, E 600		Simplo!keeilton? orsimplo!le;		1VP0000409						<sup>2</sup> om IVP-1000, July 1999

INDEPENDENT VERIFICATION OUR VER DAMESHEET

Dre Madia Sangling Calibration Expires: 23-00 01.25.00 File Index Number \_ Signature: Date: 15.3 15.7 12.3 20.1 Instrument ID Number: STSC 1331 1.5 Minute 1.5 Minute 1.5 Minute 1.5 Minute 1.5 Minute 1.5 Minute 227.07 BPS001PROSOIDIVIMESE 815-564 (#109) Operator Name: Survey Unit: Direct Static Measurement Survey Location: RFETS, 779 Cluster / Building. Eberline, HP-100 1017 1007 1001 1010 Instrument Model Number: Eberline, E 600 odol 900 6001 1 VP 0000355 1 VP 8088356 Duplius 356 Supervisory Review: \_ orm IVP-1000, July 1999 1VP0000352 IVP0000351 Detector Probe Type:\_\_ I VP8888353

98 1

Date

Signature

Print Name

PETT 770 Chiefer/ Building	eter/Buildin		Survey Unit:	127-02		Date:	01.25.10
Survey Location: NCE13.113.50	Eberline, E 600		Instrument ID Number:	115632	14331	Calibrat	Calibration Expires: 27-23-00
1	Eberline, HP-100		Operator Name:	Same	med	Signature:	. Hemile
		All and the second seco		10 10 10 10 10 10 10 10 10 10 10 10 10 1			
StimploMeetBook en:ShaploJiffe (EXitXeorateBartebook))	eull	Scinjsloer Maguedadae Mga	ାନ୍ୟପ୍ରୋକ୍ଟେମ୍ବର୍	(13) (60) (10)	9 3 3 0 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1 1	SESTINATIONS EXXXIII(6) (bin/XO9(ew)	(Inglesoccranistonoscusceognación) :- :- incelheologation coapilectio) :- :- :- incelheologation coapilection :- :- :- :- :- :- :- :- :- :- :- :- :-
VP8888357	8701	Direct Static Measurement	15518 (4014)		nute	21,9	pre Medre Jampleng metal
W   W   W   W   W   W   W   W   W   W	0801	Direct Static Measurement		1.5 Minute		20.5	
- IVP0000359	1032	Direct Static Measurement		1.5 Minute	oute .	7 1	
1 VP0000360	1034	Direct Static Measurement		1.5 Minute		20.6	
VP8088361	1034	Direct Static Measurement		1.5 Minute		2.4	
1 VP0000362	1033	Direct Static Measurement		1.5 Minute		22.0	
Duplink 362	0001	Direct Static Measurement	<b>-&gt;</b>	1.5 Minute	·	22.7	
om IVP-1000, July 1999							Till Lobert

File Index Number \_\_

Date

Signature

Print Name

Supervisory Review: \_\_

187

Survey Location: RFETS, 779 Cluster / Building	uster / Bulldl	na	Survey Unit:	10-1	Date:	00-03-10
Instrument Model NumberEbe	Eberline, E 600		Instrument ID Number:	622 (#	32/) Calibra	Calibration Expires: 23-23
			Superior Name.	Allino	Slonature:	110: Phone B
Detector Probe Type:Ebs	Eberline, HP-100	0	Operator regime.			
			Walter America	A CONTRACTOR OF THE PARTY OF TH	A Company of the Comp	
Samplehicerical)  Calciple (2)	Quily	Semilabor Merraciana Object	। विद्याली क्षेत्र (जिपणार ५	SEUTE CENTRA FIND	(GRેશ્કર્સ)/મદ્રસંગીમાં માત્ર કરાઝલાલી (લુકાર્ય/વશુરુભા)	Gomichie) (Indice) (Erangiland) (Thirtice) (Indice) (Indi
	thô!	Direct Static Measurement	78/55-64 (409) (4109)	1.6 Minute	19,4	Ore Medie Sampling metal
VP0000364	thol	Direct Static Measurement		1.5 Minute	13,4	
1 VP8688365	9601	Direct Static Measurement		1.5 Minute	29.9	
	spol	Direct Static Measurement		1.5 Minute	19.1	
I VP8688367	2501	Direct Static Measurement		1.5 Minute	22.0	airdu blok
	þsol	Direct Static Measurement		1.5 Minute	15,4	
Ruphiste 308	4501	Direct Static Measurement		1.5 Minute	14.0	
om IVP-1000, July 1999						Sedential Separation
Supervisory Review:			* Authorities		Date	

PETTS 719 Cluster Puliding	uster / Buildi	· c	Survey Unit: 727-02	7	Date:	01-25-10
Instrument Model Number:	Eberline, E 600		Instrument ID Number:	622 (H	32/) Calibra	Calibration Expires: 23-00
	Eberline, HP-100	0	Operator Name: H. SA	an street	Signature:	in: A family
			A STATE OF THE STA			
Sample de altent Gebrieben Geber (en la la constante) (en de constante de la la constante)	Sydlino Pa	Pichielands Pichielands Diffu	୍ରୀୟସ୍ତମନ୍ତର (Diffunds)	Statte Grann June	(ປີຊີອີຣີຣີໄກນັກປົກເກີຣີ ເຄີນເຂົ້າໃນ (ທີ່ສະເທີນ (ປີອີຣີເກີ)	Gomitalia (Indicationialianosaniaciania malacompanianacianiania
I VP0000369	1160	Direct Static Measurement	15564 (4109) (4109)	1.5 Minute	24.9	pre Media Samplenz Azphalt
1.VP0000370	foll	Direct Static Messurement		1.5 Minute	18.9	andu Wak
NP0000371	2011	Direct Static Measurement		1.5 Minute	36.3 187	
1 VP8088372	7111	Direct Static Measurement		1.5 Minute	29,4	
VP8888373	2200	Direct Static Messurement		1.5 Minute	24.7	asplalt
I VP80808374	<u> </u>	Direct Static Measurement		1.5 Minute	15.8	ander block
Ouphists 374 1118	1118	Direct Static Measurement	>	1.5 Minute	15.8	
*orm IVP-1000, July 1999						File Index Number

Date

Signature

Print Name

Supervisory Review: \_\_

Survey Location: REETS, 779. Cluster / Building	uster/BulldL	Da	Survey Unit:	101	Date:	0/-22-00
Instrument Model NumberEbe	Eberline, E 600		Instrument ID Number: 25/52	673 (#	62/) Calibra	Calibration Expires: 23-23-00
Delector Probe Type: Ebe	Eberline, HP-100	0	Operator Name: A.	soulth.	Signature:	110: A Same
	_		1. (V.) (1.			
Simplecation? calcidents of takingomeorgististics	an Imode	Semidoe) Vectuoinoi Vyso	ි) මූ සිය ලබාව පැතිව දැනීම සියි සියි සියි සියි සියි සියි සියි සිය	्डास्यात (ब्रह्मात	(මාලපළමුණ වැඩුම්වා) ස්ථාවක්ව (මාලපූණුණුණුණු	Gominalist (Institutionalistation) (Institution) (Institut
1 VP 0000375	4211	Direct Static Measurement	51/55/4 (#/08)	1.5 Minute	15.8	ne mala demolars
VP0000376	4760	Direct Static Measurement		1.5 Minute	かしょ	ashalf
1 VP8888377	8211	Direct Static Measurement		1.5 Minute	8.8	and block
\\ \\ \\ \\ \\ \\ \\ \\ \\ \\ \\ \\ \\	1130	Direct Static Measurement		1.5 Minute	19.1	
	1134	Direct Static Measurement		1.5 Minute	19.4	
Duplish 379	1136	Oirect Static Measurement		1.5 Minute	7.61	
		Direct Static Measurement		1.5 Minute	4.3m	
orm IVP-1000, July 1999						

190

Supervisory Review: \_

File Index Number \_

Signature

Survey Location: REETS, 779 Cluster/Building	luster/Bulld		Survey Unit:	100	Date:	00/03/00
Instrument Model NumberEbs	Eberline, E 600		Instrument ID Number:	1622 (2)	52// Calibra	Calibration Expires: 25-35-000
Detector Probe Type: Ebi	Eberline, HP-100	00	Operator Name: 794	CAMERON	Signature:	ure: Hay Cons
		第2章 · · · · · · · · · · · · · · · · · · ·	Section States (DIStance)	Section 1	GROSSIPHINIPHE	Geomitalia
STIMPLYCETION (FAURY):  OSTIMPLY CONTROLLING (FAURY):  OSTIMPL		Marting Martingson I v Wed		Cantal Shrees	(designated)	(makeeretreakkonostusteskonostusteskon
	hsö!	Direct Static Measurement	515564 4109 06/30/00		34.4 m	party moret
VP8888382	1058	Direct Static Measurement		1.5 Minute	31.1	
	1102	Direct Static Measurement		1.5 Minute	13.79	
VP0000304	3011	Direct Static Messurement		1.5 Minute	24.1	
1 VP0000305	404	Direct Static Measurement		1.5 Minute	17.5	
1 VP8688386	6111	Direct Static Measurement		1.5 Minute	13.9	
Dupliert 306	5111	Direct Static Measurement		1.5 Minute	10.6	
<sup>2</sup> om IVP-1000, July 1999		1,15m	1 Same	\	180/201	60 File Index Number
Supervisory Review:	Print Name	100	Signature		Date	

Survey Location: RFETS, 779 Cluster / Building	ster / Buildli		Survey Unit:	282-0	/	Date:		02/03/	00	
nstrument Model NumberEberline.	ine. E 600		Instrument ID Num	ID Number: 57.52	122 CH	32/) callt	Calibration Expires:	. 03.33	3-00	
Setector Probe Type:Ebet	Eberline, HP-100		Operator Name:	HASA	mitted	W.	Signature:	Hay Co	non.	
							malife a market and being			-
Sunjipileement es mappile evitkoraseorijangedpii)	Artimo M	Strugber Getterming Getkin v	्राप्ट्रेस(क) होते.	9985E)(9)(Apimede	(35.1fc (20.1h) (41n.0	બુદેલકોકોમાં કુંબોલામાં દરસ્યાણ (લુકાન/લક્ષ્ટલા)	्राक्ति। जिल्लाम्	લ્લામાં સ્ટારમાં માના છેલ્લા માર્ગ કર્માં કહ્યાં માર	iche) Kreckinzefricieraki Matapiletbioj a i	
	1119	Direct Statio Measurement	05/90 4 (0) 4 955/45	00/		24.2	S. K	e Wedi	e Donyluz Envirte	
	12/11	Direct Static Measurement			1.5 Minute	24.1				
I VP0000309	37.26	Direct Static Measurement			1.5 Minute	14,1		·		
VP8688318	1130	Direct Static Measurement			1.6 Minute	2.4				
VP000031	1134	Direct Static Measurement			1.5 Minute	6.01				
1 VP8888312	1138	Direct Static Measurement			1.5 Minute	11.1				
Aupliak 312	op. 1	Direct Static Measurement	->		1.5 Minute	8.01		$\rightarrow$		
nm IVP-1000, July 1999 upervisory Review:	SAM.	1/Sm	D.	Ham &		1 02/03	100	File Index Number	mber	_

79

			Survey Unit: 783-0	/ /	Date:	02/03/00
Survey Location: RFETS, 779 Cluster, Eviluation   Survey Location: RFETS, 779 Cluster, E 600	ne, E 600		Number:	5622 (43	Calibra	Calibration Expires: 23-23
1 1	Eberline, HP-100		Operator Name: 1999	CAMERON	Signature:	in: fry cames
		And the second s			Carlos Comments	Gennada
SimplyRestRent? of theptyles (eVitxoricesextellingeeth)	Tillino	ECINIDAN Martindado Wyjo		Shight Ships	(Gilla (Gilla) (Gilla (Gilla))	(Institute of tringst bandwruth exceptional resident bases)
	4571	Direct Statio Measurement	515564 4109 06/30/00	of i & Winute	7.7	parte Courte
1 VP0000314	1258	Direct Static Measurement		1.6 Minute	10.8	
	1302	Direct Static Measurement		1.5 Minute	17,5	
	1306	Direct Static Measurement		1.6 Minute	10.9	
1 VP0000317	1314	Direct Static Measurement		1.5 Minute	3.7	emente
	1316	Dired Static Measurement		1.5 Minute	1.1	
Ouplied 318	1317	Direct Static Measurement	>	1.5 Minute	1,7	
Supervisory Review:	SAM.	2 San	Signature		102/03.	20 File Index Number

## INDEPENDENT VERIFICATION SURVEY DATA SHEET SHOWLING 78-0/

			Survey Unit:	123-01		Date:	02/03/00	
Survey Location. Tree 1.5. (15.5) (15.5) Institute of the	Eberline, E 600		Instrument ID Number:	1981: ST 156	22 14	122) Calibr	Calibration Expires: 23-00	
	Eberline, HP-100		Operator Name:	H. Sa	milter	Signature:	ture: They Common	1
				ノイグ (	AMERO	000		1
Stappholochied Callpholochied (Anyonsocialistics)	Militoria	Scripton Scripton Street	अन्यक्षात्त्रक्षेत्रक्ष	esclOithimha	Canlin Shuces Shuces	(तहेर्डाड्डियाम् भ्रत्यात्ताम् इ.स.जनामह्ये (तृष्टेन्ट्रियाम्	(Instructional Generalis) (Instructional Control of Con	200
I VP6669319	6151	Difect Statio Measurement	105/90 50/7 59.55/13.	00	S. William	1.3	Fire Moder Samples Concrete	
1 VP 8688328	1321	Direct Static Measurement			1.5 Minute	8.0		· [
VP0000321	1323	Direct Static Measurement			1.5 Minute	4.7		T
1 VP8080322	4324	Direct Static Measurement		-	1.5 Minute	10,98		
I VP0000323	nel	Direct Static Measurement			1.5 Minute	11.5		
	1328	Direct Static Measurement			1.5 Minute	7.7		•
Rupliste 324	1329	Direct Static Measurement			1.5 Minute	7.9		
com IVP-1000, July 1999 Supervisory Review:	SAM.	Juty 1	A	Signature.		1 02/03.	60 File Index Number	1

Survey Location: REETS, 779 Cluster / Bullding	uster / Bulldl	Da	Survey Unit:	282-0		Date:	02/03/00
Instrument Model NumberEbs!	Eberline, E 600		Instrument ID Number:	Y	5622 (#	32/) Calibra	Calibration Expires: 23-23-00
Detector Probe Type: Ebe	Eberline, HP-100		Operator Name:	Mr	CAMERON	Signature:	110: Fly Canes
				A Section of the Sect			
(E) Quantification (E) (E) Quantification (E)	Tillino	Esmistoes Macturomon	16/2/2001	(ලබසින්න) ලබා රැක්කයා	લ્યાત કામાન્ક આતા	Gistaryate); Gistaryate); Gistaryate);	Gounglass (InstructivingstonoscurrecontactingsExthe incell reomposition, Extplicting 1.1.)
VP8888325	1333	Direct Statio Measurement	190	564 130/00	1.8 Minute	4.7	Pre Medra Sampling Concrete
WWW.WW.WW.WW.WW.WW.WW.WW.WW.WW.WW.WW.WW	1338	Direct Static Measurement			1.5 Minute	18,30	
VP8888327	(337	Direct Static Measurement			1.5 Minute	14.9	
VP8080328	(338	Direct Static Measurement			1.5 Minute	2.5	
1 VP 8688329	1342	Direct Static Measurement	,		1.5 Minute	1.3	
		Direct Static Measurement			1.5 Minute		
		Direct Static Measurement			1.5 Minute		
om IVP-1000, July 1999	1	J. Low		J Same		, 02/03/00	20 File Index Number

Supervisory Review:

Survey Location: RFETS, 779 Cluster / Building	luster / Buildi	lng	Survey Unit:	282-03	Date:	2/22/00
Instrument Model NumberEb	Eberline, E 600		Instrument ID Number:	15672 (2	f32/) Calibr	Calibration Expires: 23-23-00
	Eberline, HP-100		Operator Name:	Samilta.	Signature:	vre: A Samp
			\( \tau_{\tau} \)	CAMBEON	J. Z.	
Sampplicettons Capplicettons (capplications)	Stuling.	Semilabor Merraraman Merraraman	्राहरकाम्बद्धः (जिप्रामान	टाणाम सम्बद्धः अभिन्न	GROBStarreiment RALither (Genvioseen)	Gemotals (Inglesc et cerpton Grauft se enacet d'rise ann orcel reorges Hen, ce applie 1919)
I VP0000421	0939	Direct Static Measurement	00/05/00	1.5 Minute	10.4	DIE MEDIA SAMPLING
E.VP8888422	6460	Direct Static Measurement		1.5 Minute	24.1	
I VP0000423	1460	Direct Static Measurement		1.5 Minute	26.6	
	1560	Direct Static Measurement		1.5 Minute	6'9	
	8260	Direct Static Messurement		1.5 Minute	17.52	ywite
	0001	Direct Static Measurement		1.5 Minute	14.01	samped concuete
JUP 426	5001	Direct Static Measurement	<b></b>	1.5 Minute	4577	minted investe
om IVP-1000, July 1999						

196

File Index Number\_

Date

Signature

Print Name

Supervisory Review: \_\_\_

File Index Number \_\_

Oate

Signature

Print Name

Supervisory Review: \_\_

197

Calibration Expires: 03-33-00 2/2/00 Signature: INDEPENDENT VERIFICATION CORVET DATA OF EFF Date: Instrument ID Number: 575673 787-02 Operator Name: Survey Unit: Survey Location: RFETS, 779 Cluster / Building Instrument Model Number: Eberline, E 600 Detector Probe Type:\_\_\_\_

Second in Control of the Control		Para Sample Control	्राह्म (क्रियं)	नद्रशामाराज्य	Sinte	(લુસ્ક્રેસ્ટ્રિયા) ક્રમાં ભાગા	destructive entralisment unit en einstelle en tile en tile
STINGOUS FORGER PARTY		Marie Diane Water			Cinh)	(Gire/Gire))	media incelligende illem ce igglettim
I VP0000433	0501	Direct Static Measurement	60/#) (60/#) 59.55/5	00	1.6 Minute	1.1	pre modra sampling
VP0000434	hsal	Direct Static Measurement			1.5 Minute	7.71	painted unevete
I VP0000435	1058	Direct Static Measurement			1.5 Minute	21.4	painted concrete
I VP0000436	1115	Direct Static Measurement		-	1.5 Minute	(8.33	gainted concede
I VP8080437	0711	Direct Static Measurement			1.5 Minute	14.96	
VP0000438	7211	Direct Static Measurement			1.5 Minute	18.27	
Jup 433	1126	Direct Static Measurement			1.5 Minute	18,44	
com IVP-1000, July 1999 Supervisory Review:	SAMI,	15m		Minus Signature		,02/c2/cd	File Index Number
		D E					

Survey Location: RFETS, 779, Cluster / Building	ıster / Bulldlı		Survey Unit:		ij	Date:	00/
Instrument Model NumberEber	Eberline, E 600		Instrument IO Number:	- 1	S 15622 (472)		Calibration Expires: $3  23/00$
Detector Probe Type:	Eberline, HP-100		Operator Name:	5	GMPrev.	Signature:	n: Lay amer
				1			
SumpleMecation? es:samplemeation?	Mino	Stinjsker Utrijenisk Vego	) <b>:</b>	oktomiso-	OUM Hinos SHRS	GREESTINGHING CAN FILLED (GEOVEDERA)	Gennions (Instructuration of curticosing estraticalia) ricell reemperiton, coapyle (bid)
I VP0000439	1130	Direct Static Measurement	\$ 15564 (\$014)		1.5 Minute	11.8%	OFE MEDIA SAMPLING
J VP0000440	1340	Direct Static Measurement			1.5 Minute	10.41	painted uncerte
I VP000044 1	1344	Direct Static Measurement			1.5 Minute	3,55	
VP0000442	1355	Direct Static Measurement			1.5 Minute	3,43	
I VP0000443	7011	Direct Static Measurement			1.5 Minute	17.40	
	80/-1	Direct Static Measurement			1.5 Minute	20.7	
Duplante 444	0 F	Direct Static Measurement			1.5 Minute	17,35	
3upervisory Review:	7. S.A.(2),	J. (27) (C	S.	Month of Signature		1.02/03/00	File Index Number

199

03/23/00 00/00/20 Calibration Expires: Signature: MOEPENDENT VERIFICATION VILLEDA VILLEN Date: Instrument ID Number: 5' 15-622 (#32) 783-02 Operator Name: \_\_\_ Survey Unit: Survey Location: RFETS, 779.Cluster/Building Eberline, HP-100 Instrument Model Number: Eberline, E 600 Detector Probe Type:

SampleMosations	Mino	Semploor Necresonde	්) අදුරු අප් දැල් (ලා ගැන්න	SE (IC GOUN)	GRESSIBLITATIONS ROTATION	(ອົວພາກົບກະ) (matecogrammofcuriceoemacetricistes) ການຢາງເອຍພາຍວາກິຍາກຸດລະເຄີຍກຸ
(A/IIX)Onvecordibanicodaria		Dr. Kan	7/1/17	1.5 Minute		PRE MEDIA SAMPLINS
1VP0090445	hEH	Direct Static Measurement			21.0	painted uncerte
I VP0000446	1438	Direct Static Measurement	04/30/20	1.5 Minute	10,64	
I VP6686447	SHI	Direct Static Measurement		1.5 Minute	17,26	
I VP0000448	CST-1	Direct Static Measurement		1.5 Minute	4.5	mute
VP0800449	14541	Direct Static Measurement	<b>→</b>	1.5 Minute	11.18	imente
		Direct Static Measurement		1.5 Minute		
		Direct Static Measurement		1.5 Minute		
om IVP-1000, July 1999				<u> </u>		File Index Number
Supervisory Review:	omeN topo	e	Signature		Date	

Print Name

Survey Location: REETS, 779, Cluster / Bullding	uster / Bulldir		Survey Unit: 72.7-0	,	Date:	04/05/00
Instrument Model NumberEbe	Eberline. E 600		Instrument ID Number:	122 (#	Calibration Expires:	Explres: 03-23-00
	Eberline, HP-100		Operator Name: H. SH.	milan	Signature:	Lay and
Standenschilder in der State S	MACULE SEA	Seculation .	्राह्म अर्थना हुन्द्र (शिवाताक्षात्र स्था	Merican Red Billion	(स्टाइड्स) महास्वातिक । (स्टाइड्स) महास्वातिक ।	Geimignis) Bleise (Trapplemes qui Geometris)
CALCOCATION OF THE STREET OF T		PEKAL S PEKAL S		Civila	(dphysoleen)	modiseoupenilon coopilleribis
188 0000 JAP	9160	Difect Static Measurement	515564 4109 06/30/00	1.¢ Minute	17.42	
14p 0000 387	1760	Direct Static Measurement		1.5 Minute	20.7	
1VP 0000 390	2663	Direct Static Measurement		1.5 Minute	17.6	
		Direct Static Measurement		1.6 Minute		
		Direct Static Measurement		1.5 Minute		
		Direct Static Measurement		1.5 Minute		
		Direct Static Measurement		1.5 Minute		
om IVP-1000, July 1999	,		1 Am	1,00	02/13/00	S File Index Number
Supervisory Review:	Print Name	2/SAU	Signature (		Date	

#### 00 Calibration Expires: \_ Signature: INDEPENDENT VERIFICATION SURVEY DATA SHEET Survey Unit: 727-01 (1310(9,727) Date: \_ Instrument ID Number: 5 15622 (#32) J. Carreran Operator Name: Survey Location: RFETS, 779 Cluster / Building Instrument Model Number: Eberline, E 600 Detector Probe Type:\_\_

GResSilnaramenti Isabaling (Inelucio curapilono/curfecteratalistic conti (dom/loo cura) (nelucio curapilono curapilenti) i	7,56 POST Medin	4.32	70'81	4.16	14.15	17.57	24.8	
GENE COUNT MADO	1.5 Minute	1.5 Minute	1.5 Minute	1.5 Minute	1.5 Minute	1.5 Minute	1.5 Minute	
iPs/@@මන්ගේ (ලාග්යක්යා ද	00/02/10 (50/#) p955/19							10
Etinjabor Ketjuginane Maso	Direct Static . Measurement	Direct Static Measurement						
emin	1557	1556	0091	p091	1610	9191	C291 70+	
Sample Recation? or sample list. (Xiix or secretizations)	14 poco 366	: 397	398	333	704	104	704	orm IVP-1000, July 1999

File Index Number\_

Supervisory Review: \_\_

Ħ	Instrument Model Number	, , ,	Detector Probe Type:	ا ق	l Š				Contract Con	101011111111111111111111111111111111111	HODOWN THE		50100000000	) ; ;				155 1			,	£5.		•	チー				-			<i>⊃</i>		,	<del>}</del>	∴	om IVP-1000, July 1999	Supervisory Review:	
Survey Location: <u>RFETS, 779 Cluster / Building</u>	Eberline, E 600	Eberline, HP-10	Eberline, HP-100	Eberline, HP-10				-	<b>第20年 1月 1日 </b>				DC/1 8			2	2791	1 2 +			ノスジー			JC:I	aca1 1.90t			1241	\(\frac{1}{2}\)		\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\\	103 184 ×		(	250 1854		,	A. 500	2.4 1.4.0
		01	00	0						(VCC1112)(1)	2. A.	Direct Statio	Measurement		Citation Station	Messing and	Mossordinaii		Site Site Site	Messurement			Direct Static	Messurement		11-10-1-10	Messurement			Direct Static	Mossorenan		Direct Static	Measurement				1/540	
Survey Unit: 727	Instrument ID Number:	Operator Name:	Operator Name:	Operator Marries			•		COMPANY OF STREET OF STREET			1815564	400	06/30/00										•									-		<u></u>	<b>&gt;</b>	J. March	-1 - Usumb	Signature
10-	122 (#	Cameron	11/100							Splices:		T.S. Minute		-	1.5 Minute		•		1.5 Minute		•		1.5 Minute			1.5 Minute				1.5 Minute			1.5 Minute						
Date:		OM Signature:					ē	allemaneand strange case intervals	Stranberry (1616)	(Gland, Cont.)			0.0	0 , 8		<u> </u>	<u>-</u> ン	0 . 0 .		(	7 %	1.6	. !		, , ,		77			1	51.5	•	(	72.	5		00/20/20		Date
2/2/00	Expires: 12-27	": Say Cimera	2		)				(estimating)	((halietat tanjalanian kirteosanteta 1957) i i i i inteliteoiapsallaniatatajajaja i i i i i		1051 MOSKIE	STONE OF THE PARTY								-														<b>&gt;</b>		Eile Index Nimber		

## INDEPENDENT VERIFICATION SURVEY DATA SHEET Survey Unit: 727-02

Calibration Expires: 03-23-00 Signature: 4 Mm/h	Moder determinende vir evangen en met		. >:	3				
Survey Unit: 727-02 Instrument ID Number: 215622 (#321) Operator Name: A. Symmi (Myn.)	19.75264 4.099 6.15560 10.05 10.0	1.5 Minute	1.5 Minute 25.5	1.5 Minute   16, 3	1.5 Minute 29.5	1.5 Minute	1.5 Minute	
	Tilmo Strillion	Direct Static Measurement	Direct Static Of 22 Measurement	370 1106 Messurement	Direct Static Messurement	Direct Static Measurement	373 0924 Messurement	1
Survey Location: REETS, 779, Cluster / Building Instrument Model Number. Eberline, E 600  Detector Probe Type: Eberline, HP-100	Symptopie entour several symptopies (Symptopies) (Symptop	::   348	369	370	371	32	373	om IVP-1000, July 1999

File Index Number\_

Oate

Signature

Print Name

Calibration Expires: 23-23-00 Calibration Expires: 23-23-00 Signature: MMM  Calibration Expires: MMM  Calibration Expires: 23-23-00  MMM  Calibration Expires: 23-23-00  MMM  Modification Expires: 23-23-00  MMM  Modification Expires: 23-23-00  MMM  MMM  MMM  MMM  MMM  MMM  MMM	
Calibration Signature: Signature: 22 6 22 0 22 0	8.8
STSLEEZ (# 1.5 Minute	1.5 Minute 1.5 Minute 1.5 Minute
Init: Int ID Number: A IName: A I / 0 9 6 / 3 0 / 0 C	
Masurement Aeasurement Aeasurement Aeasurement	Direct Static Measurement Measurement Direct Static Measurement Measurement Measurement
Eberline, E 600 Eberline, HP-100    1/24   1/24	1139
Survey Location: REETS, 779 Cluster / Building Instrument Model Number. Eberline, E 600  Detector Probe Type: Eberline, HP-100  I V POOD 374   124   1	377 1129 378 1137 379 1138

Date

Signature

Supervisory Review: \_ 205

				782-0	>	Date:	02/03/00
Survey Location: REETS, 779 Cluster / Building	ster/Buildir		Survey Unit:				
Instrument Model Number. Eberl	Eberline, E 600		Instrument ID Number:	/ N	Sp32 (#	122/ Calibra	Calibration Expires: 23-00
1	Eberline, HP-100		Operator Name:	H SA	AMERO	Signature:	ure: Hay Comes
		•		THE PROPERTY OF THE PARTY OF TH	A Control of the Section of the Section		
Sumploficetitent? Califications to Califications (California)	) Allmo	Estmather Meetinging Wysi	्राञ्चलात्त्राम् स्वतः।	नहरू (मुक्तानार )	(55 Be (90 Be) (190 Be)	(GRESTINDEINTE RODEING (Generation)	(Inductoetranplonoscusteoshocetratiecum) noelbeoapstilon.coapiledid
108 00	rsaj	Direct Statio Measurement	105/90 4 10 9 96/30/	00	1.6 Mihitio	27.5	Jampelen Jampelen
303	0011	Dired Static Measurement			1.5 Minute	23.9	
303	tall	Direct Static Measurement			1.5 Minute	20.7	
304	1107	Direct Static Measurement			1.6 Minute	4.2	
305	1111	Direct Static Measurement			1.5 Minute	13.9	
304	UII	Direct Static Measurement			1.5 Minute	20.9	
307	1120	Direct Static Measurement			1.5 Minute	24.7	
Supervisory Review:	Print Name	Jan .		Hum Signature.		1 02/031	60 File Index Number

Survey Unit: 784-6/	10777
Eberline, HP-100 Instrument Solution (128) 1/28 Instrument Measurement Measure	S. TIS CLUSSEL   Building   Survey Unit.   PREMISER   PASS   PA
Eberline   E 800   Instrument ID Number:	### Stratument Debatic Name
Eberline   E 500   Instrument 10 Number:   A SAMI AMERICA   E 500   Instrument 10 Number:   A SAMI AMERICA   E 500   Instrument 10 Number:   A SAMI AMERICA   E 500   Instrument   A SAMI AMERICA   E 500   Instrument   A SAMI AMERICA   E 500   Instrument   Instrume	Ebarlia, #2.00   Calibration Explose   Cal
# Probe Type:	Committee   Comm
# Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#321) Calibration Epiter. (37.3)  # Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#322) Calibration Epiter. (37.3)  ## Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#322) Calibration Epiter. (37.3)  ### Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#3.4)  #### Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#3.4)  ###################################	Committee   Comm
Eberline   16.00   Instrument ID Number:	Committee   Comm
# Probe Type: Ebailing # 18400 Operation Houmber: ### ################################	Committee   Comm
# Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#321) Calibration Epiter. (37.3)  # Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#322) Calibration Epiter. (37.3)  ## Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#322) Calibration Epiter. (37.3)  ### Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#3.4)  #### Probe Type. Ebuiling E 500 Instrument ID Number. A TABLE (#3.4)  ###################################	Committee   Comm
# Probe Number Ebailing # 800	1.5 kMrote   Number   Etablina   E 600   Institute   Institu
Specifica	Ebarlia, #200   Calibration Explose   Cali
Eberfline   E 500   Instrument ID Number:   A STATE   E 22)   Calibration Expires:   A STATE   E 22)   Calibration Ex	String Cluster Building   Survey Unit   String Cluster Building   Survey Unit   Surv
Eberline   E 500   Instrument ID Number   A STATT (#23)   Calibration Expires   D 7.3	### Ebarline # £500   Instrument D Number: \$15000000000000000000000000000000000000
Eberline   E 500   Instrument ID Number   A STREET   E 523   Confination Expires   A STREET   E 523   Confination Expires   A STREET   E 523   Confination   E 523	Surge   Control   Contro
Eberline   1500   Instrument   D Number   A State   A	Sand   1.5
Specifica   E800   Instrument   D Number:	Surgicial Contest   Building   Survey Office   Contest
Eberline	S. TIS CLUSTER   Building   Survey Unit.   P. S. Minute   Superation   Survey Unit.   P. S. Minute   Superatorne
Eberline	S. TIS CLUSTER   Building   Survey Unit.   P.   P.   P.   P.   P.   P.   P.
Eberline, He-100   Instrument ID Number: STRAZZ (#32)   Calibration Expirex: D7.22   Eberline, He-100   Operation Number: STRAZZ (#32)   Calibration Expirex: D7.24   Calibration Expirex: D7.24   Calibration Expirex: D7.25   Calibration Expirex: D7.25   Calibration Expirex: D7.26   Calibration Expirex: D7.26   Calibration Expirex: D7.26   Calibration Expirex: D7.26   Calibration Expirex: D7.27   Calibration Expirex: D7.26   Calibration Expirex: D7.27   Calibration Expirex: D	Ebarline
Eberline	Ebarlina, E 500   Instrument to Number   State 2 (#321)   Calention Expires:   Operator Name:   H State 1 (#321)   Calention Expires:   Operator Name:   Ope
Separation   Driver State   Driver	Eberline   E & 20   Instrument   D Number   Stratz   F   F   F   F   F   F   F   F   F
Partine   Figo   Calibration Explose   Calibration Explose   Calibration Explose   Calibration Explose   Calibration Explose   Calibration Explose   Calibration   Calib	Etaclina   E & 200   Instrument   D Number   Strazz (
Ebaclina   E 600   Instrument ID Number   STEAR   Exact   Ex	Ebarline   E 600   Instrument   D Number   Standard   Survey Unit   Suprement   D Number   Standard   Suprement   D Number   Standard   Suprement
1.5   1.5	Exercine   E 800   Instrument   D Number   Standard   Supervice   Supervise   Supervise   Supervice   Supervise
Ebacline	Eberline, H2.00
Partine   Earline   Earl	Eberline
Calibration E 600   Institument ID Number: \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$ \$	Eberillia   E 600   Instrument 10 Number   State   Square   Squa
Eberline	Epaclina, HP.100   Control of the
Eberline   E60   Instrument   D Number   State   E221   E221   Eperline   E722   E221   Experiment   Experi	Eparline
Ebarline   E 500   Instrument ID Number: SIGNATO	Eberline
Eberline   E 500   Instrument ID Number   State   Signature   State   Signature   State   Signature   State   Signature   Signature   State   Signature   Signat	Ebarlina   H2-100   Coperation   Coperatio
Ebarilina, H2-100   Operator Name:   H STATE   H2-20   Calibration Expires:   DF-25	Eberillo
Ebarline   E 500   Instrument ID Number: S. S. S. S. S. S. Statute: P. Signature: P. S.	Ebarlina   Ebarlina   Estato   Instrument ID Number:   State   Statute:   S
Ebarlina, HP-100   Calibration Expires   Physical Parts	Eberline   E 600   Instrument ID Number:   State   Statute:   St
Eberline, E 600   Instrument 10 Number: \$ \text{Start22} (\pi \pi \pi \pi \pi \pi \pi \pi \pi \pi	Ebarline   H2-100   Carbon Linstrument ID Number:   State   Standard   Stan
Eberline	Eberline   Ebor   Instrument ID Number:
Ebarline   E 500   Instrument 10 Number:   Start   Each   Ebarline   Espature   Ebarline   Ebarline   Ebarline   E 500   Ebarline	Eberline   Eboline   E 600   Instrument ID Number:   Stratument
Eberiline   E 600   Instrument ID Number:   Start   Eberiline   Eberiline   E 600   Instrument ID Number:   Start   Eberiline   Eberilin	Eberline   E 600   Instrument ID Number:   Stratum   Signatum:
Eberline   E 600   Instrument ID Number:   Stand   Signature:   Page   Page   Signature:   Page   Signature:   Page   Signature:   Page   Page   Signature:   Page   Signature:   Page   Signature:   Page   Signature:   Page   Signature:   Page   Signature:   Page	Eberline   Eberline   E 600   Instrument ID Number   State
Eberline   E 600   Instrument ID Number:   State   Signature:   Page   Page   Signature:   Page   Signat	Ebarlina   Haristone   Harist
Eberline   E 600   Instrument ID Number:   Standard	ETS_773_Cluster_  Building   Survey Unit
Eberline   E 600   Instrument ID Number:   Stantumer	Etechine
Eberline, HP-100  Control of Name: A State of St	Eps. 773 Cluster   Building   Survey Unit
Eberline, E 800   Instrument ID Number: Stratt   Stratton Expires: 17-3	ETS. 713 Cluster / Building Survey Unit.  Eberline, HP-100  Eberline, HP-100  Operator Name: HSHIP Control Con
Eberline, E00   Instrument ID Number: \$156.22 (#32)   Calibration Expires: 27.3	ETS_713 Cluster   Building   Survey Unit   Calibration Expires   Calibration   Calibration Expires   Calibration   Calibration Expires   Calibration Exp
Eberline	ETS, 713 Cluster I Building Survey Unit.  Eparline, HP-100  Coperator Name: H. Shariff M. Signature: A. Strategies of the Stratum of the Strategies of the S
Ebacline, E 600 Instrument ID Number: STRAZZ (#321) Calibration Expires: 3.3	ETS_TISCUISITE / Building Survey Unit:
Eberline, E 500 Instrument 10 Number: State   Signature: State   Signa	ETS_TISCUISTED   E 500 Instrument ID Number. S SCAZZ (#32) Calibration Expires: 27.3  Eberline, HP-100 Operator Name: H_SMinute (Stross instrument)  Eberline, HP-100 Operator Name: H_SMinute (Stross instrument)  So   1/2 y Messurement
Eberline, E 600 Instrument ID Number: STRAZZ (#32) Calibration Expires: 37-33  Eberline, HP-100 Operator Name: H STRATCH M Signature: Stratch operator Name: HS Minute Nam	ETS, 719 Cluster / Building Survey Unit: 6 12 (1232) (1232) Calibration Expires: 87.3.  Eberline, HP-100 Operator Name: A Survey Chin. (Calibration Expires: 87.3.2)  Eberline, HP-100 Operator Name: A Survey Chin. (Calibration Committee of Calibration Committee of Calibration Committee of Calibration Committee of Calibration Calibratic Calibration Calibratic Calibration Calibratic Calibratic Calibratic Calibratic Calibratic Calibratic Calibration Calibratic C
Ebacilna, HP.100 Operator Name: A START (#32) Calibration Expires: 13-3  Ebacilna, HP.100 Operator Name: A START (#32) Calibration Expires: 13-3  Sold 1/24 Measurement	ETS_7739 Cluster / Building Survey Unit:  Eberline E 600  Instrument 1D Number: SISCA22 (#321)  Eberline HP-109  Operator Name: H State (1.6 Minute)  Sobrid   1.2 4   Direct State
Eberline, E 500 Instrument ID Number: State Hamber: State Hamber: A State Hamb	ETS, 779 Cluster / Building Survey Unit: A Survey Unite Statio
Eberline, E 600 Instrument ID Number: State (#32) Calibration Expires: 27-3  Eberline, HP-100 Operator Name: H SAMINITED (#32) Instrument Expires: 27-3  Eberline, HP-100 Operator Name: H SAMINITED (#32) Instrument (Instrument)  Sold (1/24) Measurement O6 (30) O 1.5 Minute 20.7  Sold (1/25) Measurement 1.5 Minute 1.5 M	ETS. 779 Cluster / Building Survey Unit:  Eberline, E 600  Instrument 1D Number: S/SG22 (#221)  Eberline, HP.100  Operation Name: A Superior N
Eberline, E 600   Instrument ID Number: STRAZLED   Calibration Expires: 27.3   Eberline, HP-100   Operator Name: HSM CAMPERON   Instrument   Operator Name: HSM CAMPERON   Instrument   Operator Name: HSM CAMPERON   Instrument   Object Static   Object St	ETS. 779 Cluster / Building Survey Unit:  Ebacline, H2.100  Calibration Expires: 27.3  Ebacline, H2.100  Operator Name: A State (State of State of
Febrilia E 600 Instrument ID Number: STEAR FED Calibration Expires: 07-33  Eberline HP-100 Operator Name: H-STEAR FED Calibration Expires: 07-33  Eberline HP-100 Operator Name: H-STEAR FED Committee Calibration (Institute Calibration) (Institute	ETS. 779 Cluster / Building         Survey Unit:         6 / 3 / 3 / 3 / 3 / 3 / 3 / 3 / 3 / 3 /
Eberline, E 600 Instrument ID Number: STEAZ (#32)  Eberline, HP-100 Operator Name: H. Steam (PROM. Signature: Straighton)  Straighton (Instrument of 130/00) 200 Straighton (Institute of 130/00)  Sold     24	ETS. 779 Cluster   Building   Survey Unit:
Eberline, E 500 Instrument ID Number. St. St. 22 (#32) Calibration Expires: 27-33  Eberline, HP-100 Operator Name: HEATOPING: St. 100 Operator Name: HEATOPING:	The Cluster Building Survey Unit:  Eberline, E 600  Instrument 1D Number: SCAZZ (#22)  Eberline, HP-100  Operator Name: HEAD (MARCH 100) The County C
Eberline, HP-100  Eberline, HP-100  Operator, Name: H-State Count Calibration Expires: D7-3-3  Eberline, HP-100  Operator, Name: H-State Count Calibration Expires: D7-3-3  Operator, Name: H-State Count Calibration Expires: D1-3-3  Operator, Name: H-State Calibration Calibration Expires: Calibration Calibr	ETS. 779 Cluster / Building Survey Unit:  Eberline, HP-100  Eperline, HP-100  Operator Name: H Statt Construent in Number: Statt Construent in Number: Statt Construent in Number: Statt Construent in Number: Statt Construent in Natural Constru
Eberline, E 600 Instrument ID Number: State Reconstruction Expires: 27-32  Eberline, HP-100 Operator Name: H-State Reconstruction Expires: 27-32  Eberline, HP-100 Operator Name: H-State Reconstruction of the State Reconstructi	Therefore I Building Survey Unit:  Eberline, HP-100  Charles and C
Eberline, E 600 Instrument ID Number: State Signature: Properties: 27-32  Eberline, HP-100 Operator Name: H- State Signature: Signat	The Cluster I Building Survey Unit:  Eberline, HP-100  Operator Name: H STATE H Signature: Management in the State of Samure in the Minute of Samure in the Measurement of 1.5 Minute of
Eberline, E 600 Instrument ID Number. STEAZY #321) Calibration Expires: 27-33.  Eberline, HP-100 Operator Name: #37 Am ERO.  Eberline, HP-100 Operator Name: #37 Am ERO.  Straighton: Tilling With the Name of State of Sta	ETS. TIS Cluste / Building Survey Unit:  Eberline, HP-100  Eperline, HP-100  Eperline, HP-100  Operator, Name: H. State (ARCAS)  Eperline, HP-100  Operator, Name: H. State (ARCAS)  Superator Name: H. Superator (ARCAS)  S
Eberline, E 600 Instrument ID Number: SIGNATO (#322)  Eberline, HP-100 Operator Name: A SIGNATO (#322)  Eberline, HP-100 Operator Name: A SIGNATURE:	ETS. T19 Cluster I Building Survey Unit:  Eberline. HP-199  Eperline. HP-199  Eperline. HP-199  Operator Name: H. Share Station Express: 27.23  Eperline. HP-199  Operator Name: H. Share Station Express: 27.23  Olived Statio
Eberline, E 600 Instrument ID Number: STRAZZ (#32)  Eberline, HP-100 Operator Name: H. Strand M. Signature: M. Signature: M. Strand M. Signature: M. Signatu	ETS. 179 Cluster/Building Survey Unit:  Eberline, E 600 Instrument ID Number: Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Calibration Express: 27-23  Eberline, HP-100 Operator Name: H Stand Stand Stand Stand Stand Stand Stand Stand St
The resulting E 500 Instrument ID Number: SIGNATURE AND Calibration Expires: 27-3.  Eberline, HP-100 Operator Name: H STATIST AND SIGNATURE: Si	ETS. 179 Cluster/Building  Eberline, E 600  Instrument 1D Number: STEAZ (#22)  Eberline, HP-100  Operator Name: H. Steaz (#22)  Eberline, HP-100  Operator Name: H. Steaz (#22)  Eberline, HP-100  Operator Name: H. Steaz (#22)  Signature: Strain Experiment (Instrument Committee
Therefile E 500 Instrument ID Number: STRAZ (#321) Calibration Expires: 27.3.  Eberline, HD-100 Operator Name: HSMM Strate (GROSS Instrument)  Strate (GROSS	ETS. 179 Cluster I Building  EDerline, HP-100  EDerline, HP-100  Every Name: Hardounder: State And American Express: And American Hardounder: Marketing Engineering (Inches State)  Sold
Ther. Eberline, HP-100 Instrument ID Number. Standard Manuel Manu	ETS. 179 Cluster I Building  EDecline E 500  Instrument ID Number: STEAZ (#32)  Eberline HD-100  Operator Name: H Signature: Signatu
Ther. Eberline, E 600 Instrument ID Number: Stand THAM Calibration Expires: 27-3.  Eberline, HP-100 Operator Name: H. Stand THAM Signature: M. Signature: M. Signature: Colling (Institute Cuterpiston) (Institute Cuterpiston	ETS. 779 Cluster / Building Survey Unit: 179 Character / Building E 600 Instrument ID Number: Start / Am Am Am Signature: Am Signature: Am Am Edom Signature: Am
Therefore E600 Instrument ID Number: State Community of Signature: A S	ETS. 779 Cluster / Building  Survey Unit:  Eberline, HP-100  Operator Name: H. Sharing M. Signature: P. S. Signature: P. S.
Ther. Eberline, E 600 Instrument ID Number: State Calibration Expires: 27-3.  Eberline, HP-100 Operator Name: H-100 Ay Am Elecon  Measurement A 10-9  1.5 Minute  1.5 Minute  20.3  1.24  Measurement A 10-9  1.5 Minute  20.3	ETS. 779.Cluster / Building Survey Unit:  The Company of Company o
Eberline, E 600 Instrument ID Number: Stand THAT M. Signature: P. 3.  Eberline, HP-100 Operator Name: H. Shinute County Cardina County	The Cluster / Building Survey Unit:    Eberline
Ther. Eberline, E 600 Instrument ID Number: Stand 1997 (#32) Calibration Expires: 1973.  Eberline, HP-100 Operator Name: H Stand 1998 (State of Stand 1998) (Institute retrainplement of 158 (179 Minute) (179 Minute	ETS. 773 Cluster / Building Survey Unit:  The control of the contr
Ther. Eberline, E 600 Instrument ID Number: State Community Control of Calibration Expires: 1973.  Eberline, HP-100 Operator Name: H Shifting Control of Calibration Expires: 1973.  Eberline, HP-100 Operator Name: H Shifting Control of Calibration Expires: 1973.  Soft Time Calibration	ETS. 773 Cluster / Building Survey Unit:  Eberline, E600 Instrument ID Number: State Committee of Committee o
Eberline, HP-100 Instrument ID Number: Stant THAT M Signature: PT. 2.  Eberline, HP-100 Operator Name: H Stant THAT M Signature: PT. 2.  Signature	ETS. 779 Cluster / Buildina Survey Unit:
Eberline, HP-100  Calibration Expires: 7.3.2  Eberline, HP-100  Operator Name: H. Shamfifth M. Signature: Sign	ETS. 779 Cluster / Building Survey Unit:
Eberline, E 600 Instrument ID Number: State (1922 (1932)) Calibration Expires: (23.2)  Eberline, HP-100 Operator Name: H. Shinting (1980) Synature: Similar Common (1980) Shifting (1980) Shif	ETS. 779 Cluster / Building Survey Unit: Compete Survey Unit: Calibration Expires: Calibration Cali
Eberline, HP-100  Calibration Expires: 27.2  Eberline, HP-100  Coperator Name: H. Stand Coom Straighton (Indigenerizing Commission (Indigenerizing Commission))  Cooperator Name: H. S.	ETS. 779 Cluster / Building Survey Unit: Control of the control of
Eberline, HP-100  Instrument ID Number: Stand THAT M. Signature: P. T. Signature: P. S	ETS. 779 Cluster / Building Survey Unit:  The Composition of the control of the c
Eberline, E 500 Instrument ID Number: Start And	ETS. 779 Cluster / Building Survey Unit:  Diversity Control Co
Eberline, E 600 Instrument ID Number: Start And	ETS. 779 Cluster / Building Survey Unit:
Ther. Eberline, E 600 Instrument ID Number. State And Annie Signature: Annie HP-100 Operator Name: Annie Commission (Institutional) (Instituti	ETS. 779 Cluster / Building Survey Unit:  The control of the contr
Eberline, E 600 Instrument ID Number: State Calibration Expires: 27.3.  Eberline, HP-100 Operator Name: H. State County Canal Calibration Expires: 27.3.  Eberline, HP-100 Operator Name: H. State County Canal Calibrature: State County Canal Calibrature:	ETS. 779 Cluster / Building Survey Unit:  The company of the compa
Eberline, HP-100   Instrument ID Number: State   He-100   Calibration Expires: 22.2.	ETS. 779 Cluster / Building Survey Unit:  The cluster / Building Survey Unit:  The colline in th
Ther_Eberline_E 600 Instrument ID Number: State Calibration Expires: 27-3.  Eberline_HP-100 Operator Name: H. Shart M. Signature: State Common Country Canada Calibration Expires: 27-3.  Eberline_HP-100 Operator Name: H. Shart M. Signature: M. Signature: M. Signature: Common Calibration Calibra	ETS. 779 Cluster / Building Survey Unit:
Ther. Eberline, HP-100 Instrument ID Number: State That the Signature: A County	ETS. 779 Cluster / Building Survey Unit:  The cluster / Building Survey Unit:  The cluster / Building Survey Unit:  Eberline, HP-100  Operator Name: A Survey Common Surve
Ther. Eberline, HP-100 Instrument ID Number: State (Signature: ATT CAN Calibration Expires: ATT CAN CAN ELOW Signature: ATT CAN CAN ELOW (Institute retrictibilion of the country of the c	ETS. 779 Cluster / Building Survey Unit:  The continuent in Survey Unit:  Specification Expires: 27.2  Survey Unit:  The continuent in Survey Unit:  Specification Expires: 27.2  Signature: 27.2
nber Eberline, E 600 Instrument ID Number: Statt Am Am Am Boom Signature: 27-3.  Eberline, HP-100 Operator Name: Am Am Edom Signature: Am Edom Signature: Goldmann State Communication C	ETS. 779 Cluster / Building Survey Unit:  There Eberline, E600  Instrument ID Number: State Community Canal Calibration Expires: 27-3.  Eberline, HP-100  Operator Name: H. State Community Community Canal Calibration Community Canal Calibration Ca
nber Eberline, HP-100 Instrument ID Number: State (Instrument Property Communication C	ETS. 779 Cluster / Building Survey Unit:
nber: Eberline, E 600 Instrument ID Number: State (1922)  Eberline, HP-100 Operator Name: H SHAM CAME Signature: Signature: Combounded (1920)  State (1920)  Calibration Expires: P. S.	ETS. 779. Cluster / Building Survey Unit:  Therefore Eberline, HP-100  Operator Name: H. State Common Signature: Signatur
nber Eberline, HP-100 Instrument ID Number: State Calibration Expires: 23.2.  Eberline, HP-100 Operator Name: H. Sammitte Calibration Expires: 23.2.  Eberline, HP-100 Operator Name: H. Sammitte Calibration Committee Calibration Calibr	ETS. 779. Cluster / Building Survey Unit:
nber: Eberline, E 600 Instrument ID Number: S.	ETS. 779 Cluster / Building Survey Unit:
nber. Eberline, E 600 Instrument ID Number: State (1922)  Eberline, HP-100 Operator Name: H. Sam (1920)  State (1930)	ETS. 779. Cluster / Building Survey Unit:
nber: Eberline, E 600 Instrument ID Number: S. S. S. S. S. S. S. S. S. Calibration Expires: 27-3.  Eberline, HP-100 Operator Name: H. S.	ETS. 179. Cluster / Building Survey Unit:  The control of the cont
nber Eberline, E 600 Instrument ID Number: State A State M. Signature: P. Signature: A State Com.	ETS. 779.Cluster / Building Survey Unit: 673. (432) Calibration Expires: 97.3.  The contine of t
nber Eberline, HP-100 Instrument ID Number: Statt (#321) Calibration Expires: 27-3.  Eberline, HP-100 Operator Name: Hy CAME ROW	ETS. 779.Cluster / Building Survey Unit: 673.2 (#33/)  Therefore Eberline, E600 Instrument ID Number: 5186.22 (#33/)  Eberline, HP-100 Operator Name: 797 (47) (47)
nber. Eberline, E 600 Instrument ID Number: State (#321) Calibration Expires: 27-3.  Eberline, HP-100 Operator Name: HSM CAMERON	ETS. 779. Cluster / Building Survey Unit:  The control of the cont
nber. Eberline, E 600 Instrument ID Number: State 2 (432)  Eberline, HP-100 Operator Name: Hyperator Name: Any Camber Signature:	ETS. 779. Cluster / Building Survey Unit:
nber Eberline, HP-100 Operator Name: H. Signature: Signature: And CAMERON	ETS. 779. Cluster / Building Survey Unit: 673. (#33/) Calibration Expires: 03.3. (#33/)  Therefore Eberline, E600 Instrument ID Number: 518/22/22/32/32/32/32/32/32/32/32/32/32/32/
nber: Eberline, E 600 Instrument ID Number: S.	ETS. 779.Cluster / Building Survey Unit: 673. (#32/ Calibration Expires: 03.2. (#32/ Eberline, E 500 Instrument ID Number: 57.2. (#32/ Eberline, HP-100 Operator Name: 7. 1977 CAMERON
nber: Eberline, E 600 Instrument ID Number: State 2 (#331) Calibration Expires: 03-3.  Eberline, HP-100 Operator Name: H. State Com.	ETS. 779. Cluster / Building Survey Unit: 610. Calibration Expires: 27.27. Calibration
nber Eberline, HP-100 Operator Name: Hyper CAMERON Signature: Calibration Expires: Calibration Calibration Expires: Calibration Expires	ETS. 779. Cluster / Building Survey Unit: 673.2 (#33/) Calibration Expires: 03.3.3.  Instrument ID Number: 5/5/6-22/// Calibration Expires: 03.3.2.4.2.4.  Eberline, HP-100 Operator Name: 799/CAMEROW
nber. Eberline, E 600 Instrument ID Number: S.	ETS. 179. Cluster / Building Survey Unit:  The control of the cont
nber. Eberline, HP-100  Operator Name: H. Stand M. Signature: Signature: Signature: H.	ETS. 779. Cluster / Building Survey Unit:  Therefore Eberline, HP-100  Signature:  Signature:  Signature:  Signature:
nber Eberline, HP-100 Operator Name: H. Signature: Sign	ETS. 779.Cluster / Building Survey Unit: 673. Calibration Expires: 67.3. Ca
nber. Eberline, E 600 Instrument ID Number: S. S. S. S. S. S. S. Signature: Signature: Signature: Signature: Signature: Signature: S.	ETS. 779.Cluster / Building Survey Unit: 673. (1927)  Therefore Eberline, E 600  Calibration Expires: 03-3. (1937)  Calibration Expires: 03-3. (1937)  Calibration Expires: 03-3. (1937)  Calibration Expires: 03-3. (1937)
nber Eberline, E 600 Instrument ID Number: STRAZA (#321) Calibration Expires: 23.2.	ETS. 779.Cluster / Building Survey Unit:  Therefore Eberline, E 600  Therefore Survey Unit:  The substitute of the subst
nber Eberline, E 600 Instrument ID Number: Standard Calibration Expires: 2.3.	ETS. 779.Cluster / Building Survey Unit: 673. (#32/) Calibration Expires: 03.2.
Instrument ID Number: STSP2 (#331) Calibration Expires: 03-3.	Instrument ID Number: STARTS (#321) Calibration Expires: 27-2.
Instrument ID Number: STSLAZ (#331) Calibration Expires: 23.3.	Instrument ID Number: STSLAZZ (#321) Calibration Expires: 2.3.
Instrument ID Number: STAPP (#321) Calibration Expires: 2.3.	Instrument ID Number: STSLAZ (#321) Calibration Expires: 23.2.
Instrument ID Number: STRAP (#331) Calibration Expires:	Survey Unit: Surve
Instrument ID Number: STSL22 (#321) Calibration Expires:	Survey Unit: Surve
Instrument ID Number: STAPP (#321) Calibration Expires:	Survey Unit: Surve
Instrument 10 Number S (SG 22 (#33/) Calibration Expires:	Survey Unit: Surve
(12/2) (12/2)	Survey Unit: Survey Unit: A K/ 27 / A K-22/) Cellbradio Evolves
A 101.20 (# 22.1)	Survey Unit:
	Survey Unit:
	Survey Unit:
	Survey Unit:
Survey Unit:	
Carration Carlo	196 ::: 101/4/10
14/12	19/1/4/
1/4/10 Cale: 7/4/10	17410

						1
יול פרר פדונים	eter / Buildir		Survey Unit: 782 - O	0	Date:	04/05/00
9	TENES TO BEE		S Sedmin Olympian	122 (#3	2/) Calibratic	Calibration Expires: 23-23-00
Instrument Model Number: Eber	Eberline, E 600			\	, A	On more
Detector Probe Type: Ebe	Eberline, HP-100		Operator Name: 14, 000	AMERON	Signature:	
				Control Substitute (12) Inches		(Goundont)
Their soundidates	(IIII)	Ethielber Matteriore	] इंद्र दृष्णुधेन्तुः (लिपश्चाते १ -	Operation of the control of the cont	(i))) (i)) (i))) (i)) (i))) (i))	Inglesocremelonoscusteogileetija) 🙃
The state of the s		Direct Static	%	T.S. Minute		past means
10/10000315	1304	Westername	06/30/00	•	/4./	Mensurement
11/0000 316	1308	Direct Static Measurement		1.5 Minute	7.5	
			>			
		Direct Static Measurement		1.5 Minute	•	
				40.000		
		Direct Static Measurement	•	ennum c.r		•
				1 & Minute		
		Direct Static Measurement				
		Direct Static Measurement		1.5 Minute		
		Direct Static Measurement		1.5 Minute		
Form IVP-1000, July 1999	SAL	1,15m	- Homy	7	1. 02/03/	20 File Index Number
Supervisory increase.	Print Name	me	Signature			

208

Pull   Building	170 Cli	eter / Bulldine		Survey Unit:	292-02		Date:	04/04/00
Survey Location	umber: Eberline.	dine. E 600		Instrument ID Numb	ID Number: 57.56	122 (#3	2/) Calibrat	Calibration Expires: 27-33-00
Detector Probe Type:		Eberline, HP-100		Operator Name:	A. Sa	201 / JPA-4	Signature:	10: Amy
							Special Control of the Control of th	Gemetaka
Sumplessings Pichics of Elbiostaroxings	Houth left Tresection	vilmo	Strippo: Materials Wild	ાંકાવળાં	(O)(\$(0)(\$)	Shalls Galls Galls	(mes607/ndb)	(instructuralphonofoutrecendoctification)
1VF0000421		Mbo	Difect Static Measurement	15564 16837	(4/04) (00	1.5 Minute	17,3	POST MEDIA SAMPLING
	422	0945	Direct Static Measurement			1.5 Minute	3.5	
	4.23 0450	0350	Direct Static Measurement			1.5 Minute	27.3	
	424	5560	Direct Static Measurement		-	1.5 Minute	31.6	
<del>(.</del>	426 1007	1001	Direct Static Measurement			1.5 Minute	17.33	
	427	7101	Direct Static Measurement			1.5 Minute	3.79	
>	dop	4101	Direct Static Measurement			1.5 Minute	7,26	
orm IVP-1000, July 1999	1999						,	File Index Number

Date

Signature

Print Name

Supervisory Review: \_\_

Survey Location: REETS, 779, Cluster / Building	uster / Bulld	na	Survey Unit:	10.1	Date:	
Instrument Model Number. Ebe	Eberline, E 600		Instrument ID Number:	5622 (#	32/) Calibra	Calibration Expires: 23-20
1	Eberline, HP-100		Operator Name:	Amilia	Signature:	ure: A Almy
Sampleuseanout Generalendere (chikorasekielender)	Cully	Semigloop Meet to On On USAN	ୗନ(ଲୋକ୍ଟୋଡ଼ାଏଏଲ)	Opla Shress Shress	GROSSIM neimoni. (Galevassem)	(Instructional Countries) (Instructional Instructional Architecture) (Instructional Instructional Architecture)
1VF0000430	2201	Difect Static Measurement	(60/4) (+CEA) W 1955/8	1.5 Minute	17.22	1005 MEDIA SAMPLINS
:   431	1601	Direct Static Measurement		1.5 Minute	17.81	
432 1033	1033	Direct Static Measurement		1.5 Minute	7.63	
433	2501	Direct Static Measurement		1.5 Minute	14.7	
434	1501	Direct Static Measurement		1.5 Minute	1.92	
435	001	Direct Static Measurement		1.5 Minute	ust in the	
Sur	101	Direct Static Measurement	7	1.5 Minute	14.17	
orm IVP-1000, July 1999						

90

File Index Number\_

Date

Signature

Print Name

Supervisory Review: \_

里	
7	
Ė	
₹	
K.	
$\supseteq$	
Ē	
푯	
Έ.	
Ę	
5	
T.	
RF	
ı	
F	
EN	
PEN	
TABEN !	
PERCEN	
PERCEN	
HADEN	

Survey Location. REETS. 779 Cluster / Building  Detector Probe Type: Eberline, HP-100  Detector Probe Type: Eberline, HP-100  Sample State  Sample State  WHY COUNTY HHY IN	S21) Calibration Expires: 03/23/00 Signature: Manuelli Comments and Calibration Expires: 03/23/00 Signature: Manuelli Comments application in the American Calibration of
---	---

File Index Number \_\_\_

Date

Signature

Print Name

Supervisory Review: \_\_

#### Background Data, Survey Unit 727-01

					Instrument		Background				Instrument
			E600	Probe	Operating	Channel	Compensation	Recorded		Instrument	raw count
Sample Location	Date	Time	Serial #	Serial #	Mode	0,	Mode	Value	Units	Efficiency	rate (cpm)
BACKGROUND	2/2/00	15:30:00	321	109	Scaler		Gross	7.93	dpm/100cm <sup>2</sup>	0.1949	1.5
<b>3ACKGROUND</b>	2/2/00	15:32:00	321	109	Scaler	Alpha	Gross	11.30	dpm/100cm <sup>2</sup>	0.1949	2.2
SACKGROUND	2/2/00	15:34:00	321	109	Scaler		Gross	7.75	dpm/100cm <sup>2</sup>	0.1949	1.5
<b>3ACKGROUND</b>	2/2/00	16:52:00	321	109	Scaler		Gross	8.24	dpm/100cm <sup>2</sup>	0.1949	1.6
SACKGROUND	2/2/00	16:54:00	321	109	Scaler		Gross	8.32	dpm/100cm <sup>2</sup>	0.1949	1.6
BACKGROUND	2/2/00	16:56:00	321	109	Scaler		Gross	4.76	dpm/100cm <sup>2</sup>	0.1949	6.0
<b>3ACKGROUND</b>	2/3/00	8:45:00	321	109	Scaler		Gross	4.76	dpm/100cm <sup>2</sup>	0.1949	6.0
<b>3ACKGROUND</b>	2/3/00	8:47:00	321	109	Scaler		Gross	4.46	dpm/100cm <sup>2</sup>	0.1949	6.0
BACKGROUND	2/3/00	8:49:00	321	109	Scaler		Gross	4.42	dpm/100cm <sup>2</sup>	0.1949	0.9
<b>3ACKGROUND</b>	2/3/00	9:50:00	321	109	Scaler		Gross	4.73	dpm/100cm <sup>2</sup>	0.1949	6.0
<b>3ACKGROUND</b>	2/3/00	9:52:00	321	109	Scaler		Gross	4.99	dpm/100cm <sup>2</sup>	0.1949	1.0
BACKGROUND	2/3/00	9:54:00	321	109	Scaler		Gross	7.56	dpm/100cm <sup>2</sup>	0.1949	1.5

Summary Statistics

Number of Measurements Mean	12
Log Normal Mean	6.284391 6.3
Std. Deviation CV	2.2144806 0.3354427

#### Response Data, Survey Unit 727-01

					Instrument		Background		
Sample			E600	Probe	Operating	Channel	Compensation	Recorded	
Location	Date	Time	••	Serial #	Mode		Mode	Value	Units
RESP/CHECK	2/2/00	15:36:00	321	109	Scaler		Gross	1580	dpm/100cm <sup>2</sup>
RESP/CHECK	2/2/00	15:38:00		109	Scaler		Gross	1580	dpm/100cm <sup>2</sup>
RESP/CHECK	2/2/00	15:41:00	321	109	Scaler		Gross	1560	dpm/100cm <sup>2</sup>
RESP/CHECK	2/2/00	16:58:00	321	109	Scaler		Gross	1410	dpm/100cm <sup>2</sup>
RESP/CHECK	2/2/00	16:59:00	321	109	Scaler		Gross	1570	dpm/100cm <sup>2</sup>
RESP/CHECK	2/2/00	17:01:00	321	109	Scaler		Gross	1420	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	8:51:00	321	109	Scaler		Gross	1450	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	8:52:00	321	109	Scaler		Gross	1530	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	8:54:00	321	109	Scaler		Gross	1690	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	9:58:00	321	109	Scaler		Gross	1500	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	10:00:00	321	109	Scaler		Gross	1440	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	10:07:00	321	109	Scaler	Alpha	Gross	1470	dpm/100cm <sup>2</sup>

Probe #	Response	-50%	20%
109	1631	1305	1957

Direct Static Surface Contamination Measurements, Survey Unit 727-01

#### Mean Of Replicate Measurements

Post Surface Media Sampling Direct Static Surface Measurements, Survey Unit 727-01

			E600	Probe	Instrument Operating	Channel	Background Compensation	Recorded		Instrument	Instrument raw count rate	Count
Sample Location	Date	Time	Serial #	Serial #	Mode		Mode		Units	Efficiency	(cbm)	(seconds)
IVP0000386	2/3/00	9:18:00	321	109	Scaler		Gross	17.40	dpm/100cm <sup>2</sup>	0.1949	3.4	06
IVP0000387	2/3/00	9:23:00	321	109	Scaler	Alpha	Gross	20.80	dpm/100cm <sup>2</sup>	0.1949	4.1	06
IVP0000390	2/3/00	9:32:00	321	109	Scaler	Alpha	Gross	17.60	dpm/100cm <sup>2</sup>	0.1949	3.4	06
IVP0000396	2/2/00	15:52:00	321	109	Scaler	Alpha	Gross	7.56	dpm/100cm <sup>2</sup>	0.1949	1.5	06
IVP0000397	2/2/00	15:56:00	321	109	Scaler	Alpha	Gross	4.32	dpm/100cm <sup>2</sup>	0.1949	0.8	06
IVP0000398	2/2/00	16:01:00	321	109	Scaler	Alpha	Gross	18.02	dpm/100cm <sup>2</sup>	0.1949	3.5	06
IVP0000399	2/2/00	16:06:00	321	109	Scaler	Alpha	Gross	4.16	dpm/100cm <sup>2</sup>	0.1949	0.8	6
IVP0000400	2/2/00	16:11:00	321	109	Scaler	Alpha	Gross	14.15	dpm/100cm <sup>2</sup>	0.1949	2.8	06
IVP0000401	2/2/00	16:18:00	321	109	Scaler	Alpha	Gross	17.57	dpm/100cm <sup>2</sup>	0.1949	3.4	06
IVP0000402	2/2/00	16:22:00	321	109	Scaler	Alpha	Gross	24.80	dpm/100cm <sup>2</sup>	0.1949	<b>4</b> .8	06
IVP0000403	2/2/00	16:26:00	321	109	Scaler	Alpha	Gross	8.80	dpm/100cm <sup>2</sup>	0.1949	1.7	6
IVP0000404	2/2/00	16:29:00	321	109	Scaler	Alpha	Gross	15.80	dpm/100cm <sup>2</sup>	0.1949	3.1	6
IVP0000405	2/2/00	16:32:00	321	109	Scaler	Alpha	Gross	12.30	dpm/100cm <sup>2</sup>	0.1949	2.4	06
IVP0000406	2/2/00	16:36:00	321	109	Scaler	Alpha	Gross	99.7	dpm/100cm <sup>2</sup>	0.1949	1.5	6
IVP0000407	2/2/00	16:41:00	321	109	Scaler	Alpha	Gross	7.71	dpm/100cm <sup>2</sup>	0.1949	1.5	<b>6</b>
IVP0000408	2/2/00	16:45:00	321	109	Scaler	Alpha	Gross	31.50	dpm/100cm <sup>2</sup>	0.1949	6.1	06
IVP0000409	2/2/00	16:50:00	321	109	Scaler	Alpha	Gross	25.10	dpm/100cm <sup>2</sup>	0.1949	4.9	06

#### Surface Media Sample Data, Survey Unit 727-01 Alpha Isotopic Analysis

3	Lab Sample ID #	Sample Weight	Date	Time	Units	-	Am-241	:	Pu-238	ā.	Pu-239/240	U-234		U-23\$	U-238	Total Transuranic Activity	Total Uranium Activity
		(grams)	MMDD/YY			Reported	Reported Values w/samples < MDA at 0.5 MDA	Reported	Reported Values w/samples < MDA at 0.5 MDA	Reported	Reported Values w/samples < MDA at 0.5 MDA	Reported	Reported	Reported Values w/samples < MDA at 0.5 MDA	Reported Value	dpm/100 cm²	dpm/100 cm²
MED0000386	264771	11.03	1/24/00	8:00	dpm/100 cm²	99.0	0.88	1.29	59'0	197	0.91	31.11	1,36	0.93	30.96	2.43	63.02
MED0000387	264772	13.70	1/24/00	\$0:8	dpm/100 cm²	1.36	1.36	0.51	92'0	1.07	35.0	60.67	2.61	1.31	59.34	2.15	121.32
MED0000390	264773	11.08	1/24/00	8:15	dpm/100 cm²	1.45	1.45	0.82	0.41	1.11	0.56	29.24	2.28	. 2.28	30.58	2.42	62.06
MED0000398	264774	14.91	1/24/00	8:25	dpm/100 cm²	0.57	0.57	1.16	95.0	2.03	1.02	22.18	2.55	1.28	24.68	2.17	48.12
MED0000397	264775	15.84	1/24/00	8:30	dpm/100 cm²	1.72	1.72	1.07	75'0	1.07	75'0	26.33	1.98	1.00	28.71	2.79	54.10
MED0000398	264778	13.45	1/24/00	8:35	dpm/100 cm²	1.20	09.0	0.51	92'0	1,96	0.99	19.85	11.17	0.59	19.77	1.85	40.21
MED0000399	264777	13.74	1/24/00	8:40	dpm/100 cm²	1.44	1,44	1.00	09'0	1.00	0.50	20.28	2.15	1.08	18.08	2,44	39.42
MED0000400	264778	7.52	1/24/00	8:45	dpm/100 cm²	2.23	2.23	0.57	620	0.76	0.39	13.44	141	0.71	13.26	2.91	27.41
MED0000401	264779	3.71	1/24/00	9:50	dpm/100 cm²	16.0	0.91	0.31	0.16	1.34	1.34	3.17	0.46	0.23	3.43	2.41	6.63
MED0000402	264780	15.31	1/24/00	8:55	dpm/100 cm²	1.55	1.55	0.50	0.26	1.28	0,83	19.05	2.51	87	23.17	2.43	43.48
MED0000403	264781	14.57	1/24/00	0:00	dpm/100 cm²	1.49	51.0	1.15	65.0	1.56	0.76	20.18	3.15	1.58	18.90	2.10	40.96
MED0000404	264782	8.16	1/24/00	9:10	dpm/100 cm²	2.04	2.04	0.75	96.0	1.39	1.39	6.04	1.06	16.0	6.67	3.81	13.25
MED0000405	264783	11.00	1/24/00	9:15	dpm/100 cm²	0.85	0.85	0.91	0.46	1,46	6.73	12.13	1.87	0.94	14.08	2.04	27.15
MED0000408	264784	13.10	1/24/00	9:20	dpm/100 cm <sup>2</sup>	1.63	1.63	0.44	22.0	0.91	0.46	21.99	1.69	0.85	19.92	2.31	42.78
MED0000407	264785	17.59	1/24/00	9:25	dpm/100 cm²	2.68	2.68	1.74	0.67	1.97	0.90	31.75	3.11	1.56	32.05	15.4	65.36
MED0000408	264786	10.28	1/24/00	9:30	dpm/100 cm <sup>2</sup>	0.90	0.90	1.32	0.66	1.22	0.61	16.23	1.60	0.85	14.27	2.17	31.45
MED0000409	284787	11.29	1/24/00	8:40	dpm/100 cm²	15.	1.54	1.07	0.54	1,4	0.72	19.07	1.25	0.63	15.14	2.80	34.84

#### Background Data, Survey Unit 727-02

					Instrument		Background				Instrument
			E600	Probe	Operating	Channel	Compensation	Recorded		Instrument	raw count
Sample Location	Date	Time	Serial #	Serial #	Mode	Selected	Mode	Value	Units	Efficiency	rate (cpm)
BACKGROUND	1/25/00	9:06:00	321	109	Scaler	Alpha	Gross	15.10	dpm/100cm <sup>2</sup>	0.1949	2.9
BACKGROUND	1/25/00	9:08:00	321	109	Scaler	Alpha	Gross	8.45	dpm/100cm <sup>2</sup>	0.1949	1.6
BACKGROUND	1/25/00	9:09:00	. 321	109	Scaler	Alpha	Gross	18.70	dpm/100cm <sup>2</sup>	0.1949	3.6
BACKGROUND	1/25/00	10:21:00	321	109	Scaler	Alpha	Gross	14.90	dpm/100cm <sup>2</sup>	0.1949	2.9
BACKGROUND	1/25/00	10:23:00	321	109	Scaler	Alpha	Gross	15.60	dpm/100cm <sup>2</sup>	0.1949	3.0
BACKGROUND	1/25/00	10:25:00	321	109	Scaler	Alpha	Gross	22.50	dpm/100cm <sup>2</sup>	0.1949	4.4
BACKGROUND	1/25/00	11:42:00	321	109	Scaler	Alpha	Gross	12.30	dpm/100cm <sup>2</sup>	0.1949	2.4
BACKGROUND	1/25/00	11:44:00	321	109	Scaler	Alpha	Gross	10.30	dpm/100cm <sup>2</sup>	0.1949	2.0
BACKGROUND	1/25/00	11:45:00	321	109	Scaler	Alpha	Gross	15.10	dpm/100cm <sup>2</sup>	0.1949	2.9

#### Summary Statistics

တ	14.8	14,226519	15.1	4.2152336	0.2853486
Number of Measurements	Mean	Log Normal Mean	Median	Std. Deviation	S

#### Response Data, Survey Unit 727-02

•					Instrument		Background		
Sample			E600	Probe		Channel	Compensation	Recorded	
Location	Date	Time	Serial #	Serial#		Selected	Mode	Value	Units
RESP/CHECK	1/25/00	8:58:00	321	109	Scaler	Alpha	Gross	1450	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	8:59:00	321	109		Alpha	Gross	. 1450	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	9:01:00	321	109		Alpha	Gross	1490	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	10:16:00	321	109		Alpha	Gross	1610	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	10:18:00	321	109		Alpha	Gross	1520	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	10:20:00	321	109		Alpha	Gross	1570	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	11:37:00	321	109		Alpha	Gross	1530	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	11:39:00	321	109		Alpha	Gross	1500	dpm/100cm <sup>2</sup>
RESP/CHECK	1/25/00	11:40:00	321	109		Alpha	Gross	1440	dpm/100cm <sup>2</sup>

20%	1957
-20%	1305
Response	1631
Probe #	109

# Direct Static Surface Contamination Measurements, Survey Unit 727-02

Scaler         Alpha         Gross         12.30         cpm/100cm²         Efficiency         rate (cpm)         46           109         Scaler         Alpha         Gross         12.30         cpm/100cm²         0.1949         2.4           109         Scaler         Alpha         Gross         21.50         cpm/100cm²         0.1949         4.2           109         Scaler         Alpha         Gross         21.50         cpm/100cm²         0.1949         4.2           109         Scaler         Alpha         Gross         21.50         cpm/100cm²         0.1949         4.2           109         Scaler         Alpha         Gross         21.50         cpm/100cm²         0.1949         4.3           109         Scaler         Alpha         Gross         21.50         cpm/100cm²         0.1949         4.0           109         Scaler         Alpha         Gross         21.50         cpm/100cm²         0.1949         4.1           109         Scaler         Alpha         Gross         22.50         cpm/100cm²         0.1949         4.2           109         Scaler         Alpha         Gross         22.50         cpm/100cm²         0.1949				E600 Serial	Probe	Instrument	Channel	Background Compensation	Recorded		Instrument	Instrument	Count
1755/00         9:55/00         321         109         Scaler         Apha         Gross         12.30         dpm/100cm²         0.1849         2.4           1755/00         10:00:00         321         109         Scaler         Apha         Gross         12.30         dpm/100cm²         0.1849         1.7           1755/00         10:00:00         321         109         Scaler         Apha         Gross         21.0         dpm/100cm²         0.1949         3.1           1755/00         10:00:00         321         109         Scaler         Apha         Gross         21.0         dpm/100cm²         0.1949         3.1           1755/00         10:00:00         321         109         Scaler         Apha         Gross         15.70         dpm/100cm²         0.1949         3.1           1755/00         10:29:00         321         109         Scaler         Apha         Gross         2.10         dpm/100cm²         0.1949         3.1           1755/00         10:29:00         321         109         Scaler         Apha         Gross         2.10         dpm/100cm²         0.1949         3.1           1755/00         10:37:00         321         109	e Location	Date	Time	**	Serial #	Mode	Selected	Mode	Value	Units	Efficiency	rate (cpm)	(seconds)
1/25/00         100000         321         109         Scaler         Alpha         Gross         8.82         dpm/100cm²         0.1949         1.7           1/25/00         1002:00         321         109         Scaler         Alpha         Gross         21.50         dpm/100cm²         0.1949         3.9           1/25/00         1002:00         321         109         Scaler         Alpha         Gross         21.50         dpm/100cm²         0.1949         3.9           1/25/00         1003:00         321         109         Scaler         Alpha         Gross         21.50         dpm/100cm²         0.1949         3.4           1/25/00         1003:00         321         109         Scaler         Alpha         Gross         21.90         dpm/100cm²         0.1949         3.4           1/25/00         1023:00         321         109         Scaler         Alpha         Gross         20.50         dpm/100cm²         0.1949         3.4           1/25/00         1034:00         321         109         Scaler         Alpha         Gross         20.50         dpm/100cm²         0.1949         3.4           1/25/00         1034:00         321         109	0000351	1/25/00	9:59:00	321	109	Scaler	Alpha	Gross	12.30	dpm/100cm <sup>2</sup>	0.1949	2.4	6
1/25/00         10:02:00         321         109         Scaler         Alpha         Gross         21:50         dpm/100cm²         0:1949         3.9           1/25/00         10:03:00         321         109         Scaler         Alpha         Gross         20:10         dpm/100cm²         0:1949         3.9           1/25/00         10:05:00         321         109         Scaler         Alpha         Gross         20:10         dpm/100cm²         0:1949         3.9           1/25/00         10:05:00         321         109         Scaler         Alpha         Gross         21:00         dpm/100cm²         0:1949         3.4           1/25/00         10:23:00         321         109         Scaler         Alpha         Gross         21:00         dpm/100cm²         0:1949         3.4           1/25/00         10:32:00         321         109         Scaler         Alpha         Gross         20:50         dpm/100cm²         0:1949         3.4           1/25/00         10:32:00         321         109         Scaler         Alpha         Gross         20:50         dpm/100cm²         0:1949         3.4           1/25/00         10:32:00         321         109 </td <td>0000352</td> <td>1/25/00</td> <td>10:00:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>8.82</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>1.7</td> <td>06</td>	0000352	1/25/00	10:00:00	321	109	Scaler	Alpha	Gross	8.82	dpm/100cm <sup>2</sup>	0.1949	1.7	06
1/25/00         10.04-00         321         109         Scaler         Alpha         Gross         20.10         dpm/100cm²         0.1949         3.9           1/25/00         10.086:00         321         109         Scaler         Alpha         Gross         15.70         dpm/100cm²         0.1949         3.1           1/25/00         10.086:00         321         109         Scaler         Alpha         Gross         21.30         dpm/100cm²         0.1949         3.1           1/25/00         10.28:00         321         109         Scaler         Alpha         Gross         21.30         dpm/100cm²         0.1949         4.1           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         21.30         dpm/100cm²         0.1949         4.1           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         20.50         dpm/100cm²         0.1949         4.1           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         22.00         dpm/100cm²         0.1949         4.1           1/25/00         10:34:00         321         109	0000353	1/25/00	10:02:00	321	109	Scaler	Alpha	Gross	21.50	dpm/100cm <sup>2</sup>	0.1949	4.2	8
1725/00         10:06:00         321         109         Scaler         Alpha         Gross         15.70         dpm/100cm²         0.1949         3.1           1725/00         10:06:00         321         109         Scaler         Alpha         Gross         15.70         dpm/100cm²         0.1949         3.4           1725/00         10:27:00         321         109         Scaler         Alpha         Gross         20:50         dpm/100cm²         0.1949         4.3           1725/00         10:32:00         321         109         Scaler         Alpha         Gross         20:50         dpm/100cm²         0.1949         4.1           1725/00         10:34:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0.1949         4.1           1725/00         10:34:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0.1949         4.4           1725/00         10:34:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0.1949         4.4           1725/00         10:46:00         321         109 </td <td>0000354</td> <td>1/25/00</td> <td>10:04:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>20.10</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>3.9</td> <td>8</td>	0000354	1/25/00	10:04:00	321	109	Scaler	Alpha	Gross	20.10	dpm/100cm <sup>2</sup>	0.1949	3.9	8
1/125/00         10:08:00         321         109         Scaler         Alpha         Gross         17.30         dpm/100cm²         0.1949         3.4           1/25/00         10:23:00         321         109         Scaler         Alpha         Gross         21:90         dpm/100cm²         0.1949         4.3           1/25/00         10:23:00         321         109         Scaler         Alpha         Gross         21:00         dpm/100cm²         0.1949         4.0           1/25/00         10:33:00         321         109         Scaler         Alpha         Gross         21:0         dpm/100cm²         0.1949         4.0           1/25/00         10:33:00         321         109         Scaler         Alpha         Gross         20:0         dpm/100cm²         0.1949         4.0           1/25/00         10:33:00         321         109         Scaler         Alpha         Gross         12:40         dpm/100cm²         0.1949         4.0           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         12:40         dpm/100cm²         0.1949         4.1           1/25/00         10:44:00         321         109 <td>0000355</td> <td>1/25/00</td> <td>10:06:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>15.70</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>3.1</td> <td>6</td>	0000355	1/25/00	10:06:00	321	109	Scaler	Alpha	Gross	15.70	dpm/100cm <sup>2</sup>	0.1949	3.1	6
1/25/00         10:29:00         321         109         Scaler         Alpha         Gross         21:90         dpm/100cm²         0:1949         4.3           1/25/00         10:29:00         321         109         Scaler         Alpha         Gross         21:30         dpm/100cm²         0:1949         4.0           1/25/00         10:32:00         321         109         Scaler         Alpha         Gross         20:50         dpm/100cm²         0:1949         4.0           1/25/00         10:33:00         321         109         Scaler         Alpha         Gross         20:50         dpm/100cm²         0:1949         4.0           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         12:40         dpm/100cm²         0:1949         4.0           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         13:40         dpm/100cm²         0:1949         4.0           1/25/00         10:44:00         321         109         Scaler         Alpha         Gross         22:00         dpm/100cm²         0:1949         4.0           1/25/00         10:44:00         321         109 </td <td>0000356</td> <td>1/25/00</td> <td>10:08:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>17,30</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>3.4</td> <td>8</td>	0000356	1/25/00	10:08:00	321	109	Scaler	Alpha	Gross	17,30	dpm/100cm <sup>2</sup>	0.1949	3.4	8
1/25/00         10.29:00         321         109         Scaler         Alpha         Gross         20.50         dpm/100cm²         0.1949         4.1           1/25/00         10:32:00         321         109         Scaler         Alpha         Gross         21.10         dpm/100cm²         0.1949         4.1           1/25/00         10:32:00         321         109         Scaler         Alpha         Gross         21.40         dpm/100cm²         0.1949         4.1           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0.1949         4.4           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         13.40         dpm/100cm²         0.1949         4.4           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         13.40         dpm/100cm²         0.1949         4.4           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         22.00         dpm/100cm²         0.1949         4.4           1/25/00         10:46:00         321         109 </td <td>0000357</td> <td>1/25/00</td> <td>10:27:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>21.90</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>4.3</td> <td>6</td>	0000357	1/25/00	10:27:00	321	109	Scaler	Alpha	Gross	21.90	dpm/100cm <sup>2</sup>	0.1949	4.3	6
1/25/00         10:32:00         321         109         Scaler         Alpha         Gross         21:10         dpm/100cm²         0:1949         4.1           1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         20:00         dpm/100cm²         0:1949         4.0           1/25/00         10:33:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0:1949         4.4           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0:1949         4.4           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0:1949         2.4           1/25/00         10:44:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0:1949         2.4           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         19:0         dpm/100cm²         0:1949         4.4           1/25/00         10:46:00         321         109 <td>0000358</td> <td>1/25/00</td> <td>10:29:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>20.50</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>4.0</td> <td>06</td>	0000358	1/25/00	10:29:00	321	109	Scaler	Alpha	Gross	20.50	dpm/100cm <sup>2</sup>	0.1949	4.0	06
1/25/00         10:34:00         321         109         Scaler         Alpha         Gross         20:60         dpm/100cm²         0:1949         4.0           1/25/00         10:37:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0:1949         2.4           1/25/00         10:39:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²         0:1949         2.4           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²         0:1949         2.4           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         22:00         dpm/100cm²         0:1949         2.4           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         22:00         dpm/100cm²         0:1949         2.9           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         22:00         dpm/100cm²         0:1949         2.9           1/25/00         10:60:00         321         109 </td <td>0000359</td> <td>1/25/00</td> <td>10:32:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>21.10</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>1.4</td> <td>06</td>	0000359	1/25/00	10:32:00	321	109	Scaler	Alpha	Gross	21.10	dpm/100cm <sup>2</sup>	0.1949	1.4	06
1/25/00         10:37:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0:1949         2.4           1/25/00         10:39:00         321         109         Scaler         Alpha         Gross         12.40         dpm/100cm²         0:1949         3.8           1/25/00         10:39:00         321         109         Scaler         Alpha         Gross         13.40         dpm/100cm²         0:1949         3.8           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         13.40         dpm/100cm²         0:1949         3.8           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         13.40         dpm/100cm²         0:1949         3.7           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0:1949         3.7           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0:1949         3.7           1/25/00         11:04:00         321         109 </td <td>0000000</td> <td>1/25/00</td> <td>10:34:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>20.60</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>4.0</td> <td>06</td>	0000000	1/25/00	10:34:00	321	109	Scaler	Alpha	Gross	20.60	dpm/100cm <sup>2</sup>	0.1949	4.0	06
1/25/00         10.39:00         321         109         Scaler         Alpha         Gross         122.40         dpm/100cm²         0.1949         44           1/25/00         10:33:00         321         109         Scaler         Alpha         Gross         19.40         dpm/100cm²         0.1949         3.8           1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         13.40         dpm/100cm²         0.1949         3.8           1/25/00         10:44:00         321         109         Scaler         Alpha         Gross         19.40         dpm/100cm²         0.1949         3.8           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         19.40         dpm/100cm²         0.1949         3.7           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         12.00         dpm/100cm²         0.1949         3.7           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         12.00         dpm/100cm²         0.1949         3.7           1/25/00         11:04:00         321         109 </td <td>0000361</td> <td>1/25/00</td> <td>10:37:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>12.40</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>2.4</td> <td>06</td>	0000361	1/25/00	10:37:00	321	109	Scaler	Alpha	Gross	12.40	dpm/100cm <sup>2</sup>	0.1949	2.4	06
1/25/00         10:43:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²         0.1949         3.8           1/25/00         10:44:00         321         109         Scaler         Alpha         Gross         13:40         dpm/100cm²         0.1949         2.6           1/25/00         10:46:00         321         109         Scaler         Alpha         Gross         29:90         dpm/100cm²         0.1949         2.6           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         12:00         dpm/100cm²         0.1949         2.6           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         12:00         dpm/100cm²         0.1949         2.6           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         12:00         dpm/100cm²         0.1949         2.6           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         12:00         dpm/100cm²         0.1949         2.9           1/25/00         11:26:00         321         109 </td <td>0000362</td> <td>1/25/00</td> <td>10:39:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>22.40</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>4.4</td> <td>6</td>	0000362	1/25/00	10:39:00	321	109	Scaler	Alpha	Gross	22.40	dpm/100cm <sup>2</sup>	0.1949	4.4	6
1/25/00         10-44:00         321         109         Scaler         Apha         Gross         13.40         dpm/100cm²         0.1949         2.6           1/25/00         10-46:00         321         109         Scaler         Apha         Gross         29.90         dpm/100cm²         0.1949         5.8           1/25/00         10-56:00         321         109         Scaler         Apha         Gross         19.10         dpm/100cm²         0.1949         3.7           1/25/00         10:56:00         321         109         Scaler         Apha         Gross         22.00         dpm/100cm²         0.1949         3.7           1/25/00         10:56:00         321         109         Scaler         Apha         Gross         24.90         dpm/100cm²         0.1949         3.7           1/25/00         11:04:00         321         109         Scaler         Apha         Gross         24.90         dpm/100cm²         0.1949         3.7           1/25/00         11:12:00         321         109         Scaler         Apha         Gross         24.70         dpm/100cm²         0.1949         3.7           1/25/00         11:16:00         321         109	0000363	1/25/00	10:43:00	321	109	Scaler	Alpha	Gross	19.40	dpm/100cm <sup>2</sup>	0.1949	3.8	06
1725/00         10.46:00         321         109         Scaler         Alpha         Gross         29:90         dpm/100cm²         0.1949         5.8           1725/00         10:52:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1725/00         10:56:00         321         109         Scaler         Alpha         Gross         22:00         dpm/100cm²         0.1949         4.3           1725/00         10:56:00         321         109         Scaler         Alpha         Gross         24:90         dpm/100cm²         0.1949         4.9           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         24:90         dpm/100cm²         0.1949         4.9           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         29:40         dpm/100cm²         0.1949         4.8           1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         29:40         dpm/100cm²         0.1949         4.8           1/25/00         11:24:00         321         109 </td <td>0000364</td> <td>1/25/00</td> <td>10:44:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>13.40</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>2.6</td> <td>8</td>	0000364	1/25/00	10:44:00	321	109	Scaler	Alpha	Gross	13.40	dpm/100cm <sup>2</sup>	0.1949	2.6	8
1/25/00         10:52:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         22.00         dpm/100cm²         0.1949         4.3           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         22.00         dpm/100cm²         0.1949         2.9           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0.1949         2.9           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0.1949         3.7           1/25/00         11:07:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         3.7           1/25/00         11:13:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         3.7           1/25/00         11:24:00         321         109 </td <td>0000365</td> <td>1/25/00</td> <td>10:46:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>29.90</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>5.8</td> <td>6</td>	0000365	1/25/00	10:46:00	321	109	Scaler	Alpha	Gross	29.90	dpm/100cm <sup>2</sup>	0.1949	5.8	6
1/25/00         10:54:00         321         109         Scaler         Alpha         Gross         22.00         dpm/100cm²         0.1949         4.3           1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0.1949         2.9           1/25/00         9:19:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0.1949         2.9           1/25/00         1:07:00         321         109         Scaler         Alpha         Gross         28.40         dpm/100cm²         0.1949         3.7           1/25/00         1:1:07:00         321         109         Scaler         Alpha         Gross         29.40         dpm/100cm²         0.1949         3.7           1/25/00         1:1:12:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         3.7           1/25/00         1:2:1:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.1           1/25/00         1:2:10:00         321         109<	9980000	1/25/00	10:52:00	321	109	Scaler	Alpha	Gross	19.10	dpm/100cm <sup>2</sup>	0.1949	3.7	8
1/25/00         10:56:00         321         109         Scaler         Alpha         Gross         24:90         dpm/100cm²         0.1949         2.9           1/25/00         9:19:00         321         109         Scaler         Alpha         Gross         24:90         dpm/100cm²         0.1949         2.9           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         24:90         dpm/100cm²         0.1949         3.7           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         29:40         dpm/100cm²         0.1949         7.1           1/25/00         11:12:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         3.1           1/25/00         11:16:00         321         109         Scaler         Alpha         Gross         15.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109 <td>0000367</td> <td>1/25/00</td> <td>10:54:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>22.00</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>4.3</td> <td>06</td>	0000367	1/25/00	10:54:00	321	109	Scaler	Alpha	Gross	22.00	dpm/100cm <sup>2</sup>	0.1949	4.3	06
1/25/00         9:19:00         321         109         Scaler         Alpha         Gross         24.90         dpm/100cm²         0.1949         4.9           1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         18.90         dpm/100cm²         0.1949         7.1           1/25/00         11:07:00         321         109         Scaler         Alpha         Gross         29.40         dpm/100cm²         0.1949         7.1           1/25/00         11:12:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         7.1           1/25/00         11:16:00         321         109         Scaler         Alpha         Gross         15.80         dpm/100cm²         0.1949         3.1           1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         1.7           1/25/00         11:28:00         321         109 <td>0000368</td> <td>1/25/00</td> <td>10:56:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>14.70</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>2.9</td> <td>6</td>	0000368	1/25/00	10:56:00	321	109	Scaler	Alpha	Gross	14.70	dpm/100cm <sup>2</sup>	0.1949	2.9	6
1/25/00         11:04:00         321         109         Scaler         Alpha         Gross         18:90         dpm/100cm²         0.1949         3.7           1/25/00         11:07:00         321         109         Scaler         Alpha         Gross         29:40         dpm/100cm²         0.1949         7.1           1/25/00         11:07:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         7.1           1/25/00         11:16:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         3.1           1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.7           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1/25/00         11:31:00         321         109 </td <td>6980000</td> <td>1/25/00</td> <td>9:19:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Ałpha</td> <td>Gross</td> <td>24.90</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>4.9</td> <td>06</td>	6980000	1/25/00	9:19:00	321	109	Scaler	Ałpha	Gross	24.90	dpm/100cm <sup>2</sup>	0.1949	4.9	06
1/25/00         11:07:00         321         109         Scaler         Alpha         Gross         36:30         dpm/100cm²         0.1949         7.1           1/25/00         11:12:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         7.1           1/25/00         11:25/00         11:24:00         321         109         Scaler         Alpha         Gross         15.80         dpm/100cm²         0.1949         3.1           1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         8.82         dpm/100cm²         0.1949         1.7           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         1.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         3.7           1/25/00         11:31:00         3	00000370	1/25/00	11:04:00	321	109	Scaler	Alpha	Gross	18.90	dpm/100cm <sup>2</sup>	0.1949	3.7	8
1/25/00         11:12:00         321         109         Scaler         Alpha         Gross         29:40         dpm/100cm²         0.1949         5.7           1/25/00         9:23:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         5.7           1/25/00         11:16:00         321         109         Scaler         Alpha         Gross         15.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         3.1           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         8.82         dpm/100cm²         0.1949         1.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         3.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         3.7           1/25/00         11:31:00         321         109 <td>0000371</td> <td>1/25/00</td> <td>11:07:00</td> <td>321</td> <td>109</td> <td>Scaler</td> <td>Alpha</td> <td>Gross</td> <td>36.30</td> <td>dpm/100cm<sup>2</sup></td> <td>0.1949</td> <td>7.1</td> <td>6</td>	0000371	1/25/00	11:07:00	321	109	Scaler	Alpha	Gross	36.30	dpm/100cm <sup>2</sup>	0.1949	7.1	6
1/25/00         9:23:00         321         109         Scaler         Alpha         Gross         24.70         dpm/100cm²         0.1949         4.8           1/25/00         11:16:00         321         109         Scaler         Alpha         Gross         15.80         dpm/100cm²         0.1949         3.1           1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         4.8           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         8.82         dpm/100cm²         0.1949         1.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         3.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         3.7           1/25/00         11:33:00         321         109         Scaler         Alpha         Gross         19.10         dpm/100cm²         0.1949         3.7	0000372	1/25/00	11:12:00	321	109	Scaler	Alpha	Gross	29.40	dpm/100cm <sup>2</sup>	0.1949	5.7	06
1/25/00         11:16:00         321         109         Scaler         Alpha         Gross         15:80         dpm/100cm²         0.1949         3.1           1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         15:80         dpm/100cm²         0.1949         3.1           1/25/00         9:27:00         321         109         Scaler         Alpha         Gross         8:82         dpm/100cm²         0.1949         1.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1/25/00         11:33:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7	0000373	1/25/00	9:23:00	321	109	Scaler	Alpha	Gross	24.70	dpm/100cm <sup>2</sup>	0.1949	4.8	06
1/25/00         11:24:00         321         109         Scaler         Alpha         Gross         15:80         dpm/100cm²         0.1949         3.1           1/25/00         9:27:00         321         109         Scaler         Alpha         Gross         24:80         dpm/100cm²         0.1949         4.8           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         1.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1/25/00         11:33:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.8	0000374	1/25/00	11:16:00	321	109	Scaler	Alpha	Gross	15.80	dpm/100cm <sup>2</sup>	0.1949	3.1	8
1/25/00         9:27:00         321         109         Scaler         Alpha         Gross         24.80         dpm/100cm²         0.1949         4.8           1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         1.7           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²         0.1949         3.7           1/25/00         11:33:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²         0.1949         3.8	0000375	1/25/00	11:24:00	321	109	Scaler	Alpha	Gross	15.80	dpm/100cm <sup>2</sup>	0.1949	3.1	6
1/25/00         11:28:00         321         109         Scaler         Alpha         Gross         8.82         dpm/100cm²           1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19:10         dpm/100cm²           1/25/00         11:33:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²	0000376	1/25/00	9:27:00	321	109	Scaler	Alpha	Gross	24.80	dpm/100cm <sup>2</sup>	0.1949	4.8	6
1/25/00         11:31:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²           1/25/00         11:33:00         321         109         Scaler         Alpha         Gross         19:40         dpm/100cm²	0000377	1/25/00	11:28:00	321	109	Scaler	Alpha	Gross	8.82	dpm/100cm <sup>2</sup>	0.1949	1.7	8
1/25/00 11:33:00 321 109 Scaler Alpha Gross 119.40 dpm/100cm <sup>2</sup>	90000378	1/25/00	11:31:00	321	109	Scaler	Alpha		19.10	dpm/100cm <sup>2</sup>	0.1949	3.7	06
	0000379	1/25/00	11:33:00	321	109	Scaler	Alpha		19.40	dpm/100cm <sup>2</sup>	0.1949	3.8	6

Meán Of Replicate Measurements

Post Surface Media Sampling Direct Static Surface Measurements, Survey Unit 727-02

Instrument raw count rate	(mas)	6 8	6.5	5.0	3.2	5.7		,	3.7 0.0	3.1 2.9 4.0	L 2 4 L 0 4 8	E. C. 4. 1. 4. C. 4. 80 E.	2. 2. 4. 4. 4. 4. 4. 4. 4. 4. 4. 4. 4. 4. 4.	949 3.1 90 949 2.9 90 949 4.4 90 949 1.8 90 949 1.7 90
Instrum														m² 0.1949 m² 0.1949 m² 0.1949 m² 0.1949 m² 0.1949 m² 0.1949
	Units													dpm/100cm <sup>2</sup>
														15.70 14.70 22.60 9.03 22.00 8.80 15.80
Background Compensation	Mode	Gross	Gross	Gross	Gross	Gross		Gross	Gross Gross	Gross Gross Gross	Gross Gross Gross Gross	Gross Gross Gross Gross	Gross Gross Gross Gross Gross	Gross Gross Gross Gross Gross Gross
Channel	Selected	Alpha	Alpha	Alpha	Alpha	Alpha		Alpha	Alpha Alpha	Alpha Alpha Alpha	Alpha Alpha Alpha Alpha	Alpha Alpha Alpha Alpha	Alpha Alpha Alpha Alpha Alpha Alpha	Alpha Alpha Alpha Alpha Alpha Alpha
Instrument Operating	Mode	Scaler	Scaler	Scaler	Scaler	Scaler		Scaler	Scaler Scaler	Scaler Scaler Scaler	Scaler Scaler Scaler Scaler	Scaler Scaler Scaler Scaler Scaler	Scaler Scaler Scaler Scaler Scaler	Scaler Scaler Scaler Scaler Scaler Scaler Scaler
Probe	Serial #	109	109	109	109	109	100	103	109 109	109 109 109	109 109 109	00 100 100 100 100	109 109 109 109	109 109 109 109 109
E600	Serial #	321	321	. 321	321	321	224	1 70	321	321 321	321 321 321	321 321 321	321 321 321 321	321 321 321 321 321
i	Time	10:54:00	11:00:00	9:21:00	11:05:00	11:09:00	11.14.00	20.5	9:24:00	9:24:00	9:24:00 11:22:00 11:26:00	9:24:00 11:22:00 11:26:00 9:29:00	9:24:00 11:22:00 11:26:00 9:29:00 11:29:00	9:24:00 11:22:00 11:26:00 9:29:00 11:29:00
	Date	1/25/00	1/25/00	1/25/00	1/25/00	1/25/00	1/25/00	>	1/25/00	1/25/00	1/25/00 1/25/00 1/25/00	1/25/00 1/25/00 1/25/00 1/25/00	1/25/00 1/25/00 1/25/00 1/25/00 1/25/00	1/25/00 1/25/00 1/25/00 1/25/00 1/25/00
	Sample Location	IVP0000367	IVP0000368	IVP0000369	IVP0000370	IVP0000371	IVP0000372		IVP0000373	IVP0000373 IVP0000374	IVP0000373 IVP0000374 IVP0000375	IVP0000373 IVP0000374 IVP0000375 IVP0000376	IVP0000373 IVP0000374 IVP0000375 IVP0000376	IVP0000373 IVP0000374 IVP0000375 IVP0000377 IVP0000377

#### Surface Media Sample Data, 727-02

#### Alpha Isotopic Analysis

Sample	Sample ID #	Lab Sample ID #	Sample Weight	Date	Time	Units	,	Am-241		Pu-238	ď	Pu-239/240	U-234		U-235	U-238	Total Transuranic Activity	Total Uranlum
			(grams)	MMDD/YY			Reported Value	Reported Values  w/samples < MDA at 0.5 MDA	Reported Value	Reported Values w/samples < MDA at	Reported	Reported Values w/samples < MDA at	Reported	Reported	Reported Values w/samples < MDA at	Reported	dpm/100 cm²	dpm/100 cm²
IVP0000387	MED0000367	264788	17.16	1/24/00	12:30	dpm/100 cm <sup>2</sup>	2.02	2.02	0.63	0.89	2.43	ACM 6.0	,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,,	:	0.5 MDA			
IVP0000368	MED0000388	264789	19.23	1/24/00	12:40	dpm/100 cm <sup>2</sup> i	,	200				7	16.31	)*7	2.47	37.69	3.55	62.47
NPODOGO	+-	007790			1	400014	87	07.7	83.7	0.53	1.28	190	28.33	3.10	33.1	28.55	3.37	58.43
	-	N/LOS	\$. 5. 13. 14.	1/24/00	12.43	do and	5.0	0.53	0.46	0.23	7.	0.67	0.75	1.83	29.0	22.0	1.43	2.31
IVP0000370	MED0000370	264791	16.95	1/24/00	12:50	dpm/100 cm²	2.27	2.27	0.57	0.20	2.76	2.76	35.07	2.41	2.41	37.52	6.30	2 2
IVP0000371	MED0000371	264792	11.64	1/24/00	12:55	dpm/100 cm²	1.51	1.51	0.35	0.18	1.00	0.50	23.21	1.82	8	1		3.6
IVP0000372	MED0000372	264793	15.31	1/24/00	13:05	dpm/100 cm²	1.73	1.73	8	090	18	99 6	35.30			5		46.51
IVP0000373	MED0000373	264794	7.01	1/24/00	13:10	dpm/100 cm <sup>2</sup>	0.50	0.30	9.61	T d	35			3	7,1	37.88	3.02	74.87
IVP0000374	MED0000374	264795	15.42	1/24/00	13:15	dpm/100 cm <sup>2</sup>	9	1.83	9			37	Ser.	<b>X</b>	6.42	85.	0.87	1.28
IVP0000375	MED0000375	264796	13.25	1/24/00	+	dpm/100 cm <sup>2</sup>			3	66.0	P	0.85	35.06	3.30	3.30	32.69	3.15	71.25
					╈		7	78.0	270	270	3.07	3.07	18.61	1.60	0.05	18.41	8.21	36.07
INFORMS/6	MED00003/6	284787	5.01	1/24/00	13:25	dpmv100 cm	98.0	0.10	0.40	0.20	0.45	23	0.30	5.78	0.38	9	290	3
IVP0000377	MED0000377	264798	15.08	1/24/00	13:35	dpm/100 cm²	2.53	2.53	1.00	99'0	2.11	861	2,8	8.5	- C	5		3
IVP0000378	MED0000378	264799	18.60	1/24/00	13:40	dpm/100 cm²	1.25	1.25	8.1	0.83	2.46	124	10 04	25		70.19		96.56
IVP0000379	MED0000379	264800	15.20	1/24/00	13:45	dpm/100 cm²	1.87	1.87	0.61	16.0	8	0.63	33.85	2.55	1.28	2 82	3.42	101.96
																	10.7	8.8

#### Background Data, Survey Unit 782-01

instrument raw count	0.3	0.2	6.0	6:0	6.0	6.0	0.2	5.9	10	. E.	2.4	3.1
Instrument Efficiency	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949
Units	dpm/100cm <sup>2</sup>											
2 2	1.30							•		•	•	•
Background Compensation Mode	Gross											
Channel Selected	Alpha											
Instrument Operating Mode	Scaler											
Probe Serial #	109	109	109	109	109	109	109	109	109	109	109	109
E600 Serial #	321	321	321	321	321	321	321	321	321	321	321	321
Time	10:32:00	10:34:00	10:35:00	11:41:00	11:43:00	11:45:00	12:45:00	12:47:00	12:48:00	13:59:00	14:01:00	14:02:00
Date	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00
Sample Location	BACKGROUND											

Summary Statistics

12	7.4	5.1045666	4.8	5.7247649	0.7718784
Number of Measurements	Mean	Log Normal Mean	Median	Std. Deviation	S

## Response Data, Survey Unit 782-01

Sample Location	Date	Time	E600 Serial #	Probe Serial #	Instrument Operating Mode	Channel Selected	Background Compensation Mode	Recorded Value	Units
RESP/CHECK	2/3/00	10:37:00	321	109	Scaler	Alpha	Gross	1570	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	10:39:00	321	109	Scaler	Alpha	Gross	1560	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	10:41:00	321	109	Scaler	Alpha	Gross	1550	dpm/100cm²
RESP/CHECK	2/3/00	11:46:00	321	109	Scaler	Alpha	Gross	1390	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	11:48:00	321	109	Scaler	Alpha	Gross	1530	dpm/100cm²
RESP/CHECK	2/3/00	11:50:00	321	109	Scaler	Alpha	Gross	1490	dpm/100cm²
RESP/CHECK	2/3/00	12:50:00	321	109	Scaler	Alpha	Gross	1570	dpm/100cm²
RESP/CHECK	2/3/00	12:52:00	321	109	Scaler	Alpha	Gross	1520	dpm/100cm²
RESP/CHECK	2/3/00	12:54:00	321	109	Scaler	Alpha	Gross	1670	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	13:52:00	321	109	Scaler	Alpha	Gross	1440	dpm/100cm²
RESP/CHECK	2/3/00	13:54:00	321	109	Scaler	Alpha	Gross	1460	dpm/100cm <sup>2</sup>
RESP/CHECK	2/3/00	13:56:00	321	109	Scaler	Alpha	Gross	1450	dpm/100cm²

% 20%	)5 1957
Response -20%	1631 1305
Probe #	109

Direct Static Surface Contamination Measurements, Survey Unit 782-01

Count Time (seconds)	6	06	8	6	8	6	8	6	6	<b>6</b>	8	6	8	06	6	6	6	6	6	6	6	06	06	6	6	6	<u>6</u>	6	06
Instrument raw count rate (cpm)	6.7	6.1	2.7	4.7	3.4	2.4	4.7	4.7	2.7	1.4	2.1	2.1	1.4	2.1	3.4	2.1	0.7	0.2	0.3	1.6	6.0	2.1	2.2	1.5	6.0	3.6	2.9	1.6	0.3
Instrument Efficiency	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949
Units	dpm/100cm <sup>2</sup>																												
Recorded Value	34.50	31.10	13.80	24.10	17.50	12.25	24.20	24.10	14.10	7.36	10.90	10.95	7.08	10.80	17.50	10.90	3.73	1.20	1.34	8.01	4.66	11.00	11.50	7.79	4.68	18.30	14.90	8.20	1.34
Background Compensation Mode	Gross																												
Channel Selected	Alpha																												
Instrument Operating Mode	Scaler																												
Probe Serial #	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109
E600 Serial	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321
Time	10:55:00	10:59:00	11:02:00	11:06:00	11:09:00	11:12:00	11:17:00	11:21:00	11:25:00	11:28:00	11:32:00	11:36:00	12:56:00	12:59:00	13:03:00	13:06:00	13:14:00	13:16:00	13:20:00	13:21:00	13:23:00	13:26:00	13:28:00	13:32:00	13:39:00	13:40:00	13:42:00	13:44:00	13:47:00
Date	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00
Sample Location	IVP0000301	IVP0000302	IVP0000303	IVP0000304	IVP0000305	· VP00000306	IVP0000307	IVP0000308	IVP0000309	IVP0000310	IVP0000311	* - IVP0000312	IVP0000313	IVP0000314	IVP0000315	IVP0000316	IVP0000317	- IVP0000318	IVP0000319	IVP0000320	IVP0000321	IVP0000322	IVP0000323	*** IVP0000324	IVP0000325	IVP0000326	IVP0000327	IVP0000328	IVP0000329

Mean Of Replicate Measurements

Post Surface Media Sampling Direct Static Surface Measurements, Survey Unit 782-01

Count Time	o	Ō	0	9	o	0	o	Ō	0	0	0	0	0	0	Ō	0
<u> </u>	<b>O</b>	S	σ	<u>ග</u>	O	σ	σ	σ	σ	σ	Ø	σ.	σ	O	တ	တ
Instrument raw count rate (cpm)	5.4	4.7	4.0	8.0	2.7	1.4	4.7	8.2	4.1	3.4	<b>4</b> 1.1	4.2	5.4	2.8	2.7	1.5
Instrument Efficiency	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949
Units	dpm/100cm <sup>2</sup>															
Recorded Value																
Background Compensation Mode	Gross															
Channel Selected	Alpha															
Instrument Operating Mode	Scaler															
Probe Serial #	_	_	_	_	_	_	_	_	_	109	_				_	_
E600 Serial #	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321
Time	10:57:00	11:00:00	11:04:00	11:07:00	11:11:00	11:16:00	11:19:00	11:23:00	11:27:00	11:30:00	11:33:00	11:39:00	12:58:00	13:01:00	13:04:00	13:08:00
Date	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00	2/3/00
Sample Location	IVP0000301	IVP0000302	IVP0000303	IVP0000304	IVP0000305	IVP0000306	IVP0000307	IVP0000308	IVP0000309	IVP0000310	IVP0000311	IVP0000312	IVP0000313	IVP0000314	IVP0000315	IVP0000316

# Surface Media Sample Data, Survey Unit 782-01

### Alpha Isotopic Analysis

Sample Location	Sample ID #	Lab Sample ID #	Sample Weight	Date Collected	Time Collected	Units		Am-241	- <b>-</b>	Pu-238	ď	Pu-238/240	U-234		U-23\$	U-238	Total Transuranic Activity	Total Uranium Activity
			(grams)	MM/DD/YY			Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported	Reported Values w/samples < MDA at 0.5 MDA	Reported	Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported	dpm/100 cm²	dpm/100 cm²
VP0000301	MED0000301	264870	\$.07	1/26/00	7:45	dpm/100 cm²	1.20	1.20	0.35	0.18	0.35	07.0	30.61	1.92	1.92	31.96	1.57	64.49
VP0000302	MED0000302	264871	11,30	1/28/00	7:50	dpm/100 cm²	0.73	62'0	0,10	0.05	0.72	0.36	22.99	1.84	28.0	22.87	1.14	46.68
VP0000303	MED0000303	264872	11.12	1/28/00	7:55	dpm/100 cm²	0.74	0.37	0.28	0.13	0.55	0.28	25.33	1.43	1.43	27.72	0.78	\$4.48
VP0000304	MED0000304	264873	12.10	1/26/00	9:00	dpm/100 cm <sup>2</sup>	1.18	09:0	0.21	0.11	0.53	0.27	25.81	7.02	7.02	24.13	0.97	88.88
VP0000305	MED0000305	264874	10.47	1/26/00	8:05	dpm/100 cm²	0.66	0.33	0,18	0.06	0.26	0.13	18.06	0.02	0.46	17.25	9.54	35.79
VP0000306	MED0000308	264875	8.24	1/28/00	8:10	dpm/100 cm²	0.82	0.82	90'0	0.04	0.54	6.27	28.21	1.36	1.36	27.63	1.13	57.20
VP0000307	MED0000307	264876	13.70	1/28/00	8:15	dpm/100 cm²	2.38	2.38	90:0	0.04	0.47	0.24	26.33	0.55	0.28	27.30	2.66	53.91
VP0000308	MED0000308	284877	10.62	1/28/00	8:20	dpm/100 cm <sup>2</sup>	1.04	0.52	0.29	0.15	0.74	0.37	32.45	1.59	1.59	30.82	1.04	87.88
IVP0000309	MED0000309	264878	8.18	1/28/00	8:25	dpm/100 cm²	0.70	0.70	0.15	0.06	0.28	0.13	17.89	1.18	1.18	19.37	0.91	38.44
VP0000310	MED0000310	264879	15.16	1/28/00	8:30	dpm/100 cm²	1.81	1.8.1	0.33	0.17	1.41	0.71	10.69	0.15	90'0	8.14	2.68	18.91
VP0000311	MED0000311	264880	9.24	1/28/00	\$6:8	dpm/100 cm²	1.79	1.79	0.10	0.05	0.00	0.00	21.53	2.25	2.25	18.08	1,54	42.86
VP0000312	MED0000312	264881	<u>=</u>	1/28/00	04:8	dpm/100 cm²	0.76	96'0	0.00	0:00	0.52	0.16	24.39	1.62	18.0	27.08	0.54	52.28
IVP0000313	MED0000313	264882	11,11	1/28/00	S¥:8	dpm/100 cm²	1.20	1.20	0.21	0.11	0.57	6.29	36.84	1.36	0.50	37.13	1.50	74.96
IVP0000314	MED0000314	264883	10.10	1/28/00	8:50	dpm/100 cm²	1.55	1.55	0.10	90'0	0.35	0.18	18.28	1.28	1.29	18.78	1.78	38.35
IVP0000315	MED0000315	264884	9.16	1/28/00	9:55	dpm/100 cm²	0.72	0.72	0.05	0.03	0.84	0.42	32.07	1.36	1.36	32.36	1.17	62.79
IVP0000316	MED0000316	264885	9.19	1/28/00	00:6	dpm/100 cm²	1.89	1.89	0.36	0.18	0.64	0.32	21.61	1.67	1.67	20.27	2.39	43.55
																		i

## Background Data, Survey Unit 782-02

Instrument raw count Ffficiency rate (cpm)		0.1949 2.9	0.1949 1.5	0.1949 1.6	0.1949 4.3	0.1949 4.3	0.1949 2.8	0.1949 1.5	0.1949 5.5	0.1949 1.5	0.1949 1.6		0.1949 2.2	0.1949 1.6	0.1949 1.0								
Units	dpm/100cm <sup>2</sup>	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm²	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm <sup>2</sup>														
Recorded Value	4.62	14.70	7.95	8.42	22.20	22.10	14.30	7.67	28.10	7.87	8.14	4.61	11.50	8.32	4.94								
Background Compensation Mode	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross				•			•	
Channel Selected	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha		•						
Instrument Operating Mode	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler								
Probe Serial #	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109								
E600 Serial #	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321								
Tine	9:29:00	9:31:00	9:33:00	10:42:00	10:43:00	10:45:00	11:39:00	11:40:00	11:42:00	13:19:00	13:21:00	13:22:00	15:00:00	15:01:00	15:03:00		15	11.7	9.9725138	8.3	7.2165017	0.617006	
Date	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00		nents						
Sample Location	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	BACKGROUND	Summary Statistics	Number of Measurements	Mean	Log Normal Mean	Median	Std. Deviation	S	

# Response Data, Survey Unit 782-02

<b>Units</b>	dpm/100cm <sup>2</sup>	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm²	dpm/100cm²	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm²	dpm/100cm <sup>2</sup>	dpm/100cm <sup>2</sup>	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm <sup>2</sup>	dpm/100cm²	dpm/100cm <sup>2</sup>
Recorded Value	1500	1590	1450	1410	1450	1360	1500	1500	1550	1550	1680	1510	1400	1350	1410
Background Compensation Mode	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross
Channel Selected	Alpha	Alpha	Alpha	Alpha	Aipha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha
Instrument Operating Mode	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler
Probe Serial #	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109
E600 Serial#	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321
Time	9:21:00	9:23:00	9:24:00	10:34:00	10:35:00	10:39:00	11:33:00	11:35:00	11:36:00	13:24:00	13:28:00	13:30:00	15:05:00	15:07:00	15:09:00
Date	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00
Sample Location	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK	RESP/CHECK

_	
20%	1957
-20%	1305
Response	1631
Probe #	109

Direct Static Surface Contamination Measurements, Survey Unit 782-02

E600 Serial P			Instrument Operating Mode	Channel Selected	Background Compensation Mode	ď	Units	Instrument Efficiency	Instrument raw count rate (cpm)	Count Time (seconds)
9:37:00 321 109 Scaler 9:41:00 321 109 Scaler		Scaler		Alpha	Gross	10.40 24 10	apm/100cm² dnm/100cm²	0.1949	2.0 <b>4</b> 7	3 6
321 109	_	Scaler		Alpha	Gross	20.60	dpm/100cm <sup>2</sup>	0.1949	4.0	06
9:51:00 321 109 Scaler		Scaler		Alpha	Gross	6.94	dpm/100cm <sup>2</sup>	0.1949	1.4	8
9:57:00 321 109 Scaler		Scaler		Alpha	Gross	17.50	dpm/100cm <sup>2</sup>	0.1949	3.4	06
10:00:00 321 109 Scaler		Scaler		Alpha	Gross	14.00	dpm/100cm <sup>2</sup>	0.1949	2.7	06
10:08:00 321 109 Scaler		Scaler		Alpha	Gross	3.96	dpm/100cm <sup>2</sup>	0.1949	9.0	8
10:14:00 321 109 Scaler		Scaler		Alpha	Gross	7.26	dpm/100cm <sup>2</sup>	0.1949	1.4	06
10:16:00 321 109 Scaler		Scaler		Alpha	Gross	7.15	dpm/100cm <sup>2</sup>	0.1949	4.	8
10:19:00 321 109 Scaler		Scaler		Alpha	Gross	7.23	dpm/100cm <sup>2</sup>	0.1949	4.	06
10:23:00 321 109 Scaler	_	Scaler		Alpha	Gross	11.10	dpm/100cm <sup>2</sup>	0.1949	2.2	06
10:27:00 321 109 Scaler		Scaler		Alpha	Gross	9.45	dpm/100cm <sup>2</sup>	0.1949	1.8	6
10:48:00 321 109 Scaler		Scaler		Alpha	Gross	1.10	dpm/100cm <sup>2</sup>	0.1949	0.2	6
10:54:00 321 109 Scaler		Scaler		Alpha	Gross	7.71	dpm/100cm <sup>2</sup>	0.1949	1.5 5.	8
11:01:00 321 109 Scaler		Scaler		Alpha	Gross	21.50	dpm/100cm <sup>2</sup>	0.1949	4.2	06
11:14:00 321 109 Scaler		Scaler		Alpha	Gross	18.30	dpm/100cm <sup>2</sup>	0.1949	3.6	8
11:18:00 321 109 Scaler		Scaler		Alpha	Gross	15.00	dpm/100cm <sup>2</sup>	0.1949	2.9	06
11:22:00 321 109 Scaler		Scaler		Alpha	Gross	18,35	dpm/100cm <sup>2</sup>	0.1949	3.6	06
11:28:00 321 109 Scaler		Scaler		Alpha	Gross	12.00	dpm/100cm <sup>2</sup>	0.1949	2.3	06
13:38:00 321 109 Scaler		Scaler		Alpha	Gross	10.50	dpm/100cm <sup>2</sup>	0.1949	2.0	6
13:47:00 321 109 Scaler		Scaler		Alpha	Gross	3.55	dpm/100cm <sup>2</sup>	0.1949	7.0	8
13:56:00 321 109 Scaler		Scaler		Alpha	Gross	3.44	dpm/100cm <sup>2</sup>	0.1949	0.7	6
14:03:00 321 109 Scaler		Scaler		Alpha	Gross	17.40	dpm/100cm <sup>2</sup>	0.1949	3.4	<b>6</b>
14:10:00 321 109 Scaler		Scaler		Alpha	Gross	19,05	dpm/100cm <sup>2</sup>	0.1949	3.7	<b>6</b>
14:35:00 321 109 Scaler		Scaler		Alpha	Gross	21.00	dpm/100cm <sup>2</sup>	0.1949	4.1	6
14:40:00 321 109 Scaler		Scaler		Alpha	Gross	10.60	dpm/100cm <sup>2</sup>	0.1949	2.1	6
14:47:00 321 109 Scaler		Scaler		Alpha	Gross	17.30	dpm/100cm <sup>2</sup>	0.1949	3.4	06
14:53:00 321 109 Scaler		Scaler		Alpha	Gross	4.16	dpm/100cm <sup>2</sup>	0.1949	0.8	06
14:56:00 321 109 Scaler		Scaler		Alpha	Gross	11.20	dpm/100cm <sup>2</sup>	0.1949	2.2	06

Post Surface Media Sampling Direct Static Surface Measurements, Survey Unit 782-02

Count Time (seconds)	6	8	06	8	06	06	06	06	06	06	06	8	06	8	06	06	06	06	6	06	06	06	06
instrument raw count rate (cpm)	3.4	0.7	5.3	0.9	3.4	1.1	3.4	3.5	1.5	2.9	3.5	3.5	3.5	1.6	6.0	0.7	4.1	2.1	2.6	6.0	4.	4.	0.7
instrument Efficiency	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949	0.1949
Chits	dpm/100cm <sup>2</sup>		dpm/100cm <sup>2</sup>																				
Recorded Value	17.30	3.50	27.30	31.00	17.30	5,53	17.20	17.80	7.63	14.70	17.90	18.15	18.20	8.09	4.79	3.55	20.80	10.60	13.40	4.67	7.13	7.23	3.83
Background Compensation Mode	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross	Gross
Channel Selected	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha	Aipha	Alpha	Alpha	Alpha	Alpha	Alpha	Alpha
Instrument Operating Mode	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler	Scaler
Probe Serial #	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109	109
E600 Serial #	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321	321
	9:40:00	9:43:00	9:48:00	9:53:00	10:03:00	10:10:00	10:21:00	10:25:00	10:31:00	10:49:00	10:58:00	11:05:00	11:16:00	11:20:00	11:25:00	13:41:00	13:50:00	13:59:00	14:06:00	14:29:00	14:37:00	14:43:00	14:50:00
Date	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00	2/2/00
Sample Location	IVP0000421	IVP0000422	IVP0000423	IVP0000424	IVP0000426	1/VP0000427	IVP0000430	IVP0000431	IVP0000432	IVP0000433	IVP0000434	IVP0000435	IVP0000436	IVP0000437	IVP0000438	IVP0000440	IVP0000441	IVP0000442	IVP0000443	IVP0000444	IVP0000445	IVP0000446	IVP0000447

# Surface Media Sample Data, Survey Unit 782-02

### Alpha Isotopic Analysis

											•							
Sample Location	Sample ID #	Lab Sample ID #	Sample Weight	Date Collected	Time Collected	Units		Am-241		Pu-238	ã	Pu-239/240	U-234		U-235	U-238	Total Transuranic Activity	Total Uranium Activity
			(grams)	MWDDYY			Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported Value	Reported Value	Reported Values w/samples < MDA at 0.5 MDA	Reported Value	dpm/100 cm²	dpm/100 cm²
IVP0000421	MED0000421	264688	10.49	01/27/00	12:45	dpm/100 cm²	0.83	6.47	0.10	6.05	0.15	90.06	19.75	1.45	1,45	12.24	0.59	43.44
IVP0000422	MED0000422	264669	7.87	01/27/00	12:55	dpm/100 cm <sup>2</sup>	06.0	0.45	90'0	0.04	0.44	0.24	13.79	1.18	1.18	14.73	67.0	29.70
IVP0000423	MED0000423	264890	10.61	01/27/00	13:00	dpm/100 cm²	1.09	1.09	0.14	0.07	0.92	97'0	15.48	1.37	89'0	15.05	1.62	31.22
IVP0000424	MED0000424	264891	4.52	01/27/00	13:10	dpm/100 cm²	1.33	1.33	0.07	<b>20.0</b> 4	2.15	2.15	6.74	0.81	14.0	3.41	3.52	10.56
IVP0000426	MED0000428	264892	2.82	01/27/00	13:20	dpm/100 cm²	0.68	0.88	0.00	0:00	0.72	0.36	2.64	0.30	0.20	2.45	1.24	5.29
IVP0000427	MED0000427	264893	3.82	01/27/00	13:25	dpm/100 cm²	1.08	1.08	0.12	90'0	1.30	1.30	2.31	0.12	0.12	2.23	2.44	4.86
IVP0000430	MED0000430	264894	4.49	01/27/00	13:30	dpm/100 cm²	1.88	1.88	20:0	10'0	1.03	0.52	5.24	0.07	10:01	4.15	2.43	9.43
IVP0000431	MED0000431	264895	6.62	01/27/00	13:35	dpm/100 cm²	1.75	1.75	0.08	70'0	3.00	3.00	7.10	60.0	90'0	7.40	4.79	14.55
IVP0000432	MED0000432	264896	8.24	01/27/00	13:45	dpm/100 cm²	18.0	170	0.08	80'0	0.50	625	15.36	2.06	2.06	14,61	0.69	32.03
IVP0000433	MED0000433	264897	9.21	01/27/00	13:50	dpm/100 cm <sup>2</sup>	1.23	1.23	0.17	0.00	0.04	50.05	15.59	99'0	16°0	19.37	1.34	35.30
1VP0000434	MED0000434	264898	8.15	01/27/00	13:55	dpm/100 cm²	0.48	0.24	0.03	0.02	0.72	0.36	16.33	0.92	99'0	14.82	0.62	31.61
IVP0000435	MED0000435	264899	5.99	01/27/00	14:05	dpm/100 cm <sup>2</sup>	0.05	60.03	90.0	60.03	0.17	0.00	15.43	0.88	0.88	15.49	0.14	31.80
IVP0000438	MED0000436	264900	11.24	01/27/00	14:10	dpm/100 cm <sup>2</sup>	1.36	1.36	0.30	0.15	. D.03	0.02	40.22	1.79	0.90	38.10	1.53	79.22
IVP0000437	MED0000437	264901	11.01	01/27/00	14:15	dpm/100 cm <sup>2</sup>	0.59	0:30	6.19	0.10	0.83	0.83	19.45	124	29:0	21.32	1.22	41.39
IVP0000438	MED0000438	264902	7.97	01/27/00	14:20	dpm/100 ст <sup>2</sup>	0.65	0.33	60.0	90'0	0.03	20:0	25.98	1.89	0.95	29.11	0.39	56.02
IVP0000440	MED0000440	264903	8.63	01/27/00	14:35	dpm/100 cm²	0.69	0.69	0.19	0.10	0.00	0.04	10.17	0.41	0.21	10.25	0.83	20.63
IVP0000441	MED0000441	264904	6.75	01/27/00	14:45	dpm/100 cm²	0.33	0.33	6.02	0.01	0.06	p.03	9.03	1.08	1.08	9.55	0.37	19.61
IVP0000442	MED0000442	264905	5.78	01/27/00	14;55	dpm/100 cm <sup>2</sup>	0.43	0.43	17'0	0.22	0.37	0.19	6.36	0:30	6.15	6.09	0.64	12.80
IVP0000443	MED0000443	264906	8.97	01/27/00	15:05	dpm/100 cm²	0.70	07:0	20:0	0.01	0.32	0.16	11.75	0.80	0.40	11.98	0.57	24.13
IVP0000444	MED0000444	264907	9.70	01/27/00	15:10	dpm/100 cm <sup>2</sup>	1.31	1.31	0.11	0.06	0.30	0.20	10.78	0.38	6.19	11.93	1.58	22.88
IVP0000445	MED0000445	264906	20.0	01/27/00	15:15	dpm/100 cm²	1.02	1.02	90.0	90.0	0.08	97.04	10.01	0.65	0.33	8.29	1.10	18.63
IVP0000448	MED0000448	264809	7.08	01/27/00	15:20	dpm/100 cm²	0.57	0.29	0.02	0.01	90.0	20'0	7.61	0.87	77.0	20.6	0.32	16.69
IVP0000447	MED0000447	264910	10.25	01/27/00	15:25	dpm/100 cm²	1.22	1.22	0.21	0.11	0.21	0,11	13.00	0.34	71.0	12.93	1,43	26.19

### Cross Reference Table for Blank and Spiked Samples for Survey Units 727–01, 727–02, 782–01, and 782–02

Sample Location	Sample Ticket Number (IVC)	Date Transferred	Smear Number
IVP0000261	259767	01/25/00	SMR0000261
IVP0000262	259760	01/25/00	SMR0000262
IVP0000263	259759	01/25/00	SMR0000263
IVP0000264	259730	01/25/00	SMR0000264
IVP0000265	259766	01/25/00	SMR0000265
IVP0000266	259762	01/25/00	SMR0000266
IVP0000451	259731	02/01/00	SMR0000451
IVP0000452	259732	02/01/00	SMR0000452
IVP0000453	259739	02/01/00	SMR0000453
IVP0000454	259735	02/01/00	SMR0000454
IVP0000455	259736	02/01/00	SMR0000455
IVP0000456	259737	02/01/00	SMR0000456

### Data Set Sheet

SHP100			er. \$15564		er. <u>109</u>	·	•
Eberline	E600 Pro	perty Numb	per \$15622	Serial Numb	er <u>3</u> 21		
2.2 Calib	oration Sou	rce					
	tope/	Source Serial No.	Certified Activity (dpm)	Observed Activity (dpm)	% Difference	Efficiency cpm/dpm	SOURCE GEOMETRY
	SR-90	CSL = 6041		N/A.	NA	.3838	37 mm dis
Th-230	Pu-239	GM-785	1604 dpm/100cm	1684dpm/	4.99%	-1949	150 cm 2 Pl
			ground count rate (	. ,	2.4 cpm 43 cpm		•
•	of the abov	e table.		the cell marked "4.1		hand colun	nn .
4.19 4	Decrease in Seometre, Complete	y-Deres	ency after 4 hours FOR CENTERED	NA % SEE N OVER SOURCE,	OTE *1 Reobe Fac	<i>£</i> @≈3	116"
	-			ed <u>/</u> failed e <i>e let</i> s, /		n/<	
			TION FOR US	e e ereis, T	VC 30EVE		. ·
6.	Calibrat		cer and Interv	al			
New Sti	cker Attach	ned	Due Date	130 k000			
	ed using sta			ISI 323-1978 and 10 eable to the Nationa			d
Signatu		Sunt	LIVEY		6/30/99		_
DEERA	DE MOX	use. E	ZON WELLENCE	THE FIELD O, MEMED FOR TH SHOWS THAT PI THE Z HOUR P	IRE EEG	CIRICI 1	W) DOES
PLATEA HAILAB L 1%	CRAPH LE IN TH B -> &	WAS NO E FIELD. CROSSO	T PRINTED OUT THE SCIECTE SER.	T FOR THE REC DD HIGH VOLTAGE	roed since E (1572 v	NO PRI (dc) Resu	NTER WA UTEO IN
	dibration P		tional Gas Detector	r	Calibr		STREN, Inc. am Standard

### Page 1

### EBERLINE E-600 CALIBRATION REPORT

06/30/99 10:30:36 : 321 E-600 Serial Number : E600 V3.12 Program Version : 03/23/99 to 03/23/00 Calibration Date/Due Date : 10% Scaler Precision Lower Threshold Slope : 0.9524 Lower Threshold Intercept : -0.1429 mV Upper Threshold Slope : 1.02 Upper Threshold Intercept
Alarm Editing : -0.6074 mV : Enabled : Enabled Latching Alarms : Disabled Auto Ranging Beep on Auto-Range : No Ignore E-600 Cal. Date : No Ignore Probe Cal. Date : No Ratemeter Mode Support : Enabled Integrate Mode Support : Enabled Scaler Mode Support : Enabled
Peak Hold Mode Support : Enabled
Rackersund \*\*\*\* Background Update Mode Support : Enabled Star Key Ratemeter Function : Zero Display Star Key Integrate Function : Zero Display Scaler Display Units : Rate Scaler Counting Mode Smart Probe Serial Number : 109 : HP-100 Type : 06/30/99 to 06/30/00 Calibration Date/Due Date Dead Time : 7.50 usec Surface Area : 100 cm2

: 1900 Vdc

: 80000 cps

Max High Voltage

Overrange

### Page 2

Probe HP-100 109 continued Channel 1	
Channel Type	: Alpha
Rate Units	: dpm/100cm2
Response Times	: 22,10,3 secs
High Voltage	: 1572 Vdc
Lower Threshold	: 1.00 mV
Upper Threshold	: 27.1 mV
Selected Window	
Upper Cal. Constant	<pre>: Upper : 0.1949 counts/disint.</pre>
Scaler Time	: 90 secs
Lower to Upper Crossover	: 90 Secs : 0.0012
Upper to Lower Crossover	: 0.0838
Channel 2	: 0.0636
Channel Type	: Beta
Rate Units	: dpm/100cm2
Response Times	: 22,10,3 secs
High Voltage	: 1572 Vdc
Lower Threshold	: 1.00 mV
Upper Threshold	: 27.1 mV
Selected Window	: Lower
Lower Cal. Constant	: 0.3838 counts/disint.
Scaler Time	: 90 secs
Lower to Upper Crossover	: 0.0012
Upper to Lower Crossover	: 0.0838
Channel 3	
Channel Type	: Alpha/Beta
Rate Units	: dpm/100cm2
Response Times	: 22,10,3 secs
High Voltage	: 1572 Vdc
Lower Threshold	: 1.00 mV
Upper Threshold	: 27.1 mV
Selected Window	: Both
Lower Cal. Constant	: 0.3838 counts/disint.
Upper Cal. Constant	: 0.1949 counts/disint.
Scaler Time	: 90 secs
Lower to Upper Crossover	: 0.0012
Upper to Lower Crossover	: 0.0838
Cable Length: O inches	1/2/00
Signature: Jak W Mut A	(1/ery Date: 6/30/99

After-Calibration Source Response Check Data Sheet
--

Location F	Park Ela	L IVC	Detector/I	Probe Data <i>(if</i>	applicable)		
Month Ava	Day 4	/ <sub>3</sub> /U C Year /999	Manufactu	rer EBE	RLING		
/ / / / / / / / / / / / / / / / / / /	Day	1001	Model No.	Z-600	/w SHP	1-100#109	
				ent Property No			
Survey Instrum	ent Data		Calibration	Due Date _	6-30	0-00	
Manufacturer _	EBERLIN	<u> </u>	Odiibratioi	Due Date _			
Model No	E-600		Check So	urce Data			
Government Pro	operty No. 5	5622	Isotope <u>Pu - 239</u> Source I.D. No. <u>Gm - 785</u>				
	Date 3-2:						
Instrument Scale	Source Detector Distance	Shielding/Geometry	Instrument Response	-20%	+20%	Scale Units	
NA	2 1/8"	CONTACT W/ RUBBUTA FEET	1631	1305	1957	dpm/100 cm²	
· · · · · · · · · · · · · · · · · · ·			· · · · · · · · · · · · · · · · · · ·				
		·					
Comments:	·						
:							
<del>, , , , , , , , , , , , , , , , , , , </del>						<del></del>	
A. SAA	MILJAN		1 Same	_		8/4/99	
	Performed by (print)		Performed b	y (signature)		Date	
J. L.	IVEY		WA	7		8/6/99	
•	Reviewed by (print)		Reviewed L	(signature)		Date	
GJO 1974e 11/98					File Ind	lex No.	

This appendix contains the raw ASCI text file data download directly from the E-600 instrument's memories without modification, sorting, or data reduction of any kind. The data are actually contained in three separate ASCI files, one file corresponding to each date in which information was collected. One unique aspect of this data presentation is that the data is presented exactly in the chronological order in which it was collected in the field. This provides an electronic time stamp permitting verification that time criteria included in the field operating procedures associated with the IV SAP were met.

"Normal" Time", "Probe S/N", "Log Mode", "Channel Type", "Reading", "Gross/Net", "Units", "E-600 "RESP/CHECK", "01/25/00", "08:58:00", "109", "Scaler", "Alpha", 1.45E+03, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "RESP/CHECK", "01/25/00", "08:59:00", "109", "Scaler", "Alpha", 1.45E+03, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "RESP/CHECK", "01/25/00", "09:01:00", "109", "Scaler", "Alpha", 1.49E+03, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000376", "01/25/00", "09:27:00", "109", "Scaler", "Alpha", 2.48E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000376", "01/25/00", "09:29:00", "109", "Scaler", "Alpha", 2.20E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000351", "01/25/00", "09:59:00", "109", "Scaler", "Alpha", 1.23E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000353", "01/25/00", "10:02:00", "109", "Scaler", "Alpha", 2.15E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "IVP0000354", "01/25/00", "10:04:00", "109", "Scaler", "Alpha", 2.01E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "IVP0000355", "01/25/00", "10:06:00", "109", "Scaler", "Alpha", 1.57E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "IVP0000356", "01/25/00", "10:08:00", "109", "Scaler", "Alpha", 1.58E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "IVP0000356","01/25/00","10:14:00","109","Scaler","Alpha",1.87E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2",
"RESP/CHECK","01/25/00","10:16:00","109","Scaler","Alpha",1.61E+03,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2",
"RESP/CHECK","01/25/00","10:18:00","109","Scaler","Alpha",1.52E+03,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2",
"RESP/CHECK","01/25/00","10:20:00","109","Scaler","Alpha",1.57E+03,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "BACKGROUND", "01/25/00", "09:06:00", "109", "Scaler", "Alpha", 1.51E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "BACKGROUND", "01/25/00", "10:23:00", "109", "Scaler", "Alpha", 1.56E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "BACKGROUND", "01/25/00", "10:25:00", "109", "Scaler", "Alpha", 2.15E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000357", "01/25/00", "10:27:00", "109", "Scaler", "Alpha", 2.19E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000358", "01/25/00", "10:29:00", "109", "Scaler", "Alpha", 2.05E+01, "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "IVP0000358", "dpm/100cm2", "321,0,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "321,0.0, "3 "IVP0000360","01/25/00","10:34:00","109","Scaler","Alpha",2.06E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "IVP0000361","01/25/00","10:37:00","109","Scaler","Alpha",1.24E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "IVP0000362","01/25/00","10:39:00","109","Scaler","Alpha",2.20E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", ,"01/25/00","09:23:00","109","Scaler","Alpha",2.47E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "01/25/00", "09:24:00", "109", "Scaler", "Alpha", 1.47E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", ,"01/25/00","09:08:00","109","Scaler","Alpha",8.45E+00,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2" "01/25/00", "10:00:00", "109", "Scaler", "Alpha", 8.82E+00, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "BACKGROUND", "01/25/00", "10:21:00", "109", "Scaler", "Alpha", 1.49E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "Scaler", "Alpha", 2.27E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "01/25/00","10:43:00","109","Scaler","Alpha",1.94E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2" "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "IVP0000365", "01/25/00", "10:46:00", "109", "Scaler", "Alpha", 2.99E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "IVP0000368","01/25/00","10:56:00","109","Scaler","Alpha",1.54E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "IVP0000368","01/25/00","10:58:00","109","Scaler","Alpha",1.40E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "01/25/00", "10:48:00", "109", "Scaler", "Alpha", 1.91E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "Gross"," "Gross", "Scaler", "Alpha", 1.87E+01, "01/25/00", "09:19:00", "109", "Scaler", "Alpha", 2.49E+01, "IVP0000369", "01/25/00", "09:21:00", "109", "Scaler", "Alpha", 2.55E+01, "IVP0000359", "01/25/00", "10:32:00", "109", "Scaler", "Alpha", 2.11E+01, "Scaler", "Alpha", 2.20E+01, "Scaler", "Alpha", 1.34E+01, "Scaler", "Alpha", 2.00E+01, S/N", "E-600 Address", "Stored Bkg", "Bkg Units", "Status" "01/25/00","10:40:00","109"," ,"09:09:00:0" "01/25/00","10:44:00","109", "109", "01/25/00","10:54:00","109", "01/25/00","10:52:00", "Survey Location", "Log Date", "Log "01/25/00", "BACKGROUND", "IVP0000369", "IVP0000362"

"Normal" "Normal" "Normal" "Normal" "BACKGROUND", "01/25/00", "11:45:00", "109", "Scaler", "Alpha", 1.51E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "Normal" "Normal" "Normal "Normal "109", "Scaler", "Alpha", 1.58E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "dpm/100cm2", "dpm/100cm2", "dpm/100cm2", "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "dpm/100cm2", "dpm/100cm2", "dpm/100cm2", "Gross", "dpm/100cm2", 321,0,0.0, "dpm/100cm2", "RESP/CHECK","01/25/00","11:40:00","109","Scaler","Alpha",1.44E+03,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "dpm/100cm2" "IVP0000375","01/25/00","11:24:00","109","Scaler","Alpha",1.58E+01,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", "IVP0000375","01/25/00","11:26:00","109","Scaler","Alpha",9.03E+00,"Gross","dpm/100cm2",321,0,0.0,"dpm/100cm2", ,321,0,0.0,"dpm/100cm2", "IVP0000379", "01/25/00", "11:33:00", "109", "Scaler", "Alpha", 1.94E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "dpm/100cm2" "dpm/100cm2" "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "Alpha", 3.63E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2" "Gross", "dpm/100cm2", 321, 0, 0.0, "Alpha", 8.80E+00, "Gross", "dpm/100cm2", 321, 0, 0.0, "Gross", "dpm/100cm2", 321, 0, 0.0, ,321,0,0.0, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", 321, 0, 0.0, "Scaler", "Alpha", 1.03E+01, "Gross", "dpm/100cm2", 321, 0, 0.0, "dpm/100cm2", 321, 0, 0.0, "Alpha", 8.82E+00, "Gross", "dpm/100cm2", "dpm/100cm2", "Gross", "dpm/100cm2", "IVP0000368", "01/25/00", "11:00:00", "109", "Scaler", "Alpha", 1.51E+01, "Gross", "Gross", "Scaler", "Alpha", 1.53E+03, "Gross", ,"01/25/00","11:05:00","109","Scaler","Alpha",1.63E+01, "Alpha", 2.95E+01, "Alpha", 1.58E+01, "Alpha", 1.94E+01, "Alpha", 1.23E+01, "Alpha", 2.94E+01, "109", "Scaler", "Alpha", 1.57E+01, "Alpha", 2.26E+01, "Alpha", 1.91E+01, "Alpha", 1.58E+01, "Alpha", 1.23E+01, "Scaler", "Alpha", 1.50E+03, "01/25/00", "11:04:00", "109", "Scaler", "Alpha", 1.89E+01, "109", "Scaler", "Scaler", "Scaler", "Scaler", "01/25/00", "11:07:00", "109", "Scaler", "Scaler", "Scaler", "Scaler", "Scaler", "Scaler", "Scaler", "Scaler", "109", "01/25/00","11:12:00","109", "01/25/00","11:18:00","109", ,1001, ,"109", "109", "01/25/00","11:39:00","109", "109", "109", "01/25/00","11:38:00","109", "01/25/00","11:44:00","109", "109" "01/25/00","11:09:00", "01/25/00", "11:14:00", "01/25/00", "11:16:00", "01/25/00", "11:28:00", "01/25/00","11:29:00", "11:32:00", ,"11:35:00", "01/25/00", "11:31:00", ,"11:42:00", "01/25/00", "11:37:00", "01/25/00","11:22:00", "01/25/00", "01/25/00", "01/25/00", "IVP0000370", "BACKGROUND", "IVP0000378", "IVP0000378", "IVP0000379", "RESP/CHECK", "IVP0000374", "IVP0000374", "IVP0000374" "IVP0000377", "IVP0000370" "IVP0000371" "IVP0000372" "IVP0000372" "IVP0000377" "IVP0000379" "RESP/CHECK" "BACKGROUND"

"IVP0000430","02/02/00","10:19:00","109","Scaler","Alpha",7.23E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000430","02/02/00","10:21:00","109","Scaler","Alpha",1.72E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000431","02/02/00","10:23:00","109","Scaler","Alpha",1.78E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000431","02/02/00","10:25:00","109","Scaler","Alpha",1.78E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "IVP0000425","02/02/00","09:57:00","109","Scaler","Alpha",1.75E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000426","02/02/00","10:03:00","109","Scaler","Alpha",1.73E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000426","02/02/00","10:03:00","109","Scaler","Alpha",1.73E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000426","02/02/00","10:04:00","109","Scaler","Alpha",1.40E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "IVP0000427","02/02","10:08:00","109","Scaler","Alpha",3.96E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000427","02/02/00","10:10:00","109","Scaler","Alpha",3.80E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000427","02/02/00","10:11:00","109","Scaler","Alpha",7.26E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "RESP/CHECK", "02/02/00", "09:21:00", "109", "Scaler", "Alpha", 1.50E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "02/02/00", "09:23:00", "109", "Scaler", "Alpha", 1.59E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "02/02/00", "09:43:00", "109", "Scaler", "Alpha", 3.50E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "IVP0000428","02/02/00","10:14:00","109","Scaler","Alpha",7.26E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "IVP0000429","02/02/00","10:16:00","109","Scaler","Alpha",7.15E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" ,"02/02/00","10:27:00","109","Scaler","Alpha",1.12E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "Normal" "RESP/CHECK","02/02/00","10:34:00","109","Scaler","Alpha",1.41E+03,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal""RESP/CHECK","02/02/00","10:35:00","109","Scaler","Alpha",1.45E+03,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" Time", "Probe S/N", "Log Mode", "Channel Type", "Reading", "Gross/Net", "Units", "E-600 "RESP/CHECK", "02/02/00", "09:24:00", "109", "Scaler", "Alpha", 1.45E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "BACKGROUND", "02/02/00", "09:29:00", "109", "Scaler", "Alpha", 4.62E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "BACKGROUND", "02/02/00", "09:31:00", "109", "Scaler", "Alpha", 1.47E+01, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "IVP0000424", "02/02/00", "09:53:00", "109", "Scaler", "Alpha", 3.10E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "09:33:00", "109", "Scaler", "Alpha", 7.95E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "09:37:00", "109", "Scaler", "Alpha", 1.04E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "dpm/100cm2", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "dpm/100cm2", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "dpm/100cm2", "IVP0000434", "02/02/00", "10:54:00", "109", "Scaler", "Alpha", 7.71E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "IVP0000434", "02/02/00", "10:58:00", "109", "Scaler", "Alpha", 1.79E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", ,"02/02/00","10:29:00","109","Scaler","Alpha",7.69E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2", "RESP/CHECK", "02/02/00", "10:39:00", "109", "Scaler", "Alpha", 1.36E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "dpm/100cm2", "BACKGROUND", "02/02/00", "10:45:00", "109", "Scaler", "Alpha", 2.21E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "IVP0000433","02/02/00","10:48:00","109","Scaler","Alpha",1.10E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2", ',321,0,52.2,"dpm/100cm2", "02/02/00", "09:41:00", "109", "Scaler", "Alpha", 2.41E+01, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", 321, 0, 52.2, ,"02/02/00","10:42:00","109","Scaler","Alpha",8.42E+00,"Gross","dpm/100cm2",321,0,52.2,","02/02/00","10:43:00","109","Scaler","Alpha",2.22E+01,"Gross","dpm/100cm2",321,0,52.2," "Scaler", "Alpha", 1.47E+01, "Gross", "dpm/100cm2", "02/02/00", "09:45:00", "109", "Scaler", "Alpha", 2.06E+01, "Gross", "02/02/00", "09:48:00", "109", "Scaler", "Alpha", 2.73E+01, "Gross", " "Gross", "09:40:00", "109", "Scaler", "Alpha", 1.73E+01, "IVP0000424", "02/02/00", "09:51:00", "109", "Scaler", "Alpha", 6.94E+00, "Survey Location","Log Date","Log Time","Probe S/N","Log S/N","E-600 Address","Stored Bkg","Bkg Units","Status" ,"10:49:00","109", "02/02/00"," "02/02/00", "02/02/00", "IVP0000422", "BACKGROUND", "IVP0000423", "IVP0000421", "IVP0000423", "IVP0000432", "IVP0000432", "BACKGROUND" "IVP0000432" "BACKGROUND"

"BACKGROUND", "02/02/00", "16:54:00", "109", "Scaler", "Alpha", 8.32E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal"
"BACKGROUND", "02/02/00", "16:56:00", "109", "Scaler", "Alpha", 4.76E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal"
"RESP/CHECK", "02/02/00", "16:59:00", "109", "Scaler", "Alpha", 1.57E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal"
"RESP/CHECK", "02/02/00", "16:59:00", "109", "Scaler", "Alpha", 1.42E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal"
"RESP/CHECK", "02/02/00", "17:01:00", "Scaler", "Alpha", 1.42E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal" "BACKGROUND", "02/02/00", "16:52:00", "109", "Scaler", "Alpha", 8.24E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal"

"IVP0000388","02/03/00","09:23:00","109","Scaler","Alpha",2.08E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000388","02/03/00","09:26:00","109","Scaler","Alpha",2.03E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000389","02/03/00","09:28:00","109","Scaler","Alpha",1.84E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000399","02/03/00","09:32:00","109","Scaler","Alpha",1.76E+01,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000390","02/03/00","09:34:00","109","Scaler","Alpha",1.90E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000392","02/03/00","09:37:00","109","Scaler","Alpha",8.94E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000392","02/03/00","09:39:00","109","Scaler","Alpha",8.96E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000392","02/03/00","09:39:00","109","Scaler","Alpha",8.96E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000392","02/03/00","09:39:00","109","Scaler","Alpha",8.87E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "BACKGROUND", "02/03/00", "08:49:00", "109", "Scaler", "Alpha", 4.42E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "08:51:00", "109", "Scaler", "Alpha", 1.53E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "08:52:00", "109", "Scaler", "Alpha", 1.53E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "08:54:00", "109", "Scaler", "Alpha", 1.69E+03, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal" "IVP0000394","02/03/00","09:42:00","109","Scaler","Alpha",2.02E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"IVP0000395","02/03/00","09:46:00","109","Scaler","Alpha",1.91E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"BACKGROUND","02/03/00","09:50:00","109","Scaler","Alpha",4.73E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal"
"BACKGROUND","02/03/00","09:52:00","109","Scaler","Alpha",4.99E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "02/03/00", "09:02:00", "109", "Scaler", "Alpha", 8.94E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Normal" ,"02/03/00","09:14:00","109","Scaler","Alpha",1.85E+00,"Gross","dpm/100cm2",321,0,52.2,"dpm/100cm2","Normal" "IVP0000386", "02/03/00", "09:16:00", "109", "Scaler", "Alpha", 1.14E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal"
"IVP0000386", "02/03/00", "09:19:00", "109", "Scaler", "Alpha", 1.74E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal"
"IVP0000386", "02/03/00", "09:19:00", "109", "Scaler", "Alpha", 7.40E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal"
"IVP0000387", "02/03/00", "09:21:00", "109", "Scaler", "Alpha", 2.09E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "BACKGROUND", "02/03/00", "09:54:00", "109", "Scaler", "Alpha", 7.56E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "09:58:00", "109", "Scaler", "Alpha", 1.50E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "10:00:00", "109", "Scaler", "Alpha", 1.44E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "10:07:00", "109", "Scaler", "Alpha", 1.47E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "BACKGROUND", "02/03/00", "10:32:00", "109", "Scaler", "Alpha", 1.30E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "02/03/00", "10:34:00", "109", "Scaler", "Alpha", 1.13E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "IVP0000301", "02/03/00", "10:55:00", "109", "Scaler", "Alpha", 3.45E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "IVP0000301", "02/03/00", "10:57:00", "109", "Scaler", "Alpha", 2.76E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" "RESP/CHECK", "02/03/00", "10:41:00", "109", "Scaler", "Alpha", 1.55E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Normal" Time", "Probe S/N", "Log Mode", "Channel Type", "Reading", "Gross/Net", "Units", "E-600 "BACKGROUND", "02/03/00", "08:45:00", "109", "Scaler", "Alpha", 4.76E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "BACKGROUND", "02/03/00", "08:47:00", "109", "Scaler", "Alpha", 4.46E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "IVP0000384", "02/03/00", "09:12:00", "109", "Scaler", "Alpha", 5.05E+00, "Gross", "dpm/100cm2", 321,0,52.2, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "BACKGROUND", "02/03/00", "10:35:00", "109", "Scaler", "Alpha", 4.80E+00, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2", "dpm/100cm2", "RESP/CHECK", "02/03/00", "10:39:00", "109", "Scaler", "Alpha", 1.56E+03, "Gross", "dpm/100cm2", 321, 0, 52.2, " "02/03/00","10:37:00","109","Scaler","Alpha",1.57E+03,"Gross", "02/03/00", "09:06:00", "109", "Scaler", "Alpha", 2.04E+00, " "02/03/00", "09:10:00", "109", "Scaler", "Alpha", 5.46E+00, S/N", "E-600 Address", "Stored Bkg", "Bkg Units", "Status" "Survey Location", "Log Date", "Log "IVP0000381", "IVP0000383", "IVP0000382", "IVP0000385"

Scaler", "Alpha", 3.11E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.07E+011, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.07E+011, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.07E+011, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.3E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.3E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.3E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.3E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.0E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.0E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.0E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 2.0E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "dpm/100cm2", 331,0, 52.2, "dpm/100cm2", "Normal" 'Scaler", "Alpha", 1.1E+01, "Gross", "d	Scaler", "Alpha", 1.41E+01, "Gross", "dpm/100cm2", 321, 0, 52.2, "dpm/100cm2",
	. E.
	E+01,
Alpha"  Alpha"	"Alpha",
	=
110:59:00", "109" 111:02:00", "109" 111:04:00", "109" 111:04:00", "109" 111:11:00", "109" 111:11:00", "109" 111:11:00", "109" 111:12:00", "109" 111:28:00", "109" 111:38:00", "109" 111:38:00", "109" 111:38:00", "109" 111:38:00", "109" 111:38:00", "109" 111:36:00", "109" 111:36:00", "109" 111:36:00", "109" 111:36:00", "109" 111:50:00", "109" 111:50:00", "109" 111:50:00", "109" 111:50:00", "109" 111:50:00", "109" 111:50:00", "109" 112:50:00", "109" 112:50:00", "109" 112:50:00", "109"	3:04:00","10
02/03/00", 02/03/00",	3/00/1
"IVP0000302", "IVP0000303", "IVP0000303", "IVP0000304", "IVP0000304", "IVP0000306", "IVP0000306", "IVP0000306", "IVP0000301", "IVP0000311", "IVP0000312", "IVP0000312", "IVP0000312", "IVP0000311", "IVP0000311", "IVP0000312", "IVP0000311", "IVP0000312", "IVP0000312", "IVP0000313", "IVP0000313", "IVP0000313", "IVP0000313", "IVP0000313", "RESP/CHECK", "R	I. <del></del> .